



Artisan Technology Group is your source for quality new and certified-used/pre-owned equipment

- FAST SHIPPING AND DELIVERY
- TENS OF THOUSANDS OF IN-STOCK ITEMS
- EQUIPMENT DEMOS
- HUNDREDS OF MANUFACTURERS SUPPORTED
- LEASING/MONTHLY RENTALS
- ITAR CERTIFIED SECURE ASSET SOLUTIONS

SERVICE CENTER REPAIRS

Experienced engineers and technicians on staff at our full-service, in-house repair center

*InstraView*SM REMOTE INSPECTION

Remotely inspect equipment before purchasing with our interactive website at www.instraview.com ↗

WE BUY USED EQUIPMENT


Sell your excess, underutilized, and idle used equipment. We also offer credit for buy-backs and trade-ins. www.artisanng.com/WeBuyEquipment ↗

LOOKING FOR MORE INFORMATION?

Visit us on the web at www.artisanng.com ↗ for more information on price quotations, drivers, technical specifications, manuals, and documentation

Contact us: (888) 88-SOURCE | sales@artisanng.com | www.artisanng.com

PIGGYBACK INFUSION



1. Prepare IV fluids per set instructions. Primary sets must have check valves
2. Connect distal end of Piggyback set into the uppermost injection site of the Primary set
3. Hang the Primary fluid container at least 8 inches below the Piggyback fluid container
4. Press **HOLD** to stop the Primary infusion

5. Press **PIGGYBACK**
6. To enter Piggyback Rate (mL/hr) and Piggyback Volume To Be Delivered (mL), press
7. Verify all entries are correct
8. Press **PIGGYBACK** to start

NOTE: Inserting the piggyback connector into an injection site distal to the device will cause uncontrolled fluid delivery.

NOTE: To minimize or prevent fluid flow from the primary container during a piggyback infusion (sympathetic flow), it may be necessary to lower the primary bag more than 8 inches or clamp off the primary tubing. Sympathetic flow increases significantly when the piggyback rate is greater than 125 mL/hr and clamping the primary tubing is recommended at rates greater than 125 mL/hr. When clamping the primary tubing, it is advisable to activate the Piggyback Callback feature on the device to notify the clinician when the piggyback infusion is complete so the primary line can be unclamped.

ALARM GUIDE

Red Text	Screen	Cause
Air in Line	Air In Line	Air in downstream tubing Tubing improperly inserted in Air In Line detector
Bag Empty	Container Is Empty	Check for air in cassette IV container empty Upstream clamp closed Cassette improperly installed
Check Set	CLOSE ROLLER CLAMP	Tubing missing or cassette improperly loaded.
Door Open	Door Open	The door must be closed to deliver fluid.
Error	System Alarm or Repair Instrument	Turn pump off and return to Biomed for service
Hold	Hold Time Exceeded	3 minute hold time exceeded
Occlusion	Downstream Occlusion or Upstream Occlusion	Clamp closed or filter blocked Tubing kinked IV positional or infiltrated Catheter or vein too small Occlusion Limit too low or Inadequate fluid flow to device. Upper clamp closed
Occl Warning	Downstream tubing occluded	
PLUG IN	Low Battery or Battery Very Low	Battery low: plug into AC immediately
Flashing Rate and VTBD	KVO: Primary (or Piggyback) Complete	Infusion is complete

Rx only. Outlook is a registered trademark and DoseScan is a trademark of B. Braun Medical Inc.
© B. Braun Medical Inc. 2004

B | BRAUN
SHARING EXPERTISE

B. Braun Medical Inc.
Bethlehem, PA
800-854-6851
www.bbraunusa.com

OUTLOOK®

SAFETY INFUSION SYSTEM

General Use




Quick Reference Guide

For Clinical and Technical Support, call 800-854-6851

B | BRAUN
SHARING EXPERTISE

CASSETTE PRIMING

1. Gravity prime set while inverting cassette (pointed end up)
2. Ensure all air is expelled by tapping the inverted cassette, injection sites, and check valve
3. Close roller clamp



CASSETTE LOADING

1. Close blue slide clamp on the Free Flow Protector Clip
2. Position cassette onto alignment pins with metal disc toward the device
3. Insert and press gray Free Flow Protector Clip into green receptacle
4. Close door and open all clamps

WARNING: *Unrestricted fluid flow may occur if the Free Flow Protector Clip is improperly installed into the receptacle. For intentional flow, press blue slide clamp on the Free Flow Protector Clip.*

PRIMARY INFUSION

1. Prime and load cassette
2. Press **POWER**
3. Press PRIMARY
4. To enter Rate (mL/hr) and Volume To Be Delivered (mL), press
5. Verify all entries are correct
6. Press PRIMARY **RUN** to start

TITRATION (CHANGING THE RATE WHILE INFUSING)

1. To enter new rate, press
2. Press **ENTER** or **RUN** to validate data and begin infusing at the new rate

NOTE: If **ENTER** or **RUN** key is not pressed within 4 seconds, error beeps will sound and rate change will be cancelled. The device will continue to sound an intermittent beep and display "Rate change aborted"

To silence the intermittent beeps, either re-attempt the titration, or, if the displayed rate is acceptable, press **RUN**

CASSETTE REMOVAL

1. Press **HOLD** to stop the infusion
2. Close downstream roller clamp
3. Open door
4. Grasp blue slide clamp on the Free Flow Protector Clip and pull it from the receptacle with one hand while the other hand gently pulls the tubing from the left side of the device

NOTE: If gravity infusion is desired, open the blue slide clamp on the Free Flow Protector Clip and adjust the rate with the roller clamp

MENU ACCESS

The Outlook® Safety Infusion System contains two menus (Main and Alternate) which list multiple infusion modes and operating features including:

- DoseScan™
- DOSE Mode
- q(x) Schedule
- Profile Mode
- Program Mode
- Clear Total Infused
- Occlusion Limit
- Change Alarm Volume
- Micro Rate/Limit
- Site Trend
- Calculate BSA
- Adjust Display
- Piggyback Callback

NOTE: Some items may have been disabled by your institution and will not appear in the menu system. DoseScan is not included with the Outlook 100 model.

1. Press **MENU**
2. To highlight mode or feature, use the
3. Press **ENTER**

OCCUSION LIMIT

1. Press **HOLD** to stop the infusion
2. Press **MENU**
3. To highlight "Occlusion Limit", use
4. Press **ENTER**
5. To increase or decrease the occlusion limit, press
6. Press **ENTER**
7. Press **RUN** to restart the infusion

NOTE: When infusing at rates greater than or equal to 400 mL/hr, the occlusion limit will automatically set to 400 mmHg to accommodate fluid resistance



Artisan Technology Group is your source for quality new and certified-used/pre-owned equipment

- FAST SHIPPING AND DELIVERY
- TENS OF THOUSANDS OF IN-STOCK ITEMS
- EQUIPMENT DEMOS
- HUNDREDS OF MANUFACTURERS SUPPORTED
- LEASING/MONTHLY RENTALS
- ITAR CERTIFIED SECURE ASSET SOLUTIONS

SERVICE CENTER REPAIRS

Experienced engineers and technicians on staff at our full-service, in-house repair center

*InstraView*SM REMOTE INSPECTION

Remotely inspect equipment before purchasing with our interactive website at www.instraview.com ↗

WE BUY USED EQUIPMENT

Sell your excess, underutilized, and idle used equipment. We also offer credit for buy-backs and trade-ins. www.artisanng.com/WeBuyEquipment ↗

LOOKING FOR MORE INFORMATION?

Visit us on the web at www.artisanng.com ↗ for more information on price quotations, drivers, technical specifications, manuals, and documentation

Contact us: (888) 88-SOURCE | sales@artisanng.com | www.artisanng.com