


VXI Technology VT1509A

## 8-Channel x64 Gain and 7 Hz Fixed Filter SCP Module



**In Stock**

**Used and in Excellent Condition**

**Open Web Page**

<https://www.artisanng.com/88897-3>

All trademarks, brandnames, and brands appearing herein are the property of their respective owners.



Your **definitive** source  
for quality pre-owned  
equipment.

**Artisan Technology Group**

(217) 352-9330 | [sales@artisanng.com](mailto:sales@artisanng.com) | [artisanng.com](http://artisanng.com)

- Critical and expedited services
- In stock / Ready-to-ship

- We buy your excess, underutilized, and idle equipment
- Full-service, independent repair center

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.



# **Eight-Channel Fixed Filter and Amplifier Signal Conditioning Plug-on VT1509A**

---

## **User's Manual**

The VT1509A manual also applies to Agilent/HP E1413s as Agilent/HP E1413 Option 19.

Enclosed is the User's Manual for the VT1509A Signal Conditioning Plug-on. Insert this manual in your VT1413C and Agilent/HP E1313 manual behind the "Signal Conditioning Plug-ons" divider.



Copyright © VXI Technology, Inc., 2003



# VT1509A

## Eight-Channel Fixed Filter and Amplifier Signal Conditioning Plug-on

---

### Introduction

The VT1509A is a Signal Conditioning Plug-on that provides eight fixed low-pass filters with a 3 dB cutoff frequency of 7 Hz and eight amplifiers with gain of 64. Also provided is input over-voltage protection and open transducer detection on each channel.

### About this Manual

This manual shows you how to control the Signal Conditioning Plug-on (SCP) using SCPI commands as well as Register-Based commands, and explains the capabilities of this SCP. Finally, it covers specifications for this SCP. The contents of this manual are:

Installation .....	3
Identifying the Plug-on .....	3
Connecting To The Terminal Module .....	4
Programming With SCPI Commands .....	6
Programming With Register Commands .....	8
Specifications .....	10

### Installation

Installation for this Plug-on is common to several others and is covered in Chapters 1 and 2 of your VT1413C or Agilent/HP E1313 manual.

### Identifying the Plug-on

You'll find the VXI Technology part number on the connector side of the SCP to the left of the serial number bar code. For the VT1509A, the part number is : VT1509A.

# Connecting To The Terminal Module

This section shows how to make connections to the Terminal Module.

The SCP connections for the Terminal Modules are shown on the stick-on labels that came with the SCP. Use the appropriate label for the type of Terminal Module you have. The connections and appropriate stickers are as follows:

For VT1413C and above Terminal Modules, use stickers for VT1509A SCPs. The connections are shown in Figure 1.

For Agilent/HP E1313 Terminal Modules, use stickers for VT1509A SCPs. The connections are shown in Figures 2 and 3.

For Agilent/HP E1413B and below Terminal Modules, use stickers for HP E1413 Option 19 SCPs. The connections are shown in Figure 4.




Figure 1 VT1509A C-Size Terminal Module Connections




Figure 2 VT1509A B-size Terminal Module Connections (Ch 00-31)




Figure 3 VT1509A B-size Terminal Module Connections (Ch 32-63)




Figure 4 Agilent/HP E1413B Option 19 Terminal Module Connections

## Programming With SCPI Commands

The SCPI commands shown here are covered in Chapters 3 and 5 of your VT1413C or Agilent/HP E1313 manual. This section will relate those commands to the parameter values which are specific to this Plug-on.

## Checking the ID of the SCP

To verify the SCP type(s) installed on the VT1413C or Agilent/HP E1313 use the **SYSTem:CTYPE?** (**@<channel>**) command.

The channel parameter specifies a single channel in the channel range covered by the SCP of interest. The first channel number for each of the eight SCP positions are; 0,8,16,24,32,40,48 and 56.

The value returned for the SCP in an Agilent/HP E1413B is:  
HEWLETT-PACKARD,E1413 Opt 19 8-Channel Fixed Gain-Filter SCP,0.0

The value returned for the SCP in a VT1413C or an Agilent/HP E1313A is:  
HEWLETT-PACKARD.E1509 8-Channel Fixed Gain-Filter SCP.0.0

To determine the type of SCP installed on channels 0 through 7 send

SYST:CTYP? (@100)                      query SCP type @ ch 0  
enter statement here

## Querying the Filter Cutoff Frequency

While the the VT1509A does not provide programmable cutoff frequency the filter frequency can be queried. The response to this query will always be 7. To query any channel for its cutoff frequency use the INP:FILTer[:LPASs]:FREQuency? (@<channel>) command. The INP:FILT:FREQ? command returns the numeric cutoff value currently set for the channel specified.

The channel parameter must specify a single channel.

To query the cutoff frequency of channel 6 send

INP:FILT:FREQ? (@106)                      query channel 6  
enter statement here

## Querying the Filter State

While the VT1509A does not allow controlling whether the filters are enabled or disabled, this state can be queried. The response to this query will always be 1. To query any channel to determine if it is enabled or disabled use the `INP:FILTer[:LPASs][:STATe]? (@<channel>)` command. The `INP:FILT?` command returns a 0 if the channel is OFF or a 1 if the channel is ON.





---

## NOTES

1) When OTD is enabled, the inputs have up to 0.2  $\mu$ A injected into them. If this current will adversely affect your measurement, but you still want to check for open transducers, you can enable OTD, make a single scan, check the CVT for bad measurements, then disable OTD and make your regular measurement scans. The specifications apply only when OTD is off.

2) Allow 5 minutes before checking for open transducers to allow filter capacitors to charge.

---

To enable or disable Open Transducer Detection, use the DIAGnostic:OTDetect <enable>, (@<ch\_list>) command.

The enable parameter can specify ON or OFF

An SCP is addressed when the ch\_list parameter specifies a channel number contained on the SCP. The first channel on each SCP is:  
0, 8, 16, 24, 32, 40, 48 and 56

To enable Open Transducer Detection on all channels on SCPs 1 and 3:

DIAG:OTD ON, (@100,116)                      0 is on SCP 1 and 16 is on SCP3

To disable Open Transducer Detection on all channels on SCPs 1 and 3:

DIAG:OTD OFF, (@100,116)

## Register Based Programming

The register-based commands shown here are covered in Appendix D of the VT1413C or Agilent/HP E1313 manual. You should read that section first to become familiar with accessing registers and executing Register-Based Commands. This section will relate those commands to the parameter values which are specific to this Plug-on.

When Register Programming an SCP most communication is through the Signal Conditioning Bus. For that you will use the Register Commands:

SCBWRITE <regaddr> <regvalue>  
and  
SCBREAD? <regaddr>

### VT1509A Register Map

Read (returned value)	Write( <regvalue>)	SCP Register	<regaddr> Value
SCP ID (9020 <sub>16</sub> )		Whole SCP Reg 0	00ppp000000 <sub>2</sub>

ppp=Plug-on  
ccc=SCP  
channel

In addition you will access bits in the Card Control register to control Open Transducer Detection.

### Checking ID of SCP

To query an SCP for its ID value, write the following value to Parameter Register 1:

(SCP number) 40<sub>16</sub>

Then write the opcode for SCBREAD? (0800<sub>16</sub>) to the Command Register. The ID value will be returned to the Query Response Register.

### Detecting Open Transducers

Open Transducer Detection (OTD) is controlled by bits in the Card Control Register. For more information on OTD see Figure 1.

#### Card Control Register

(Base + 12<sub>16</sub>)

15	14	14-13	12	11	10-8	7-0
PSI Pwr Reset	FIFO Mode	unused	FIFO Clear	VPPEN	A24 Window	Open Transducer Detect

Writing a one (1) to a bit enables open transducer detect on that signal conditioning module. Writing a zero (0) to a bit disables open transducer detect.

Bit 7	Bit 6	Bit 5	Bit 4	Bit 3	Bit 2	Bit 1	Bit 0
SCP 7	SCP 6	SCP 5	SCP 4	SCP 3	SCP 2	SCP 1	SCP 0

# Specifications

These specifications for the VT1509A reflect the combined performance of the VT1413C or Agilent/HP E1313 and the VT1509A Signal Conditioning Plug-on. These specifications are not to be added to those presented in the VT1413C or Agilent/HP E1313 User's Manual.

## General Specifications

Measurement ranges						
DC Volts	±3.9 mV to ±0.25 V FS					
Temperature	Thermocouples - -200 to +1700°C Thermistors - (Opt 15 required) -80 to +160°C RTD's - (Opt 15 required) -200 to +850°C					
Resistance	(Opt 15 required) 128 to 131 k FS					
Strain	25,000 μe or limit of linear range of strain gage					
Maximum input voltage (Normal mode plus common mode)						
			Operating: < ±16 V peak    Damage level: > ±42 V peak			
Maximum common mode voltage						
			Operating: < ±16 V peak    Damage level: > ±42 V peak			
Normal mode rejection						
			@ 7 Hz -3 dB, @ 60 Hz >-25 dB			
Common mode rejection						
			0 - 60 Hz -100 dB			
Input impedance						
			100 M ±10% (each differential input to ground)			
Maximum tare cal offset						
			(Maximum tare offset depends on A/D range and SCP gain)			
A/D range ±V F.Scale	16	4	1	0.25	0.0625	
Max Offset	0.04970	0.01220	0.00297	0.00055	N/A	

**Measurement accuracy  
DC Volts**

(90 days) 23°C±1°C (with \*CAL? done after 1 hr warm up and CAL:ZERO? within 5 min.). If autoranging is ON, add ±0.02% FS to accuracy specifications.  
**For Agilent/HP E1313, multiply Noise Spec. by 1.4.**

Gain 64	Range ±V FS	Linearity % of reading	Offset Error μV (10 Hz)	Noise μV 3 sigma	Noise μV* 3 sigma
	0.0039	0.01	2.3	1.7	1.4
	0.0156	0.01	2.4	2.5	2.2
	0.0625	0.01	3.0	7.0	5.7
	0.25	0.01	8.0	28	23

\* [SENSe:]FILTer[:LPASs][:STATe] ON (max scan rate - 100 rdgs/sec/channel)

Temperature Coefficients: Gain - 15 ppm/°C. Offset - (0 - 30°C) 0.16 μV/°C, (30 - 40°C) 0.18 μV/°C,  
(40 - 55°C) .39 μV

**Measurement accuracy  
Temperature**

(90 days) 23°C±1°C (with \*CAL? done after 1 hr warm up and CAL:ZERO? within 5 min.). If autoranging is ON, add ±0.02% FS to accuracy specifications.

(simplified specifications, see temperature accuracy graphs in VT1413C or Agilent/HP E1313 manual for details)

The temperature accuracy specifications include instrument and firmware linearization errors. The linearization algorithm used is based on the IPTS-68(78) standard transducer curves. Add your transducer accuracy to determine total measurement error.

**Thermocouples**

<b>Type E</b>	A/D Filter	-200 to 0°C	0 to 200°C	200 to 400°C	400 to 800°C
	OFF ON*	1.25°C 1.20°C	0.10°C 0.095°C	0.12°C 0.10°C	0.125°C 0.11°C
<b>Type EExtended</b>	A/D Filter	-200 to 0°C	0 to 200°C	200 to 800°C	800 to 1000°C
	OFF ON*	13.4°C 13.3°C	0.30°C 0.25°C	0.20°C 0.15°C	0.35°C 0.30°C
<b>Type J</b>	A/D Filter	-200 to 0°C	0 to 280°C	280 to 600°C	600 to 775°C
	OFF ON*	1.50°C 1.47°C	0.10°C 0.11°C	0.15°C 0.15°C	0.17°C 0.15°C
<b>Type K</b>	A/D Filter	-200 to 0°C	0 to 375°C	375 to 800°C	800 to 1400°C
	OFF ON*	2.25°C 2.70°C	0.15°C 0.15°C	0.20°C 0.17°C	0.25°C 0.25°C
<b>Type R</b>	A/D Filter	0 to 100°C	100 to 200°C	200 to 600°C	600 to 1000°C
	OFF ON*	1.40°C 1.40°C	0.75°C 0.70°C	0.45°C 0.40°C	0.30°C 0.30°C

\* [SENSe:]FILTer[:LPASs][:STATe] ON (max scan rate - 100 rdgs/sec/channel)

**Measurement accuracy  
Temperature (cont.)**

(simplified specifications, see temperature accuracy graphs in VT1413C or Agilent/  
HP E1313 manual for details)

**Thermocouples (cont.)**

<b>Type S</b>	A/D Filter	0 to 100°C	100 to 200°C	200 to 800 C	800 to 1750°C
	OFF ON*	2.85°C 2.85°C	1.35°C 1.80°C	0.70°C 0.65°C	0.65°C 0.55°C
<b>Type T</b>	A/D Filter	-200 to -100°C	-100 to 0°C	0 to 200°C	200 to 400°C
	OFF ON*	1.35°C 1.35°C	0.25°C 0.22°C	0.10°C 0.10°C	0.15°C 0.13°C

**5 k Reference Thermistor**

A/D Filter	45 to 65°C	65 to 85°C
OFF ON*	0.0061°C 0.0055°C	0.011°C 0.010°C

**100 Reference RTD**

A/D Filter	-125 to 70°C
OFF ON*	0.080°C 0.080°C

**100 RTD**

A/D Filter	-200 to 75°C	75 to 300°C	300 to 600°C	600 to 970°C
OFF ON*	0.08°C 0.07°C	0.21°C 0.18°C	0.27°C 0.25°C	0.37°C 0.35°C

**2252 Thermistor**

A/D Filter	63 to 70°C	70 to 83°C	83 to 100°C
OFF ON*	0.008°C 0.007°C	0.009°C 0.008°C	0.014°C 0.012°C

**5 k Thermistor**

A/D Filter	15 to 30°C	30 to 70°C	70 to 85°C
OFF ON*	0.011°C 0.010°C	0.013°C 0.013°C	0.017°C 0.017°C

**10 k Thermistor**

A/D Filter	30 to 60°C	60 to 90°C	90 to 115°C
OFF ON*	0.015°C 0.014°C	0.016°C 0.015°C	0.022°C 0.020°C

# Artisan Technology Group is an independent supplier of quality pre-owned equipment

## Gold-standard solutions

Extend the life of your critical industrial, commercial, and military systems with our superior service and support.

## We buy equipment

Planning to upgrade your current equipment? Have surplus equipment taking up shelf space? We'll give it a new home.

## Learn more!

Visit us at [artisanTG.com](https://www.artisanTG.com) for more info on price quotes, drivers, technical specifications, manuals, and documentation.

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

**We're here to make your life easier. How can we help you today?**

(217) 352-9330 | [sales@artisanTG.com](mailto:sales@artisanTG.com) | [artisanTG.com](https://www.artisanTG.com)

