

New Brunswick Scientific BioFlo 110
Modular Fermentation System



**Limited Availability
Used and in Excellent Condition**

Open Web Page

<https://www.artisanng.com/53185-1>

All trademarks, brandnames, and brands appearing herein are the property of their respective owners.



Your **definitive** source
for quality pre-owned
equipment.

Artisan Technology Group

(217) 352-9330 | sales@artisanng.com | artisanng.com

- Critical and expedited services
- In stock / Ready-to-ship

- We buy your excess, underutilized, and idle equipment
- Full-service, independent repair center

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.




Guide to Operations

BioFlo 110 Modular Benchtop Fermentor

MANUAL NO: M1273-0054

Revision H

August 8, 2007



NEW BRUNSWICK SCIENTIFIC CO., INC.

BOX 4005 • 44 TALMADGE ROAD • EDISON, NJ 08818-4005

Telephone: 1-732-287-1200 • 1-800-631-5417

Fax: 732-287-4222 • Telex: 4753012 NBSCO

Internet: <http://www.nbsc.com> • E-mail: bioinfo@nbsc.com

INTERNATIONAL OFFICES:

BELGIUM

New Brunswick Scientific NV-SA
Stationsstraat 180/4
3110 Rotselaar
België/Belgique
Tel: 32 (0)16 56 28 31
Fax: 32 (0)16 57 27 53
E-mail: sales@nbsnv-sa.be

CHINA

New Brunswick Scientific Co., Inc.
A903, 904 Yin Hai Building
No. 250, Cao Xi Road
Shanghai 200235, P.R. China
Tel: 86 21 6484 5955 or 5966
Fax: 86 21 6484 5933
E-mail: nbschc@online.sh.cn

FRANCE

New Brunswick Scientific SARL
12-14, Rond Point des Champs Elysées
75008 Paris
France
Tel: 33 (0)1 5353 1511
Fax: 33 (0)1 5353 1557
E-mail: sales@nbssarl.fr

GERMANY

New Brunswick Scientific GmbH
In Der Au 14
D-72622 Nürtingen
Deutschland
Tel: 49 (0)7022 932490
Fax: 49 (0)7022 32486
E-mail: sales@nbsgmbh.de

THE NETHERLANDS

New Brunswick Scientific BV
Kerkenbos 1101, 6546 BC Nijmegen
P.O Box 6826, 6503 GH Nijmegen
Nederland
Tel: 31 (0)24 3717 600
Fax: 31 (0)24 3717 640
E-mail: sales@nbsbv.nl

UNITED KINGDOM

New Brunswick Scientific (UK) Ltd.
17 Alban Park
St. Albans, Herts. AL4 0JJ
United Kingdom
Tel: 44 (0)1727 853855 or 0800 581331
Fax: 44 (0)1727 835666
E-mail: bioinfo@nbsuk.co.uk
Web: www.nbsuk.co.uk

**WARNING!**

High voltage.

Always make sure this equipment is properly grounded.

**WARNING!**

This product is not designed to contain gases within the range of their lower explosion limit (LEL) and their upper explosion limit (UEL). If your process requires or produces gases, be sure to verify their LEL and UEL concentration range (available online).

**CAUTION!**

This equipment *must* be operated as described in this manual. If operational guidelines are not followed, equipment damage and personal injury *can* occur. Please read the entire User's Guide before attempting to use this unit.

Do not use this equipment in a hazardous atmosphere or with hazardous materials for which the equipment was not designed.

New Brunswick Scientific Co., Inc. (NBS) is not responsible for any damage to this equipment that may result from the use of an accessory not manufactured by NBS.

Copyright Notice

New Brunswick Scientific Company, Inc.
Box 4005 44 Talmadge Road
Edison, New Jersey 08818-4005

© Copyright 2007 New Brunswick Scientific Co., Inc.

All Rights Reserved.

Reproduction, adaptation, or translation without prior written permission from New Brunswick Scientific is prohibited.

Disclaimer Notice

New Brunswick Scientific Co., Inc. reserves the right to change information in this document without notice. Updates to information in this document reflect our commitment to continuing product development and improvement.

Manual Conventions



NOTE:

Notes contain essential information that deserves special attention.



CAUTION!

Caution messages appear before procedures which, if caution is not observed, could result in damage to the equipment.



WARNING!

Warning messages alert you to specific procedures or practices which, if not followed correctly, could result in serious personal injury.

Bold

Text in boldface type emphasizes key words or phrases.



This particular *Warning* message, whether found in the manual or on the unit, means **HOT SURFACE**—and therefore represents a potential danger to touch.



CRUSH WARNING!

Crush Warning messages alert you to specific procedures or practices regarding heavy objects which, if not followed correctly, could result in serious personal injury .





New Brunswick Scientific



DECLARATION OF CONFORMITY

New Brunswick Scientific, Hereby declares that the product(s) listed below conform to the European Union directive and standards identified in this declaration.

Product(s)

BF110 Thermal Mass Flow Controller module M1273-3109

EU Directive(s)

Low Voltage (73/23/EEC/93/68/EEC)

Electromagnetic Compatibility (89/336/EEC/93/68/EEC)

Standard(s)

IEC61010-1: 1990 +A1 1992 +A2 1995

EN61010-1: 1993 +A2: 1995

EN61326: 1997+A1: 1998


The conformity assessment procedures were performed at the following Testing Lab.:


Intertek Testing Services, 40 Commerce Way, Totowa, NJ 07512

The technical documentation relevant to the above equipment will be held at:

New Brunswick Scientific Company
PO Box 4005
44 Talmadge Road
Edison, New Jersey 08818-4005 U.S.A
Tel. (732) 287-1200
Fax. (732) 287-4222


Lee Epstein
VP of Science & Technology


Date



Intertek Testing Services
40 B Commerce Way, Totowa, NJ 07512

EMC VERIFICATION No. J20046647

EQUIPMENT UNDER TEST

Type of equipment: Four Pump Module

Type/Model: BioFlo 110

Manufacturer: New Brunswick Scientific Co., Inc.

Tested by request of: Mr. Michael Gut

STANDARDS:

EN 55014-1: 1997 - Electromagnetic Compatibility - Requirements for household appliances, electric tools and similar apparatus, Part 1. Emission;

EN 61000-3-2:1995 Electromagnetic compatibility Part 3 Limits Section 2. Limits for harmonic current emissions;

EN 61000-3-3:1995 Electromagnetic compatibility Part 3 Limits Section 3. Limits of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current ≤ 16 A.

TEST REPORT No. J20046647

SUMMARY OF RESULTS

We confirm that the product tested, and our review of the above numbered report, without reasonable doubt will fulfill the requirements concerning electromagnetic compatibility according to the above mentioned standards harmonized with the EMC Directive 89/336/EEC//93/68/EEC.

EMC Department
Date of Issue: 05/04/01

Signature:  Luis M. Bas, Operations Manager

TABLE OF CONTENTS

1	OVERVIEW.....	1
1.1	DESIGN CONCEPT.....	1
1.2	PROCESS CONTROL.....	1
1.3	COMPUTER-READY.....	2
1.4	ENHANCING YOUR FERMENTOR.....	2
2	INSPECTION & UNPACKING.....	3
2.1	INSPECTION OF BOXES.....	3
2.2	PACKING LIST VERIFICATION.....	3
3	PREPARING THE LOCATION.....	4
3.1	PHYSICAL LOCATION.....	4
3.2	ENVIRONMENT.....	4
3.3	UTILITIES.....	4
4	INSTALLATION, ASSEMBLY & SET-UP.....	5
4.1	INTRODUCTION TO CONTROL MODULE REAR PANELS.....	5
4.1.1	Primary Control Unit (PCU).....	5
4.1.2	Power Controller.....	6
4.1.3	dO ₂ /pH Controller.....	7
4.1.4	Level Controller.....	8
4.1.5	Gas Mix Controller.....	8
4.1.6	Pump Module.....	9
4.1.7	Other Optional Modules.....	10
4.2	STACKING MODULES.....	10
4.3	INTERCONNECTING THE MODULES.....	12
4.4	**IMPORTANT WARNINGS**	14
4.5	VESSEL ASSEMBLY: NON-JACKETED.....	16
4.5.1	Headplate.....	18
4.5.2	Install Heat Blanket.....	20
4.5.3	Install Vessel in Vessel Stand.....	21
4.5.4	Install Baffle (14.0L Fermentation Vessels ONLY).....	21
4.5.5	Install Impeller(s).....	22
4.5.6	Install Cooling Coil.....	23
4.5.7	Install Sparger (3.0L, 7.5L & 14.0L Vessels).....	23
4.5.8	Install Harvest Tube.....	24
4.5.9	Install Sampler Tube.....	24
4.5.10	Install Thermowell.....	24
4.5.11	Install Foam Probe.....	24
4.5.12	Install Foam Exhaust Tube.....	25
4.5.13	Install Level Probe(s).....	25
4.5.14	Install Addition Tube(s).....	25
4.5.15	Install pH Probe.....	25
4.5.16	Install dO ₂ Probe.....	27
4.5.17	Install Exhaust Condenser.....	28
4.5.18	Install Sampler.....	30
4.5.19	Install Foam Trap.....	32
4.5.20	Plug Unused Ports.....	33
4.5.21	Install 1.3L, 3.0L or 7.5L Fermentation Vessel Baffle.....	34
4.5.22	Install Headplate.....	34
4.6	VESSEL ASSEMBLY: WATER-JACKETED.....	35
4.6.1	Headplate.....	37

4.6.2	Install Headplate Clamping Ring	40
4.6.3	Install Vessel on Base Plate.....	40
4.6.4	Install Baffle (14.0L Fermentation Vessels ONLY)	41
4.6.5	Install Impeller(s)	41
4.6.6	Install Sparger	42
4.6.7	Install Harvest Tube	43
4.6.8	Install Sampler Tube.....	43
4.6.9	Install Thermowell.....	43
4.6.10	Install Foam Probe	43
4.6.11	Install Foam Exhaust Tube	44
4.6.12	Install Level Probe(s).....	44
4.6.13	Install Addition Tube(s)	44
4.6.14	Install pH Probe.....	44
4.6.15	Install Dissolved Oxygen Probe.....	46
4.6.16	Install Exhaust Condenser	47
4.6.17	Install Sampler.....	50
4.6.18	Install Foam Trap	52
4.6.19	Install Antifoam System.....	53
4.6.20	Plug Unused Ports.....	53
4.6.21	Install 1.3L, 3.0L or 7.5L Fermentation Vessel Baffle.....	53
4.6.22	Install Headplate.....	54
4.7	INSTALL THE AGITATION MOTOR.....	54
4.8	INSTALL THE ROTAMETER.....	55
4.9	INSTALL LIQUID ADDITION SYSTEMS.....	56
4.10	MAKING CONNECTIONS	57
4.10.1	Power	57
4.10.2	Water.....	57
4.10.3	Probes and Sensors.....	58
4.10.4	Pumps	61
4.10.5	Air and Optional Gases	62
4.11	INSTALL CABLE GUIDES.....	63
4.12	VOLTAGE SELECTION AND FUSING	64
4.13	CALIBRATING THE pH PROBE.....	65
5	BEFORE FIRST USE.....	69
5.1	SELECT PCU DISPLAY LANGUAGE	69
5.2	ASSIGN VESSEL ADDRESS	69
5.3	FOR PC SUPERVISION	70
5.3.1	Connect to Supervisory System.....	70
5.3.2	Select Communications Protocol.....	71
5.3.3	Set Base Addresses	72
5.3.4	Set Chart Recorder for Data Logging	72
6	STERILIZATION & REINSTALLATION	75
6.1	INITIAL PREPARATION FOR AUTOCLAVING	76
6.1.1	Filling the Water Jacket.....	76
6.2	ADDITIONAL PREPARATION FOR AUTOCLAVING.....	76
6.3	AUTOCLAVING THE VESSEL	77
6.3.1	Sterilization Time and Temperature	78
6.4	REINSTALLING THE SYSTEM.....	79
6.5	CONFIRMING pH CALIBRATION.....	80
6.6	POLARIZING & CALIBRATING THE DO2 PROBE.....	80
6.7	POWERING THE JACKET WATER HEATER	82
7	OPERATION	84
7.1	PRIMARY CONTROL UNIT DISPLAY & CONTROLS.....	84

7.2	SELECTING THE APPLICATION	86
7.3	MAIN SCREEN	87
7.4	SET THE TEMPERATURE LOOP.....	89
7.4.1	Control Mode.....	89
7.4.2	Setpoint.....	91
7.4.3	Graphic Display	92
7.5	SETTING THE AGITATION LOOP.....	93
7.6	SETTING THE pH LOOP	93
7.6.1	Setting a Deadband for pH control.....	94
7.6.2	Setting the Output Multiplier for pH Control	95
7.6.3	Setting P & I Gains for pH Control.....	97
7.7	SETTING THE DO2 LOOP	98
7.8	SETTING DO2 CONTROL CASCADES.....	98
7.8.1	Cascades for Fermentation without Gas Mix Controller	98
7.8.2	Cascades for Fermentation with Gas Mix Controller	100
7.8.3	Cascades for Cell Culture without Gas Mix Controller	102
7.8.4	Cascades for Cell Culture with Gas Mix Controller	102
7.9	SETTING P & I GAINS FOR DO2 CONTROL	105
7.10	SETTING THE GAS MIX FOR FERMENTATION.....	106
7.11	SETTING THE GAS MIX FOR CELL CULTURE.....	106
7.11.1	4-Gas Mode.....	107
7.11.2	Demand Mode.....	107
7.11.3	100% Feed Mode.....	107
7.12	SETTING THE PUMPS	108
7.12.1	Pump Operation.....	108
7.12.2	Pump Setpoint and Multiplier	108
7.12.3	Control Mode.....	109
7.12.4	Setting Pumps	109
7.13	SETTING THE PUMP CONTROL MODE: ANTIFOAM, WET ON & WET OFF	110
7.14	INOCULATION.....	112
7.15	STERILE SAMPLING	112
7.16	AUTOMATIC FOAM CONTROL	112
7.17	HARVESTING	113
8	MAINTENANCE.....	114
8.1	CLEANING THE MODULES	114
8.2	pH PROBE MAINTENANCE AND STORAGE	114
8.3	DO2 PROBE MAINTENANCE AND STORAGE.....	114
8.4	CLEANING THE VESSELS	115
8.4.1	14.0L Vessel Disassembly.....	115
8.4.2	Vessel Cleaning Methods.....	115
8.5	PORT & ADAPTER O-RING REPLACEMENT	116
8.6	HEADPLATE O-RING REPLACEMENT	117
8.7	BEARING HOUSING CAP REPLACEMENT.....	117
8.8	REPLACING PCU FUSES	117
9	SERVICE.....	118
9.1	TROUBLESHOOTING	118
10	REPLACEMENT PARTS & ACCESSORIES	119
10.1	REPLACEMENT PARTS	119
10.2	REPLACEMENT VESSELS	122
10.3	MISCELLANEOUS PARTS & OPTIONAL ACCESSORIES.....	122
10.4	MODULE KITS	123
11	APPENDIX A: SPECIFICATIONS.....	124

11.1	MODULES	124
11.2	PUMP FLOW RATES	126
11.3	TUBING	127
11.4	VESSELS	127
11.4.1	Non-Jacketed Vessel Dimensions for Autoclaving (Vertical)	127
11.4.2	Non-Jacketed Vessel Dimensions for Autoclaving (Inclined)	128
11.4.3	Water-Jacketed Vessel Dimensions for Autoclaving (Vertical)	129
11.4.4	Non-Jacketed Vessel Dimensions for Autoclaving (Inclined)	129
12	APPENDIX B: TECHNICAL INTERVENTIONS.....	131
12.1	REPLACING POWER CONTROLLER FUSES	131
12.2	REPLACING PCU BATTERIES	132
12.3	VOLTAGE SELECTION.....	133
12.4	CHANGING PCU FUSING	134
12.4.1	Converting from Single Fuse to Double Fuse.....	134
12.5	P&I CONTROL EQUATIONS	136
12.5.1	pH	136
12.5.2	dO2	137
13	APPENDIX C: COMMUNICATIONS PROTOCOLS.....	138
13.1	COMMUNICATIONS SETTINGS.....	138
13.1.1	AFS/ModBus Pin Connections.....	139
13.2	AFS (BioCOMMAND) COMMUNICATIONS PROTOCOL	139
13.2.1	General Specifications	139
13.2.2	Protocol	140
13.2.3	Requests & Response Headers.....	140
13.2.4	Change Commands	141
13.2.5	Message Formats	141
13.3	MODBUS COMMUNICATIONS PROTOCOL	141
13.3.1	Register Map.....	142
14	APPENDIX D: THERMAL MASS FLOW CONTROLLER.....	143
14.1	INSTALLATION	143
14.2	LOADING SOFTWARE UPDATE.....	145
14.3	OPERATION	146
14.3.1	Fermentation Mode Setup.....	147
14.3.2	Fermentation Mode Cascades	149
14.3.3	Cell Culture Mode Setup.....	150
14.3.4	Cell Culture Mode Cascades	151
14.4	CLEANING	152
14.5	SPECIFICATIONS	153
14.6	TROUBLESHOOTING	153
15	APPENDIX E: DRAWINGS.....	154
15.1	MODULE SCHEMATICS	154
15.2	LIST OF DRAWINGS	162
16	APPENDIX F: PORTS & ADAPTERS	164
17	INDEX	165

1 OVERVIEW

1.1 *Design Concept*

The BioFlo 110 is a modular fermentation system for microbial and cell culture applications. It features a modern, easy-to-use control interface, plus an available family of vessels, controller modules, and other accessories that address the diverse needs of the bioprocessing community.

A complete range of water-jacketed and non-jacketed fermentation and cell culture vessels is available to cover 0.4 through 14.0 liters. For temperature control, non-jacketed vessels use an external heating blanket and an internal cooling coil. Jacketed vessels use an external jacket water temperature controller.

Modularity means you can position components to suit your workspace and working style. The controller modules can be vertically stacked to conserve bench space. Modularity also makes it easy and economical to adapt your system for different tasks at any time.

As an autoclavable fermentor, the BioFlo 110 requires an external sterilizing apparatus that is large enough to accept the complete vessel assembly.

1.2 *Process Control*

The heart of the BioFlo 110 is the Primary Control Unit (PCU). The PCU serves as the operator interface for one to four vessels. Its bright graphic display and clearly marked keypad serve as the control center for all attached vessels and their associated loop controllers. Four-channel output to a strip-chart recorder is available from the PCU.



NOTE:

No more than 16 modules (any combination) can be supported by one PCU.

The BioFlo 110 Power Controller is required for temperature and agitation control. It also provides power for liquid pumps and other control modules. The rotameter and cooling water valve are mounted on the outside of the Power Controller.

Additional control modules and sensors enable expanding a basic temperature/agitation/airflow fermentor to include control of pH, dissolved oxygen (dO₂), antifoam addition, removal and addition of nutrients and other liquids, and blending inlet air with up to three other gases.

1.3 *Computer-Ready*

The BioFlo 110 is compatible with both ModBus and New Brunswick Scientific's *BioCommand* communications protocols. Optional *BioCommand* software adds automatic data logging, centralized monitoring and control, and process programming capabilities to one or more fermentors.

1.4 *Enhancing Your Fermentor*

Other accessories are available and new ones will be added to extend the capabilities of the BioFlo 110. Whenever your bioprocessing needs or applications change, contact your New Brunswick Scientific sales representative for a current description of BioFlo 110 products.

2 INSPECTION & UNPACKING

2.1 *Inspection of Boxes*

After you have received your order from New Brunswick Scientific, inspect the boxes carefully for any damage that may have occurred during shipping. Report any damage to the carrier and to your local NBS Sales Order Department immediately.

2.2 *Packing List Verification*

Unpack your order, saving the packing materials for possible future use. Also be sure to save the User's Guide, for instruction and reference.

Verify against your NBS packing list that you have received the correct materials, and that nothing is missing. The part numbers on the packing list correspond to part numbers and/or manufacturing numbers. You will find the number affixed to each item, either on a paper sticker on the packaging or on the serial number identification sticker on the equipment itself.

If any part of your order was damaged during shipping or is missing, or fails to operate, please fill out Customer Satisfaction Form 6300 (packed in the envelope with your warranty card) and return it by fax or mail. You can also call the New Brunswick Scientific Service Department.

3 PREPARING THE LOCATION

3.1 *Physical Location*

The surface where you place the BioFlo 110 should be smooth, level and sturdy. Ensure that there is enough space around the back and the front of the combined system for proper operation and access.

3.2 *Environment*

The BioFlo 110 operates best under the following conditions:

- Ambient temperature range of 10°C to 35° C
- Relative humidity up to 80% non-condensing

3.3 *Utilities*

Utility		Requirements (minimum-maximum)
Water (50 µm filtration)		5-10 PSIG (0.34-0.68 kg/cm ²)
Air		3-10 PSIG (0.20-0.68 kg/cm ²)
Gases (optional)	O2	3-10 PSIG (0.20-0.68 kg/cm ²)
	CO2	3-10 PSIG (0.20-0.68 kg/cm ²)
	N2	3-10 PSIG (0.20-0.68 kg/cm ²)
Electricity	100V	90-100 VAC 50/60 Hz 10 Amp
	120V	108-132 VAC 50/60 Hz 10 Amp
	220V	198-242 VAC 50/60 Hz 6 Amp
	240V	216-264 VAC 50/60 Hz 6 Amp



NOTE:


All gases supplied should be medical grade.

4 INSTALLATION, ASSEMBLY & SET-UP

4.1 Introduction to Control Module Rear Panels

4.1.1 Primary Control Unit (PCU)

Figure 1: Primary Control Unit (PCU) Rear Panel




Refer to Figure 1 above to visually locate the following features on the rear panel:

- FLOPPY DRIVE** An internal A: drive for field upgrades to the interface software.
- POWER SWITCH** On/Off rocker switch.
- AC CONNECTION** Inlet for power cord (supplied), either for connection to Power Controller outlet or for electrical service outlet.
- FUSE HOLDER** Easy access to replace 250V 2-amp Slo-Blo[®] fuse.

- **AFS/MODBUS CONNECTOR** 25-pin D connector for any supervisory software you may add.
- **CONTROLLER BUS CONNECTORS** Two RS-485 9-pin D connectors, one above the other. Use one to connect the Power Controller to the PCU with the cable supplied.
- **RECORDER CONNECTOR** One 9-pin D connector, reserved for use with any analog recorder you may add. 2.5V output supplies data records.

4.1.2 Power Controller

Figure 2: Power Controller Rear Panel




Refer to Figure 2 above to visually locate the following features:

- **WATER SOLENOID VALVE** Located on the side panel (not visible in drawing). Provides water from outside source for cool temperature control within the vessel and to the optional exhaust condenser.
- **WATER SOLENOID CONNECTOR** Located on the rear panel. Control connection for the water solenoid valve.
- **VOLTAGE SELECTOR** Located in the fuse drawer in the rear panel, next to the power cord connection.
- **AC CONNECTION** Located on the rear panel. Input connection, using the power cord supplied, to the electrical service.
- **PCU CONNECTION** Located on the rear panel. Output connection, using the power cord supplied, to power the PCU.
- **CONTROLLER BUS CONNECTORS** Located on the rear panel. Two RS-485 9-pin D connectors, one above the other. Use one to connect the Power Controller to the PCU with the cable supplied.
- **UNIT ADDRESS SELECTOR** Located on the rear panel. Four two-position (on/off) switches used to identify the vessel address, uniquely associating this module to one particular vessel.


4.1.3 dO2/pH Controller

Figure 3: dO2/pH Controller – Rear Panel




4.1.4 Level Controller

Figure 4: Level Controller – Rear Panel



4.1.5 Gas Mix Controller

Figure 5: Gas Mix Controller Rear Panel



4.1.6 Pump Module

A four-pump module is available to stack with the other BioFlo 110 modules, or to stand freely, adjacent to the module stack and vessel.




NOTE:

Unlike the other modules, the pump module is wired for either 110V or 220V electric service only. Be sure that your module is correct for your electrical supply.

Refer to Figure 6 below to locate the following feature:

- POWER INLET PLUGS** Located on the rear panel. Each pump is individually connected to the Power Controller by cables provided.

**Figure 6: 4-Pump Module
Rear View**



4.1.7 Other Optional Modules

Currently, an optional Thermal Mass Flow Controller is available. See Section 14 for all pertinent details (installation, operation, maintenance, troubleshooting and specifications).

4.2 *Stacking Modules*

The Power Controller, the largest and heaviest of the BioFlo 110 modules, is intended to be the base unit of any stack. You may stack any of the other modules on top of it, in any order. Every module has indentations on top to accommodate the feet of any module stacked on it.

Each Power Controller and any of the control modules (dO₂/pH, level, gas mix and/or 4-pump) associated with it should be stacked beside the vessel they support. The modules should be within easy reach of the sensor cables and other leads that will plug into the stack.

If you are running more than one vessel from the PCU, you may wish to place the PCU at the top of the stack most convenient to the operator.

NOTE:

Figure 7 below is an illustration of a sample installation of two vessels with their control modules. For easy access to the touchpad and display screen, and for your working comfort, we suggest that you place the PCU at the top of the stack closest to the operator.


The illustration below is more schematic than realistic.

NOTE:

No more than 16 modules (any combination) can be supported by one PCU.

Figure 7: Sample BioFlo 110 Installation

(For clarity, cables and tubing are not shown in this drawing.)



4.3 Interconnecting the Modules

All BioFlo 110 control modules are interconnected, to share power and communications.

**WARNING!**

Make sure the Power Controller and Primary Control Unit (PCU) are switched off and unplugged from the power source before connecting or disconnecting any equipment.

**NOTE:**


The PCU comes with two Module Bus terminator plugs that cap its 9-pin D-connectors on the rear panel. As you plug RS-485 (ModBus) cables into these connectors, move the terminator plugs to the open end of the cables, to close them off.

To connect the modules to each other, with the modules stacked in the preferred order:

1. Use a short RS-485 communication cable (provided) to connect the Power Controller from its top controller bus 9-pin connector to the bottom controller bus connector of the next module above. When you remove the terminator plug from the connector, set it aside for reuse.
2. In similar fashion, continue to daisy chain each module to the next, using the additional RS-485 cables provided.
3. Cap off the final module's unused controller bus connector with the terminator plug.

You can interconnect as many as three additional module stacks, each with its own Power Controller. Through this interconnection, all stacks will communicate with the PCU, giving it control over all four associated vessels. *See Figure 8 on the following page for a sample installation.*

Figure 8: Interconnecting the Modules



To connect the first stack to another:

1. Use a long RS-485 communication cable. Connect it to either the topmost controller bus on the first stack (i.e., the upper connector on the PCU) or to the bottommost controller bus (i.e., the lower connector on the Power Controller). Save the terminator plug from the controller bus for reuse.
2. Connect the other end to either the topmost or the bottommost controller bus on the next stack.
3. Using short RS-485 cables, continue to daisy chain each module to the next in the second stack, as you did before. Be sure that a terminator plug is on each unused controller bus connector.

Repeat steps 1-3 above for stacks three and four, if appropriate.



CAUTION!

A terminator plug must be fastened on both unused Controller Bus connectors in the system for the communication links to work properly.

4.4 ****Important Warnings****

Before you begin to assemble or operate your vessel, whether it is non-jacketed or water-jacketed, be sure to read this section, for it contains essential information, cautions and warnings to protect your safety and the safety of your equipment.



WARNING!

NEVER PRESSURIZE A GLASS VESSEL!

- Always use eye protection, and exercise caution in the vicinity of glass. If the vessel exhaust becomes blocked, pressure can build up, possible shattering the vessel and endangering personnel.
- As soon as you open the airflow valve(s), verify by feel that air is flowing freely from the exhaust. If not, immediately close the valve(s) or turn off the air/gas supplies.
- Never intentionally block the exhaust to raise vessel pressure.
- Use the minimum air/gas pressure that will provide adequate airflow for the application. Never exceed the maximum air pressure of 10 psi. This maximum pressure is necessary only to obtain the highest gas flow rates. It is capable of breaking the vessel if the vessel is not properly vented.



CAUTION!


To protect the integrity of your glass vessel and to avoid damage, familiarize yourself with these cautions:

- Never allow hot glass to touch cold water or a cold surface.
- Never rest the vessel on an uneven surface.
- Never drag or roll the vessel across any surface.
- Avoid metal-to-glass contact. With the exception of occasional contact with baffles inside a vessel used for fermentation, avoid touching the glass with any metal object.
- Use non-abrasive cleaners only, and clean with soft brushes (no sharp ends or bristles).
- Any surface that comes into contact with any portion of the vessel must be clean and non-abrasive.
- Only finger-tighten the knurled headplate bolts and port adapters. Over-tightening puts undesirable pressure on the glass.
- Keep the glass free from contact with any diamond material (diamond jewelry, industrial diamonds or diamond dust from grinding wheels).

Whenever you assemble or disassemble the vessel components, if you need to lay the drive assembly aside while it is still attached to the headplate and the agitation impeller shaft, note that there is a correct and an incorrect way to position the assembly on a flat surface.


The wrong way, which is resting the headplate and impeller shaft on a surface (see *Figure 9a below*) puts the impeller shaft at risk for damage:

Figure 9a: WRONG Handling of Drive Assembly



The correct way, which is resting the drive assembly and headplate on the surface (see *Figure 9b below*), protects the impeller shaft from bearing weight. Naturally, you will have to take care not to hit the shaft as you work around it

Figure 9b: CORRECT Handling of Drive Assembly




4.5 Vessel Assembly: Non-Jacketed

One of the most versatile features of the BioFlo 110 is the wide variety of glass vessels available. There are two types of vessels, non-jacketed—which are provided with the basic models—and water-jacketed. Each type of vessel is available in four sizes: 1.3 liters, 3.0 liters, 7.5 liters and 14.0 liters (total volume; *for more detail, see Specifications in Appendix A*). **If you have a water-jacketed vessel, skip to Section 0.**

Every single-walled, non-jacketed vessel comes with a stainless steel stand from which the vessel is suspended. The stand has four rubber feet to provide stability. An electric heat blanket provides temperature control for the contents of the vessel. The blanket has two large viewing windows so the culture remains visible for inspection. (*See Figure 10*)

Figure 10 below shows a typical installation of the non-jacketed vessel, in its vessel stand, with the most commonly used accessory equipment. To provide a full view of how the internal components are arranged, the heat blanket is not shown.

Figure 10: Non-Jacketed Vessel Assembly



Familiarize yourself with the arrangement of the headplate ports, as shown in the following diagrams, before proceeding with the vessel assembly. You may find it more practical to change the arrangement; the variety of ports and adapters will easily accommodate your needs.

For easy reference, see Appendix E for a mix and match summary of port sizes and the types of adapters and inserts available.

4.5.1 Headplate

Figure 11: 1.3L Headplate




Figure 12: 3.0L Headplate





Figure 13: 7.5L & 14.0L Headplate



4.5.2 Install Heat Blanket

1. Wrap the heat blanket *as snugly as possible* around the vessel, taking care to leave one of the viewing windows facing forward. You will probably want to orient the blanket so the power cord connection is out of the way.
2. Secure the blanket by overlapping the Velcro strips, and pressing them together.



WARNING!


NEVER cut any portion of the heat blanket.
NEVER fold the heat blanket or place any weight upon it.
 For storage, always lay the heat blanket flat.

4.5.3

Install Vessel in Vessel Stand

1. Place the clamping ring on the vessel stand: align the clamping ring holes with the vessel stand pillars, then slide it into place. It will come solidly to rest on the shoulder of each pillar.
2. Place sections of U-shaped rubber bumper *equidistantly* around the inside of the clamping ring: there are three pieces for 1.3L & 3.0L vessels, and two larger pieces for 7.5L and 14.0L vessels. Press each section securely against the inner edge of the ring.

Figure 14: Upper Vessel Bumper Installation



3. Gently lower the glass vessel through the center of the clamping ring, until the vessel flange rests snugly against the rubber bumpers.
4. Orient the vessel so the gradations on the glass are clearly visible at the front, facing the user, and situated between two vessel stand pillars.

4.5.4

Install Baffle (14.0L Fermentation Vessels ONLY)

For installation of the 1.3L, 3.0L and 7.5L vessel baffle, see Section 4.5.21.

If you are using a 14.0L vessel, install the baffle assembly inside the glass vessel:

1. Gently compress the baffle ring at its ends (to avoid scratching the vessel walls). You may find it convenient to squeeze the tab with your thumb.
2. Slide the assembly inside, with the tab facing up, until it comes to rest at the bottom of the vessel.

3. Orient the baffle so the opening is opposite the gradations on the vessel, and the tab is aligned with the back vessel stand pillar.

4.5.5 Install Impeller(s)

Install the impeller(s) as follows:

- A. **For Cell Culture:** Slide the impeller onto the agitation drive shaft (from the bearing housing). Position the impeller at least 10 mm above the sparger. Clamp it down in place.



NOTE:

It is normal for the agitation impeller shaft to be very resistant to turning by hand. The shaft seal resistance ensures sterile operation.

- B. **For Fermentation:** Slide one impeller onto the agitation drive shaft (from the bearing housing). Position this lower impeller according to the table below. Clamp it down in place. Then install the second (upper) impeller in the same manner.

<i>Distance from Bottom of Headplate to Top of Impeller Blade</i>				
	1.3L	3.0L	7.5L	14.0L
Lower Impeller	4 1/8 in. 105 mm	6 11/16 in. 170 mm	8 7/8 in. 225 mm	12 in. 305 mm
Upper Impeller	2 5/8 in. 67 mm	4 in. 102 mm	6 1/2 in. 165 mm	9 1/4 in. 235 mm



NOTE:

The distances indicated above provide a recommended starting point. As working volumes and agitation rates change, you may wish to adjust the impeller location(s).



NOTE:

It is good practice to lightly lubricate all O-rings, port threads and adapter threads with silicone grease before you install equipment in the headplate. Also inspect the headplate O-ring to be sure it is securely seated in its groove.

4.5.6 Install Cooling Coil

1.3L Vessel Cooling Coil/Sparger Assembly

The cooling coil and sparger connections are welded into one special 12mm tri-port assembly.

1. From beneath the headplate, insert the assembly into the appropriate ports (*see Figure 11*).
2. From above the headplate, lock the assembly in place with a knurled 12mm to 12mm adapter. Finger tighten.
3. There are three set screws in the adapter. If you need to raise or lower the adapter/tri-port assembly, use the Allen key provided to adjust the set screw that is easiest to access. You only need to adjust one.

3.0L, 7.5L & 14.0L Vessel Cooling Coil

1. From beneath the headplate, insert both ends of the coil into the Cooling Coil (In) port and the Cooling Coil (Out) port.
2. From above the headplate, finger tighten the knurled adapter on each side of the cooling coil.

4.5.7 Install Sparger (3.0L, 7.5L & 14.0L Vessels)

1. From *beneath the headplate*, insert the sparger tube into the sparger port (*see Figures 12 & 13 for reference*).
2. Finger tighten the knurled adapter on the sparger, then use the Allen key provided to tighten the set screw. Do not overtighten.



CAUTION!

Finger tighten only any adapter that has a white Teflon ferrule (tapered, cone-shaped insert under the Teflon washer). The ferrule can deform under too much pressure.

4.5.8 Install Harvest Tube

1. Working from beneath the headplate, install the harvest tube in the harvest port (*see Figures 11, 12 & 13 for reference*). If you are using the 1.3L vessel, the harvest tube and sampler tube are welded into the same tri-port to save space. When the headplate is in place on the vessel, the bottom of the harvest tube should rest at the bottom of the vessel.
2. Finger tighten the knurled adapter on the harvest tube, then use the Allen key provided to tighten the set screw. Do not overtighten.

4.5.9 Install Sampler Tube

1. Working from beneath the headplate, install the optional sampler tube in the sample port (*see Figures 11, 12 & 13 for reference*). If you are using the 1.3L vessel, the sampler tube and harvest tube are welded into the same tri-port to save space.
2. Finger tighten the knurled adapter on the sampler tube, then use the Allen key provided to tighten the set screw.

4.5.10 Install Thermowell

1. Working from *above the headplate*, insert the thermowell tube into the RTD port (*see Figures 11, 12 & 13 for reference*).



CAUTION!

Make sure that the thermowell does not touch the cooling coil.

2. Finger tighten the knurled adapter on the thermowell.

4.5.11 Install Foam Probe

If you are using a foam sensor with a foam trap kit:

1. Working from above the headplate, insert the foam sensor into the appropriate port (*see Figures 11, 12 & 13 for reference*).
2. Finger tighten the knurled adapter.

4.5.12 Install Foam Exhaust Tube

If you are using a foam trap, install the foam exhaust tube:

1. Working from beneath the headplate, insert the foam exhaust tube into the appropriate port, close to a headplate clamping nut (*see Figures 11, 12 & 13 for reference*) where you will later mount the foam trap.
2. Finger tighten the knurled adapter. If you need to raise or lower the tube at any time, use the Allen key provided to adjust the adapter's set screw.

4.5.13 Install Level Probe(s)

If you are using a level probe as part of the antifoam system and/or a level probe to detect media level, one at a time:

1. Working from above the headplate, insert the level probe into the appropriate port (*see Figures 11, 12 & 13 for reference*).
2. Finger tighten the knurled adapter.

4.5.14 Install Addition Tube(s)

Insert addition tubes and/or tri-ports in the appropriate ports for any or all of the following additions: media, nutrients, acid, base, antifoam. For each insertion:

1. Finger tighten the knurled addition or tri-port adapter.
2. Working from above the headplate, insert the addition tube or tri-port into the appropriate port (*see Figures 11, 12 & 13 for reference*).

4.5.15 Install pH Probe



NOTE:

Prior to installation, any pH probe you are using should be inspected for damage, and replaced if necessary.



NOTE:

To avoid damage to the probes during operation, be sure that there is no interference between the probes and the baffle assembly, impeller blades, or cooling coil.


1. Wear protective gloves to protect yourself in case of accidental breakage.
2. Lightly coat the pH probe with glycerol.



CAUTION!

Do not install the pH port adaptor in the headplate before inserting the probe. Follow the steps below to fit the pH port adaptor onto the probe first, then insert the probe and adapter into the headplate.

Figure 15: pH Probe with Port Adapter (exploded)



With reference to Figure 15 above:

3. Gently slide the top portion of the knurled port adapter (part of the probe kit) onto the probe.
4. Slide the two white ferrules onto the probe, the narrower one on top of the deeper, cup-shaped one.
5. Gently slide the bottom portion of the port adapter onto the probe, taking care to orient the longer threaded section toward the top of the probe.
6. Remove the two O-rings installed in the pH port; first slide the white Teflon O-ring onto the probe, then follow with the black 12mm port adapter O-ring.
7. Do not yet close up all the elements of the port adapter.
8. Gently insert the probe into the appropriate port (*see Figures 11, 12 & 13 for reference*), allowing the O-rings to seat fully into the port.



NOTE:

The fit may be snug. Gently rotate the probe as you press it into the port to avoid breakage.

9. Finger tighten the bottom portion of the port adapter into the port.
10. Adjust the probe to the desired height; then, nesting the ferrules, close the top portion of the adapter onto the bottom portion.
11. Finger tighten the knurled adapter assembly.

4.5.16 Install dO2 Probe

 NOTE:

Prior to installation, any dissolved oxygen probe you are using should be inspected for damage and replaced if necessary.

 NOTE:

To avoid damage to the probes during operation, be sure that there is no interference between the probes and the baffle assembly, impeller blades or cooling coil.


1. Wear protective gloves to protect yourself in case of accidental breakage.
2. Lightly coat the dO2 probe with glycerol.



CAUTION!

Do not install the dO2 port adaptor in the headplate before inserting the probe. Follow the steps below to fit the dO2 port adaptor onto the probe first, then insert the probe and adapter into the headplate.

Figure 16: dO2 Probe with Port Adapter (exploded)



With reference to Figure 16 above:

3. Gently slide the top portion of the knurled port adapter (part of the probe kit) onto the probe.
4. Slide the two white ferrules onto the probe, the narrower one on top of the deeper, cup-shaped one.
5. Gently slide the bottom portion of the port adapter onto the probe, taking care to orient the longer threaded section toward the top of the probe.
6. Remove the two O-rings installed in the dO2 port; first slide the white Teflon O-ring onto the probe, then follow with the black 12mm port adapter O-ring.
7. Do not yet close up all the elements of the port adapter.
8. Gently insert the probe into the appropriate port (*see Figures 11, 12 & 13 for reference*), allowing the O-rings to seat fully into the port.



NOTE:

The fit may be snug. Gently rotate the probe as you press it into the port to avoid breakage.

9. Finger tighten the bottom portion of the port adapter into the port.
10. Adjust the probe to the desired height; then, nesting the ferrules, close the top portion of the adapter onto the bottom portion.
11. Finger tighten the knurled adapter assembly.

4.5.17

Install Exhaust Condenser



WARNING!

Never intentionally block the exhaust to raise vessel pressure.

If you are using the optional exhaust condenser:


1. Unscrew the spare/exhaust port plug from the headplate, saving it for reuse.
2. Place the 12mm exhaust condenser adapter into the port.
3. Place the exhaust condenser inlet (*see Figures 18a & 18b below*) into the port, and finger tighten the knurled adapter.
4. Tighten it with the Allen key provided, until it is secure.
5. Attach the exhaust filter (respecting the direction of flow if stamped on the filter) to the condenser outlet. Secure the filter with a plastic tie.
6. Connect silicone tubing to the inlet port of the exhaust condenser. Secure with a plastic tie.
7. Connect the other end of the tubing to the water inlet (*see NOTE below and Figure 17 on the following page*), split from the water solenoid valve on the side of the Power Controller. Secure with a plastic tie.

NOTE:


The water solenoid valve's brass body has two hose barbs, identified as "1" and "2" by numbers stamped on the body, just above the barb. Hose barb 1 is the water inlet. Hose barb 2 is the water outlet. The valve body swivels to facilitate access.

8. Connect silicone tubing to the top outlet port of the exhaust condenser. Allow enough tubing to run to the water drain. Secure the tubing with a plastic tie.

Figure 17: Water Solenoid Valve




**Figure 18a: Exhaust Condenser
(1.3L, 3.0L & 7.5L Vessels)**



**NOTE:**

If the weight of the exhaust filter kinks the tubing, fasten a short length of stiffening material to the tubing, using rubber bands or tie wraps, to support the filter.

**Figure 18b: Exhaust Condenser
(14.0L Vessel Only)**



4.5.18 Install Sampler

The optional BioFlo 110 sampler system is designed to aseptically remove batch samples from the vessel. The entire installation is easily autoclaved in place on the vessel. If you are using the sampler, install the kit as follows, using Figures 19a & 19b below for reference:

1. Remove a headplate clamping nut adjacent to the location of the sampler tube.
2. Mount the metal sampler bottle holder arm on the clamping screw, and secure it in place with the clamping nut.

Figure 19a: Sampler/Harvest System (1.3L Vessel)





Figure 19b: Sampler System (3.0L, 7.5L & 14.0L Vessels)



3. Connect a length of silicone tubing to the sampler tube on the headplate. Secure it in place with a plastic tie.
4. Slip a thumb clamp onto the tubing.
5. Connect the other end of the tubing to the tall sampler inlet pipe. Secure it in place with a plastic tie.
6. Connect a short length of silicone tubing to the short sampler outlet pipe. Secure it in place with a plastic tie.
7. Connect the sterile syringe filter to the other end of the tubing, taking care to respect the direction of flow if stamped on the filter. Secure the tubing in place with a plastic tie.
8. Insert the tip of the sampler syringe as far as it will go into the open end of the filter. Although the syringe will lodge there and hang freely in place, you can add a plastic tie for security.
9. Close the plunger.
10. Remove the cap from one of the sample bottles and screw the bottle into the metal holder.
11. Position the entire assembly to your satisfaction, then finger tighten the clamping nut.


4.5.19 Install Foam Trap

If you are using a foam trap kit (see Figure 20 below):

1. Unscrew the headplate clamping nut (or base clamping nut, if you prefer to mount the trap at the base of the vessel) closest to the foam exhaust tube.
2. Mount the foam trap bottle holder on the clamping screw, using the hole at the end of the holder's mounting arm.
3. Secure the holder in place with the clamping nut. Leave the nut loose enough to swivel the holder.
4. Firmly place the foam trap bottle (250 ml or 500 ml) in the holder.

5. With the bottle cap in place, aseptically install a sterile (0.2 μ) filter on the shorter tube that penetrates the cap. Be sure to respect the proper flow direction if stamped on the filter.
6. Connect a length of silicone tubing to the longer tube in the other bottle cap penetration. Secure the tubing with a plastic tie, and clamp it off on the top.
7. Connect the tubing, securing it with a plastic tie, to the foam exhaust tube in the headplate.
8. After autoclaving, you will position the bottle holder where you want it, then finger tighten the clamping nut.

Figure 20: Foam Trap



4.5.20 Plug Unused Ports

Close off unused ports:

1. Install a blind plug (without a hole) in any headplate port that will not be used.
2. Install silicone tubing, secured with a plastic tie and clamped shut, on any access tube (i.e., harvest tube) that will not immediately be used.

**NOTE:**

It is good practice to lightly lubricate the underside of the headplate with silicone before installing it on the vessel.

4.5.21 Install 1.3L, 3.0L or 7.5L Fermentation Vessel Baffle

1. Gently place the baffle, tab facing up, around all of the other instruments protruding from the headplate, including the cooling coil.
2. Position the tab between the two uprights of the cooling coil.

Hold the baffle in place with two fingers when you lift the headplate assembly.

4.5.22 Install Headplate

1. Orient the cooling coil uprights toward the back, opposite the gradations marked on the vessel glass. If this is a 1.3L, 3.0L or 7.5L fermentation vessel, squeeze and hold the baffle in place (opening toward the back) with thumb and forefinger. You may find it convenient to squeeze the tab with your thumb.
2. Carefully lower the headplate, easing all of its attachments into the vessel without hitting the glass (or the baffle inside, if this is a 14.0L fermentation vessel).

**NOTE:**

If you are using a baffle, after installing the headplate, insert any convenient length of wood, plastic or stainless steel (do not use any other kind of metal) through an unused port to push the baffle down as far as it will go.

The baffle is stainless steel; repeated installations may cause it to retain a compressed position. Expand it before you squeeze it for installation, so it will spring back against the vessel walls.

3. Align the headplate holes with the vessel stand pillars, then slide it down until it rests securely against the vessel flange.
4. Finger tighten each clamping nut a little at a time to secure the headplate on the vessel stand, working diagonally from one to another (rather than working around the circle) to apply equal pressure.

*See important **NOTE** on following page.*

**CAUTION!**

To avoid vessel stress cracks, especially during autoclaving, never overtighten vessel clamping nuts.


4.6 Vessel Assembly: Water-Jacketed

One of the most versatile features of the BioFlo 110 is the wide variety of glass vessels available. There are two types of vessels, non-jacketed—which are provided with the basic models—and water-jacketed. Each type of vessel is available in four sizes: 1.3 liters, 3.0 liters, 7.5 liters and 14.0 liters (maximum working volume; *for more detail, see Specifications in Appendix A*). **If you have a non-jacketed vessel, and have already completed Section 4.5 above, skip down to Section 4.7.**

Water-jacketed vessels need no stand; the water jacket, which is part of the vessel, is flared and flat at the bottom to provide secure, stable support. At the bottom is a metal base plate, to provide additional security against breakage. In operation, the jacketed vessel sits on the Jacket Water Heater. The jacket water heater is designed so that the vessel water inlet and outlet fit in a notch at the rear, and the vessel feet fit into the four holes at the perimeter of the heater plate.

Figure 21 on the following page shows a typical installation of the double-walled, water-jacketed vessel, with the most commonly used accessory equipment.

Figure 21: Water-Jacketed Vessel Assembly



Familiarize yourself with the arrangement of the headplate ports, as shown in the following diagrams, before proceeding with the vessel assembly. You may find it more practical to change the arrangement; the variety of ports and adapters will easily accommodate your needs.

For easy reference, see Appendix E for a mix and match summary of port sizes and the types of adapters and inserts available.

4.6.1 Headplate

Figure 22: 1.3L Headplate




Figure 23: 3.0L Headplate





Figure 24: 7.5L & 14.0L Headplate




4.6.2 Install Headplate Clamping Ring

The clamping ring that secures the headplate to the vessel is split in half to facilitate installation under the vessel flange. They are joined with two rectangular mounting plates.

1. As shown in Figure 25 below, install one mounting plate with two Phillips head screws (provided) on the end of one ring half so that the plate extends beyond the ring.

Figure 25: Installing Headplate Clamping Ring




2. In the same manner, install the second mounting plate on the other end of the ring half.
3. Bring the two halves of the headplate clamping ring together under the vessel flange, with the mounting plates on the bottom for easy access from below.
4. Align the mounting plates with their corresponding holes on the other ring half, and drop in the remaining Phillips head screws. Tighten the screws to fasten the ring in place.

4.6.3 Install Vessel on Base Plate

1. Place the base plate on a level surface.
2. Lightly lubricate the base plate O-ring, and seat it securely in its groove.
3. Fit the one-piece water jacket guard (rubber gasket) around the outside of the bottom vessel flange, against the water jacket (*see Figure 26 on the following page*).
4. With the clamping screws in place on the ring, fit the bottom clamping ring onto the base plate.

Figure 26: Water Jacket Guard Installation (top view)



5. With the gradations marked on the glass facing front (toward the user), slide the vessel into the bottom clamping ring, until it rests securely against the base plate. **Make sure the water inlet tube stands free** (not kinked) inside the water jacket.
6. Finger tighten the six knurled thumb screws, to securely attach the clamping ring to the base plate. This seals the water jacket.

4.6.4 Install Baffle (14.0L Fermentation Vessels ONLY)

For installation of the 1.3L, 3.0L & 7.5L vessel baffle, see Section 4.6.21.

If you are using a 14.0L vessel, install the baffle assembly inside the glass vessel:

1. Gently compress the baffle ring at its ends (to avoid scratching the vessel walls). You may find it convenient to squeeze the tab with your thumb.
2. Slide the assembly inside, with the tab facing up, until it comes to rest at the bottom of the vessel.
3. Orient the baffle so the opening is opposite the gradations on the vessel.

4.6.5 Install Impeller(s)

Install the impeller(s) as follows:

- A. **For Cell Culture:** Slide the impeller onto the agitation drive shaft (from the bearing housing). Position the impeller at least 10 mm above the sparger. Clamp it down in place.

**NOTE:**

It is normal for the agitation impeller shaft to be very resistant to turning by hand. This resistance ensures sterile operation.

- B. **For Fermentation:** Slide one impeller onto the agitation drive shaft (from the bearing housing). Position this lower impeller *according to the table below*. Clamp it down in place. Then install the second (upper) impeller in the same manner.

<i>Distance from Bottom of Headplate to Top of Impeller Blade</i>				
	1.3L	3.0L	7.5L	14.0L
Lower Impeller	4 1/8 in. 105 mm	6 11/16 in. 170 mm	8 7/8 in. 225 mm	12 in. 305 mm
Upper Impeller	2 5/8 in. 67 mm	4 in. 102 mm	6 1/2 in. 165 mm	9 1/4 in. 235 mm

**NOTE:**

The distances indicated above provide a recommended starting point. As working volumes and agitation rates change, you may wish to adjust the impeller location(s).

**NOTE:**

It is good practice to lightly lubricate all O-rings, port threads and adapter threads with silicone grease before you install equipment in the headplate. Also inspect the headplate O-ring to be sure it is securely seated in its groove.

4.6.6**Install Sparger**

1. From *beneath the headplate*, insert the sparger tube into the sparger port (see *Figures 22, 23 & 24 for reference*).
2. Finger tighten the knurled adapter on the sparger, then use the Allen key provided to tighten the set screw. Do not overtighten.

**CAUTION!**

Finger tighten only any adapter that has a white Teflon ferrule (tapered, cone-shaped insert under the Teflon washer). The ferrule can deform under too much pressure.

4.6.7 Install Harvest Tube

1. Working from beneath the headplate, install the harvest tube in the harvest port (*see Figures 22, 23 & 24 for reference*). If you are using the 1.3L vessel, the harvest tube and sampler tube are welded into the same tri-port to save space. When the headplate is in place on the vessel, the bottom of the harvest tube should rest at the bottom of the vessel.
2. Finger tighten the knurled adapter on the harvest tube, then use the Allen key provided to tighten the set screw. Do not overtighten.

4.6.8 Install Sampler Tube

1. Working from beneath the headplate, install the optional sampler tube in the sample port (*see Figures 22, 23 & 24 for reference*). If you are using the 1.3L vessel, the sampler tube and harvest tube are welded into the same tri-port to save space.
2. Finger tighten the knurled adapter on the sampler tube, then use the Allen key provided to tighten the set screw.

4.6.9 Install Thermowell

1. Working from *above the headplate*, insert the thermowell tube into the RTD port (*see Figures 22, 23 & 24 for reference*).



CAUTION!

Make sure that the thermowell does not touch the cooling coil.

2. Finger tighten the knurled adapter on the thermowell.

4.6.10 Install Foam Probe

If you are using a foam sensor with a foam trap kit:

1. Working from above the headplate, insert the foam sensor into the appropriate port (*see Figures 22, 23 & 24 for reference*).
2. Finger tighten the knurled adapter.

4.6.11 Install Foam Exhaust Tube

If you are using a foam trap, install the foam exhaust tube:

1. Working from beneath the headplate, insert the foam exhaust tube into the appropriate port, close to a headplate clamping nut (*see Figures 22, 23 & 24 for reference*) where you will later mount the foam trap.

Finger tighten the knurled adapter. If you need to raise or lower the tube at any time, use the Allen key provided to adjust the adapter's set screw.

4.6.12 Install Level Probe(s).

If you are using a level probe as part of the antifoam system and/or a level probe to detect media level, one at a time:

1. Working from above the headplate, insert the level probe into the appropriate port (*see Figures 22, 23 & 24 for reference*).
2. Finger tighten the knurled adapter.

4.6.13 Install Addition Tube(s)

Insert addition tubes and/or tri-ports in the appropriate ports for any or all of the following additions: media, nutrients, acid, base, antifoam. For each insertion:

1. Working from above the headplate, insert the addition tube or tri-port into the appropriate port (*see Figures 20, 21 & 22 for reference*).
2. Finger tighten the knurled addition or tri-port adapter.

4.6.14 Install pH Probe



NOTE:

Prior to installation, any pH probe you are using should be inspected for damage and replaced if necessary.



NOTE:


To avoid damage to the probes during operation, be sure that there is no interference between the probes and the baffle assembly or impeller blades.

1. Wear protective gloves to protect yourself in case of accidental breakage.
2. Lightly coat the pH probe with glycerol.

**CAUTION!**

Do not install the pH port adaptor in the headplate before inserting the probe. Follow the steps below to fit the pH port adaptor onto the probe first, then insert the probe and adapter into the headplate.

Figure 27: pH Probe with Port Adapter (exploded)



With reference to Figure 27 above:

3. Gently slide the top portion of the knurled port adapter (part of the probe kit) onto the probe
4. Slide the two white ferrules onto the probe, the narrower one on top of the deeper, cup-shaped one.
5. Gently slide the bottom portion of the port adapter onto the probe, taking care to orient the longer threaded section toward the top of the probe.
6. Remove the two O-rings installed in the pH port; first slide the white Teflon O-ring onto the probe, then follow with the black 12mm port adapter O-ring.
7. Do not yet close up all the elements of the port adapter.
8. Gently insert the probe into the appropriate port (*see Figures 22, 23 & 24 for reference*), allowing the O-rings to seat fully into the port.

**NOTE:**

The fit may be snug. Gently rotate the probe as you press it into the port to avoid breakage.

9. Finger tighten the bottom portion of the port adapter into the port.

10. Adjust the probe to the desired height; then, nesting the ferrules, close the top portion of the adapter onto the bottom portion
11. Finger tighten the knurled adapter assembly.

4.6.15 Install Dissolved Oxygen Probe

**NOTE:**

Prior to installation, any dissolved oxygen (dO₂) probe you are using should be inspected for damage and replaced if necessary.

**NOTE:**


To avoid damage to the probes during operation, be sure that there is no interference between the probes and the baffle assembly or impeller blades.

1. Wear protective gloves to protect yourself in case of accidental breakage.
2. Lightly coat the dO₂ probe with glycerol.

**CAUTION!**

Do not install the dO₂ port adaptor in the headplate before inserting the probe. Follow the steps below to fit the dO₂ port adaptor onto the probe first, then insert the probe and adapter into the headplate.

Figure 28: dO₂ Probe with Port Adapter (exploded)



With reference to Figure 28 above:

3. Gently slide the top portion of the knurled port adapter (part of the probe kit) onto the probe.
4. Slide the two white ferrules onto the probe, the narrower one on top of the deeper, cup-shaped one.
5. Gently slide the bottom portion of the port adapter onto the probe, taking care to orient the longer threaded section toward the top of the probe.
6. Remove the two O-rings installed in the dO₂ port; first slide the white Teflon O-ring onto the probe, then follow with the black 12mm port adapter O-ring.
7. Do not yet close up all the elements of the port adapter.
8. Gently insert the probe into the appropriate port (*see Figures Figures 22, 23 & 24 for reference*), allowing the O-rings to seat fully into the port. If you are using an Ingold probe in a 1.3L, 3.0L or 7.5L vessel, be sure to use the seal washer provided.



NOTE:

The fit may be snug. Gently rotate the probe as you press it into the port to avoid breakage.

9. Finger tighten the bottom portion of the port adapter into the port.
10. Adjust the probe to the desired height; then, nesting the ferrules, close the top portion of the adapter onto the bottom portion.
11. Finger tighten the knurled adapter assembly.

4.6.16

Install Exhaust Condenser



WARNING!

Never intentionally block the exhaust to raise vessel pressure.

If you are using the optional exhaust condenser:

1. Unscrew the spare/exhaust port plug from the headplate, saving it for reuse.
2. Place the 12mm exhaust condenser adapter into the port.
3. Place the exhaust condenser inlet (*see Figures 30a & 30b below*) into the port, and finger tighten the knurled adapter.
4. Tighten it with the Allen key provided, until it is secure.

5. Attach the exhaust filter (respecting the direction of flow if stamped on the filter) to the condenser outlet. Secure the filter with a plastic tie.
6. *With reference to Figures 30a and 30b*, connect a length of silicone tubing between the lower (cooling water inlet) port of the exhaust condenser and the T-connector (upstream of the water solenoid valve). Secure at both ends with a plastic tie.
7. Connect the T-connector inlet (see Figure 30b) to hose barb 2 (see **NOTE** and Figure 29 below) on the water solenoid valve. The valve is on the side of the Power Controller. Secure with a plastic tie.

**NOTE:**

The water solenoid valve's brass body has two hose barbs, identified as "1" and "2" by numbers stamped on the body, just above the barb. Hose barb 1 is the water inlet. Hose barb 2 is the water outlet. The valve body swivels to facilitate access.

8. Connect silicone tubing to the top (cooling water outlet) port of the exhaust condenser. Allow enough tubing to run to the water drain. Secure the tubing with a plastic tie.

Figure 29: Water Solenoid Valve






Figure 30a: Exhaust Condenser (1.3L, 3.0L & 7.5L Vessels)



NOTE:

If the weight of the exhaust filter kinks the tubing, fasten a short length of stiffening material to the tubing, using rubber bands or tie wraps, to support the filter.

Figure 30b: Exhaust Condenser (14.0L Vessel Only)



4.6.17 Install Sampler

The optional BioFlo 110 sampler system is designed to aseptically remove batch samples from the vessel. The entire installation is easily autoclaved in place on the vessel. If you are using the sampler, install the kit as follows, using Figures 31a and 31b for reference:

1. Remove a headplate clamping nut adjacent to the location of the sampler tube.
2. Mount the metal sampler bottle holder on the clamping screw, and secure it in place with the clamping nut.

Figure 31a: Sampler/Harvest System (1.3L Vessel)





Figure 31b: Sampler System (3.0L, 7.5L & 14.0L Vessels)




3. Connect a length of silicone tubing to the sampler tube outlet on the headplate. Secure it in place with a plastic tie.
4. Slip a thumb clamp onto the tubing.
5. Connect the other end of the tubing to the tall sampler inlet pipe. Secure it in place with a plastic tie.
6. Connect a short length of silicone tubing to the short sampler outlet pipe. Secure it in place with a plastic tie.
7. Connect the sterile syringe filter to the other end of the tubing, taking care to respect the direction of flow if stamped on the filter. Secure the tubing in place with a plastic tie.
8. Insert the tip of the sampler syringe as far as it will go into the open end of the filter. Although the syringe will lodge there and hang freely in place, you can add a plastic tie for security. Close the plunger.
9. Remove the cap from one of the sample bottles and screw the bottle into the metal holder.
10. Position the entire assembly to your satisfaction, then finger tighten the clamping nut.

4.6.18 Install Foam Trap

If you are using a foam trap kit (*see Figure 32 below*):

1. Unscrew the headplate clamping nut (or base clamping nut, if you prefer to mount the trap at the base of the vessel) closest to the foam exhaust tube.
2. Place a washer on the clamping screw, then mount the foam trap bottle holder on the screw, using the hole at the end of the holder's mounting arm.
3. Secure the holder in place with the clamping nut. Leave the nut loose enough to swivel the holder.
4. Firmly place the foam trap bottle (250 ml or 500 ml) in the holder.
5. With the bottle cap in place, aseptically install a sterile (0.2 μ) filter on the shorter tube that penetrates the cap. Be sure to respect the proper flow direction if stamped on the filter.

Figure 32: Foam Trap



6. Connect a length of silicone tubing to the longer tube in the other bottle cap penetration. Secure the tubing with a plastic tie, and clamp it off on the top.

7. Connect the tubing, securing it with a plastic tie, to the foam exhaust tube in the headplate.
8. After autoclaving, you will position the bottle holder where you want it, then finger tighten the clamping nut.

4.6.19 Install Antifoam System

If you are using a level probe to activate the addition of chemical defoamer, install the antifoam addition bottle following the procedure for installing a foam trap, with the following differences:

- Install the bottle holder close to the antifoam addition tube in the headplate.
- Connect the longer tube that penetrates the bottle cap to the antifoam addition tube in the headplate.

4.6.20 Plug Unused Ports

Close off all unused ports:

1. Install a blind plug (without a hole) in any headplate port that will not be used.
2. Install silicone tubing, secured with a plastic tie and clamped shut, on any access tube (i.e., harvest tube) that will not immediately be used.



NOTE:

It is good practice to lightly lubricate the headplate O-ring with silicone before installing it on the vessel.

4.6.21 Install 1.3L, 3.0L or 7.5L Fermentation Vessel Baffle

1. Gently place the baffle, tab facing up, around all of the other instruments protruding from the headplate.
2. Position the tab and baffle opening toward the back of the headplate.



CAUTION!

Hold the baffle in place with two fingers when you lift the headplate assembly.

4.6.22 Install Headplate

1. If this is a 1.3L, 3.0L or 7.5L fermentation vessel, be sure to squeeze and hold the baffle in place (opening toward the back) with thumb and forefinger. You may find it convenient to squeeze the tab with your thumb.
2. Carefully lower the headplate, easing all of its attachments into the vessel without hitting the glass walls (or the baffle inside if this is a 14.0L fermentation vessel).

**NOTE:**

If you are using a baffle, after installing the headplate, as indicated below, insert any convenient length of wood, plastic or stainless steel (do not use any other kind of metal) through open ports to push the baffle down as far as it will go.

The baffle is stainless steel; repeated installations may cause it to retain a compressed position. Expand it before you squeeze it for installation, so it will spring back against the vessel walls.

3. Align the headplate holes with clamping ring holes, then slide it down until it rests securely against the vessel flange.
4. Finger tighten each clamping screw a little at a time to secure the headplate, working diagonally from one to another (rather than working around the circle) to apply equal pressure.


**CAUTION!**

To avoid vessel stress cracks, especially during autoclaving, never overtighten vessel clamping screws.

4.7 Install the Agitation Motor

1. Make sure the motor coupling O-ring is lubricated.
2. Position the motor sleeve on the bearing housing (*see Figure 33*), taking particular care to align the side slots with the shoulder studs as you ease the sleeve down until it rests securely on the bearing housing. You may have to walk it down by rotating it slightly left and right.

Figure 33: Motor Installation

**WARNING!**

Never remove or install the motor while it is running.

**NOTE:**

Periodically check the condition of the plastic cushioning (#P0740-3111) as you remove or reinstall the motor. Replace it when it begins to wear.


4.8 Install the Rotameter

**CAUTION!**

Verify that the source of air is regulated at 10 PSI MAXIMUM.

1. Hang the rotameter on the left or right side of the Power Controller or PCU, whichever is most convenient. Using a Phillips head (+) screwdriver, mount the meter bracket using the screws provided.
2. Use silicone tubing to connect the air source to the bottom air inlet hose barb on the rotameter (see Figure 34).

Figure 34: Rotameter—Side View



3. Attach another piece of silicone tubing to the top air outlet hose barb (see Figure 34 above). Insert the other end of this tubing into the inlet sterile filter on the top of the vessel headplate.
4. Verify that the exhaust port is connected to a filter. If you are using the optional exhaust condenser, the filter should be attached to its outlet.
5. Read the airflow at the top flat edge of the float.

To increase airflow, turn the adjusting knob counterclockwise. To decrease airflow, turn it clockwise.

Rotameter Maximum Flow Rate				
Application	Vessel Size			
	1L	3.0L	7.5L	14.0L
Fermentation	2.5 Lpm	5 Lpm	11.4 Lpm	20 Lpm
Cell Culture	0.5 Lpm	1.2 Lpm	2.5 Lpm	5 Lpm

NOTE:

Adjust the airflow according to your specific application. A commonly suggested airflow for a typical culture is 1 VVM (vessel volume per minute) or less for fermentation, and 0.05 VVM or less for cell culture; unnecessarily high airflow can induce foaming.

4.9 Install Liquid Addition Systems

Addition ports are used to feed base, acid, nutrients and media into the vessel. To install an addition system:

1. Install a sterile (0.2 μ) filter on one of the two penetrations on the addition bottle cap. Be sure to respect the proper flow direction if stamped on the filter.
2. Connect a length of silicone tubing, securing it with a plastic tie, to the addition tube in the other penetration of the bottle cap. Clamp it off on the top.
3. **After autoclaving**, thread the tubing through the desired peristaltic feed pump.
4. Aseptically connect the tubing, securing it with a plastic tie, to the appropriate addition port on the headplate.
5. Remove the clamp.

4.10 Making Connections

4.10.1 Power

Confirm that the voltage selection of your BioFlo 110 matches the voltage rating of your electrical service by reading the voltage selection in the window of the fuse drawer (*see Figure 45*). Also confirm that the power requirements of all control modules (marked on each module's serial number sticker) are met by your electrical service.

Plug in the Power Controller: connect power cord P0720-2053 to the Power Controller's power inlet. Do not plug into your electrical service until you are ready to run the unit.

Plug in the Primary Control Unit (PCU):

1. **To support a single vessel assembly**, connect power cable P0720-2054 to the Power Controller's rear panel power outlet, labeled "PCU", and plug the other end into the PCU's rear panel power inlet.
2. **To support multiple vessel assemblies** (with multiple Power Controllers), either:
 - (a) follow step 1 above, choosing the most convenient Power Controller, or
 - (b) use the supplied power cord to plug the PCU directly into your electrical supply, thus making it independent of the Power Controllers.

4.10.2 Water

The BioFlo 110 water solenoid valve is located on the side of the Power Controller. Water pressure should be 10 PSIG maximum, with 50 μ m filtration.

Be sure to secure all tubing connections for both the water and the drain lines with the fiberglass-reinforced hose clamps provided. Attach the ¼ inch ID braided PVC tubing provided (see **NOTE** below) to the water solenoid valve outlet (look for a **2** stamped in the brass valve body, just above the hose barb—see *Figure 17 or 29*), and to the vessel cooling coil inlet (for non-jacketed vessels) or the vessel water jacket inlet.

 NOTE:

There is a distinct, visible difference between this PVC tubing and the silicone tubing used elsewhere; be careful not to confuse them. The PVC tubing is clear, with white cross-hatched reinforcement braiding. Silicone tubing is milky.

PVC tubing is NOT autoclavable and is not for use with hot liquids; it will melt.

1. Attach sufficient PVC tubing to the cooling coil outlet (for non-jacketed vessels) or the vessel water jacket outlet to reach the drain.
2. Connect to the water supply:
 - **If you are using the optional exhaust condenser**, connect tubing from the valve splitter outlet (marked **2** on the brass valve body) to the exhaust condenser water inlet. Connect pressure-regulated clean water to the water splitter inlet (marked **1** on the brass valve body).
 - **If you are not using the exhaust condenser**, connect pressure-regulated clean water directly to the inlet of the water solenoid valve (marked **1** on the brass valve body).
3. If you are using the optional Water Filter/Regulator kit (see *Section 10.3*), install it according to the kit instructions.

4.10.3 Probes and Sensors


Power Controller

Familiarize yourself with the connections and controls on the front panel of the Power Controller, as illustrated in Figure 35 below:

- **POWER SWITCH** On/Off rocker switch.
- **OPERATING LIGHT** Lights when the Power Controller is on.
- **AGITATION CONNECTOR** Connection for agitation motor.

- **TEMPERATURE** Connection for RTD temperature probe, to sense temperature inside the vessel.
- **HEATER CONNECTOR** Connection for the heat blanket (non-jacketed vessels) or for the jacket water heater (water-jacketed vessels).

Figure 35: Power Controller Front Panel



Note the following features and connections on the side of the Power Controller:

- **ROTAMETER** Can be mounted on either side. Manually adjusted rotameter regulates the flow of air/gases into the vessel.
- **WATER INLET & OUTLET** Via the water solenoid valve, located on the side panel, which provides the connection between the water supply and the vessel water inlet (either cooling coil or water jacket). Inlet is marked 1 and outlet 2.

- **SPARGE AIR INLET** Via the rotameter mounted on the side panel or the Thermal Mass Flow Controller, either of which provides the connection between the air/gas supply and the sparger.

Attach all cables to their appropriate probes and sensors, then attach the cables to the appropriate module (not necessarily in this order):

Into the front panel of the Power Controller:

- Plug the RTD (temperature probe) cable where it is marked “Temperature”.
- Plug the agitation motor cable where it is marked “Agitation”.



WARNING!

Agitation and Heater connection sites are high voltage. Never insert anything but the appropriate cable plug.


You will also plug into the front panel, when you are ready to begin operation:

- For non-jacketed vessels: the heat blanket cable where it is marked “Heater”.
- For water-jacketed vessels: the jacket water heater cable where it is marked “Heater”, **but only after** the jacket is filled with water, the vessel is filled with media, and the vessel is properly installed on the heater.

dO₂/pH Controller

If you are using the BioFlo 110 dO₂/pH Controller, familiarize yourself with the connections and controls on the front panel, as illustrated in Figure 36 below:

Figure 36: dO₂/pH Controller – Front Panel




Plug the following into the front panel of the dO2/pH Controller:

- The dissolved oxygen probe cable (“dO2”).
- The pH probe cable (right connector marked “pH”), and the pH probe ground (left connector marked “pH”). Attach the ground clip to the top of a headplate screw.

Level Controller

If you are using the BioFlo 110 Level Controller, familiarize yourself with the connections and controls on the front panel, as illustrated in Figure 37 below:

Figure 37: Level Controller – Front Panel



Plug the following into the front panel of the Level Controller:


- Foam probe (left connector marked “Foam”). When this probe gets wet, its associated pump turns *on*.
- Wet On level sensor (middle connector marked “Wet On”). When this probe gets wet, its associated pump turns *on*.
- Wet Off level sensor (right connector marked “Wet Off”). When this probe gets wet, its associated pump turns *off*.

4.10.4 Pumps

If you are using the BioFlo 110 4-Pump Module, familiarize yourself with these features (*see Figure 38 below*):

- **ON/OFF SWITCH** An individual power switch for each pump is located on the front panel.
- **PUMP DOOR** Behind each hinged door is access to a peristaltic pump.

**Figure 38: 4-Pump Module
Front View**



Place the pump module adjacent to the vessel and its Power Controller. Install the tubing according to the instructions in the pump manual. You may want to consult the Pump Flow Rates listed in Section 11.2.

The Power Controller also provides power to the peristaltic pumps. Plug the pumps into the back panel of the Power Controller (see Figure 2):

1. The pumps used for pH control have dedicated sockets labeled “Acid” and “Base”.
2. Other pumps can be plugged into the sockets labeled “Pump A”, “Pump B” and “Pump C”.



CAUTION!

**Plug the pump cables ONLY into power controller sockets marked “Pump”.
DO NOT use the socket marked “PCU”.**

For further details, see Section 7.12, *Setting the Pumps*.

4.10.5 Air and Optional Gases

All gases should be medical grade. Gas pressure should not exceed 10 PSIG. Be sure to secure all tubing connections with plastic ties.

If you are using the optional Air Filter/Regulator kit (see Section 10.3), install it according to the kit instructions.

Without a 4-Gas controller:


1. Connect pressure-regulated clean air to the inlet port of the rotameter using 5/16-inch (7.9 mm) OD silicone tubing, NBS #P0740-2505.
2. Connect the outlet of the rotameter to a sterile filter, taking care to observe the flow direction requirements if stamped on the filter.
3. Connect the filter outlet to the sparger inlet.

With the Gas Mix Controller:

Familiarize yourself with the front panel of the controller and its features:

- **GAS INLETS** Four push-in inlets for delivery of Oxygen, Carbon Dioxide, Air and Nitrogen from your source.
- **GAS STREAM OUTLET** One push-in outlet for sequential output of various individual gases, as controlled by the Gas Mix Controller, and for delivery to the rotameter or Thermal Mass Flow Controller (TMFC) inlet. The rotameter or TMFC outlet delivers air/gases to the sparger via the vessel headplate.

Figure 39: Gas Mix Controller Front Panel



1. Attach pressure-regulated clean air and other gases to the push-in inlets of the 4-Gas controller using the blue polyurethane tubing supplied (NBS #P0470-3111, 0.125 inch ID/0.250 inch OD).
2. Connect the gas stream outlet port of the 4-Gas controller to the inlet of the rotameter.
3. Connect the outlet of the rotameter to a sterile filter, taking care to observe the flow direction requirements if stamped on the filter.
4. Connect the filter outlet to the sparger inlet.

4.11 Install Cable Guides

Bolt-on cable guide clamps are provided to dress probe and sensor cables from the front of their control modules toward the back of the stack, where they naturally hang from the top of the vessel. Their use is optional. You may prefer to tuck the cables under their module, directing them toward the back of the stack, or to simply allow them to hang.

The cable guides can be mounted on either side of each module, wherever they will best suit your needs. They can also be mounted with the top jaw facing up or down. The two jaws of the guide snap together, so the tip of the top jaw zips along the tracks on the bottom jaw to expand or contract the cable enclosure. The guide functions rather like a plastic tie, with the difference that it can be opened and closed repeatedly. As shown in the drawing below, to open a clamp:


1. Place your thumb and forefinger on either side of the top jaw, at the boss.
2. Grip tightly and pull up.



NOTE:

Because the guide clamps down tightly, you may have to encourage unclamping with the tip of a screwdriver under the leading edge. Do not use undue force, however, to avoid damage to the plastic clamp or, if already present, to the cable inside.

Figure 40: Cable Guide



4.12 Voltage Selection and Fusing



NOTE:


If you move your BioFlo 110 to a new location that requires a change in voltage and/or fusing, consult your NBS representative.

It is possible to change the voltage on your BioFlo 110. It is also possible to change the fusing arrangement from a single fuse to a double fuse. These procedures, however, should only be performed by a qualified technician, referring to Appendix B in this manual.

4.13 Calibrating the pH Probe

If you are using a pH probe, be sure to calibrate it before autoclaving the vessel assembly:








1. Check that the Power Controller and PCU are plugged in and powered on.
2. Make sure that the pH electrode is plugged into the pH connector on the dO2/pH Controller front panel, using the appropriate cable. Also check that the pH ground wire is plugged into the adjacent ground connector, and the ground clip is grounded.
3. When the PCU is powered on, the display says: New Brunswick Scientific BioFlo 110 Scanning Hardware
4. When the system detects the hardware, the display shows the following screen:



The default language setting is *English*. If you prefer to use another screen language, press the selector button next to your preferred language.


Section 5, *Before First Use*, and Section 7, *Operation*, explain the Application (*Use*) settings and the choices offered by the *More...* selection. They also explain how to use all other screens and available settings.

5. Press the  button to open the Main Screen (see sample screen below):

					Loop Control Mode
Loop Name	Current Value	Loop Setpoint			
Name	Value	Setpoint	Control		
Temp	23.4	30.0 °C	Auto		
Agit	134	350 rpm	Auto		
Pump A	0.0	0.0 %	Manual		
Pump B	0.0	0.0 %	Manual		
Pump C	0.0	0.0 %	Manual		
pH	6.68	7.00 pH	Auto		
dO2	77.1	0.0 %	Auto		
					

The above screen is only one sample. The loop names will reflect the specific modules in your system.

Press the selector button in the *pH* row. The pH Loop Details screen opens:



6. In the pH Loop Details screen, press the *Calibrate* selector button. The Calibrate pH screen opens:

B i o F i o 1 1 0	Calibrate pH 6.52	
	Zero Std: 7.00	Set Zero
	Raw Input: 5500	Set Span
		Reset

7. Rinse the pH electrode with distilled water, then immerse it in a pH 7.00 buffer solution. Allow a few minutes for the system to equilibrate, as indicated by a nearly stable Raw Input reading.
8. Press the Set Zero selector button. A highlighted edit box opens:

B i o F i o 1 1 0	Calibrate pH 6.52	
	Zero Std: 7.00	Set Zero
	Raw Input: 5500	Set Span
		Reset

Select "Reset" at any time in the Calibration process to return to the factory-set default Zero Standard and Span settings.

9. Using the touchpad, key in **7.0**, the pH of the buffer.
10. Reconfirm that the Raw Input reading is stable, then press the *Enter* key to save the value (or press **X** to cancel).
11. Press the Set Span selector button. As above, a highlighted edit box will open in the Set Span row.
12. Rinse the electrode with distilled water, then immerse it in a second pH buffer solution which is several pH units below pH 7.00 (e.g., 4.00) and allow a few minutes for the system to equilibrate, as indicated by a nearly stable Raw Input.

13. Using the touchpad, enter the new Span value, which is the buffer pH.
14. Reconfirm that the Raw Input is stable, then press the *Enter* key to save the value (or press **X** to cancel).
15. To ensure accuracy, repeat the above steps, using the same two buffer solutions.
16. After the values have been set, press the ☐⇒ button to return to the Main Screen.

**CAUTION!**

Do not use hydrochloric acid (HCl) with the BioFlo 110 for pH control or any other purpose, because HCl corrodes stainless steel. Over time, HCl will damage the stainless steel components.

Phosphoric and sulfuric (10% maximum concentration) acids are acceptable and are commonly used for pH control.

5 BEFORE FIRST USE


This section provides critical pre-operation instructions.

The BioFlo PCU display screens are available in English, French, German and Spanish. Before you operate your BioFlo 110 for the first time, be sure to select the preferred language for your PCU display screens.

Be sure also to configure the PCU and all other control modules so they can communicate with each other and, if desired, with an outside computer.

5.1 Select PCU Display Language

When the PCU is powered on, the display shows the following screen:



The default language setting is *English*. If you prefer to use another screen language, this is your opportunity to change the setting, if you have not already done so:

1. Press the selector button next to your preferred language.

5.2 Assign Vessel Address

Assign one address to all control modules associated with a single vessel. As the PCU can control as many as four vessels, it is essential that each vessel's control module stack have one unique address.

The address is set using the unit address selector switches on the back panel of each module. The module is shipped with all four switches On (facing down). As you can see in the reference table below, which is also posted on the back panel of each module next to the unit address switches), this is the address for Unit 1 (the first vessel and its associated control modules).

UNIT	SW1	SW2	SW3	SW4
1	↓	↓	↓	↓
2	↓	↓	↓	↑
3	↓	↓	↑	↓
4	↓	↓	↑	↑

Make sure that each module is set to the same address for Unit 1. A unit address switch can be flipped up or down with a pencilpoint, or the tip of a screwdriver.

If you have a second stack with a second vessel, set those modules for the Unit 2 address. Do the same for Units 3 and 4, if present.

5.3 For PC Supervision

5.3.1 Connect to Supervisory System

Connect the AFS/ModBus 25-pin connector on the PCU, using the appropriate cable, to the PC's serial port. Be sure to consult the user's manual that accompanies the supervisory software program and any interface box or converter you may need to complete the communications link.

The AFS/ModBus connector pin-outs are assigned as follows for RS-232 or RS-422 input and output:

Pin #	Signal	
2	RS-232	Transmit Data
3	RS-232	Receive Data
7	RS-232/ RS-422	Ground
21	RS-422	Enable (Jumper to Pin #7 to enable RS-422)
13	RS-422	Transmit Data +
25	RS-422	Transmit Data -
24	RS-422	Receive Data +
12	RS-422	Receive Data -

5.3.2 Select Communications Protocol

If a computer will be used for automatic data logging or process supervision, select an appropriate digital communications protocol, either ModBus or, if you are using NBS *BioCommand* supervisory software, AFS.

1. Press the *More...* selector button from the Application/Language screen, or press the  key (once from the Main Screen, or twice from the opening Language Selection Screen) to open the System Details screen:

S e t u p	System Details	
	Unit Base Address: 0	BaseAdd
	Unit 1 ID: 0	
	Unit 2 ID: 1	
	Unit 3 ID: 2	
	Unit 4 ID: 3	
	Comm. Protocol: AFS	Comm
	Recorder Status: Off	Recorder

2. Press the selector button next to Comm. The Communication Settings screen will open:

S e t u p	Comm. Settings	
	Comm. Protocol: AFS	Mod Bus
	Baud Rate: 9600	AFS
	Data Bits: 8	
	Parity: Even	
	Stop Bits: 2	
	Flow Control: None	

3. Press the selector button for your choice: AFS or ModBus.

There is no need to press the Enter key. Use the **✕** key to return to the Systems Detail screen, or the  key to return to the Main Screen.

5.3.3 Set Base Addresses

You must set base addresses, or ID numbers, for AFS or your other supervisory software, so the software recognizes each fermentor group as a distinct unit.

1. Press  to open the Systems Details screen (see sample screen above).
2. Press the selector button next to BaseAdd. The Base Address screen will open:

S e t u p	System Details	0
	Unit Base Address: 0	4
	Unit 1 ID: 0	8
	Unit 2 ID: 1	12
	Unit 3 ID: 2	16
	Unit 4 ID: 3	20
	Comm. Protocol: AFS	24
	Recorder Status: Off	28

3. Press the selector button for the number you wish to assign to the first unit. The system will automatically assign the next consecutive numbers to the other units you may be running from this PCU. In this way, if you choose to give Unit 1 the address of 8, for example, Unit 2 will be automatically assigned #9, Unit 3 will be #10 and Unit 4 will be #11.

NOTE:

Regardless of the actual number of fermentors being used, all four addresses are occupied by the PCU. Do not assign any other fermentor in your system a multidrop address within the range used by the PCU.

4. Press the **X** key to return to the Systems Detail screen, or press the  key to return to the Main Screen.

5.3.4 Set Chart Recorder for Data Logging

You may choose to use an optional analog data recorder, which you connect via user-supplied cable to the PCU's 9-pin male "D" recorder connector. Four outputs (per PCU) are available to send data to the supervisory system for logging. Following are the pin designations for those four outputs:

Data Logging Pin Outputs

From PCU	To Recorder	
Pin 2	+	Channel 1
Pin 1	-	
Pin 3	+	Channel 2
Pin 5	-	
Pin 4	+	Channel 3
Pin 7	-	
Pin 6	+	Channel 4
Pin 8	-	
Pin 9	Shield	

To choose assign the outputs to the loops of choice:

1. Press  to open the Systems Details screen.
2. Press the selector button next to Recorder. The Recorder screen will open:

S
e
t
u
p

Recorder

Recorder Status: Off

Output 1: None

Out 1

Output 2: None

Out 2

Output 3: None

Out 3

Output 4: None

Out 4

3. Press the selector button for Output 1. A screen with eight available loops opens:

S
e
t
u
p

Recorder

Recorder Status: Off

Output 1: None

None

Output 2: None

Temp-U1

Output 3: None

Agit-U1

Output 4: None

PumpA-U1

PumpB-U1


PumpC-U1

pH-U1

More...

Initially, all of the Output are assigned to “None”.

4. Press the selector button next to the loop you wish to assign to Output 1. If you do not see the loop you want, press the selector button next to More... to see the next group of available loops. In this example, we have selected Temp-U1, which is the Temperature loop from Unit 1. The previous screen returns, with the selected loop now assigned to Output 1:



5. Press the selector button next to On to turn Output 1 on.
6. Repeat steps 4 and 5 above to assign Output 2, Output 3 and Output 4.
7. Press the  key to return to the Main Screen.

In future, should you wish to change an output assignment or to return it to None, follow steps 3-5 above to make new selections.

6 STERILIZATION & REINSTALLATION

 NOTE:

Before proceeding, consult the dimensions outlined in Sections 11.4.1 and 11.4.2 to be sure your autoclave is large enough to accommodate the vessel with its various components.



WARNING!

During autoclaving, the vessel exhaust filter *and*, if present, the water jacket drain must be vented to avoid explosion.



WARNING!

Use protective gloves when handling hot components.



CAUTION!

During sterilization:

- The bearing housing cap must be installed on the motor bearing housing, to keep steam from damaging the internal bearings.
- On water-jacketed vessels, the jacket must be half-filled with water.



CAUTION!

Never autoclave PVC tubing (clear with white braiding).

There are four objectives to preparing a vessel for sterilization:

- A. To minimize pressure differences throughout the sterilization process by ensuring that the air can transfer freely between the inside and the outside of the vessel;
- B. To ensure that minor pressure differences do not expel liquid from the vessel by clamping off all penetrations that go below liquid level;
- C. To protect hydrophobic filters from blockage, which would occur if condensation were allowed to wet and block the filter surface;
- D. To protect susceptible vessel assembly components from steam damage.

The first objective is met by leaving at least one vessel port open, the second by clamping shut flexible tubing attached to immersed penetrations, and the third by wrapping filters with a protective cap of aluminum foil. Use protective caps on probes and bearings to meet the fourth objective.

6.1 *Initial Preparation for Autoclaving*

If this is a **water-jacketed vessel**, the jacket must be **half-filled with water** prior to autoclaving. Follow the steps in Section 6.1.1 below.

If this is a non-jacketed vessel, proceed to Section 6.2.

6.1.1 *Filling the Water Jacket*

To fill the water jacket:

1. After the tubing and water supply are connected, make sure the solenoid valve cable and the RTD cable are plugged into the Power Controller.
2. Set the temperature control mode to *Off* (for instructions, see Section 7.4.1).
3. Check that the temperature reading is higher than 5°C.
4. Allow water to enter the piping system; it will stop at the solenoid valve.
5. Set the temperature loop control mode to *Auto*.
6. Enter a temperature setpoint (**SP**) that is at least 12°C below the current value (**CV**). The controller will respond to the call for cooling by opening the solenoid valve, filling the jacket with water.
7. When the jacket is halfway filled, set the temperature control mode to *Off*.

6.2 *Additional Preparation for Autoclaving*

To continue preparing the vessel for sterilization:

1. Remove the motor from the top of the vessel and carefully put it aside.
2. Lubricate the vinyl bearing housing cap with silicone grease to facilitate sliding the cap securely onto the housing.
3. Place the bearing housing cap on the top of the bearing housing.
4. Disconnect the air and/or gas lines from the inlet filter on the sparger.
5. Disconnect the water lines. Remove all PVC tubing.
6. Clamp off the harvest tube, the sample tube and all other penetrations that are immersed in the media.
7. Remove the RTD from the thermowell.
8. Disconnect all probes and sensors, and remove their cables.
9. If you are using pH and dO₂ probes, install each probe's shorting cap (provided in the probe kit).

10. Before placing the vessel into the autoclave, loosen the glass sample bottle by ½ turn.
11. Wrap all filters with aluminum foil to protect them from steam.
12. Attach a piece of tubing, wrapped with some non-absorbent material (such as glass wool or non-absorbent cotton) to one of the addition ports. Wrap foil around the end of the tubing, shaped like a funnel, to allow the vessel to vent more easily during autoclaving. Place a clamp on the tubing.

 NOTE:

Be sure to leave one clamp open during autoclaving to equalize pressure. If this is a water-jacketed vessel, also leave the jacket water inlet clamp open.

If you have addition, foam trap or harvest bottles mounted at the base of the vessel, you can autoclave them with the vessel. Without detaching their tubing from the headplate:

13. Remove the bottle holder(s) and reinstall each on one of the headplate clamping screws.
14. Reinsert the bottle and turn the holder until the bottle and holder are positioned over the headplate, rather than extended over the edge.
15. Finger tighten the knurled nut.
16. Clamp off the tubing, and, where appropriate, remove it from the pump.


Probe tips must be moist during sterilization:

- If you will be doing batch fermentation, be sure the vessel is filled with media so the media will also be sterilized.
- If you will be doing cell culture, use at least 100 ml of a balanced salt solution (such as phosphate-balanced saline solution). Sterilize the media separately, after autoclaving the vessel.

6.3 *Autoclaving the Vessel*

1. If you have a 7.5L or 14.0L vessel assembly that is too tall for your autoclave, carefully lay the vessel, still mounted in its stand if present, in the optional angled autoclave rack (*part number M1273-9266—see Figure 41 below*). Secure it in place with the strap.
2. Insert the entire vessel assembly (glass jar, vessel stand if present, headplate and all headplate components) into an autoclave and sterilize. **Do not sterilize the jacket water heater.**
3. When you remove the vessel from the autoclave, immediately crimp the foil funnel on the addition port and close off the vent tubing to maintain sterility.

Figure 41: Angled Autoclave Rack Option for 7.5L & 14.0L Vessels



6.3.1

Sterilization Time and Temperature

Sterilization time varies with autoclave characteristics, temperature settings, vessel size and contents (i.e., media properties). As a starting point, **autoclave for 25 minutes after the autoclave reaches 121° C.**



CAUTION!

During autoclaving, the vessel and the water jacket (if present) must be vented at all times.

Release autoclave pressure only when the temperature has dropped below 90° C. Use slow exhaust (30-60 minutes).

Autoclave should be on liquid cycle pressure release, if it has such a feature.



NOTE:

Filter manufacturers generally advise limiting filter sterilization to 30 minutes, but the longer time required for slow exhaust is essential to protecting the vessel integrity. NBS' long experience has shown no adverse effects at all on filters exposed to longer autoclaving times.

Adjust the time and temperature as needed. If you have a water-jacketed vessel and the jacket is not half-filled, the vessel may not reach sterile temperature.

If, after autoclaving, most of the liquid has left the vessel, the autoclave is exhausting too quickly. Adjust the autoclave to exhaust more slowly.

6.4 **Reinstalling the System**

1. Position the vessel back on the benchtop. **If this is a water-jacketed vessel**, place the vessel on the jacket water heater: carefully align the water jacket inlet and outlet with the heater plate's rear notch, then place the vessel on the heater so the vessel's rubber feet fit into the holes provided for them.
2. Connect the heater to the heater connector on the front of the Power Controller.
3. Reconnect the water lines to the cooling coil (non-jacketed vessel) or the water jacket inlet and to the exhaust condenser, if present.



WARNING!

Cold water and hot glass is a potentially dangerous mix! Be sure to let the vessel cool for a few minutes before reconnecting the water line.

4. If present, remove the pH probe shorting cap and reconnect the pH cable to the probe and the dO2/pH Controller.
5. If present, remove the dO2 probe shorting cap and reconnect the dO2 cable to the probe and the dO2/pH Controller.
6. Remove the bearing housing cap (inspect for wear and replace for next use, if necessary), and reinstall the motor.
7. Reconnect the temperature probe and any other probe/sensor cables.
8. Remove addition, foam trap and/or harvest bottle assemblies from the headplate, and reinstall them on the base. Be sure to finger tighten all nuts.
9. Reinstall the addition, foam trap and/or harvest tubing in the appropriate pump(s) and release the clamp(s).
10. Reconnect the air and/or gas lines.
11. Add 1-2 ml of glycerin to the thermowell and reinstall the RTD temperature probe. Connect the cable to the Power Controller.



WARNING!

If the temperature probe is plugged into its controller but not inserted into the thermowell, the vessel will overheat, resulting in risk to the operator and to the culture.

6.5 Confirming pH Calibration

Autoclaving can alter the zero characteristics of pH probes, typically by 0.1 to 0.3 pH. To check, and to compensate for any discrepancy, you will need an accurate external pH meter. If one is not available, skip this section.

1. Following sterilization, with the media at room temperature, note the pH value on the BioFlo 110 Main screen.
2. Take a sample of media, and measure the pH on the external meter.
3. If the two values disagree, return to the BioFlo 110 Calibrate pH screen and set Zero to the value reported by the external meter. Do not change the Span or you will invalidate the entire calibration.

The pH value will now agree with the external meter's reading.


6.6 Polarizing & Calibrating the dO2 Probe

NOTE:

If you are using a polarographic dO2 probe, it must be polarized after autoclaving and before calibration.

Polarographic dissolved oxygen probes must be polarized for several hours after autoclaving and before calibration, according to the manufacturer's instructions. Polarization can be accomplished by plugging the dO2 probe cable into the dO2/pH Controller and leaving it on for several hours. Alternatively, battery-powered polarizers are available for this purpose.

From the Main Screen of the appropriate fermentor, press the selector button in the dO2 row. The dO2 Loop Details screen opens:



The procedures to set the Control Mode, the Setpoint, and the Graphic Display High and Low parameters are the same as for the Temperature and Agitation Loops.

1. After autoclaving, wait until the media reaches operating temperature. You can speed the cooling process by entering the Temperature setpoint, agitating and sparging.
2. In the dO2 Loop Details screen, press the *Calibrate* selector button. The dO2 Loop Calibration screen opens:

BIOFLO
110

Calibrate dO2 0

Zero Std: 0.00

Raw Input: 1022

Set Zero
Set Span
Reset

3. Press the Set Zero selector button. A highlighted edit box opens:

BIOFLO
110

Calibrate dO2 0

Zero Std: 0.00

Raw Input: 1022

Set Zero
Set Span
Reset

4. Using the touchpad, key **0** into the highlighted edit box. **Do not press *Enter* yet.**
5. Establish a reference “zero” signal from the dissolved oxygen probe, using one of the following choices:
 - (a) More accurate: Sparge with nitrogen until the dO2 Raw Input stops decreasing.
 - (b) Widely used alternate: Momentarily disconnect the dO2 cable from the probe.

6. When the Raw Input stops decreasing, press the *Enter* key to save the value (or press **X** to cancel).
7. Press the Set Span selector button. A highlighted edit box opens:

B i o F l o 1 1 0	Calibrate dO2 0	
	Zero Std: 0.00	Set Zero
	Raw Input: 1022	Set Span
		Reset

8. Using the touchpad, key **100** for the new Span value into the highlighted edit box. **Do not press *Enter*** yet.
9. Establish a reference 100% level of dissolved oxygen in the media: sparge with air at the appropriate flow rate for your application.
10. When the Raw Input stops increasing, press the *Enter* key to save the 100% dO2 reference value.
12. After the values have been set, press the ☐ ⇨ button to return to the Main Screen.

6.7 Powering the Jacket Water Heater

If you have a non-jacketed vessel, skip this section. After sterilization, you should have installed the vessel on the jacket water heater and connected the heater to the Power Controller (see Section 6.4). *Never set the temperature control loop to the desired working temperature (refer to Section 7.4.1) until all water connections have been made, the jacket is filled with water, the vessel is filled with media, and the RTD has been installed in the thermowell and connected to the Power Controller.*

Never plug the jacket water heater into the controller before the vessel is filled. When the heater is connected properly and the PCU is turned on, the jacket water stirrer should spin freely.



WARNING!

When the jacket water heater is plugged in, do not touch the top surface. It is very hot.

**NOTE:**


Always unplug the jacket water heater when not in use.

Whenever you remove the vessel, be sure to turn the temperature control *Off* (refer to Section 7.4.1) and unplug the jacket water heater. If you do not, the jacket water heater could reach excessive temperatures, posing a burn danger to personnel and causing the heater's internal safety cutoff switch to interrupt power to the heater.

If the safety cutoff is activated, power will not be restored until you do the following:

1. Wait for all surfaces to cool.
2. Turn the jacket water heater over and press the Thermostat Reset button, located on the bottom (see Figure 42 below).
3. Reinstall the vessel (jacket filled with water and vessel with media; temperature probe in the thermowell and plugged into the Power Controller).

Figure 42: Jacket Water Heater—Bottom View



7 OPERATION

NOTE:


Before operating a water-jacketed vessel, make sure it is properly installed on the vessel jacket water heater, and that the heater is properly connected (see Section 6.4) before you set the desired working temperature.

7.1 Primary Control Unit Display & Controls

The Primary Control Unit (PCU) serves as the operator interface. A single PCU can control up to four sets of control modules (but no more than sixteen total) for operating four vessels simultaneously. If you plan to run more than four vessels, another PCU must be added for each group of four.

The PCU has universal AC power input that can be set to accommodate electrical services between 100 and 240 Volts. It can be plugged into the outlet on the back of the Power Controller (see Figure 2), or plugged directly into your electrical service. The PCU is equipped with battery back-up to safeguard memory in case of power loss.


Figure 43: PCU Front Panel




The user interface touchpad has 12 numeric keys. As you can see in the drawing below, these keys are the numbers 0 through 9, a decimal point, and a +/- sign.

There are also 8 selector keys, which allow the user to choose specific loop detail screens from the main screen, and to select specific interactive screens for operational input.

Figure 44: PCU Touchpad



There are also four function keys on the touchpad:




This key allows you to **toggle** the display from Main screen to Main screen, if you have more than one fermentor on the system. It also allows you to return to the Main screen from any other screen.



This key allows you to **retrieve** the Language screen or the Systems Detail screen.



This key allows you to **go back** to the previous screen. It also allows you to **cancel** any numerical input (e.g., setpoint, etc) that has not yet been entered.




This key allows you to **enter** any touchpad input that you wish to save.


7.2 Selecting the Application

The default Application selection is *Fermentation*. The *Use...* selection on the initial setup screen will allow you to verify and/or change the application selected for each vessel (identified as Unit 1, Unit 2, etc.):



1. Press the *Use...* selector button.








2. The Application screen appears. To change the application for any vessel, press the *Use* button next to that unit. For this example, we will change Unit 2 from *Fermentation* to *Cell Culture*.



3. Press Unit 2's *Use* selector button. The display changes to Application selection screen (see sample screen below). Note that the alignment of choices (*Fermentation* or *Cell Culture*) on this screen does not relate to the units that appear on the left. You have already selected Unit 2, therefore your choice in this new screen will only affect Unit 2.

S e t u p	Application		
	Unit 1: Fermentation	Fermentation Cell Culture	
	Unit 2: Fermentation		
	Unit 3: Fermentation		
	Unit 4: Fermentation		
			
			
			

4. Press the *Cell Culture* selector button. The display automatically returns to the previous screen, showing Unit 2's application as Cell Culture.

S e t u p	Application		
	Unit 1: Fermentation	Use Use Use Use	
	Unit 2: Cell Culture		
	Unit 3: Fermentation		
	Unit 4: Fermentation		
			
			
		...Back	

Follow the same steps any time you wish to verify or change application for a vessel ("Unit").

7.3 Main Screen

1. Press the  button to open the Main Screen (see sample screen below).

Loop Name		Current Value	Loop Setpoint	Loop Control Mode		
B i o F l o 1 1 0	Name	Value	Setpoint	Control		
	Temp	23.4	30.0 °C	Auto		
	Agit	134	350 rpm	Auto		
	Pump A	0.0	0.0 %	Off		
	Pump B	0.0	0.0 %	Manual		
	Pump C	0.0	0.0 %	Off		
	pH	6.68	7.00 pH	Auto		
	dO2	77.1	0.0 %	Auto		

The above screen is only one sample. The loop names will reflect the modules in your system.

- If you have more than one fermentor running off the PCU, press the button again to open the Main Screen for Unit 2 (see sample screen below).
- Press again for Unit 3, and again for Unit 4.

U n i t 2	Name	Value	Setpoint	Control
	Temp	25.7	30.0 °C	Auto
	Agit	253	350 rpm	Auto
	Pump A	0.0	0.0 %	Manual
	Pump B	0.0	0.0 %	Manual
	Pump C	0.0	0.0 %	Manual
	pH	6.59	7.00 pH	Auto
	dO2	81.2	0.0 %	Auto

- Press once more to return to the Unit 1 (labeled “BioFlo 110”) Main Screen.

7.4 Set the Temperature Loop








All loops have three parameters that you can specify through the PCU. They are the Control Mode, Setpoint and Graphic Display. As you go through the following steps to set the Temperature loop, you will learn how to set the others as well.

7.4.1 Control Mode


To change the control mode of the temperature loop, starting from the Main Screen (see above) of the appropriate fermentation unit:

1. Press the selector button in the *Temp* row (see sample screen below).

Press here


B i o F i o 1 1 0	Name	Value	Setpoint	Control	
	Temp	23.4	30.0 °C	Off	
	Agit	134	350 rpm	Auto	
	Pump A	0.0	0.0 %	Manual	
	Pump B	0.0	0.0 %	Manual	
	Pump C	0.0	0.0 %	Manual	
	pH	6.68	7.00 pH	Auto	
	dO2	77.1	0.0 %	Auto	
					

The Temperature Loop Details screen appears (see sample below):




If the *Control* mode is *Off*, the PCU is not controlling temperature. To turn the control mode on:

- Press the selector button next to *Control*.




The Temperature Loop Control Mode screen opens:



- Press the selector button next to *Auto*.


The control mode is automatically updated as the Temperature Loop Details screen returns (see sample screen below):




7.4.2 Setpoint

To set the temperature setpoint:

1. Press the selector button next to *Setpoint*. A highlighted edit box appears:




2. Using the numeric touchpad, enter the new setpoint:



3. Press the *Enter* key to save the new setpoint.

The Temperature Loop Details screen returns, showing the new setpoint on the graphic display's X axis:



7.4.3 Graphic Display


To change the range of the Y axis on the graphic display:

1. Press the *Disp Hi* selector button. A highlighted edit box appears.
2. Using the touchpad, enter the desired High value.
3. Press the *Enter* key to save the new value.
4. Press the *Disp Lo* selector button. A highlighted edit box appears.
5. Using the touchpad, enter the desired Low value.
6. Press the *Enter* key to save the new value.

Press the  button to return to the Main Screen.

7.5 Setting the Agitation Loop

From the Main Screen of the appropriate fermentor, press the selector button in the *Agit* row. The Agitation Loop Details screen appears:



Set the Control Mode, the Setpoint, and the Display High and Low range in the same way as you did for the Temperature loop.




CAUTION!

Never exceed 200 rpm unless at least one impeller is immersed in liquid.

7.6 Setting the pH Loop

From the Main Screen of the appropriate fermentor, press the selector button in the *pH* row. The pH Loop Details screen opens:

The pH Loop Details screen opens:



The procedures to set the Control Mode, the Setpoint, and the Graphic Display High and Low parameters are the same as for the Temperature and Agitation Loops.




NOTE:

Be sure to calibrate the pH probe, following the procedure outlined in Section 4.13, before autoclaving it with the vessel.

7.6.1 Setting a Deadband for pH control

A deadband is a tight plus and minus tolerance around the setpoint. This is a useful feature, for example, to trigger an alarm via supervisory software.

1. In the pH Loop Details screen, press the *Setup* selector button. The pH Setup screen opens:



2. Press the *Deadband* selector button. A highlighted text box opens:

B i o F l o 1 1 0	pH 6.53 pH		Defaults
	Deadband: 0.00		Deadband
	Output Multiplier: 100.00		Multiplier
	<u>Advanced Control</u>		
	P-Gain: 140.00		P Gain
	I-Gain: 14.00		I Gain

3. Using the touchpad, enter the desired deadband value:

B i o F l o 1 1 0	pH 6.53 pH		Defaults
	Deadband: 0.00	0.20	Deadband
	Output Multiplier: 100.00		Multiplier
	<u>Advanced Control</u>		
	P-Gain: 140.00		P Gain
	I-Gain: 14.00		I Gain

4. Press the *Enter* key (or press **X** to cancel). The new value appears onscreen.

7.6.2

Setting the Output Multiplier for pH Control

Later, when you begin your batch, you may notice that pH remains steady at the setpoint, yet the acid and base pumps are continually alternating in making additions. This is an indication that the controller is overcompensating for minor fluctuations in pH. *Output Multiplier* is a feature that attenuates controller output to the acid and base pumps and the CO₂ gas line, providing more nuanced control of additions to maintain pH.

We recommend that you begin by implementing a multiplier of 25%. This means that if the controller's output to the base pump, for example, is 100%, then the 25% multiplier will reduce pump output to 25%. If the controller's output to the pump is 50%, the 25% multiplier factor will reduce pump output to 12.5%.

If, after applying an Output Multiplier of 25%, you find the results are attenuated but the controller seems unable to maintain the setpoint, increase the Multiplier by small increments until the controller is able to maintain setpoint.

To set the Multiplier:

1. In the pH Loop Details screen, press the *Setup* selector button. The pH Setup screen opens:

B i o F l o 1 1 0	pH 6.53 pH		
	Deadband: 0.20	Defaults	
	Output Multiplier: 100.00	Deadband	
		Multiplier	
	<u>Advanced Control</u>		
	P-Gain: 140.00	P Gain	
	I-Gain: 14.00	I Gain	

2. Press the *Multiplier* selector button. A highlighted text box opens:

B i o F l o 1 1 0	pH 6.53 pH		Defaults
	Deadband: 0.20		Deadband
	Output Multiplier: 100.00		Multiplier
	<u>Advanced Control</u>		
	P-Gain: 140.00		P Gain
	I-Gain: 14.00		I Gain

3. Using the touchpad, enter the desired Multiplier value (note that you can set this percentage to the hundredths):

B i o f i o 1 1 0	pH 6.53 pH		Defaults	
	Deadband: 0.20		Deadband	
	Output Multiplier: 100.00		25.00	Multiplier
	Advanced Control			
	P-Gain: 140.00		P Gain	
	I-Gain: 14.00		I Gain	

4. Press the Enter key to save the new value (or press **X** to cancel). The new value appears onscreen.

7.6.3 Setting P & I Gains for pH Control

NOTE:

We recommend that you do not change the factory P & I settings unless you are an advanced user of P & I algorithms.

P and I values are numbers that determine how the fermentor responds to changing growth conditions and new setpoints. Because growth characteristics of bacterial and cell cultures are very different, they require some different P and I values for proper control. Incorrect P and I values can cause poor control, which may actually mimic component failure. For this reason, we recommend that you use the default factory-set P and I gains.

If you are very experienced with P and I algorithms and wish to change the default values, use this procedure:

1. In the pH Loop Setup screen, press the *P Gain* selector button.
2. Using the touchpad, enter the new *P Gain* value in the highlighted edit box.
3. Press the Enter key to save the new value (or press **X** to cancel). The new value appears onscreen.
4. Press the *I Gain* selector button. A highlighted edit box will open.
5. Using the touchpad, enter the new *I Gain* value.
6. Press the Enter key to save the new value (or press **X** to cancel). The new value appears onscreen.

7. After the values have been set, press the  button to return to the Main Screen.

Anytime you wish to return to the default P & I Gain settings, open the pH Loop Setup screen, and press the *Defaults* selector button.

See Section 12.5.1 for the pH P & I control equation.

7.7 Setting the dO2 Loop

When your system includes the dO2/pH Controller, the PCU can supervise dissolved oxygen (dO2) in the media. If you are using a dO2 probe, you will need to set the dO2 loop. You will calibrate the probe after autoclaving.



NOTE:

After autoclaving, be sure to calibrate the dO2 probe *before* setting the dO2 loop. If you are using a polarographic dO2 probe, it must be polarized before calibration, but after autoclaving. To polarize and calibrate the dO2 probe, see Section 6.6.

7.8 Setting dO2 Control Cascades

Cascading brings several systems together to work jointly to achieve your goal. There are four possibilities for Fermentation, and four other choices for Cell Culture, if you are using the Gas Mix Controller. Without the Gas Mix Controller, there are two cascades available to either Fermentation or Cell Culture. In all cases, when you first set up the system, the default choice is set to *None* (no cascade).

7.8.1 Cascades for Fermentation without Gas Mix Controller

For Fermentation, you have two choices when you are not using the Gas Mix Controller: *None* or *Agitation*. (If you have the optional TMFC module, see Section 14 for details.)

Cascading to Agitation is designed to control dissolved oxygen through automatically controlled agitation speed and oxygen output (via P & I Gains—see Section 7.9 below). This is how it functions: when the actual dO2 value rises above the dO2 setpoint, the agitation speed will automatically decrease until the dO2 setpoint is reached. Conversely, when the actual dO2 value drops below the setpoint, the cascade system acts to bring it back up.

The default setting is *Cascade To: None*, which means that dissolved oxygen will be controlled only by the addition of oxygen when the actual dO2 level drops below setpoint.

To cascade dissolved oxygen to *Agitation*:

1. In the dO2 Details screen, press the *Setup* selector button. The Setup screen opens:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: None		Defaults
	Agit Hi Limit: 1200		Cascade
	Agit Lo Limit: 250		Agit Hi
			Agit Lo
	<u>Advanced Control</u>		
	P-Gain: 0.20		P Gain
	I-Gain: 1.00		I Gain

 NOTE:

Selecting *Defaults* at any time (during setup or after settings have been changed) will return all settings EXCEPT *Cascade To:* to the default setting. Cascades remain in force until you select otherwise.

2. Press the *Cascade* selector button. The Cascade screen opens:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: None		Agit
	Agit Hi Limit: 1200		None
	Agit Lo Limit: 250		
	<u>Advanced Control</u>		
	P-Gain: 0.20		
	I-Gain: 1.00		

3. Press the *Agit* selector button. The Setup screen returns, showing the new cascade parameter:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: Agit Agit Hi Limit: 1200 Agit Lo Limit: 250		Defaults Cascade Agit Hi Agit Lo
	<u>Advanced Control</u> P-Gain: 0.20 I-Gain: 1.00		P Gain I Gain

4. Press **X** to return to the dO2 Details screen or ☐ $\Rightarrow$ to return to the Main Screen.

7.8.2 Cascades for Fermentation with Gas Mix Controller

When you are using the Gas Mix Controller, you have four choices to cascade to for Fermentation: *Agitation*, *Oxygen*, *Agitation/Oxygen* or *None*. (If you have the optional TMFC module, these screens are different. *See Section 14 for details.*)

Cascading to:

- *Agitation* controls dissolved oxygen through automatically (via P & I Gains—*see Section 7.9 below*) controlled agitation speed and oxygen output. This is how it functions: when the actual dO2 value rises above the dO2 setpoint, the agitation speed will automatically decrease until the dO2 setpoint is reached. Conversely, when the actual dO2 value drops below the setpoint, the cascade system acts to bring it back up.
- *Oxygen* controls dissolved oxygen by automatically adjusting the mix of air and oxygen. (This is not available without the Gas Mix Controller.)
- *Agitation/Oxygen* controls dissolved oxygen by automatically running Agitation to the maximum speed, then, if dissolved oxygen still has not reached the setpoint, increasing the oxygen flow through the sparger. This cascade is the most frequently used in fermentation. (This is not available without the Gas Mix Controller.)
- The default setting is *Cascade To: None*, which means that dissolved oxygen will be controlled only by the addition of oxygen when the actual dO2 level drops below setpoint.

To cascade dissolved oxygen to *Agitation*:

1. In the dO2 Details screen, press the *Setup* selector button. The Setup screen opens:

B i o f l o 1 1 0	dO2 73.1 %		
	Cascade To: None Agit Hi Limit: 1200 Agit Lo Limit: 250		Defaults Cascade Agit Hi Agit Lo
	<u>Advanced Control</u>		
	P-Gain: 0.20 I-Gain: 1.00		P Gain I Gain

2. Press the *Cascade* selector button. The Cascade screen opens:

B i o f l o 1 1 0	dO2 73.1 %		
	Cascade To: None Agit Hi Limit: 1200 Agit Lo Limit: 250		Agit O2 Agit/O2 None
	<u>Advanced Control</u>		
	P-Gain: 0.20 I-Gain: 1.00		

3. Press the *Agit* selector button. The Setup screen returns, showing the new cascade parameter:

B i o F l o 1 1 0	dO2 73.1 %		Defaults
	Cascade To: Agit Agit Hi Limit: 1200 Agit Lo Limit: 250		Cascade
			Agit Hi
			Agit Lo
	<u>Advanced Control</u>		
	P-Gain: 0.20		P Gain
	I-Gain: 1.00		I Gain

4. If you wish to change the high and low limits for Agitation speed, use the *Agit Hi* and *Agit Lo* selector buttons to open an edit box for each, use the touchpad to input a value, and press the *Enter* button to save the new value to memory. The Hi Limit may not exceed 1200 and the Lo Limit may not go below 250.
5. Press **✕** to return to the dO2 Details screen or **☐⇒** to return to the Main Screen.

To cascade to Oxygen: follow steps 1-4 above, but select *O2* instead of *Agitation*.

To cascade to Agitation/O2: follow steps 1-4 above, but select *Agit/O2* instead of *Agitation*.

7.8.3 Cascades for Cell Culture without Gas Mix Controller

For Cell Culture, without the Gas Mix Controller, you have the same two choices explained for Fermentation without the Gas Mix Controller: *None* or *Agitation*. See Section 7.8.1 above. (If you have the optional TMFC module, see Section 14 for details.)

7.8.4 Cascades for Cell Culture with Gas Mix Controller

When you are using the Gas Mix Controller, you have four choices to cascade to for Cell Culture: *Agitation*, *4 Gas*, *Demand* or *None*. (If you have the optional TMFC module, see Section 14 for details.)

Cascading to:

- *Agitation* controls dissolved oxygen through automatically (via P & I Gains—see Section 7.9 below) controlled agitation speed and oxygen output. This is how it functions: when the actual dO2 value rises above the dO2 setpoint, the agitation speed will automatically decrease until the dO2 setpoint is reached.

Conversely, when the actual dO₂ value drops below the setpoint, the cascade system acts to bring it back up.

- *4 Gas* controls dissolved oxygen by automatically adjusting the feed of all gases introduced to the culture. (This is not available without the Gas Mix Controller.)
- *Demand* controls dissolved oxygen by automatically adjusting the feed of carbon dioxide, nitrogen and oxygen. (This is not available without the Gas Mix Controller.)
- The default setting is *Cascade To: None*, which means that dissolved oxygen will be controlled only by the addition of oxygen when the actual dO₂ level drops below setpoint.

To cascade dissolved oxygen to *Agitation*:

6. In the dO₂ Details screen, press the *Setup* selector button. The Setup screen opens:

B i o f i o 1 1 0	dO₂ 73.1 %		
	Cascade To: None Agit Hi Limit: 300 Agit Lo Limit: 25		Defaults Cascade Agit Hi Agit Lo
	<u>Advanced Control</u> P-Gain: 0.20 I-Gain: 1.00		P Gain I Gain

7. Press the *Cascade* selector button. The Cascade screen opens:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: None Agit Hi Limit: 300 Agit Lo Limit: 25		Agit 4 Gas Demand None
	<u>Advanced Control</u>		
	P-Gain: 0.20		
	I-Gain: 1.00		

8. Press the *Agit* selector button. The Setup screen returns, showing the new cascade parameter:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: Agit Agit Hi Limit: 300 Agit Lo Limit: 25		Defaults Cascade Agit Hi Agit Lo
	<u>Advanced Control</u>		
	P-Gain: 0.20		
	I-Gain: 1.00		P Gain I Gain

9. If you wish to change the high and low limits for Agitation speed, use the *Agit Hi* and *Agit Lo* selector buttons to open an edit box for each, use the touchpad to input a value, and press the *Enter* button to save the new value to memory. The Hi Limit may not exceed 300 and the Lo Limit may not go below 25.
10. Press **X** to return to the dO2 Details screen or ☐ ⇨ to return to the Main Screen.

To cascade to 4 Gas: follow steps 1-4 above, but select *4 Gas* instead of *Agitation*.

To cascade to Demand: follow steps 1-4 above, but select *Demand* instead of *Agitation*.

 NOTE:

After a Cascade To: Gas Mix or to Cascade To: Demand has been selected, if the Gas Mix function is turned off through the O2 Loop Gas Mix screen, the cascade will be overridden.

7.9 Setting P & I Gains for dO2 Control

 NOTE:

We recommend that you do not change the factory P & I settings unless you are an advanced user of P & I algorithms.

P and I values are numbers that determine how the fermentor responds to changing growth conditions and new setpoints. Because growth characteristics of bacterial and cell cultures are very different, they require some different P and I values for proper control. *For example, for cell culture, the P value should be 2.00 and the I value 0.67, while the factory-set values for fermentation are $P = 0.20$ and $I = 1.00$.*

Incorrect P and I values can cause poor control, which may actually mimic component failure. For this reason, we recommend that you use the default factory-set P and I gains.

If you are very experienced with P and I algorithms and wish to change the default values, use this procedure:


1. In the dO2 Loop Setup screen, press the *P Gain* selector button.
2. Using the touchpad, enter the new *P Gain* value in the highlighted edit box.
3. Press the Enter key to save the new value (or press **X** to cancel). The new value appears onscreen.
4. Press the *I Gain* selector button. A highlighted edit box will open.
5. Using the touchpad, enter the new *I Gain* value.
6. Press the Enter key to save the new value (or press **X** to cancel). The new value appears onscreen.
7. After the values have been set, press the  button to return to the Main Screen.

Anytime you wish to return to the default P & I Gain settings, open the dO2 Loop setup screen, and press the *Defaults* selector button.

See Section 12.5.2 for the dO2 P & I control equation.

7.10 Setting the Gas Mix for Fermentation

If you are using **Gas Mix Controller**, your system's Main Screen will include an oxygen (O₂) loop. From the Main Screen of the appropriate fermentor, press the selector button in the O₂ row to open the O₂ loop screen:



1. Press the *Control* selector button.
2. In the next screen, press the *Manual* selector button.
3. **To cascade O₂ to your dO₂ loop**, see Section 7.8.2.
4. **To set a manual O₂ percentage without cascading**, press the Setpoint selector button in the O₂ loop screen (see sample screen above) and enter the percentage of oxygen needed.

The controller will automatically feed oxygen as needed to maintain the setpoint you have entered.

7.11 Setting the Gas Mix for Cell Culture

If you are using **Gas Mix Controller**, your system's Main Screen will include a Gas loop. From the Main Screen of the appropriate fermentor, press the selector button in the Gas row to open the Gas Mix Screen. "Gas Mix" for this controller is not actually a mixing process; it is an automatically sequenced feeding of various gases.

Four control modes are possible: *4-Gas*, *Demand*, *100% Feed* and *Off*. Selecting *Off* will turn the Gas Mix Controller off.

7.11.1 4-Gas Mode

Because your system has the Gas Mix Controller, when you select *4-Gas* as the control mode the PCU will control the levels of all gases (carbon dioxide, nitrogen, oxygen and air) being introduced to the culture, to maintain dO₂ and pH at their respective setpoints.

B i o f l o 1 1 0	Gas Mix		4 Gas
	Control: 4 Gas		Demand
			Off
	Gas	%	
	CO₂	20.0	100% CO₂
	N₂	49.7	100% N₂
	O₂	0.0	100% O₂
	Air	30.3	100% Air

Select this control mode by pressing the *4-Gas* selector button. There are no further choices to program; in this mode, the 4-Gas controller will automatically control the sequential introduction of gases according to need.

7.11.2 Demand Mode

Selecting *Demand* as the control mode will provide automatic “on demand” sparging of carbon dioxide, nitrogen and/or oxygen to maintain dO₂ and pH at their respective setpoints.

Press the *Demand* selector button. There are no further choices to program.

7.11.3 100% Feed Mode

If you need to feed 100% of any one of the four gases for a period of time, you can manually override the 4-Gas controller using this feature.

To override the Gas Mix Controller’s control of gas sequencing in order to feed 100% of one gas: in the Gas Mix screen, press the selector button next to the gas of choice. The screen reflects your choice (in this case, CO₂):

B i o F l o 1 1 0	Gas Mix		4 Gas
	Control: CO2		Demand
			Off
	Gas	%	
	CO2	100.0	100% CO2
	N2	0.0	100% N2
	O2	0.0	100% O2
	Air	0.0	100% Air

 NOTE:

The gas will continue to feed at 100% until you manually select another gas, return to automatic 4-Gas Mode, to *Demand* mode or choose *Off*.

 NOTE:

If, at any time after establishing a dO2 cascade, you turn the Gas Mix controller *Off* using the Gas Mix screen, the cascade will be overridden.

7.12 *Setting the Pumps*

7.12.1 **Pump Operation**

The BioFlo 110 supports five peristaltic pumps. On the rear panel of the Power Controller, the pump outlets labeled Acid and Base are integrated with the pH control loop. This is why they do not appear as individual pumps on the Main Screen. Assignable pumps A, B and C, do appear individually on the display. Each of these can be operated manually, by assigning a setpoint, or be set to respond automatically to signals from conduction probes.

7.12.2 **Pump Setpoint and Multiplier**

Each of the three assignable pumps (A, B & C) accepts a setpoint. The setpoint determines the percentage of On time during each ten-second period while the pump is set to On. For example, if Pump A has a setpoint of 40%, every ten seconds it will turn on for four seconds.

Instead of a setpoint, the Acid and Base Pumps use a Multiplier value which can be assigned in the pH Control screen. A multiplier of 50%, for example, has the base pump run for 5 seconds every 10 seconds if the pH controller is calling for base addition.

7.12.3 Control Mode

Pumps A, B and C accept a user-assigned Control Mode. The available control modes are Manual, Wet On, Wet Off and Antifoam. *Wet On*, *Wet Off* and *Antifoam* are the designations given, for convenience, to the probes that are plugged into the socket of the same name on the Level Module. Control modes affect pump operation in the following ways:

- **Manual:** pump operates at the setpoint
- **Wet On:** when the probe is immersed in media or foam, the pump turns on at its assigned setpoint. This is useful to remove media when additions accumulate to an undesired level, for example.
- **Wet Off:** when the probe is not immersed in media or foam, the pump turns on at its assigned setpoint. This is useful for media addition.
- **Antifoam:** when the probe is immersed in media or foam, the pump turns on at its assigned setpoint. This is useful to add chemical defoamer.

7.12.4 Setting Pumps

To set the Multiplier for pH pumps (Acid and Base), see Section 7.6.2. Be sure also to set the Control Mode, Setpoint and Graphic Display (Hi and Lo limits) for those pumps.


Following the instructions below:

1. Set the Control Mode,
2. Enter a Setpoint,
3. Select the Hi and Lo limits of the Graphic Display presented in the Loop Details screen

for each assignable Pump A, B and C.

To set up the first peristaltic pump (Pump A) under the PCU's supervision:

1. In the Main Screen, press the selector button in the Pump A row to open this screen:



The procedures are generally the same as for the Temperature and Agitation Loops, to set the Graphic Display Hi and Lo parameters, the Setpoint, and the Control Mode (if you wish to choose Manual—which allows the operator to manually set pump output—or Off).


For the three automatic Control Modes unique to the pump loops (Antifoam, Wet On and Wet Off), *see Section 7.13 below for details*.

2. Repeat the setup procedure for Pump B and Pump C.


7.13 Setting the Pump Control Mode: Antifoam, Wet On & Wet Off

If your system includes the Level Controller, the PCU can supervise the three level loops identified as Antifoam, Wet On and Wet Off, each of which is actually a control mode for the assigned pump, as mentioned in the section above.

1. To set up the defoamer pump, from the pump's Details screen, press the *Control* selector button for its designated pump. The Control screen opens:



2. Press the Antifoam selector button. The pump Details screen returns; after a short delay, the control mode will reflect your choice. Note that a new parameter, Probe Sensitivity, also appears:



When the probe is activated, CV is current percent output (either 0.0 or the setpoint).

3. Probe sensitivity reflects the density of foam necessary to activate the antifoam pump. To set a sensitivity value (we recommend starting with 75.0), press the *Sensitivity* selector button, input the desired value using the touchpad, and press the *Enter* key.
4. Press the ☐ ⇨ button to return to the Main Screen.
5. Follow the same procedure to set the Wet On pump and the probe sensitivity value, and the Wet Off pump and probe sensitivity.

7.14 *Inoculation*

When you are ready to inoculate the media to start a batch, observe aseptic techniques to do the following:

1. Using a syringe, inject the inoculum through the self-sealing rubber seal of the septum port.



NOTE:

Always inspect the condition of the septum seal and the bearing housing cap prior to autoclaving. Replace as necessary.

7.15 *Sterile Sampling*

To remove a sample from an active batch:

1. Make sure the sampler syringe plunger is completely closed.
2. Confirm that the sample bottle is in place and finger-tight against its holder.
3. Open the thumb clamp.
4. Pull back on the syringe plunger to fill the bottle.
5. Close the thumb clamp. Close the plunger.
6. Observe aseptic techniques to remove the filled bottle and cap it.
7. Immediately install a new sample bottle in the holder.

7.16 *Automatic Foam Control*

If you have installed a foam probe and set its control mode to *Antifoam*, the addition of antifoam is automatic. When the probe detects the presence of foam, its signal prompts the Level Controller to signal the Power Controller. The Power Controller activates the Antifoam pump to add chemical defoamer from the addition bottle through an addition port. When the probe no longer detects foam, the pump will be turned off.

7.17 Harvesting



WARNING!

NEVER PRESSURIZE A GLASS VESSEL!

- **Never** intentionally block the exhaust to raise vessel pressure.
- **Never** exceed the maximum pressure specified in this manual. This maximum pressure is necessary only to obtain the highest gas flow rates. It is capable of breaking the vessel if the vessel is not properly vented.

At the end of a batch, to harvest the entire culture:

1. Aseptically prepare an autoclaved harvest bottle large enough to hold the entire contents of the vessel. The bottle cap should be equipped with a sterile filter on one penetration. There should also be sufficient silicone tubing on the other penetration to reach from the bottle, placed on the floor below the benchtop, to the harvest tube outlet. Secure both the tubing and the filter with plastic ties.
2. Aseptically install the harvest tubing on the harvest tube outlet, with a thumb clamp on the tubing. Close the clamp. Secure the tubing on the harvest tube with a plastic tie.
3. Set the harvest bottle on the floor below the vessel.
4. There are two ways to prime the harvest process: (a) with a pump, or (b) without a pump.
 - a. Thread the tubing through one of the peristaltic pumps, unclamp the tubing and manually turn the pump on. Harvesting will begin.
 - b. Install a rubber sampler bulb on the other end of the harvest bottle filter. Unclamp the harvest tubing, then squeeze and release the bulb. Harvesting will begin.
5. Once harvesting has been primed, there are two ways to complete the process:
 - a. Allow the pump to operate until the vessel is exhausted.
 - b. Allow gravity to do the work.

Note that a small amount of residue will probably remain in the vessel no matter how you harvest.

8 MAINTENANCE

Preventive maintenance keeps your equipment in proper working condition. When performed routinely, maintenance results in longer life for your equipment. It also reduces time lost due to equipment failure.

**WARNING!**

Always disconnect the power cord from the Power Controller and the PCU before performing maintenance.

Never immerse any BioFlo 110 module in liquid of any kind.

8.1 *Cleaning the Modules*

Use a sponge dampened with mild detergent and water to wipe the exterior of each BioFlo 110 module. Rinse the sponge in clean water, then wipe the exterior again.

8.2 *pH Probe Maintenance and Storage*

The probe should be stored standing upright, with the electrode tip immersed in a solution of 3 molar KCl or a buffer solution between pH 4.00 and pH 7.00.

**CAUTION!**

**Never let a pH probe rest on its tip.
Never leave a pH probe in DI water.**

8.3 *dO₂ Probe Maintenance and Storage*

Use soft facial tissue to clean the dO₂ probe.

Check the probe's Teflon membrane to be sure there are no punctures, puckers or wrinkles. If there are, the probe should be replaced.

When it is not in use in the vessel, the dO₂ probe should be stored standing upright with the shorting cap in place and the membrane isolated from the air environment. **At no time should the probe be allowed to rest on its membrane.**

**CAUTION!**

Never let a dissolved oxygen probe rest on its tip.

8.4 *Cleaning the Vessels*

After each use, the vessel and the vessel components should be thoroughly cleaned. If necessary, decontaminate the vessel by autoclaving. If you are autoclaving your vessel, fill it with water to the maximum working volume. This will allow for easier clean-up later.

**CAUTION!**

Never allow the drive assembly to rest on the impeller shaft (see Section 4.4).

8.4.1 14.0L Vessel Disassembly

Special care must be taken when you remove the headplate assembly from the 14.0L vessel because the vessel narrows at the neck:

1. After you remove the headplate nuts and washers, lift the headplate assembly a few inches, and hold it up with one hand.
2. With the other hand, reach into the vessel, squeeze the baffle (between thumb and forefinger) tightly enough to clear the vessel neck.
3. Lift the assembly straight up until the entire assembly is clear.

8.4.2 Vessel Cleaning Methods

There are three ways to clean your vessel:

Method 1	1. Drain all the liquid from the vessel. 2. Disassemble and wash the headplate, glass vessel, impellers, baffles and other components in hot water and an appropriate laboratory detergent. 3. If necessary, use a very small test tube brush to clean the ports and orifices. A soft brush may also be used to remove dried-on material from the glass. 4. Rinse all components in tap water. 5. Rinse twice in purified (deionized, RO, distilled water). Allow everything to dry thoroughly before reassembling the vessel.
Method 2	1. Drain the liquid from the vessel. 2. Fill the vessel to the maximum working volume with water. 3. Reassemble the vessel and components, and reconnect the agitation shaft to the motor. 4. Press the Agitation key On. 5. After approximately ten minutes, press the Agitation key Off. 6. Remove the vessel and disassemble its components. 7. Wash all elements in detergent and hot water.
Method 3	1. Drain the liquid from the vessel. 2. Fill the vessel to just above the maximum working volume with hot water and the recommended amount of a suitable detergent. 3. Allow the vessel with its component parts to soak (for a short time or overnight). 4. Drain the vessel. 5. Disassemble and wash the headplate, glass vessel, impellers, baffles and other components in hot water and an appropriate laboratory detergent. 6. If necessary, use a very small test tube brush to clean the ports and orifices. A soft brush may also be used to remove dried-on material from the glass. 7. Rinse all components in tap water. 8. Rinse twice in purified (de-ionized, RO, distilled water). Allow everything to dry thoroughly before reassembling the vessel.

**NOTE:**

For easy cleaning, the ring sparger has a removable plug at the tip.

8.5 Port & Adapter O-Ring Replacement

Each time you clean the vessel assembly, check all port and adapter O-rings for wear, distortion, rips or any other damage to their integrity. Replace them as needed.

8.6 Headplate O-Ring Replacement

When it needs to be replaced, pull the old O-ring out with a knife or a flat-bladed screwdriver. Lubricate the new O-ring with a light coat of silicone grease, then insert it, making sure it is well-seated in the groove.

8.7 Bearing Housing Cap Replacement

Check the bearing housing cap after each use for distortion, tears or any other damage to its integrity. Replace the cap as needed.

8.8 Replacing PCU Fuses

To replace a fuse in the PCU:




CAUTION!

Always disconnect the unit from the power source first.

1. Remove the cover/fuse block located on the rear of the unit with a small screwdriver.
2. Remove the old fuse.
3. Insert a new fuse of the same type.
4. Replace the cover/fuse block into the Fuse Holder/Power Inlet (see Figure 45 below).

Figure 45: Fuse Holder/Power Inlet



9 SERVICE

If any problems occur with your BioFlo 110 system or its individual components, do not attempt to perform any service on it. Unauthorized servicing may void the warranty. Please contact your local NBS Service Department or your local NBS distributor.

In any correspondence with NBS, please refer to the Model Number (BioFlo 110), Manufacturing Part Number(s) and Serial Number (s) of the problem module. This information is located on the bottom of each unit.

9.1 *Troubleshooting*

If no information appears on the display, check the following *before you call for service*:

1. Verify that the power cord is correctly connected to the Power Controller and to an outlet of appropriate output.
2. If the PCU is powered via the Power Controller, make sure the power cord is correctly connected between both modules. If not, verify that the power cord is correctly connected to the PCU and to an outlet of appropriate output.
3. Verify that the Power Controller's power switch is turned On.
4. Confirm that the PCU power switch is turned On.
5. Check that all control modules are correctly connected to each other.
6. Make sure that all cables are attached properly.
7. Verify that each module's unit address selector switches are correctly assigned (per Section 5.2).

10 REPLACEMENT PARTS & ACCESSORIES

When you order replacement parts, accessory parts, or when you request service information, please provide the Model Number (BioFlo 110), Manufacturing Part Number(s) and Serial Number(s) of your unit. This information is located on the bottom of each module.

10.1 Replacement Parts

Vessel Assembly		NBS Part Number
FERMENTATION VESSEL ASSEMBLY: glass vessel, headplate, clamping nuts, baffle, thermowell, RTD probe, O-rings, port plugs & adapters, exhaust tube, sparge tube, harvest tube, direct drive motor & agitation shaft, bearing housing, set of bearing housing caps, rotameter, 2 six-bladed Rushton-type impellers. Non-jacketed vessel assemblies also include cooling coil, heater blanket, vessel stand and water valve. Jacketed vessel assemblies also include jacket water heater.		
Non-Jacketed	1.3L	M1273-1001
	3.0L	M1273-1002
	7.5L	M1273-1005
	14.0L	M1273-1010
Water-Jacketed	1.3L	M1273-1011
	3.0L	M1273-1012
	7.5L	M1273-1015
	14.0L	M1273-1020
CELL CULTURE VESSEL ASSEMBLY: glass vessel, headplate, clamping nuts, baffle, thermowell, RTD probe, O-rings, port plugs & adapters, exhaust tube, sparge tube, harvest tube, motor & agitation shaft, bearing housing, set of bearing housing caps, rotameter, 1 pitched blade impeller. Non-jacketed vessel assemblies also include cooling coil, heater blanket, vessel stand and water valve. Jacketed vessel assemblies also include jacket water heater.		
Non-Jacketed, with direct-drive motor	1.3L	M1273-1041
	3.0L	M1273-1042
	7.5L	M1273-1045
	14.0L	M1273-1050
Non-Jacketed, with magnetic drive motor	1.3L	M1273-1051
	3.0L	M1273-1052
	7.5L	M1273-1055
	14.0L	M1273-1060
Water-Jacketed, with direct drive motor	1.3L	M1273-1021
	3.0L	M1273-1022
	7.5L	M1273-1025
	14.0L	M1273-1030
Water-Jacketed, with magnetic drive motor	1.3L	M1273-1031
	3.0L	M1273-1032
	7.5L	M1273-1035
	14.0L	M1273-1040

Angled Autoclave Rack	NBS Part Number
For 7.5L & 14.0L vessels	M1273-9266

Batteries & Fuses	NBS Part Number
Power Controller Fuse, 250V, 1.25 Amps	P0380-3750
Power Controller Fuse, 250V, 3.15 Amps	P0380-3751
PCU Fuse, Slo-Blo [®] , 250V, 2 Amps	P0380-3533
Battery, Alkaline, 1.5 VDC	P0480-9030

Exhaust Condenser	NBS Part Number
Exhaust Condenser, 1.3L, 3.0L & 7.5L	M1273-9945
Exhaust Condenser, 14.0L	M1273-9957

Foam Trap Kits	NBS Part Number
Foam Trap Kit, 500 mL bottle (all vessels)	M1273-9943
Foam Trap Kit, 250 mL bottle (all vessels)	M1273-9942

Headplate Port Fittings	NBS Part Number
6.35mm port to 4.76mm tube Adapter (compression)	M1273-5042
12mm port to 4.76mm tube Adapter (compression)	M1273-5055
6.35mm port to 6.35mm tube Adapter (set screw)	M1273-5054
12mm port to 6.35mm tube Adapter (set screw)	M1273-5056
12mm Probe Adapter (compression)	M1273-5040
12mm to 12mm Adapter (set screw)	M1273-5058
6.35mm Plug	M1273-9405
12mm Plug	M1273-9406
19mm Plug	M1273-9407

Heaters & Heat Blankets	NBS Part Number
Jacket Water Heater, 1.3L & 3.0L vessels	M1273-3107
Jacket Water Heater, 7.5L & 14L vessels	M1273-3108
Heat Blanket, 1.3L vessel	M1273-9931
Heat Blanket, 3.0L vessel	M1273-9932
Heat Blanket, 7.5L vessel	M1273-9930
Heat Blanket, 14.0L vessel	M1230-9933

Impellers	NBS Part Number
6-bladed Rushton-type, 52 mm, 1.3L & 3.0L	M1273-9291
6-bladed Rushton-type, 59 mm, 7.5L	M1273-9292
6-bladed Rushton-type, 74 mm, 14.0L	M1273-9293
Pitched Blade, 1.3L & 3.0L	M1273-9290
Pitched Blade, 7.5L & 14.0L	M1230-9212
Spin Filters: see following page	

Motors	NBS Part Number
Direct Drive Agitation Motor, Cell Culture, all vessels	M1273-3135
Direct Drive Agitation Motor, Fermentation, 1.3L & 3.0L	M1273-3120
Direct Drive Agitation Motor, Fermentation, 7.5L & 14.0L	M1273-3125
Magnetic Drive Agitation Motor, all vessels	M1273-3130

Spin Filters		NBS Part Number
Spin Filter, Suspension Cells	1.3L	M1273-3201
	3.0L	M1273-3202
	7.5L	M1273-3205
	14.0L	M1273-3210
Spin Filter, Microcarriers	1.3L	M1273-3211
	3.0L	M1273-3212
	7.5L	M1273-3215
	14.0L	M1273-3220

Tubing	NBS Part Number
Blue Polyurethane Tubing, 1/8-inch (3.2 mm) ID, 1/4-inch (6.35 mm) OD, 50 feet	P0740-3111
White Silicone Tubing, 3/16-inch (4.8 mm) ID, 5/16-inch (7.9 mm) OD, 25 feet	P0740-2505
White Silicone Tubing, 1/16-inch ID (1.6 mm), 3/16-inch (4.8 mm) OD, 50 feet	P0740-2396
White, Braided PVC Tubing, 1/4-inch (6.35 mm) ID, 0.438-inch (11 mm) OD, 15 feet	P0740-1631

Probes & Sensors	NBS Part Number
dO ₂ Probe Kit, Ingold polarographic, 1.3L	M1273-9974
dO ₂ Probe Kit, Ingold polarographic, 3.0L	M1273-9879
dO ₂ Probe Kit, Ingold polarographic, 7.5L	M1273-9886
dO ₂ Probe Kit, Ingold polarographic, 14.0L	M1273-9888
dO ₂ Probe, Ingold polarographic, 1.3L	P0720-6281
dO ₂ Probe, Broadley James polarographic, 1.3L	P0720-6370
dO ₂ Probe, Ingold polarographic, 3.0L	P0720-6282
dO ₂ Probe, Broadley James polarographic, 3.0L	P0720-6371
dO ₂ Probe, Ingold polarographic, 7.5L	P0720-6283
dO ₂ Probe, Broadley James polarographic, 7.5L & 14.0L	P0720-6372
dO ₂ Probe, Ingold polarographic, 14.0L	P0720-6284
Foam Level Probe Kit, 1.3L	M1273-9947
Foam Level Probe Kit, 3.0L	M1273-9951
Foam Level Probe Kit, 7.5L & 14.0L	M1273-9960
pH Probe Kit, Ingold gel-filled, 1.3L	M1273-9970
pH Probe Kit, Ingold gel-filled, 3.0L	M1273-9977
pH Probe Kit, Ingold gel-filled, 7.5L	M1273-9982
pH Probe Kit, Ingold gel-filled, 14.0L	M1273-9985
pH Probe, Ingold gel-filled, 1.3L	P0720-5582
pH Probe, Broadley James gel-filled, 1.3L	P0720-5741
pH Probe, Ingold gel-filled, 3.0L	P0720-5584
pH Probe, Broadley James gel-filled, 3.0L	P0720-5747
pH Probe, Ingold gel-filled, 7.5L	P0720-5580
pH Probe, Broadley James gel-filled, 7.5L	P0720-5742
pH Probe, Ingold gel-filled, 14.0L	P0720-5583
pH Probe, Broadley James gel-filled, 14.0L	P0720-5743
RTD Assembly, all vessels	M1273-8019

Sampler System	NBS Part Number
Sampler Kit, 1.3L	M1273-9946
Sampler Kit, 3.0L	M1273-9949
Sampler Kit, 7.5L	M1273-9953
Sampler Kit, 14.0L	M1273-9956
Syringe, 60 mL	P0440-0054
Disposable Filter	P0200-0970

Spargers	NBS Part Number
Sparger Ring, 1.3L	M1273-9259
Sparger Ring, 3.0L	M1273-9256
Sparger Ring, 7.5L	M1273-9246
Sparger Ring, 14.0L	M1273-9251
Porous Sparger, 1.3L, non-jacketed vessel	M1273-5007
Porous Sparger, 1.3L, water-jacketed vessel	M1273-5003
Porous Sparger, 3.0L, jacketed or non-jacketed vessel	M1273-5004
Porous Sparger, 7.5L, jacketed or non-jacketed vessel	M1273-5005
Porous Sparger, 14.0L, jacketed or non-jacketed vessel	M1273-5006

10.2 Replacement Vessels

Part Description	NBS Part Number
1.3L Glass Vessel, non-jacketed	M1273-9907
1.3L Glass Vessel, water-jacketed	M1273-9908
3.0L Glass Vessel, non-jacketed	M1273-9909
3.0L Glass Vessel, water-jacketed	M1273-9915
7.5L Glass Vessel, non-jacketed	M1273-9916
7.5L Glass Vessel, water-jacketed	M1273-9917
14.0L Glass Vessel, non-jacketed	M1273-9918
14.0L Glass Vessel, water-jacketed	M1273-9919

10.3 Miscellaneous Parts & Optional Accessories

Description	NBS Part Number
Addition/Harvest Bottle Kit, 250 mL	M1273-9989
Addition/Harvest Bottle Kit, 500 mL	M1273-9990
Air Filter /Regulator Kit (single manifold for one vessel)	M1230-3030
Air Filter /Regulator Kit (four manifolds for up to four vessels)	M1230-5002
Bearing Housing Cap, disposable (for autoclaving), pack of 10	M1273-9936
O-Ring Kit (washers & O-rings, including headplate), ALL vessels	M1273-9900
Septum Kit for 12mm port	M1273-3031
Spare Parts Kit, 1.3L – 3.0L non- jacketed vessel	M1273-9991
Spare Parts Kit, 7.5L – 14.0L non- jacketed vessel	M1273-9992
Spare Parts Kit, 1.3L – 3.0L water-jacketed vessel	M1273-9998
Spare Parts Kit, 7.5L – 14.0L water-jacketed vessel	M1273-9999
Water Regulator/Filter Kit (single manifold for one vessel)	M1117-2040
Water Regulator/Filter Kit (four manifolds for up to four vessels)	M1273-5001

10.4 Module Kits

Power Controller	NBS Part Number
Kit contains the power controller, 4 cable guides, power cord.	
100-120V	M1273-3100
200-240V	M1273-3110

Primary Control Unit (PCU)	NBS Part Number
Kit contains PCU, display & keypad, floppy disk drive, battery back-up, 2 cable guides, power line.	M1273-3101

dO2/pH Controller	NBS Part Number
Kit contains the dO2/pH controller, 1 Ingold gel pH probe & 1 Ingold polarographic dO2 probe, 2 headplate probe adapters, 2 probe cables, 2 probe shorting caps, module entry cable.	
For 1.3L Vessel	M1273-9850
For 3.0L Vessel	M1273-9851
For 7.5L Vessel	M1273-9852
For 14.0L Vessel	M1273-9853
dO2 /pH Controller only	M1273-3102
Level Controller	NBS Part Number
Kit contains the level controller, 1 level sensor, 1 probe cable, 1 headplate probe adapter, 2 cable guides, module entry cable.	
For 1.3L Vessel	M1273-9841
For 3.0L Vessel	M1273-9842
For 7.5L or 14.0L Vessel	M1273-9843
Level Controller only	M1273-3103

Gas Mix Controller	NBS Part Number
Kit contains gas mix controller, module entry cable, polyurethane tubing, 2 cable guides.	M1273-3104

4-Pump Module	NBS Part Number
Kit contains the pump housing box, 4 peristaltic pumps, power line, 2 cable guides.	
100-120V	M1273-3106
200-240V	M1273-3116

Thermal Mass Flow Controller	NBS Part Number
Kit contains the thermal mass flow controller, module entry cable, 24VDC power line with transformer, polyurethane tubing, 2 cable guides.	
Fermentation application, 20 SLPM	M1273-3109
Cell Culture application, 5 SLPM	M1273-3112

11 APPENDIX A: SPECIFICATIONS

11.1 Modules

BioFlo 110 PCU			
Weight, Overall	5 lb. (2.3 kg)		
Dimensions	Height	6 in. (15.3 cm)	
	Depth*	13¾ in. (35 cm)	
	Width	12 in. (30.5 cm)	
	* allow 3 inches (7.6 cm) more, front and rear, for cable access		
Display	3.4" x 4.5" (86.4 mm x 114.3 mm) electroluminescent monochrome, bright white (on blue screen), ¼ VGA resolution		
Control • Per vessel, up to four vessels total • Up to 32 loops total	Loop Name	Display Range	Control Modes
	Agitation	0-1200 rpm (fermentation) 0-300 rpm (cell culture)	Off, Automatic
	Temperature	0-70.0° C	Off, Automatic
	pH (with dO2/pH Module)	2-12.0 pH units	Off, Automatic: PID, adjustable deadband, liquid additions, 4-gas blending
	dO2 (with dO2/pH Module)	0-100%	Off, Automatic: PID, agitation cascade, sparge-on-demand, 4-gas blending
	Pump 1	0-100.00%	Off, Manual, Automatic (with conduction probes)
	Pump 2	0-100.00%	Off, Manual, Automatic (with conduction probes)
	Pump 3	0-100.00%	Off, Manual, Automatic (with conduction probes)
	Pump 4	0-100.00%	Off, Manual, Automatic (with conduction probes)
	O2 (with Gas Mix Controller)	0-100%	Off, Manual, Automatic
Touchpad		User interface: 24 touch-sensitive buttons	
Internal A: drive		Supports flash e-prom upgrades	
Computer connectivity		Serial port, supports AFS and ModBus protocols	
Recorder connectivity		Four 0-2.5-volt outputs	
Power		85-265 VAC 50/60 Hz, 100 VA	
Power Switch		Located on back.	
Memory Back-Up		Two AA alkaline batteries	
Fusing		2 Amp, 250V Slo-Blo®	
Power Indicator Light		Located on front.	
Other		• Dual Rotameter hangers • Cabling/tubing organizers	

BioFlo 110 Power Controller		
Weight	16 lb. (7.2 kg)	
Dimensions	Height	8 9/16 in. (21.8 cm)
	Depth*	12 in. (30.5 cm)
	Width	12 in. (30.5 cm)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Agitation	Drive	• Lift-off electric motor. • Self-aligning motor and vessel shafts. • Direct drive with surface seal for fermentation and cell culture. • Magnetic drive for cell culture.
	Speed Range	• Fermentation: 50-1200 rpm • Cell culture: 30-300 rpm
	Impellers	Standard Rushton; optional pitched blade
Temperature	Range	5°C above ambient temperature to 70°C
	System Accuracy	20-40°C ± 0.1°C (± 0.5°C at other temperatures)
	Control Stability	0.1°C
	Sensor	One 100 platinum RTD, input located on front of console
	Control	P-I control algorithm
	SSR PWM Outputs	For heat blanket and for cooling solenoid valve
Electrical	115/230 VAC, 50-60 Hz, 1500 VA	
Output Connectors	• Five controlled (PWM) IEC outlets for pumps • Vessel Heater • Vessel Cooling Solenoid Valve • One switched mains outlet for PCU • Agitation motor	
Inputs	• Tachometer (joint with agitation output) • RTD (temperature probe)	
Power Switch	Located on front.	
Pilot Light	Located on front. Illuminates when power is on.	
Other	• Hangers for 2 Rotameters. • Hanger for cooling water valve. • Cable/tubing organizers.	

BioFlo 110 dO2/pH Controller		
Weight	5 lb. (2.3 kg)	
Dimensions	Height	2 5/8 in. (6.7 cm)
	Depth*	12 in. (30.5 cm)
	Width	12 in. (30.5 cm)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Inputs	pH Probe	NBS-supplied cable
	Dissolved Oxygen Probe	Polarographic electrode with NBS-supplied cable.
Control Options	See PCU Specification	
Power	12 VDC (supplied by PCU)	

BioFlo 110 Antifoam/Level Controller		
Weight	4 lb. (1.8 kg)	
Dimensions	Height	4 in. (10.2 cm)
	Depth*	12 in. (30.5 cm)
	Width	12 in. (30.5 cm)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Inputs	Three 3-pin DIN connectors for conduction probes.	<i>Operating Modes</i>
		Antifoam additions
		Wet ON: pump turns on when probe is wetted
		Wet OFF: pump turns off when probe is wetted
Power	12 VDC (supplied by PCU)	

BioFlo 110 Four-Pump Module		
Weight	14 lb. (6.4 kg)	
Dimensions	Height	12 in. (30.5 cm)
	Depth*	12 in. (30.5 cm)
	Width	12 in. (30.5 cm)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Pumps	Four peristaltic type pumps, independently powered	
Power	Line voltage (as labeled), supplied through Power Controller	

BioFlo 110 Gas Mix Controller		
Weight	5 lb. (2.3 kg)	
Dimensions	Height	4 in. (10.2 cm)
	Depth*	10 in. (25.4 cm)
	Width	10 in. (25.4 cm.)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Range	0-100%	
Valves	4 solenoid valves	
Inputs	4 gas inputs (oxygen, air, carbon dioxide and nitrogen)	
Output	1 gas stream output (to sparger)	
Power	12 VDC (supplied by PCU)	
Maximum Gas Pressure	10 PSI	

11.2 Pump Flow Rates

The following table lists pump flow rates according to tubing size, pump frequency and rotation speed.

Tubing ID (silicone only)	50 Hz / 12 RPM	60 Hz / 14.4 RPM
1/50 inch / 0.5 mm	0.25 ml/min	0.30 ml/min
1/32 inch / 0.8 mm	0.59 ml/min	0.71 ml/min
1/16 inch / 1.6 mm	2.62 ml/min	3.14 ml/min
1/8 inch / 3.2 mm	9.60 ml/min	11.5 ml/min
3/16 inch / 4.8 mm	19.6 ml/min	23.5 ml/min

11.3 Tubing

The following tubing sizes are provided with your BioFlo 110 system:

Description	ID	OD	Length	NBS Part Number
Blue polyurethane	1/8 inch 3.2 mm	¼ inch 6.35 mm	50 feet	P0740-3111
White silicone	3/16 inch 4.8 mm	5/16 inch 7.9 mm	25 feet	P0740-2505
White silicone	1/16 inch 1.6 mm	3/16 inch 4.8 mm	50 feet	P0740-2396
White braided PVC	¼ inch 6.35 mm	0.438 inch 11.13 mm	15 feet	P0740-1631

11.4 Vessels

BioFlo 110 Non-Jacketed Vessels						
Vessel Volume	Total Volume		1.3 L	3.0 L	7.5 L	14.0 L
	Working Volume	Max.	1.0 L	2.2 L	5.6 L	10.5 L
		Min.	0.4 L	0.8 L	1.5 L	4.0 L
Vessel Weight	with Headplate, Stand, Drive, Inserts & Plugs		15 lb. (6.8 kg)	20.5 lb. (9.3 kg)	39.5 lb. (18 kg)	43 lb. (19.5 kg)
Stand	Non-Jacketed Vessels		Stainless steel, four legs for stability, high friction feet to eliminate "walking".			
Aeration	Max. Flow Rate		Fermentation: 2 VVM Cell Culture: 0.5 VVM			
	Sparger		Stainless steel ring sparger with removable EPDM plug			
	Maximum Sparge Pressure		10 PSIG			
	Filters		0.2µ absolute filters, resterilizable			
Exhaust	Condenser		Optional, mounted in headplate			
	Filter		0.2µ absolute filter, resterilizable			

11.4.1 Non-Jacketed Vessel Dimensions for Autoclaving (Vertical)

Non-Jacketed Vessel Assembly with supplied vertical stand & as prepared for autoclaving				
Total Vessel Volume	1.3 L	3.0 L	7.5 L	14.0 L
Height <i>without</i> Exhaust Condenser (to top of bearing housing)	16.5 in. (42 cm)	16.5 in. (42 cm)	19.5 in. (49.5 cm)	24 in. (61 cm)
Diameter <i>without</i> Exhaust Condenser	8.5 in. (22 cm)	8.5 in. (22 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)

...continued...

Non-Jacketed Vessel Assembly with supplied vertical stand & as prepared for autoclaving				
Total Vessel Volume	1.3 L	3.0 L	7.5 L	14.0 L
Height <i>with</i> Exhaust Condenser (to top of bent-over exhaust filter)	22 in. (56 cm)	22 in. (56 cm)	25.5 in. (65 cm)	29 in. (74 cm)
Width <i>with</i> Exhaust Condenser	8.5 in. (22 cm)	8.5 in. (22 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)
Length <i>with</i> Exhaust Condenser	9.5 in. (24.1 cm)	9.5 in. (24.1 cm)	14.5 in. (37 cm)	11.5 in. (29 cm)

The above dimensions do not include optional addition bottles or optional sampler, which may be mounted to the vessel assembly for autoclaving; they can also be autoclaved unattached.

- If addition bottles are attached, add the following approximate dimensions to vessels *without* exhaust condenser:
 - +4 cm to diameter
 - +8 cm to height of 1L and 3L vessels
 - +10 cm to height of 7.5L and 14L vessels.
- Addition bottles add no height to vessels with condenser.
- Sampler adds approximately +4 cm to diameter.

11.4.2 Non-Jacketed Vessel Dimensions for Autoclaving (Inclined)



NOTE:

For use with an Angled Autoclave Rack, the attachment of addition bottles is not recommended.

Non-Jacketed Vessel Assembly with optional Angled Autoclave Rack (M1273-9266) & as prepared for autoclaving		
Total Vessel Volume	7.5 L	14 L
Height <i>with</i> Exhaust Condenser	18 in. (46 cm)	20 in. (51 cm)
Length <i>with</i> Exhaust Condenser	28 in. (71 cm)	31.5 in. (80 cm)
Width <i>with</i> Exhaust Condenser	11.5 in. (29 cm)	11.5 in. (29 cm)

11.4.3 Water-Jacketed Vessel Dimensions for Autoclaving (Vertical)

Water-Jacketed Vessel Assembly as prepared for autoclaving				
Total Vessel Volume	1.3 L	3.0 L	7.5 L	14.0 L
Height <i>without</i> Exhaust Condenser (to top of bearing housing)	16 in. (41 cm)	18 in. (46 cm)	20.5 in. (52 cm)	27 in. (69 cm)
Diameter <i>without</i> Exhaust Condenser	9.25 in. (23.5 cm)	9.25 in. (23.5 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)
Height <i>with</i> Exhaust Condenser (to top of bent-over exhaust filter)	20.5 in. (52 cm)	22.5 in. (57 cm)	24.5 in. (62 cm)	31 in. (79 cm)
Width <i>with</i> Exhaust Condenser	9.5 in. (24.1 cm)	9.5 in. (24.1 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)
Length <i>with</i> Exhaust Condenser	11.5 in. (29 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)	11.5 in. (29 cm)



CAUTION!

Never autoclave the jacket water heater. Unplug and remove it.

The above dimensions do not include optional addition bottles or optional sampler, which may be mounted to the vessel assembly for autoclaving; they can also be autoclaved unattached.

- If addition bottles are attached, add the following approximate dimensions to vessels *without* exhaust condenser:
 - +4 cm to diameter
 - +8 cm to height of 1L and 3L vessels
 - +10 cm to height of 7.5L and 14L vessels.
- Addition bottles add no height to vessels with condenser.
- Sampler adds approximately +4 cm to diameter.

11.4.4 Non-Jacketed Vessel Dimensions for Autoclaving (Inclined)



NOTE:

For use with an Angled Autoclave Rack, the attachment of addition bottles is not recommended.

Water-Jacketed Vessel Assembly with optional Angled Autoclave Rack (M1273-9266) & as prepared for autoclaving		
Total Vessel Volume	7.5 L	14 L
Height <i>with</i> Exhaust Condenser	18.5 in. (47 cm)	20.5 in. (52 cm)
Length <i>with</i> Exhaust Condenser	26 in. (66 cm)	28.5 in. (72 cm)
Width	11.5 in. (29 cm)	11.5 in. (29 cm)

**CAUTION!**

Never autoclave the jacket water heater. Unplug and remove it.

12 APPENDIX B: TECHNICAL INTERVENTIONS



NOTE:

Only a qualified technician should perform the procedures contained in this chapter.



WARNING!

Always disconnect the power cord from the Power Controller and the PCU before undertaking any of the following procedures.

12.1 Replacing Power Controller Fuses

There are eight replaceable fuses inside the power controller:

Fuse Number	Item Controlled	Fuse Number	Item Controlled
F1	Pump A	F5	Acid Pump
F2	Pump B	F6	Water Solenoid Valve
F3	Pump C	F7	Heater
F4	Base Pump	F8	Heater

To replace one of these fuses:

1. Unplug all power cords from the system.
2. Disconnect the Power Controller from any other modules.
3. Unplug the water solenoid valve from the rear panel.
4. Unplug the motor, RTD and heater cables from the front panel.
5. Loosen the two screws that secure the water solenoid valve assembly in place, then remove the assembly by sliding the mounting plate up until the slotted holes allow you to lift the assembly away from the screws.
6. Unscrew the ten side mounting screws that hold the Power Controller cover in place. Retain them for reuse.
7. Carefully pull the cover straight up until it separates from the base. There will be some resistance until it comes unclipped. Be careful not to pull on the ground wire that is attached to the front of the left side (from the rear).
8. Rest the cover on its front panel, adjacent to the base.

9. Grasp the fuse and pull it straight up until it separates from its base.
10. Insert a new fuse, taking care to align its two pins with the pinholes in the base. (If you are replacing one of the heater fuses, use a continuity meter to determine which fuse is blown.)
11. Ease the cover back in place, lowering it straight down, until it fits tightly. Take care to tuck the ground wire inside, clear of both the cover and the lip of the base.
12. Reinstall the side mounting screws. Tighten securely.
13. Reinstall the water solenoid valve assembly, tightening the screws against the mounting plate.
14. Reconnect the motor drive, RTD and heater to the front panel.
15. Reconnect the water solenoid valve to the rear panel.
16. Reconnect the Power Controller to the other module(s).
17. Reconnect the power cord.

**WARNING!**

Always disconnect the power cord from the Power Controller and PCU before undertaking the following procedures.

12.2 Replacing PCU Batteries

1. Unplug all power cords from the system.
2. Disconnect the PCU from any other modules.
3. Unscrew the eight side mounting screws that hold the cover in place. Retain them for reuse.
4. Carefully pull the cover straight up until it separates from the base. Be careful not to pull on the ground wire that is attached to the front of the left side (from the rear), or on the ribbon cables that are attached to the front of the cover.
5. Rest the cover on its front panel, adjacent to the base.
6. The two AA alkaline batteries are housed in the battery holder at the front of the base. Cut the plastic tie wrap. Remove, then replace the batteries, taking care to align them properly (the polarity is marked in the bottom of the battery holder).
7. Replace the tie wrap (this is optional).
8. Ease the cover back in place, lowering it straight down, until it fits tightly. Take care to tuck the ground wire and the ribbon cables inside, clear of both the cover and the lip of the base.
9. Reinstall the side mounting screws. Tighten securely.

**WARNING!**

Always disconnect the power cord from the Power Controller and PCU before undertaking any of the following procedures.

12.3 Voltage Selection

NOTE:


If you move your BioFlo 110 to a new location that requires a change in voltage and/or fusing, be sure to consult your NBS representative.

Locate the fuse drawer on the back panel of the Power Controller, next to the power plug. The voltage selection is displayed in a small window in the middle of the fuse drawer. The Power Controller is preset at 115 Volts.

If the voltage needs to be changed to 230V:

1. Using a narrow-tipped, flat-bladed screwdriver, introduce the blade into the small slot to the left of the fuse drawer.

Figure 46: Voltage Selection



2. Use the blade to pry the lip of the plastic lock from left to right, then gently lift up toward you. The drawer will pop out.
3. Pull the entire drawer assembly out and set it aside.
4. Use the screwdriver blade to pull the selector cube out (from the left side, insert the blade into the hollow of the cube).
5. Turn it 180° to display **115**, then snap it back into place. Be sure to press it firmly until it snaps.
6. Reinsert the fuse drawer, and snap it into place.

**NOTE:**

If you are using pumps (with or without the Pump Module), you will need to replace them with pumps configured for the new voltage.

**NOTE:**

Although the voltage selector is located in a housing termed “fuse drawer”, it contains no fuses. The fuse drawer contains a spring and a blank. Be sure to keep these intact and in place when you reinstall the fuse drawer.

12.4 Changing PCU Fusing

If you are changing the voltage on your Power Controller, it may also be necessary to modify the fusing arrangement on the PCU.



**NOTE:**

Be sure to discuss this with your NBS representative before a qualified technician makes any modifications to the unit.

12.4.1 Converting from Single Fuse to Double Fuse

Your BioFlo 110 PCU was initially set up with a single fuse arrangement. To change to a double fuse:

1. Open the cover of the power module using a small blade screwdriver and remove the cover/fuse block assembly (*see Figure 47 below*).
2. Loosen the Phillips screw behind the fuse block 2 full turns(*see Figure 48 below*).
3. Remove the fuse block by sliding it up and away from the screw shaft and lifting it off the pedestal (*see Figure 48 below*).
4. Invert the fuse block and slide it back onto the Phillips screw and pedestal. Tighten the Phillips screw.

Figure 47: Fuse Holder/Power Inlet**Figure 48: Fuse Block/Cover Assembly**

5. Verify the correct fusing arrangement (*see Figures 49a and 49b below*).
6. Insert fuses.
7. Replace the cover unit into the fuse holder.

Figure 49a: Double Fusing Arrangement





Figure 49b: Single Fusing Arrangement



12.5 P&I Control Equations



NOTE:

The following information is provided for the use of advanced P & I algorithm users only.

12.5.1 pH

As mentioned in Section 7.6.3, following is the P & I control equation for pH:

$$\text{Out}\% = P * \frac{100\%(SP - CV)}{14\text{pH}} + I * \frac{100\%}{14\text{pH}} \int (SP - CV) dt$$

Where P is dimensionless and I is in min^{-1} .

12.5.2 dO2

As mentioned in Section 7.9, following is the P & I control equation for dO2:

$$\text{Out\%} = P * \frac{100\%}{100dO_2} (\text{SP-CV}) + I * \frac{100\%}{100dO_2} \int (\text{SP-CV}) dt$$

Where P is dimensionless and I is in min⁻¹.

13 APPENDIX C: COMMUNICATIONS PROTOCOLS



CAUTION!

This section is provided FOR ADVANCED PROGRAMMERS ONLY. Do not attempt to apply or change any of this information unless you are professionally trained in these matters. Interventions by unskilled operators may severely compromise the operation of your BioFlo 110 system and may also void the warranty.

13.1 Communications Settings

The onboard computer has two RS-232 serial ports, labeled Comm1 and Comm2. Both ports are connected to the keyboard interface board (KIB), which converts them to RS-485.

Comm1 is set up for connection to an NBS AFS (BioCommand) supervisory software interface box. (See also Section 13.2.)

Comm1 can also be used for connection to a computer equipped with an RS-485 (ModBus) port. This allows the operator to control the BioFlo 110 with the remote computer. (See also Section 13.3.)

Comm1 Settings	For AFS Use	For ModBus Use
Baud Rate	9600	19200
Data Bits	8	8
Stop Bits	2	1
Parity	Even	None

Comm2 is set up to communicate with the other BioFlo 110 control modules, which are daisy-chained together. Terminators **must** be installed on any and all unused connectors.

Comm2 Settings	
Baud Rate	38400
Data Bits	8
Stop Bits	1
Parity	None

13.1.1 AFS/ModBus Pin Connections

The AFS/ModBus 25-pin “D” connector on the rear panel is pinned out as follows:

Pin Number	Signal	Observations
13	RS422 Tx+	
25	RS422 Tx-	
24	RS422 Rx+	
12	RS422 Rx -	
7	Ground	
2	RS232 Tx	
3	RS 232 Rx	
21	Mode	• Open for RS232. • Connect to 7 for RS422

13.2 AFS (BioCommand) Communications Protocol

This RS422 communications protocol specification is for “Slave” equipment. “Slave” refers to any instrument that is interfaced to the NBS BioCommand (also called AFS) system, and “Master” refers to the BioCommand supervisory interface module.

13.2.1 General Specifications

- All line drivers and receivers must conform to IEEE RS-422 or RS-485 standards.
- Communications settings are listed in Section 13.1 above. There is no CTS/RTS or XON/XOFF handshaking.
- The Master initiates all transactions.
- A Slave transmit driver must be held in a high impedance state (tri-stated) at all times, except when it is responding to a transaction initiated by the Master.
- All transactions are carried out sequentially (one at a time).
- All transactions are prefixed with a single byte that indicates the multidrop ID of the Slave station involved.
- Valid Slave unit ID numbers are 0 through 99.
- All transmissions from the Master are terminated with a carriage return (0DH) character.
- All responses to data requests are terminated with a carriage return (0DH) character.
- The Slave shall respond to a Change command from the Master with an ACK (06H) character.
- If the Slave detects an error in the transmission from the Master, it shall make no response.
- If the Slave successfully receives a transmission from the Master, it must begin the appropriate response in less than 200 mSec.

13.2.2 Protocol

- All Slaves must respond to the Requests for data shown in Section 13.2.3 below.
- The Response Header for each Request must be as shown in Section 13.2.3 below.
- All Headers must contain 9 characters; where necessary, spaces are used to complete the 9 characters. In Section 13.2.3, a space is represented by “(Sp)”.
- The format of the Response must be as shown in Section 13.2.5 below.
- All Messages consist of a Header plus 10 Fields.
- All Fields must contain 6 characters; where necessary, spaces are used to complete the 6 characters.
- If there is no data for a particular Field, the Slave must fill that Field with space (20H) characters.
- All Slaves must respond to Change commands as shown in Section 13.2.4 below.
- The Headers for each Change command are shown in Section 13.2.4 below.
- The Slave response to a Change command shall be an ACK (06H) character.
- Even if a Change command is not applicable or not implemented by the Slave, the Slave must still acknowledge the command.
- When a Field in a Change command is blank, the Slave shall not change the settings for the loop or parameter associated with that Field.
- Numeric values in Fields must be ASCII decimal format.

13.2.3 Requests & Response Headers

Request Type	Request Format	Response Header†
Get Loop Names *	(MD#)RA(CR)	Loop(Sp)(Sp)(Sp)(Sp)(Sp)
Get Current Values *	(MD#)RC(CR)	Cur:Value
Get Setpoints	(MD#)RD(CR)	OP Points
Get Controller Outputs	(MD#)RE(CR)	Output(Sp)%(Sp)
Get Control Modes *	(MD#)RJ(CR)	Control(Sp)(Sp)
Get Units *	(MD#)RL(CR)	Main(Sp)Unit

* All controllers must respond to these four requests or they will not be recognized by BioCommand.

† All Headers must contain a total of 9 characters

(MD#) = Multidrop ID Number (CR) = Carriage Return (Sp) = Space

13.2.4 Change Commands

Change Command	Command Header†
Change Setpoints	MS(Sp)(Sp)(Sp)(Sp)(Sp)(Sp)(Sp)
Change Controller Output	MO(Sp)(Sp)(Sp)(Sp)(Sp)(Sp)(Sp)

† All Headers must contain a total of 9 characters (Sp) = Space

13.2.5 Message Formats

Response and Change command message format:

MD#	Header 9 chars	S p a c e	Field 1 6 chars	S p a c e	Field 2 6 chars	S p a c e	Field 3 6 chars	S p a c e	. . .	Field 9 6 chars	S p a c e	Field 10 6 chars	S p a c e	CR
-----	-------------------	-----------------------	--------------------	-----------------------	--------------------	-----------------------	--------------------	-----------------------	-------	--------------------	-----------------------	---------------------	-----------------------	----

All Headers must contain a total of 9 characters.

All Fields must contain a total of 6 characters.

(MD#) = Multidrop ID Number Chars = Characters (CR) = Carriage Return

Request format:

MD#	Request 2 chars	CR
-----	--------------------	----

(MD#) = Multidrop ID Number Chars = Characters (CR) = Carriage Return

Change command acknowledge format:

MD#	ACK
-----	-----

(MD#) = Multidrop ID Number ACK = Acknowledgement character

13.3 ModBus Communications Protocol

- All BioFlo110 parameters are represented by Register assignments (see Register Map in Section 13.3.1 below).
- The BioFlo 110 implements Modbus functions 03, 06 and 16 decimal.
- All registers are transmitted high byte first then low byte (standard Modbus format).
- All floating point values are stored in two-Register pairs, with the higher half in the first (even) Register and the lower half in the second (odd) Register. Therefore, floating point values are transmitted as four bytes, starting with the highest and ending with the lowest.

- Character strings are stored in registers, with one character per register. The high byte of the returned Register value will contain 0 and the low byte will contain the character code.
- Integer values are stored in a single Register.

CONTROL TYPE VALUES	
Control Type	Value
OFF	= 0X00
Auto	= 0X01
Manual	= 0X02
ON/OFF	= 0X03
P-DB*	= 0X04
Total	= 0X0B

* P-DB means Proportional with Deadband

13.3.1 Register Map



NOTE:

Programmers take note that all of the following are decimal addresses.

Loop Number:	0	1	2	...	31	Formula
Parameters	Decimal Address					
1-Curr_val:	0	2	4	...	62	$0 + (N \text{ Loop} * 2)$
2-Setpoint:	64	66	68	...	126	$64 + (N \text{ Loop} * 2)$
3-Control_out:	128	130	132	...	190	$128 + (N \text{ Loop} * 2)$
4-Integral:	192	194	196	...	254	$192 + (N \text{ Loop} * 2)$
5-Derivative:	256	258	260	...	318	$256 + (N \text{ Loop} * 2)$
6-Raw_float:	320	322	324	...	382	$320 + (N \text{ Loop} * 2)$
7-Cal_gain:	384	386	388	...	446	$384 + (N \text{ Loop} * 2)$
8-Zero_std_off:	448	450	452	...	510	$448 + (N \text{ Loop} * 2)$
9-Eng_units:	512	516	520	...	636	$512 + (N \text{ Loop} * 4)$
10-Raw_input:	642	643	644	...	673	$642 + (N \text{ Loop})$
11-Cal_offset:	676	677	678	...	707	$676 + (N \text{ Loop})$
12-Loop_name:	740	756	772	...	1236	$740 + (N \text{ Loop} * 16)$
13-Control_mode:	1268	1276	1284	...	1516	$1268 + (N \text{ Loop} * 8)$



NOTE:

Parameters 1 through 8 use two registers each, Parameter 9 uses four registers, Parameters 10 & 11 use one register each, Parameter 12 uses sixteen registers and Parameter 13 uses eight registers.

14 APPENDIX D: THERMAL MASS FLOW CONTROLLER

The Thermal Mass Flow Controller (TMFC) replaces the rotameter. The TMFC will give you the ability to display gas flow rate on the PCU. This digital gas flow rate can also be logged to data acquisition software, such as *BioCommand Lite* or *BioCommand Plus*.

When you use it with the dO2/pH Controller, you can automatically adjust the gas flow into the system, using either direct operator interaction, setpoint manipulation, or cascade control from the DO loop.

 NOTE:

In Section 1 of this manual, you read that the PCU serves as the operator interface for one to four vessels. Because the PCU cannot support more than a total of 16 modules, adding the Thermal Mass Flow Controller limits the PCU to a maximum of three vessels.

Section 2 (*Inspection & Unpacking*), **Section 3** (*Preparing the Location*) in this manual provide essential information; be sure to read them prior to installing this optional module.



WARNING!

Be sure you are familiar with all warnings and precautions in the body of this manual, which are provided for your safety and for the protection of your BioFlo 110 system.

14.1 Installation

The Thermal Mass Flow Controller (TMFC) has the same footprint as the Level Controller. You can place it on your module stack, or beside other modules.

With reference to Figure 8 (*Inconnecting the Modules*), repeated on the following page for reference, and Figure 50 (*TMFC Module Rear Panel*), daisy chain the TMFC module to the others in your stack. Remember that it is powered by the Power Controller and controlled by the PCU, so it must be daisy chained.

Figure 8: Interconnecting the Modules





Figure 50: Thermal Mass Flow Controller – Rear Panel


**NOTE:**

Be sure to check the setting of the Unit Address Selector on the rear panel, to be sure it is the same as the other control modules for this system. For details, see Section 5.2.

With reference to Figure 51 below:

1. **If you have the Gas Mix module:** connect the Gas Mix module output to the TMFC Gas Input; **if you do not have the Gas Mix module:** connect your air supply to the TMFC Gas Input.
2. Connect the TMFC Gas Output connector to the sparger inlet with the tubing provided.

Figure 51: TMFC Front View




14.2 Loading Software Update

If you purchased the optional TMFC module with the rest of your BioFlo 110 system, the supervisory software is fully configured for its presence; **skip to Section 14.3, Operation.**

If you are adding the TMFC module to a pre-established BioFlo 110 system:

1. In the PCU's floppy drive (*see Figure 1*), load the update disk shipped with the TMFC.
2. After the system boots, then follow any instructions that appear on the display screen.
3. When the initial setup screen appears (*see sample on the following page*), you are ready for operation.



14.3 Operation

Before proceeding through the following instructions, be sure to familiarize yourself with the PCU display, selector buttons, and the various interface screens described in Section 7 (*Operation*) of this manual.

With a dO₂/pH Controller is present in the system, there are four possible controller combinations that determine the operation of the eighth loop on the fermentor, as recapped in the following table:





Module(s) Present		8th Loop	Operation
Gas Mix	TMFC		
NO	NO	N/A	Only option for dO ₂ control is a cascade to the Agitation loop.
YES	NO	"O ₂ "	As explained in the Operation section of this manual.
NO	YES	"Gas"	Pressing the selector key for the 8th loop opens the <i>Gas Loop Details</i> screen, where you can set the Control mode to Auto or Off ; enter Setpoint (0-20 SLPM in Fermentation mode or 0-5 SLPM in Cell Culture mode) & set Hi and Lo display limits. The <i>dO₂ Setup</i> screen shows Gas Flow cascade Hi and Lo limits in addition to the above information. Press the appropriate selector keys to change these limits. Pressing the Cascade key in the <i>dO₂ Setup</i> screen reveals dO ₂ cascade selections of Agitation ("Agit"), Gas Flow ("Gas"), Agitation/Gas Flow ("Agit/Gas"), or None .

...continued...

Module(s) Present		8th Loop	Operation
Gas Mix	TMFC		
YES	YES	"Gas"	<u>In Fermentation mode</u>, the system is configured for gas flow control with oxygen supplementation. Press the Gas loop selector key to open the <i>Gas Control</i> screen, where you can use the selector keys to set these parameters for gas flow: Control mode (Auto or Off), Setpoint (0-20 SLPM), oxygen enrichment (O2 Mix) Control mode and output percentage (O2%). The <i>dO2 Setup</i> screen shows Gas Flow cascade Hi and Lo limits in addition to the above information. Press the appropriate selector keys to change these limits. Pressing the Cascade key in the <i>dO2 Setup</i> screen reveals dO2 cascade selections of Agitation ("Agit"), Gas Flow ("Gas"), Gas Flow/Oxygen Enrichment ("Gas/O2"), Agitation/Gas Flow ("Agit/Gas"), Agitation/Gas Flow/Oxygen Enrichment ("Ag/Ga/O2") or None.
YES	YES	"Gas"	<u>In Cell Culture mode</u>, the system is configured for dO2/pH control using the NBS gas mixing algorithm (see <i>Section 7.9</i>) and separate gas flow control. Press the Gas loop selector key to open the <i>Gas Mix</i> screen. This screen works as explained in <i>Section 7.11.1</i>, but there is also one additional element: Gas Flow control. Pressing the <i>Gas Flow</i> selector key allows you to select the gas flow Control mode and Setpoint (0-5 SLPM). The <i>dO2 Setup</i> screen shows Gas Flow cascade Hi and Lo limits in addition to the above information. Press the appropriate selector keys to change these limits. Pressing the Cascade key in the <i>dO2 Setup</i> screen reveals the same dO2 cascade possibilities as with the Gas Mix module alone: 4-Gas ("4 Gas"), On Demand ("Demand"), Agitation ("Agit") or None (see <i>Section 7.8.4</i>).

14.3.1 Fermentation Mode Setup

1. In the main screen (see sample screen on the following page), select the *Gas* loop, which represents oxygen and air:

B i o F l o 1 1 0	Name	Value	Setpoint	Control	
	Temp	23.4	30.0 °C	Off	
	Agit	134	350 rpm	Auto	
	Pump A	0.0	0.0 %	Manual	
	Pump B	0.0	0.0 %	Manual	
	Pump C	0.0	0.0 %	Manual	
	PH	6.68	7.00 pH	Auto	
	dO2	77.1	0.0 %	Auto	
	Gas	0.3	20.0 SLPM	Auto	

Press here 

- In the *Gas* loop screen that opens, set the *O2 Mix* in *Auto* Mode, as reflected in the resulting screen:

B i o F l o 1 1 0	Gas 0.0		SLPM	
	Control: Off		Control	
	Setpoint: 0.0		Setpoint	
	Output: 0.0			
	O2 Mix: Auto		O2 Mix	
	Output: 0.0		Set O2%	

- The default gas flow setpoint is 20. If you wish to change it, select *Setpoint*, then type in the desired value (from 0 to 20 SLPM—the system will automatically limit the gas setpoint to 20).
- Select *Set O2%* to set the percentage of oxygen.

 NOTE:

After you set the *O2%*, the PCU will provide the complementary percentage of air. For example, if you set *O2%* at 25, the PCU will provide 75% air.

B i o F l o 1 1 0	Gas 0.0		SLPM	
	Control: Off Setpoint: 18.0 Output: 0.0			Control Setpoint
	O2 Mix: Auto Output: 100.0			
				O2 Mix Set O2%

- Finally, press the *Control* selector button, then set the *Gas* loop control mode to *Auto*. You will hear the solenoid click, and gas will come through the sparger.

 NOTE:

It is important to set the O2 Mix and O2% *before* turning the Gas loop control on, to avoid a sudden burst of large bubbles into the vessel.

14.3.2 Fermentation Mode Cascades

Before you set cascades, first review the table at the top of Section 14.3 to identify the cascades available, according to the combination of modules in your system.

If you have the TMFC without the Gas Mix Module, the *dO2 Cascade* screen will look like this:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To: None		Agit
	Agit Hi Limit: 1000		Gas
	Agit Lo Limit: 250		Agit/Gas
	Gas Hi Limit : 20.0		
	Gas Lo Limit: 5.0		None
	P-Gain: 0.05		
	I-Gain: 0.25		

If you have the TMFC with the Gas Mix Module, the *dO2 Cascade* screen will look like this:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To:	None	
	Agit Hi Limit:	1000	
	Agit Lo Limit:	250	
	Gas Hi Limit :	20.0	
	Gas Lo Limit:	5.0	
	P-Gain:	0.05	
	I-Gain:	0.25	
			Agit
			Gas
			Gas/O2
			Agit/Gas
			Ag/Ga/O2
			None


Make your choices as explained in Section 7.8.2.

14.3.3 Cell Culture Mode Setup

1. In the main screen, select the *Gas* loop. The *Gas* loop screen opens:

B i o F l o 1 1 0	Gas 0.0		SLPM	Gas Flow
	Control:	Off		4 Gas
	Gas Mix:	Off		Demand
	Gas	%		Off
	CO2	0.0		100% CO2
	N2	0.0		100% N2
	O2	0.0		100% O2
	Air	0.0		100% Air

2. While the *Control* mode and *Gas Mix* are *Off*, select *%Air* (or any other gas) to change the gas mix.
3. Select *Gas Flow*, then use the following screen to select a new gas flow *Setpoint*, if so desired:



4. The default setpoint is 5. If you wish to change it, select *Setpoint*, then type in the desired value (from 0 to 5 SLPM—the system will automatically limit the gas setpoint to 5).
5. If you wish to adjust the display's Y axis, use the *Disp Hi* and *Disp Lo* selections (as explained in Section 7.4.3).
6. Finally, set *Control* to *Auto* to turn on the gas flow. You will hear the solenoid valve click, and gas will come through the sparger.

NOTE:

It is important to set the *O2 Mix* and *O2%* **before** turning the Gas loop control on, to avoid a sudden burst of large bubbles into the vessel.

7. Press the **X** key to return to the Gas loop screen. The *Demand* and *4 Gas* options work as explained in Sections 7.11.1 and 7.11.2. The presence of the TMFC does not affect these controls.

14.3.4 Cell Culture Mode Cascades

Before you set cascades, first review the table at the top of Section 14.3 to identify the cascades available, according to the combination of modules in your system.

If you have the TMFC without the Gas Mix Module, the *dO2 Cascade* screen will look like this (see sample screen on following page):

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To:	None	Agit
	Agit Hi Limit:	150	Gas
	Agit Lo Limit:	25	Agit/Gas
	Gas Hi Limit :	5.0	None
	Gas Lo Limit:	0.0	
	P-Gain:	0.05	
	I-Gain:	0.25	

If you have the TMFC with the Gas Mix Module, the *dO2 Cascade* screen will look like this:

B i o F l o 1 1 0	dO2 73.1 %		
	Cascade To:	None	4 Gas
	Agit Hi Limit:	150	Demand
	Agit Lo Limit:	25	Agit
	<u>Advanced Control</u>		None
	P-Gain:	0.05	
	I-Gain:	0.25	

Make your choices as explained in Section 7.8.4.

14.4 Cleaning

See Section 8.1 for instructions on cleaning the module.

14.5 Specifications

BioFlo 110 Thermal Mass Flow Controller		
Weight	4 lb. (1.8 kg)	
Dimensions	Height	4 in. (10.2 cm)
	Depth*	12 in. (30.5 cm)
	Width	12 in. (30.5 cm.)
	*allow 3 inches (7.6 cm) more, front and rear, for cable access	
Range	0-100%	
Input	1 gas input	
Output	1 gas output	
Power	24 VDC power transformer (supplied)	
Maximum Gas Pressure	10 PSI	

14.6 Troubleshooting

If no information from the Thermal Mass Flow Controller appears on the display, check the following *before you call for service*:

1. Verify that the 24VDC power transformer cable is properly connected to the TMFC and to a grounded electrical outlet.
2. Confirm that the TMFC is properly connected (daisy-chained with RS-485 cables) to the Power Controller, dO₂/pH Controller and the PCU. There must be a dO₂/pH Controller or the TMFC will not be recognized by the PCU.
3. Check that the Power Controller's power switch is turned On.
4. Verify that the PCU power switch is turned On.
5. Verify that the TMFC's unit address selector switches are correctly assigned (per Section 5.2).



NOTE:

If gas **Output**, which is now controlled by the TMFC, does not stabilize, check the levels in your supply tanks.

15 APPENDIX E: DRAWINGS

15.1 Module Schematics

Figure 52: PCU Schematics




Figure 54: dO2/pH Controller Schematics




Figure 55: Level Controller Schematics




Figure 56: Gas Mix Controller Schematics





Figure 57: 4-Pump Module Schematics (110V)



Figure 58: 4-Pump Module Schematics (220V)

Figure 59: Thermal Mass Flow Controller Schematics



15.2 List of Drawings

Figure Number	Caption	Page
1	PCU Rear Panel	5
2	Power Controller Rear Panel	6
3	dO2/pH Controller Rear Panel	7
4	Level Controller Rear Panel	8
5	Gas Mix Controller Rear Panel	8
6	4-Pump Module (Rear View)	9
7	Sample BioFlo 110 Installation	11
8	Interconnecting the Modules	13, 144
9a	WRONG Handling of Drive Assembly	15
9b	CORRECT Handling of Drive Assembly	15
10	Non-Jacketed Vessel Assembly	17
11	1.3L [non-jacketed] Headplate	18
12	3.0L [non-jacketed] Headplate	19
13	7.5L & 14.0L [non-jacketed] Headplate	20
14	Upper Vessel Bumper Installation	21
15	pH Probe with Port Adapter (exploded)	26
16	dO2 Probe with Port Adapter (exploded)	27
17	Water Solenoid Valve	29
18a	Exhaust Condenser (1.3L, 3.0L & 7.5L Vessels)	29
18b	Exhaust Condenser (14.0L Vessel only)	30
19a	Sampler/Harvest System (1.3L Vessel)	31
19b	Sampler System (3.0L, 7.5L & 14.0L Vessels)	31
20	Foam Trap	33
21	Water-Jacketed Vessel Assembly	36
22	1.3L [jacketed] Headplate	37
23	3.0L [jacketed] Headplate	38
24	7.5L & 14.0L [jacketed] Headplate	39
25	Installing the Headplate Clamping Ring	40
26	Water Jacket Guard Installation (top view)	41
27	pH Probe with Port Adapter (exploded)	45
28	dO2 Probe with Port Adapter (exploded)	46
29	Water Solenoid Valve	48
30a	Exhaust Condenser (1.3L, 3.0L & 7.5L Vessels)	49
30b	Exhaust Condenser (14.0L Vessel only)	49
31a	Sampler/Harvest System (1.3L Vessel)	50
31b	Sampler System (3.0L, 7.5L & 14.0L Vessels)	51
32	Foam Trap	52
33	Motor Installation	55
34	Rotameter—Side View	56
35	Power Controller Front Panel	59
36	dO2/pH Controller Front Panel	60
37	Level Controller Front Panel	61
38	4-Pump Module Front View	62

Figure Number	Caption	Page
39	Gas Mix Controller Front Panel	63
40	Cable Guide	64
41	Angled Autoclave Rack Option for 7.5L & 14.0L Vessels	78
42	Jacket Water Heater—Bottom View	83
43	PCU Front Panel	84
44	PCU Touchpad	85
45	Fuse Holder/Power Inlet (PCU)	117
46	Voltage Selection	133
47	Fuse Holder/Power Inlet (PCU)	135
48	Fuse Block/Cover Assembly (PCU)	135
49a	Double Fusing Arrangement (PCU)	136
49b	Single Fusing Arrangement (PCU)	136
50	Thermal Mass Flow Controller Rear Panel	144
51	Thermal Mass Flow Controller Front View	145
52	PCU Schematics	154
53	Power Controller Schematics	155
54	dO ₂ /pH Controller Schematics	156
55	Level Controller Schematics	157
56	Gas Mix Controller Schematics	158
57	4-Pump Module Schematics (110V)	159
58	4-Pump Module Schematics (220V)	160
59	Thermal Mass Flow Controller Schematics	161

16 APPENDIX F: PORTS & ADAPTERS

One of the flexibility features of the BioFlo 110 is the variety of adapters that allow you to set up the headplate to suit your needs. The summary below will help you determine any adjustments you may wish to make to the recommended headplate arrangement.

PORT SIZE	APPLICATION	PART NUMBER
PLUGS		
6.35mm	Unused port	M1273-9405
12mm	Unused port	M1273-9406
19mm	Unused port	M1273-9407
INSERTS		
12mm	Septum	M1273-3031
12mm	Thermowell/RTD	M1273-9200 (1.3L) M1273-9201 (3.0L) M1273-9202 (7.5L) M1273-9203 (14.0L)
12mm	Tri-port	M1273-9961
SET SCREW ADAPTERS:		
 CAUTION! Never use a set screw on a probe.		
12mm port to 12mm tube	For Exhaust Tube or Condenser	M1273-5058
6.35mm port to 6.35mm tube	For Sparger, Dip Tube, Harvest, Cooling Coil	M1273-5054
12mm port to 6.35mm tube	To make 12mm port universal for 12mm & 6.35mm diameters	M1273-5056
COMPRESSION ADAPTERS:		
Compression adapters allow you to easily raise or lower the probe or tube, while maintaining a safe seal.		
6.35mm port to 4.76mm tube	For Sample Tube, Foam Probe, Level Probe	M1273-5042
12mm port to 4.76mm tube	To make 12mm port universal for 12mm, 6.35mm & 4.76mm diameters	M1273-5055
12mm port to 12mm tube	For dO ₂ & pH probes	M1273-5040

17 INDEX

1

100% Gas Feed
Setting up, 107

4

4-Gas Mode, 107

A

Accessories
Part Numbers, 122
Addition/Harvest Kit
Part Numbers, 122
AFS Communications Protocol, 139
AFS Supervisory Software
Settings for, 138
AFS/ModBus Pin Connections, 139
Agitation Motor Connection, 60
Air Filter/Regulator Kit
Installation of, 63
Part Numbers, 122
Angled Autoclave Rack
Part Numbers, 120
Antifoam, 110
Application
Choosing the, 86
Autoclave Caps
Replacement Part Numbers, 122
Autoclaving
Preparing for, 76
Autoclaving Cap Replacement, 117
Autoclaving Dimensions, 127
Autoclaving the Vessel, 77
Automatic Gas Mix
Overriding the, 107
Automatic Gas Sequencing
Setting the, 107

B

Baffle
Installation of, 21, 34, 41, 53
Base Plate
Installing Vessel on, 40
Batteries
Part Numbers, 120
Battery Back-Up, 84
Battery Replacement, 132
Bearing Housing Cap
Replacement Part Numbers, 122
Bearing Housing Cap Replacement, 117
BioCommand Communications Protocol, 139

Bottom Clamping Ring
Installing the, 40

C

Cascade To
4 Gas, 102, 104
Agitation, 98, 100, 101, 102, 103
Agitation/O₂, 100, 102
Demand, 102, 104
None, 98, 100, 102
O₂, 100, 102
Cascades
for Cell Culture, 102
for Fermentation, 98, 100
CAUTION
Symbol for, v
CE Declaration of Conformity, vii
Change Command, 139, 141
Change Command Acknowledgement, 141
Checklist
for Basic Microbial Kit, 3
Choosing the Application
Cell Culture, 86
Fermentation, 86
Choosing the Language, 69
Cleaning
the Modules, 114
the Vessel, 115
Comm1, 138
Comm1 Settings, 138
Comm2, 138
Comm2 Settings, 138
Communications Protocol
for AFS (BioCommand), 139
for ModBus, 141
Communications Settings, 138
Connections, 12
Control Equation
for pH P & I Gains, 136
Control Type Values, 142
Cooling Coil
Installation of, 23
Copyright Notice, v

D

Data Logging, 72
Demand Mode, 107
Dimensions for Autoclaving, 127
Disclaimer Notice, v
Dissolved Oxygen. *See* dO₂
dO₂ Control
Setting Cascades for, 98, 100, 102
Setting P & I Gains, 105
dO₂ Probe

- Calibrating the, 81
- Installation of, 27, 46
- Polarizing the, 80
- dO₂ Setpoint
 - Setting the, 98, 105
- Drawings
 - 1.3L Headplate, 18, 37
 - 3.0L Headplate, 38
 - 7.5L & 14.0L Headplate, 20, 39
 - Cable Guide, 64
 - dO₂/pH Controller, Front, 60
 - dO₂/pH Controller, Rear, 7
 - Exhaust Condenser, 48, 49
 - Foam Trap, 52
 - Fuse Block/Cover Assembly, 135
 - Fuse Holder/Power Inlet, 135
 - Fusing Arrangement, 136
 - Gas Mix Controller, Rear, 8
 - Harvest System, 31, 50
 - Index of, 162
 - Level Controller, Front, 61
 - Level Controller, Rear, 8
 - Motor Installation, 55
 - Non-Jacketed Vessel, 17
 - PCU Touchpad, 85
 - PCU, Front, 84
 - PCU, Rear, 5
 - Power Controller Front Panel, 59
 - Power Controller Rear Panel, 6
 - Pump Module, Rear, 9
 - Sample BioFlo 110 Installation, 11
 - Sampler System, 31, 50, 51
 - Stacked Modules & Vessel, 11
 - TMFC, Front, 145
 - Vessel Autoclave Rack, 78
 - Vessel Bumper Installation, 21
 - Voltage Selection, 133
 - Water Solenoid Valve, 48
 - Water-Jacketed Vessel Assembly, 36
- Drive Assembly Handling, 15

E

- Essential Warnings, 14
- Exhaust Condenser
 - Part Numbers, 120

F

- Filling the Water Jacket, 76
- Foam Control, 112
- Foam Exhaust Tube
 - Installing the, 25, 44
- Foam Probe
 - Installing the, 24, 43
- Foam Trap Kit
 - Part Numbers, 120
- Fuse Replacement, 117
- Fuses
 - Part Numbers, 120

G

- Gas Mix
 - Control Modes for, 106
 - Setting the, 106

H

- Handling Tips, 15
- Harvest Tube
 - Installation of, 24, 43
- Headplate
 - Arrangement of, 164
 - Installation of, 34, 54
- Headplate Clamping Ring
 - Installation of, 40
- Headplate O-Ring Replacement, 117
- Headplate Port Fittings
 - Part Numbers, 120
- Headplates
 - 1.3L, 18, 37
 - 3.0L, 19, 38
 - 7.5L & 14.0L, 39
- Heat Blanket, 16
 - Installation of, 20
- Heat Blanket Connection, 60
- Heater Connection, 60
- How to set the pumps, 109

I

- Impellers
 - Installation of, 21, 41
 - Part Numbers, 120
- Important Warnings, 14
- Inoculation, 112
- Inspection
 - of Boxes, 3
 - of Equipment, 3
- Interconnecting Modules, 12
- International Offices
 - List of, iii
- Introduction, 1

J

- Jacket Temp. Control, 60
- Jacket Water Heater, 35, 82
 - Reset Button, 82
 - Safety Cutoff, 82
- Jacket Water Heater Connection, 60

L

- Language
 - Choosing the, 69
- Level Probe(s)
 - Installing the, 25, 44
- Location
 - Environment for, 4
 - Physical, 4

Preparing the, 4

M

Maintenance, 114
Manual Conventions, v
Message Formats, 141
ModBus
 Settings for, 138
Module Kits
 Replacement Part Numbers, 123
Motors
 Part Numbers, 120
Multidrop Number, 140

N

NOTE
 Symbol for, v

O

Option Part Numbers
 Air Filter/Regulator Kit, 122
 Spin Filter, 122
 Water Filter/Regulator Kit, 122
Optional Accessories
 Part Numbers, 122
O-Ring Kit
 Replacement Part Numbers, 122
O-Ring Replacement, 116
Overview, 1

P

P & I Gains
 Control Equation for pH, 136
Part Numbers
 Accessories, 122
 Module Kits, 123
 Port Adapters, 164
pH Control
 Setting a Deadband, 94
 Setting an Output Multiplier, 95
 Setting P & I Gains, 97
pH Probe
 Calibrating the, 65, 94
 Installation of, 25, 44
Port & Adapter O-Ring Replacement, 116
Power Back-Up, 84
Probes & Sensors
 Part Numbers, 121
Pt 100, 60
Pump Control Mode, 109
Pump Flow Rates, 126
Pump Module, 9
Pump Multiplier, 108
Pump Operation, 108
Pump Setpoint, 108

R

Recorder Output
 Pin Designations for, 72
Register Addresses, 142
Register Map, 142
Replacement Part Numbers
 Bearing Housing Cap, 122
 Miscellaneous, 122
 O-Ring Kit, 122
Replacement Parts
 Addition/Harvest Kit, 122
 Angled Autoclave Rack, 120
 Batteries, 120
 Exhaust Condenser, 120
 Foam Trap Kit, 120
 Fuses, 120
 Headplate Port Fittings, 120
 Impellers, 120
 Motors, 120
 Part Numbers, 119
 Probes & Sensors, 121
 Sparger, 122
 Tubing, 121
 Vessel Assemblies, 119
 Vessels, 122
Replacing Batteries, 132
Replacing Bearing Housing Cap, 117
Replacing Fuses, 117
Replacing Headplate O-Ring, 117
Replacing Port & Adapter O-Rings, 116
Request for Data, 140
Request Format, 141
RTD, 24, 43
RTD Connection, 60

S

Sampler
 Installation of, 30
 Installing the, 50
Sampler Tube
 Installation of, 24, 43
Sampling, 112
Schematics
 4-Pump Module, 159
 dO₂/pH Controller, 156
 Gas Mix Controller, 158
 Level Controller, 157
 Thermal Mass Flow Controller, 161
Service, 118
Setting Setpoints, 98, 105
Setting the Pumps, 109
Sparger
 Installation of, 23, 42
 Part Numbers, 122
Specifications
 4-Pump Module, 126
 Antifoam/Level Controller, 126
 Dimensions for Autoclaving, 127, 128, 129
 dO₂/pH Controller, 125
 Gas Mix Controller, 126

- Non-Jacketed Vessels, 127
- PCU, 124
- Power Controller, 125
- TMFC, 153
- Spin Filter
 - Option Part Numbers, 122
- Stacking Modules, 10
- Sterilization
 - Preparing for, 76
- Sterilization Temperature, 78
- Sterilization Time, 78

T

- Table of Contents, ix
- Temperature Probe Connection, 60
- Terminators
 - for Unused Connectors, 138
- Thermostat Reset Button, 82
- Thermowell
 - Installation of, 24, 43
- TMFC
 - Cascades for Cell Culture, 151
 - Cascades for Fermentation, 149
 - Cleaning, 152
 - Loading Software for, 145
 - Operating the, 146
 - Setup for Cell Culture, 150
 - Setup for Fermentation, 147
 - Specifications, 153
 - Troubleshooting, 153
- Troubleshooting, 118
 - TMFC, 153
- Tubing
 - Part Numbers, 121
- Tubing Specifications, 127

U

- Unpacking boxes, 3
- Unused Ports, 33, 53
- Utilities
 - Requirements for, 4

V

- Verification
 - against Packing List, 3
- Vessel Assemblies
 - Replacement Part Numbers, 119
- Vessel Assembly
 - Non-Jacketed, 16
 - Water-Jacketed, 35, 40
- Vessel Bumpers, 21, 40
- Vessel Stand, 16, 21
- Vessels
 - Part Numbers, 122
- Voltage Selection, 133

W

- WARNING
 - Symbol for, v
- Warranty, vi
- Water Filter/Regulator Kit
 - Installation of, 58
 - Part Numbers, 122
- Water Jacket
 - Filling the, 76
- Wet Off, 61, 110, 111
- Wet On, 61, 110, 111, 112

Artisan Technology Group is an independent supplier of quality pre-owned equipment

Gold-standard solutions

Extend the life of your critical industrial, commercial, and military systems with our superior service and support.

We buy equipment

Planning to upgrade your current equipment? Have surplus equipment taking up shelf space? We'll give it a new home.

Learn more!

Visit us at [artisanng.com](https://www.artisanng.com) for more info on price quotes, drivers, technical specifications, manuals, and documentation.

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

We're here to make your life easier. How can we help you today?

(217) 352-9330 | sales@artisanng.com | [artisanng.com](https://www.artisanng.com)

