

Dreufs 3080/2
Lighted Rocker Switch



© Artisan Technology Group

**Limited Availability
Used and in Excellent Condition**

Open Web Page

<https://www.artisanng.com/80405-3>

All trademarks, brandnames, and brands appearing herein are the property of their respective owners.



Your **definitive** source
for quality pre-owned
equipment.

Artisan Technology Group

(217) 352-9330 | sales@artisanng.com | artisanng.com


- Critical and expedited services
- In stock / Ready-to-ship

- We buy your excess, underutilized, and idle equipment
- Full-service, independent repair center

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

TA Instruments

Thermal Analysis & Rheology



TGA 2950

Thermogravimetric Analyzer

Operator's Manual

PN 925602.001 Rev. D (Text and Binder)
PN 925602.002 Rev. D (Text Only)
Issued July 2000

©1995, 1996, 1997, 2000 by
TA Instruments, Inc.
109 Lukens Drive
New Castle, DE 19720

Notice

The material contained in this manual is believed adequate for the intended use of this instrument. If the instrument or procedures are used for purposes other than those specified herein, confirmation of their suitability must be obtained from TA Instruments. Otherwise, TA Instruments does not guarantee any results and assumes no obligation or liability. This publication is not a license to operate under or a recommendation to infringe upon any process patents.

TA Instruments Operating Software and Module, Data Analysis, and Utility Software and their associated manuals are proprietary and copyrighted by TA Instruments, Inc. Purchasers are granted a license to use these software programs on the module and controller with which they were purchased. These programs may not be duplicated by the purchaser without the prior written consent of TA Instruments. Each licensed program shall remain the exclusive property of TA Instruments, and no rights or licenses are granted to the purchaser other than as specified above.

Table of Contents

Notes, Cautions, and Warnings	<i>xii</i>
Hotline Numbers	<i>xiv</i>
Safety	<i>xv</i>
Using This Manual	<i>xviii</i>
CHAPTER 1: Introducing the TGA 2950	1-1
Introduction.....	1-3
Components	1-4
The 2950 Instrument	1-5
2950 Display	1-5
2950 Keypad	1-6
HEATER Switch	1-10
POWER Switch	1-10
Accessories	1-11
Gas Switching Accessory	1-11
Other Accessories	1-11
Specifications	1-12

Table of Contents

(continued)

CHAPTER 2: Installing the TGA 2950	2-1
Unpacking/Repacking the 2950	2-3
Unpacking the 2950	2-3
Repacking the 2950	2-6
Installing the Instrument	2-7
Inspecting the System	2-8
Choosing a Location	2-9
Filling the Heat Exchanger	2-10
Connecting Cables and Gas Lines	2-12
Heat Exchanger Cable and Water Lines	2-12
GPIB Cable	2-15
Purge Lines	2-17
Cooling Gas Line	2-18
Power Cable	2-19
Unpacking the Balance	2-20
Installing the Hang-Down Wires	2-22
Aligning the Sample Hang-Down Wire	2-25
Adjusting the Sample Platform	2-29
Starting the 2950	2-30
Shutting Down the 2950	2-32

Table of Contents

(continued)

CHAPTER 3: Running Experiments 3-1

Overview 3-3

Before You Begin 3-3

Calibrating the TGA 3-5

Temperature Calibration 3-5

Weight Calibration 3-6

Sample Platform Calibration 3-6

Running a TGA Experiment 3-7

Experimental Procedure 3-7

Preparing Samples 3-8

Selecting Sample and Tare Pans 3-8

Taring the Sample Pan 3-9

Automatic Tare 3-10

Manual Tare 3-10

Loading the Sample 3-11

Setting Up an Experiment 3-14

Setting Up Accessories 3-15

Using the Air Cool Option..... 3-16

Using a Purge Gas 3-16

Using the Gas

Switching Accessory 3-19

Starting an Experiment 3-20

Table of Contents

(continued)

Stopping an Experiment	3-21
Unloading the Sample	3-21
Use in an Oxygen-Free Atmosphere	3-22
Purge Gas System	3-22
Instrument Set-Up	3-23
CHAPTER 4: Technical Reference	4-1
Description of the TGA 2950	4-3
Components	4-5
Balance	4-6
Sample Loading Assembly	4-8
Furnace	4-9
TGA Standard Furnace	4-9
EGA Furnace	4-11
Cabinet	4-13
Heat Exchanger	4-15
Theory of Operation	4-16
Status Codes	4-17

Table of Contents

(continued)

CHAPTER 5: Maintenance and Diagnostics	5-1
Overview	5-3
Routine Maintenance	5-4
Inspection	5-4
Cleaning the Instrument	5-4
Cleaning the Furnace Housing	5-5
TGA Standard Furnace Only	5-5
Heat Exchanger	5-6
Maintaining Heat Exchanger Coolant	5-6
Draining and Refilling the Water Reservoir	5-6
Replacing the Thermocouple	5-9
Removing and Reinstalling the Furnace	5-11
Furnace Removal	5-11
Furnace Replacement	5-14
Diagnosing Power Problems	5-16
Fuses	5-16
Power Failures	5-18

Table of Contents

(continued)

TGA Test Functions	5-19
The Confidence Test	5-19
Replacement Parts	5-21
Appendix A: Ordering Information	A-1
Appendix B: High Resolution	
TGA Option	B-1
Overview	B-5
Option Installation	B-6
Using Hi-Res™ TGA	B-7
Background	B-7
The TA Instruments Hi-Res™	
Technique	B-8
The Hi-Res™ Ramp Segment	B-9
Calcium Oxalate Example Scans	B-13
Advanced Hi-Res™ Techniques	B-18
What Can Be Resolved and	
What Cannot?	B-18
Unresolved Transitions	B-19
Selecting a Hi-Res™ Technique	B-22

Table of Contents

(continued)

Dynamic Rate Hi-Res™ Ramp.....	B-22
Constant Reaction Rate Hi-Res™ Ramp.....	B-23
Weight Gain Experiments	B-26
Signature Analysis	B-27
Sample Quantity and Orientation	B-28
Exposed Surface Area	B-28
Bubble Formation	B-29
Thermocouple Placement	B-32
Data Analysis Effects	B-32
Derivative Plots	B-33
Adjusting Heating Rate	B-34
Adjusting Resolution Setting	B-37
Useful Resolutions Settings	B-39
Temperature Calibration	B-41
Hi-Res™ Transition Temperatures ...	B-42
Hi-Res™ Sensitivity Segment	B-43

Table of Contents

(continued)

Understanding Sensitivity Setting	B-44
Adjusting Sensitivity in Dynamic Rate Mode	B-45
Adjusting Sensitivity in Constant Reaction Rate Mode	B-47
Abort Segment	B-51
Stepwise Isothermal Heating	B-53
Hi-Res™ TGA Examples	B-58
Mixture of Bicarbonates	B-58
Dynamic Rate Scans	B-59
Varying Resolution Setting	B-61
Varying Sensitivity Setting	B-62
Constant Reaction Rate Scans	B-63
Stepwise Isothermal Scans	B-65
Monosodium Glutamate	B-68
Banana Taffy	B-69
Plastic Laboratory Tubing	B-71
References	B-75

Table of Contents

(continued)

APPENDIX C: TGA 2950	
Autosampler Option	C-1
Introducing the Auto TGA	C-3
Getting Started	C-4
The TGA 2950 Instrument	C-6
Calibrating the Auto TGA	C-10
Running Experiments	C-11
Preparing the Samples	C-12
Selecting Sample	
and Tare Pans	C-12
Taring the Sample Pans	C-13
Loading the Samples	C-16
Setting Up an Experiment	C-18
Manual Operation	C-19
Tracking the TGA	
Autosampler Status	C-19
Interrupting a Run	C-20

Table of Contents

(continued)

Appendix D: TGA 2950	
EGA Furnace Option	D-1
Introducing the EGA Furnace	D-3
EGA Furnace Specifications	D-5
Installing the EGA Furnace	D-6
First-Time Installation	D-6
Removing and Reinstalling the	
EGA Furnace	D-14
EGA Furnae Removal.....	D-14
EGA Furnace Installation	D-16
Connecting the Spectrometer	D-18
Using the EGA Furnace	D-21
Cleaning the Quartz Furnace Tube	D-22
Index.....	I-1

Notes, Cautions, and Warnings

This manual uses NOTES, CAUTIONS, and WARNINGS to emphasize important and critical instructions.

NOTE:

A NOTE highlights important information about equipment or procedures.

◆ CAUTION:

A CAUTION emphasizes a procedure that may damage equipment or cause loss of data if not followed correctly.

! WARNING

A WARNING indicates a procedure that may be hazardous to the operator or to the environment if not followed correctly.

Hotlines

To TA Instruments

For Technical Assistance (302) 427-4070

To Order Instruments
and Supplies (302) 427-4040

For Service Inquiries (302) 427-4050

Sales (302) 427-4000

Safety

Electrical Safety

You must unplug the instrument *before* doing any maintenance or repair work; voltages exceeding 120 volts AC are present in this system.



High voltages are present in this instrument. If you are not trained in electrical procedures, do not remove the cabinet covers unless specifically instructed to do so in the manual. Maintenance and repair of internal parts must be performed only by TA Instruments or other qualified service personnel.



After transport or storage in humid conditions, this equipment could fail to meet all the safety requirements of the safety standards indicated. Refer to the NOTE on page 2-9 for the method used to dry out the equipment before use.

Chemical Safety

Use only the purge gases listed in Table 1.4 in Chapter 1. Use of other gases could cause damage to the instrument or injury to the operator.



Do not use hydrogen or any other explosive gas in the TGA 2950 furnace or the TGA 2950 EGA furnace.



Oxygen can be used as a purge gas in the TGA 2950. However, the furnace must be kept clean so that volatile hydrocarbons (which might combust) are removed.

Safety

(continued)



The TGA 2950 furnace assembly contains a layer of refractory ceramic fiber (RCF) insulation. This insulation is completely encapsulated within the ceramic subassembly, which is not meant to be disassembled. If the subassembly should break in such a way as to expose the RCF insulation, we recommend that you dispose of it as you would any refractory material.



If you are routinely evaluating materials in the TGA that lose a large amount of volatile hydrocarbons (*e.g.*, lubricating oils), you need to clean the furnace more frequently to prevent dangerous buildup of debris in the furnace.



If you are using samples that may emit harmful gases, vent the gases by placing the instrument near an exhaust.



The TGA 2950 EGA furnace assembly also contains refractory ceramic fiber (RCF) insulation. This insulation is enclosed within the furnace housing. The furnace housing should only be disassembled for replacement of EGA furnace sample tube or furnace assemblies. Refer to instructions provided with the sample tube or furnace replacement kits for procedures for handling RCF insulation.

Thermal Safety

After running an experiment, allow the open furnace and thermocouple to cool down before you touch them.



! WARNING

During a sample run, the furnace base (see Figure 1.1) can be hot enough to burn skin. Avoid contact with the furnace base during experiments.

Mechanical Safety



! WARNING

Keep your fingers and all other objects out of the path of the furnace when it is moving. The furnace seal is very tight.

Using This Manual

Chapter 1	Describes your TGA 2950 instrument and its accessories and specifications.
Chapter 2	Describes how to unpack and install the TGA and how to connect the TGA to the rest of your system.
Chapter 3	Describes how to run TGA experiments and set up the accessories.
Chapter 4	Provides technical information and explains principles of TGA operation.
Chapter 5	Describes how to perform routine maintenance, replace the thermocouple, remove and reinstall the furnace, and diagnose power problems; also provides an explanation of the confidence test.
Appendix A	Lists TA Instruments offices that you can contact to place orders, receive technical assistance, and request service.
Appendix B	Describes the High Resolution option including installation, useage, applications, etc.

Using This Manual

(continued)

Appendix C

Provides instructions needed to operate the TGA Autosampler option to automatically load and run samples.

Appendix D

Provides instructions needed to install and use the EGA furnace with the TGA 2950.





CHAPTER 1: Introducing the TGA 2950

Introduction.....	1-3
Components	1-4
The 2950 Instrument.....	1-5
2950 Display	1-5
2950 Keypad	1-6
HEATER Switch	1-10
POWER Switch	1-10
Accessories	1-11
Gas Switching Accessory	1-11
Other Accessories	1-11
Specifications	1-12

Introducing the TGA 2950

Introduction

Your TA Instruments Thermogravimetric Analyzer (TGA) 2950 is a thermal weight-change analysis instrument, used in conjunction with a TA Instruments thermal analysis controller and associated software, to make up a thermal analysis system.

The Thermogravimetric Analyzer 2950 measures the amount and rate of weight change in a material, either as a function of increasing temperature, or isothermally as a function of time, in a controlled atmosphere. It can be used to characterize any material that exhibits a weight change and to detect phase changes due to decomposition, oxidation, or dehydration. This information helps the scientist or engineer identify the percent weight change and correlate chemical structure, processing, and end-use performance.

Your controller is a computer that performs the following functions:

- Provides an interface between you and the analysis instruments
- Enables you to set up experiments and enter constants
- Stores experimental data
- Runs data analysis programs.

Components

The TGA 2950 has five major components, illustrated in Figure 1.1:

- The balance, which provides precise measurement of sample weight. The balance is the key to the TGA system.
- The sample platform, which loads and unloads the sample to and from the balance.
- The furnace, which controls the sample atmosphere and temperature.
- The cabinet, where the system electronics and mechanics are housed.
- The heat exchanger, which dissipates heat from the furnace.




Figure 1.1
TGA 2950 Components

The 2950 Instrument

The parts of the TGA 2950 instrument that provide for operator control are:

- The instrument display
- The instrument keypad.




Figure 1.2
TGA 2950
Display and Keypad

2950 Display

The TGA instrument display is the lighted area of the keypad (Figure 1.2). It contains two rows of 20 characters each.

During normal operation, the display is segmented into three areas. The left eight characters on the upper line show the instrument status; the right nine characters show the sample temperature; and the bottom line is a realtime signal display (*e.g.*, weight).

A pound sign (#) after the weight signal indicates that the balance reading has not yet stabilized. When the weight stabilizes, the pound sign will disappear.

2950 Keypad

The instrument keypad (see Figure 1.2) contains the keys found in Table 1.1 and the HEATER and POWER switches

NOTE:

Experiment information and instrument constants are entered from the controller keyboard, not the instrument keypad.

Table 1.1
TGA 2950 Keypad
Function Keys

Key/Function	Explanation
SCROLL	Scrolls the realtime signals shown on the bottom line of the display. For more details on the experiment, refer to status and signal displays on the controller.
TARE	Zeros the displayed weight of an empty sample pan: automatically loads the pan from the sample platform, raises the furnace to protect the pan from air currents, weighs the pan, stores the weight as an offset, and then unloads the pan. (table continued)

Table 1.1
TGA 2950 Keypad
Function Keys
(continued)

Key/Function	Explanation
	Loads a sample pan from the sample platform onto the balance.
	Unloads the sample pan from the balance onto the sample platform.
	Toggles between the furnace closed (up) and furnace open (down) functions, depending on where the furnace is when you press the key. This key can be pressed while the furnace is moving to reverse the direction of movement.
	Begins the experiment. This is the same function as Start on the controller. <i>Forced Start</i> can be done by pressing the START key while the status line displays "Set Up." Forced start begins collecting data during instrument setup. <i>(table continued)</i>

Table 1.1
TGA 2950 Keypad
Function Keys
(continued)

Key/Function	Explanation
STOP	If an experiment is running, this key ends the method normally, as though it had run to completion; <i>i.e.</i>, the method-end conditions go into effect and the data that has been generated is saved. This is the same function as Stop on the controller. If an experiment is not running (the instrument is in a stand-by or method-end state), the STOP key will halt any activity (air cool, all mechanical motion, <i>etc.</i>).
REJECT (Hold down SCROLL and press STOP)	If an experiment is running, SCROLL-STOP ends the method normally, as though it had run to completion; <i>i.e.</i>, the method-end conditions go into effect and the data that has been generated is <i>discarded</i>. This is the same function as Reject on the controller. <i>(table continued)</i>

Table 1.1
TGA 2950
Keypad Function Keys
(continued)

Key/Function	Explanation
NOTE:	The SCROLL key operates normally (scrolls the realtime signals) until the STOP key is pressed. If an experiment is not running, SCROLL-STOP works like the STOP key.
AUTO SELECT	This key appears only on instruments with an autosampler installed. See Appendix C for details.

Automatic Keypad Functions

Some of the TGA instrument keys automatically perform additional functions under certain conditions:

- **START** automatically loads the sample pan and closes the furnace, if necessary, before beginning the experiment.
- **TARE**, **LOAD**, and **UNLOAD** automatically open the furnace if necessary.
- **START** can be pressed while a sample **LOAD** is in progress.

HEATER Switch

The HEATER on/off switch (see Figure 1.2) turns the power to the instrument heater on and off. This switch should be in the ON (1) position before you start an experiment.

NOTE:

|| The light in the HEATER switch will glow only when an experiment is in progress.

POWER Switch

The POWER switch (see Figure 1.2) turns the power to the instrument on and off.

Accessories

Gas Switching Accessory

The TA Instruments Gas Switching Accessory can be used to turn the purge gas on and off or to switch between two different purge gases during TGA experiments.

Evolved Gas Analysis Furnace

The TGA 2950 EGA furnace is an accessory to the instrument which allows you to perform combined TGA and evolved gas analysis experiments.

Other Accessories

The TGA can be used with many standard analytical accessories offered by various manufacturers, including vacuum, FTIR, mass spectrometers, gas chromatographs, and evolved gas analyzers. Consult the appropriate local instrument manufacturer for further information.

Specifications

Tables 1.2 through 1.4 contain the technical specifications for the TGA 2950.

Table 1.2
TGA 2950
Operating Parameters

Temperature range	25°C to 1000°C
Thermocouple	Platinel II*
Heating rate with standard furnace with EGA furnace	0.1 to 100°C/min 0.1 to 50°C/min
*Platinel II is a registered trademark of Engelhard Industries.	

Table 1.3
TGA 2950
Instrument
Characteristics

Operating line voltage	115 volts, 50/60 Hz
Energy consumption	1.5 kVA

Table 1.4
Sampling System

Sample pans Types	Platinum, Alumina (Al ₂ O ₃), Aluminum
Volume capacity	Platinum: 50 μ L, 100 μ L Alumina: 100 μ L, 250 μ L, 500 μ L Aluminum 100 μ L
Weighing capacity ¹	1.0 gm
Balance measurement ² Resolution Accuracy	0.1 μ g $\leq \pm 0.1\%$
Ranges	100 mg range: 0.1 μ g - 100 mg 1000 mg range: 1 μ g - 1000 mg

◆ **CAUTION:**

¹ The total mechanical capacity of the balance is 5 gm. In order to avoid damaging the balance assembly, never allow the total weight of the sample, tare weight, hang-down wires, and pans to exceed 5 gm.

² The TGA balance mechanism is sensitive to changes in the surrounding room temperature. For optimum accuracy, you must regulate the ambient temperature.

(table continued)

Table 1.4
(continued)

Furnace Atmosphere Purge gases Purge rate	Helium, nitrogen, oxygen, air, argon³ Up to 100 cc/min
 WARNING ³ Do not use hydrogen or any other explosive gas in the TGA 2950 furnace or the TGA 2950 EGA furnace. Oxygen may be used. However, the furnace must be kept clean of volatile hydrocarbons to prevent combustion.	

CHAPTER 2: Installing the TGA 2950

Unpacking/Repacking the 2950	2-3
Unpacking the 2950	2-3
Repacking the 2950	2-6
Installing the Instrument	2-7
Inspecting the System	2-8
Choosing a Location	2-9
Filling the Heat Exchanger	2-10
Connecting Cables and Gas Lines	2-12
Heat Exchanger Cable and Water Lines	2-12
GPIB Cable	2-15
Purge Lines	2-17
Cooling Gas Line	2-18
Power Cable	2-19
Unpacking the Balance	2-20
Installing the Hang-Down Wires	2-22
Aligning the Sample Hang-Down Wire	2-25
Adjusting the Sample Platform	2-29
Starting the 2950	2-30
Shutting Down the 2950	2-32

Installing the 2950

Unpacking/Repacking the 2950

NOTE:

These instructions are also found as separate unpacking instructions in the shipping box.

You may wish to retain all of the shipping hardware, the plywood, and boxes from the instrument in the event you wish to repack and ship your instrument.

Unpacking the 2950

Refer to Figures 2.1 to 2.3 while unpacking your instrument.



Have an assistant help you unpack this unit. Do not attempt to do this alone.




Figure 2.1
Shipping Boxes

1. Open the shipping carton and remove the accessory box.
2. Remove the cardboard packing insert.
3. Stand at one end of the box with your assistant facing you at the other end. Lift your end of the unit out of the box as your assistant lifts his/her end.
4. Place the unit on a lab bench with one side hanging over the edge of the bench (see Figure 2.2). **Someone must be holding onto the unit at all times while it is in this position.**




Figure 2.2
Removing the Plywood
Board

5. While your assistant holds the unit, use a wrench to remove the two nuts and washers from the bottom. Then lift and rotate the unit so that the other end hangs over the edge of the bench. **Someone must hold onto the unit at all times while it is in**

this position. While your assistant holds the unit, remove the two nuts and washers from the other side.

6. Slide the unit completely onto the lab bench. Have your assistant hold one side up while you unscrew and remove the black rubber shipping feet from the bottom. Then rotate the unit and remove the shipping feet from the other side in the same manner.
7. Have your assistant lift the entire unit while you slide the plywood board out from under the unit.
8. Have your assistant lift one side of the unit while you use a wrench to install two mounting feet on the other side (see Figure 2.3). Rotate the unit and install the two remaining mounting feet in the same manner.
9. Remove the furnace clamp before turning on the power to the unit.




Figure 2.3
Installing the
Mounting Feet

Repacking the 2950

To pack and ship your instrument, use the hardware retained during unpacking and reverse the instructions found on pages 2-3 to 2-5.

Installing the Instrument

Before shipment, the TGA 2950 instrument is inspected both electrically and mechanically so that it is ready for operation upon proper installation. Installation involves the following procedures, described in this chapter:

- Inspecting the system for shipping damage and missing parts
- Filling the heat exchanger
- Connecting the TGA to the TA Instruments controller
- Connecting the heat exchanger cable and water lines, purge gas lines, accessories, and power cable
- Unpacking the balance
- Installing the hang-down wires
- Leveling the instrument and aligning the hang-down wires
- Adjusting the sample platform.

If you wish to have your TGA installed by a TA Instruments Service Representative, call for an installation appointment when you receive your instrument.

◆ **CAUTION:**

|| To avoid mistakes, read this entire chapter before you begin installation.

Inspecting the System

When you receive your TGA 2950, look over the instrument and shipping container carefully for signs of shipping damage, and check the parts received against the enclosed shipping list.

- If the instrument is damaged, notify the carrier and TA Instruments immediately.
- If the instrument is intact but parts are missing, contact TA Instruments.

A list of TA Instruments phone numbers can be found in Appendix A of this manual.

Choosing a Location

Because of the sensitivity of TGA experiments, it is important to choose a location for the instrument using the following guidelines. The TGA should be:

- In* ... a temperature-controlled area.
- ... a clean, vibration-free environment.
- ... an area with ample working and ventilation space.
- On* ... a stable work surface.
- Near* ... a power outlet (115 volts AC, 50 or 60 Hz, 15 amps). A step up/down line transformer may be required if the unit is operated at a higher or lower line voltage.
- ... your TA Instruments thermal analysis controller.
- ... compressed lab air and purge gas supplies with suitable regulators and flowmeters.
- Away from* ... dusty environments.
- ... exposure to direct sunlight.
- ... direct air drafts (fans, room air ducts).
- ... poorly ventilated areas.
- ... noisy or mechanical vibrations.

NOTE:

Drying out the instrument may be needed, if it has been exposed to humid conditions. It is important to be certain that the instrument ground is adequately connected to the facilities ground for safe operation.

Run the following method to dry out the instrument (refer to Chapter 4 for further information):

- 1 Ramp at 10°C/min to 400°C
- 2 Isothermal for 30 min.

Filling the Heat Exchanger

The heat exchanger contains a liquid reservoir that supplies the instrument with coolant to dissipate heat from the furnace. The coolant exits the heat exchanger through the supply line, circulates to the furnace, and comes back to the reservoir via the return line as seen in Figure 2.4 (for instructions on how to connect the water lines, turn to page 2-12). To fill the heat exchanger, follow the directions given below.




Figure 2.4
Rear Panel of
Heat Exchanger

1. Unscrew the water reservoir cap on the heat exchanger (see Figure 2.4).



Figure 2.5
Heat Exchanger
Bottle

2. Add TA Instruments TGA Conditioner (PN 952377.901) into the water reservoir bottle. Refer to the instructions on the bottle for the amount of conditioner to add to the reservoir. Then fill the bottle to the inner rim (see Figure 2.5) with distilled water.

NOTE:

After the system has been started, recheck the level of water in the reservoir bottle and refill to the inner rim if necessary.



Do not put any liquid other than distilled water in the heat exchanger reservoir.

3. Replace and tighten the water reservoir cap.

Connecting Cables and Lines

To connect the cables and water and gas lines, you will need access to the TGA instrument's rear panel. All directional descriptions are written on the assumption that you are facing the back of the instrument.

NOTE:

Connect all cables before connecting the power cords to outlets. Tighten the thumbscrews on all computer cables.

◆ **CAUTION:**

Whenever plugging or unplugging power cords, handle them by the plugs, not by the cords.



WARNING

Protect power and communications cable paths. Do not create tripping hazards by laying the cables across accessways.

Heat Exchanger Cable and Water Lines

1. Locate the cooling accessory connector on the left rear of the instrument cabinet (Figure 2.6).
2. Connect the heat exchanger cable to the connector. The heat exchanger cable is the only cable that fits into this connector.




Figure 2.6
TGA 2950
Connector Panel

3. Remove the water lines from the packing.
4. Connect one end of the water line marked “SUPPLY” to the connector labeled “SUPPLY” on the right side of the instrument cabinet.
5. Connect the other end of the water line marked “SUPPLY” to the connector labeled “SUPPLY” on the heat exchanger.
6. Connect one end of the unmarked water line to the connector labeled “RETURN” on the right side of the instrument cabinet.
7. Connect the other end of the unmarked water line to the connector labeled “RETURN” on the heat exchanger.

Figure 2.7 illustrates the correct water line connections for the TGA and heat exchanger.




Figure 2.7
Heat Exchanger Water
Line Connections

8. Plug in the power cable and turn on the power switch for the heat exchanger.

NOTE:

Air trapped in the heat exchanger system must be purged before starting the first run. After installation of the TGA is complete, turn on the instrument by placing the HEATER and POWER switches in the ON position. Then start the heat exchanger pump by turning on Air Cool from the controller. Refill the coolant reservoir as needed. Repeat this process until all the air has been purged from the system and the instrument stops reporting an "Err 119."

GPIB Cable

1. Locate the GPIB connector on the right rear of the TGA instrument (see Figure 2.6).
2. Connect the GPIB cable to the connector. The GPIB cable is the only cable that fits into the connector.
3. Tighten the hold-down screws on the connector.
4. Connect the other end of the GPIB cable to the controller or to the GPIB cable of another TA Instruments instrument connected to the controller. (For more information, see your controller manual.)
5. Select an address from 1 to 9. Then use the binary address switches on the TGA connector panel to set the desired address (See Table 2.1). Figure 2.8 on the next page shows a instrument address of 7.

NOTE:

If you have a multi-instrument system, each instrument must have a different address.

If you change the address after the TGA is powered on, you must press the TGA's Reset button to enter the new address. Wait until the instrument completes its startup displays, then reconfigure the instrument with the controller to bring the instrument back online.

NOTE:

The instrument's GPIB address is displayed during start up and can also be viewed on the instrument's status display.

Table 2.1
Binary Address Settings

Address	Switch Pattern
	1 2 3 4 5
1	0 0 0 0 1
2	0 0 0 1 0
3	0 0 0 1 1
4	0 0 1 0 0
5	0 0 1 0 1
6	0 0 1 1 0
7	0 0 1 1 1
8	0 1 0 0 0
9	0 1 0 0 1
*0 = OFF; 1 = ON	




Figure 2.8
Binary Address Switches
(Showing an Address of 7)

Purge Lines



Do not use any liquid in the purge lines.

1. Locate the FURNACE PURGE and BALANCE PURGE fittings on the right side of the TGA instrument back (Figure 2.9).




Figure 2.9
TGA PURGE Fittings

2. Make sure that the pressure of your purge gas source does not exceed the manufacturers' recommended pressures for flowmeters and other regulated devices you are using.

NOTE:

If you are using laboratory purge, rather than bottled purge, you will need to install an external drier.

CAUTION:

The use of corrosive gases is not recommended.

! WARNING

Use of an explosive gas as a purge gas is dangerous and is not recommended for this instrument. For a list of the purge gases that can be used with the TGA instrument, see Chapter 1. Oxygen may be used as a purge gas, but the furnace must be kept clean of volatile hydrocarbons to prevent combustion.

3. Connect a length of 1/4-inch I.D. flexible tubing from each of the PURGE fittings to a flowmeter (consult your compressed gas vendor for specific requirements). Then connect each flowmeter to the purge gas source.
4. The recommended setting for the purge rate is 100 mL per minute or less. The flow distribution should be as follows: (a) for the standard furnace, 40 percent to the balance chamber and 60 percent to the furnace, and (b) for the EGA furnace, 10 percent to the balance and 90 percent to the furnace.

Cooling Gas Line

1. Locate the COOLING GAS fitting, a 1/4-inch compression fitting on the left side of the TGA cabinet back, marked with a 120 psig maximum warning label (Figure 2.10).




Figure 2.10
TGA COOLING
GAS Fitting

2. Make sure your compressed lab air source is regulated to between 25 and 120 psig and is free of oil and water vapors.
3. Connect a compressed lab air line to the COOLING GAS fitting.

NOTE: Nitrogen may also be used as a cooling gas.

Power Cable

1. Make sure the TGA POWER switch (Figure 2.11) is in the Off (0) position.




Figure 2.11
TGA POWER Switch

2. Plug the power cable into the TGA.

◆ **CAUTION:**

Before plugging the TGA power cable into the wall outlet, make sure the instrument is compatible with the line voltage. Check the label on the back of the unit to verify the voltage.

3. Plug the power cable into the wall outlet.

Unpacking the Balance

◆ **CAUTION:**

When unpacking the balance, be careful not to damage the balance arm or hang-down loops.

1. Using the 7/64-inch ball driver supplied in your TGA accessory kit, loosen and remove the six screws securing the balance chamber faceplate to the instrument.
2. Take off the faceplate.
3. Loosen and remove the thumbscrew holding the balance cover on the sample (left) side of the balance mechanism (Figure 2.12), and take off the cover.




Figure 2.12
Interior of Balance
Chamber Before Unpacking

4. Using tweezers, remove the foam insert from around the screw hole (Figure 2.13):
 - a. Gently compress the foam with the tweezers, being careful not to touch the balance.
 - b. Remove the foam insert from the balance chamber.




Figure 2.13
Removal of Foam
Insert from Balance
Chamber

5. Replace the sample side cover and screw.
6. Repeat the procedure to remove the foam insert in the tare (right) side of the balance.

Installing the Hang-Down Wires

◆ **CAUTION:**

During installation, take care not to bend the hang-down wires or damage the hang-down loops.

1. Turn on the TGA instrument.
2. Press the FURNACE key to lower the furnace.
3. Locate the sample hang-down wire in your TGA Accessory Kit.
4. Hold the wire in your hand so that the doubly bent top hook is pointing to the left and the bottom hook is pointing to the right.
5. Carefully insert the bottom of the hangdown wire into the top of the furnace far enough so that you can insert the top of the wire into the thermocouple tube without bending the wire (Figure 2.14).




Figure 2.14
Installing the Sample Hang-Down Wire

6. Thread the hang-down wire up through the thermocouple tube into the balance chamber, and hook the top of the wire over the top of the tube (see Fig. 2.14).

NOTE:

To make the hang-down loops easier to see, we suggest sliding a piece of white paper into the balance chamber behind each loop before you hook the hang-down wire into it. (Do not forget to remove the paper when finished.)

7. Grasp the top hook of the hang-down wire with brass tweezers. Being careful to keep the top hook pointing to the left, pass the double bend through the hang-down loop so the wire is hanging from the loop.
8. Unscrew and remove the tare tube.
9. Locate the tare hang-down wire in your accessory kit.
10. Hold the wire in your hand so that the doubly bent top hook is pointing to the left and the bottom hook is pointing to the right.
11. Using brass tweezers, insert the tare hang-down wire into the balance chamber on the tare side and down through the hole above the tare tube connection, taking care not to bend the wire (Figure 2.15 on next page).
12. Being careful to keep the top hook pointing to the left, pass the double bend through the hang-down loop so the wire is hanging from the loop.

13. Select the sample pan you will use in your experiments, and load one of the same size and type onto the tare hang-down wire.
14. Replace the tare tube and finger-tighten it to compress the O-ring seal.

You are now ready to align the hang-down wires.




Figure 2.15
Installing the Tare
Hang-Down Wire

Aligning the Sample Hang-Down Wire

To avoid weight signal noise, the TGA instrument must be level so that the sample pan and hang-down wire hang inside the furnace and thermocouple tube without touching them. The angle at which the pan hangs is very sensitive to slight irregularities in benchtop surfaces, so it is important that you select a sturdy table or bench for your TGA.

Once you have your TGA in a satisfactory location, adjust the top and bottom of the sample hang-down wire and level the instrument using the following procedures.

To align the top of the sample hang-down wire:

1. Place an empty sample pan on the sample platform.
2. Press the LOAD key on the instrument keypad. The TGA will automatically lower the furnace (if necessary), move the sample platform over to the furnace, and load the pan onto the balance.

If the pan will not automatically load, place pan manually (using brass tweezers) on the sample hang-down wire and continue with the procedure. Use the Instrument Control Sample Platform Adjust procedure to correct loading after completing sample hang-down wire alignment.

3. Check to see whether the top end of the sample hang-down wire is hanging freely and roughly centered within the top of the thermocouple tube inside the balance chamber.
4. If the wire is not roughly centered inside the thermocouple tube, turn the balance adjustment screw (Figure 2.16) with the 7/64-inch ball driver until the wire is centered.

Turning the balance adjustment screw clockwise will move the wire backwards; turning the screw counterclockwise will move the wire frontwards.




Figure 2.16
Location of Balance
Adjustment Screw

To align the bottom of the hang-down wire:

1. Press the FURNACE key to raise the furnace just to the bottom of the sample pan, and press STOP.
2. Check the alignment of the sample pan within the furnace. It should hang freely, roughly centered, and should not be touching the sides of the furnace or the thermocouple tube (Figure 2.18).
3. If the sample pan is not centered and hanging freely within the furnace, level the TGA instrument by adjusting the feet on the bottom. Turn the feet clockwise to lengthen or counterclockwise to shorten the legs. Continue adjusting until the pan hangs correctly.




Figure 2.18
Aligning the Sample
Pan in the Furnace

4. Press the FURNACE key to lower the furnace.
5. Press the UNLOAD key to remove the sample pan from the furnace.
6. Replace the balance chamber faceplate and its 6 screws.

If you had to load the sample pan manually in order to align it in the furnace, you should now adjust the sample platform using the procedure described on the next page.

Adjusting the Sample Platform

If the sample hang-down wire fails to pick up a sample pan during an automatic loading procedure, you will need to adjust the position of the sample platform, using the Sample Platform Adjust procedure. This procedure is part of the Instrument Control software program. Refer to the online help and documentation for further information on adjusting the sample platform shown in Figure 2.19.

NOTE:

This procedure assumes that the instrument has been properly levelled (see page 2-26) and that the sample hang-down wire is straight.




Figure 2.19
Sample Platform
Assembly

Starting the 2950

1. Check all connections between the TGA 2950 and the controller. Make sure each component is plugged into the correct connector.
2. Press the instrument POWER and HEATER switch to the ON (1) position. The instrument will run an internal confidence test, which is run each time you power on the unit.

NOTE:

|| The HEATER and POWER indicator lamps may flicker under low AC voltage conditions.

3. Watch the instrument display during the confidence test for any error messages that may be indicated. If an error occurs, make a note of the test number in which the error occurred, and call TA Instruments for service.

After the confidence test, the screen will briefly display the system status, indicating the amount of data storage memory available and the GPIB address. Next follows the copyright display, and then the standby display, shown in Figure 2.20.




Figure 2.20
TGA 2950
Standby Display

4. Bring the instrument online with the TA controller.

NOTE:

|| Allow the TGA to warm up for at least 30 minutes before performing an experiment.

Shutting Down the 2950

Before you decide to power down your instrument, consider the following:

- All of the components of your thermal analysis system are designed to be powered on for long periods.
- The electronics of the TGA and the controller perform more reliably if power fluctuations caused by turning units on and off are minimized.

For these reasons, turning the system and its components on and off frequently is discouraged.

When you finish running an experiment on your instrument and wish to use the thermal analysis system for some other task, leave the instrument on; it will not interfere with whatever else you wish to do.

If you do need to power down your instrument for any reason, simply press the POWER and HEATER switches to the OFF position.

CHAPTER 3: Running Experiments

Overview	3-3
Before You Begin	3-3
Calibrating the TGA	3-5
Temperature Calibration	3-5
Weight Calibration	3-6
Sample Platform Calibration	3-6
Running a TGA Experiment	3-7
Experimental Procedure	3-7
Preparing Samples	3-8
Selecting Sample and Tare Pans	3-8
Taring the Sample Pan	3-9
Automatic Tare	3-10
Manual Tare	3-10
Loading the Sample	3-11
Setting Up an Experiment	3-14
Setting Up Accessories	3-15
Using the Air Cool Option	3-16
Using a Purge Gas	3-16
Using the Gas	
Switching Accessory	3-19

Running Experiments

Starting an Experiment	3-20
Stopping an Experiment	3-21
Unloading the Sample	3-21
Use in an Oxygen-Free Atmosphere	3-22
Purge Gas System	3-22
Instrument Set-Up	3-23

Overview

This chapter provides step-by-step instructions on how to run TGA experiments. Explanations of terminology and how the instrument operates are given in Chapter 4, “Technical Reference.”

All of your TGA experiments will have the following general outline. In some cases, not all of these steps will be performed.

- Entering experiment information through the TA controller (sample and instrument information)
- Creating and selecting the thermal method
- Selecting and taring the sample pan
- Loading the sample
- Setting the purge gas flow rate
- Starting the experiment
- Unloading the sample at the end of the experiment.

To obtain accurate results, follow procedures carefully and check calibration periodically (once a month).

Before You Begin

Before you set up an experiment, ensure that the TGA and the controller have been installed properly. Make sure you have:

- Made all necessary cable connections between the TGA and the controller
- Connected heat exchanger cable and water lines

Running Experiments

- Connected all gas lines
- Powered on each unit
- Installed all appropriate options
- Configured the instrument online with the controller
- Become familiar with controller operations
- Calibrated the TGA, if necessary.

Calibrating the TGA

To obtain accurate experimental results you should calibrate the instrument when you first install it. For the best results, however, you should recalibrate periodically.

Three types of calibration are needed for the TGA 2950: temperature, weight, and sample platform calibration. All three calibration procedures are performed through the Instrument Control software. This section provides a brief description of these types of calibration, for details refer to the online help and documents.

Temperature Calibration

Temperature calibration is useful for TGA experiments in which precise transition temperatures are essential. To temperature calibrate the TGA, you need to analyze a high-purity magnetic standard for its curie temperature, and then enter the observed and correct values in the temperature calibration table (see the online help and documentation for further information). The standard most often used is nickel with a curie temperature of 354.4°C (NIST Certificate for GM761). The observed and correct temperatures correspond to the experimental and theoretical transition temperatures (*e.g.*, curie temperature) of the calibrant. From one to five temperature calibration points (pairs of observed and correct temperature points) can be entered in the calibration table. A multiple-point calibration is more accurate than a one-point calibration.

Weight Calibration

Weight calibration should be performed on the TGA at least once a month. The weight calibration procedure calibrates both the 100 mg and 1 gram weight ranges. The calibration parameters are stored internally in the instrument.

NOTE:

You must be sure to determine the exact weight of the calibration weights before they are used to calibrate the instrument.

The Instrument Control weight calibration functions guide you through the calibration procedure step-by-step, see the online help and documentation for further information.

Sample Platform Calibration

The sample platform adjustment procedure is used if the sample hang-down wire fails to pick up a sample pan during an automatic loading procedure. The sample platform must be adjusted so that the instrument can properly load and unload the sample pans.

To avoid weight signal noise, the TGA must be level so that the sample pan and hang-down wire hang inside the furnace and thermocouple tube without touching them. The Instrument Control sample platform adjust functions take you step-by-step through the procedure, see the online help and documentation for further information.

Running a TGA Experiment

Experimental Procedure

All of your TGA experiments will have the following general outline. In some cases, not all of these steps will be performed.

- Selecting the pan type and material
- Loading the pan
- Taring the empty sample pan
- Loading the sample into the pan
- Entering experiment information through the TA controller (sample and instrument information).
- Creating and selecting the thermal method on the controller.
- Attaching and setting up external accessories as required (*e.g.*, purge gas, Gas Switching Accessory).
- Starting the experiment.

Preparing Samples

TGA experiments utilize different types of sample pans, depending upon the type of material that you are analyzing.

Selecting Sample and Tare Pans

Three kinds of sample pans are available for the TGA 2950, platinum, alumina ceramic, and aluminum. The platinum pans come in 50 and 100 μL sizes, the ceramic pans come in 100, 250, and 500 μL sizes, and the aluminum pans are 100 μL in size. The criteria for choosing a sample pan are as follows:

- For most experiments, platinum is the desirable choice. It is easy to clean and does not react with most organics and polymers. Ceramic pans are more porous and are therefore more easily contaminated. There are some conditions, however, in which other types pans are desirable, as explained below.
- Use ceramic pans for samples that might amalgamate or react with platinum (*e.g.*, metals, corrosives, inorganics).
- Use aluminum pans when disposability is desired. Aluminum pans are meant for one-time use in experiments that do not go above 600°C and for samples that do not react with aluminum.

- If your sample will melt during the experiment, use a pan that is deep enough to prevent spilling (the deepest pan is the 500 μ L ceramic pan).

The platinum and ceramic pan types are reusable. To clean between experiments, use a Bunsen burner or a propane torch, or run the pan through a hot thermal program in the TGA to burn out any residue. Aluminum pans are disposable, do not attempt to clean and reuse them.

Once you have selected the proper sample pan, remove the tare tube, and *using brass tweezers*, put the same type and size pan on the tare hook. Use the first step in the weight calibration procedure to mechanically tare the balance.

Taring the Sample Pan

Taring the sample pan ensures that the weight measured by the balance reflects the weight of the sample only. You should tare the sample pan before each experiment, even if you use the same pan in consecutive experiments.

When you tare a pan, the TGA reads the weight of the empty pan and then stores the weight as an offset, which is subtracted from subsequent weight measurements. For optimum accuracy, the weight reading must be stable before it is accepted as an offset. If you use the automatic tare procedure, the TGA will determine when the weight reading is sufficiently stable; or you can determine the acceptability of the weight reading by taring the system manually. Both tare procedures are explained here.

Automatic Tare

Because the TGA 2950 has two weight ranges, taring is done for both ranges. The tare weight is stored by the instrument for the appropriate weight range.

1. Place the empty sample pan on the sample platform.
2. Press the TARE key on the instrument keypad. The TGA will automatically load the pan, raise the furnace (to protect the pan from air currents), weigh the pan, store the weight as the offset for each weight range, and unload the pan.

Manual Tare

Manual tare operates in the weight range indicated, by storing the current reading as an offset, and estimates the tare weight for the other weight range (typically the 1-gm range). The estimate is accurate if the TGA 2950 has been tared or weight-calibrated recently.

1. Place the empty sample pan on the sample platform.
2. Press the LOAD key to load the pan onto the balance.
3. Press the FURNACE key to close the furnace, to protect the pan from air currents.
4. Observe the weight reading on the controller's Signal Display window (Signal A Weight).
5. Wait for the Signal A Weight to stabilize, and then choose **Auto Zero** to store the displayed weight as the offset.

Loading the Sample

After taring the sample pan, load the sample into the TGA furnace as follows:

1. Place the sample in the sample pan, and position the pan on the sample platform (Figure 3.1).

The wire on the bottom of the sample pan should align with the groove in the panhole, so that the sample pan can be picked up by the sample hang-down wire.

NOTE:

Always use brass tweezers to handle the sample pans.

CAUTION:

Manually loading the sample pan onto the hang-down wire may damage the balance mechanism.




Figure 3.1
Sample Pan
Ready to Load

2. Press the LOAD key. The TGA will automatically load the sample pan onto the balance.
3. Position the thermocouple at the edge of the sample pan, rather than in the middle, for best results (Figure 3.2).

NOTE:

The position of the thermocouple should be the same as it was during temperature calibration.




Figure 3.2
*Adjusting the
Thermocouple*

4. Press the FURNACE key to close the furnace by moving it up around the sample (Figure 3.3).




Figure 3.3
Furnace in
Closed Position

Setting Up an Experiment

Once you have prepared the sample, the next step in your experiment is to enter the needed information in the TA controller. All of the controller functions described in this section are accessed through the Instrument Control. Refer to the online help and documentation to learn how to perform the following steps.

1. **Select the Instrument.**
2. **Select the Instrument Mode.**
3. **Enter Sample Information.**
4. **Enter Instrument Information.**
5. **Create and Select Thermal Methods.**

The first time you use your TGA you will need to create at least one thermal *method* to control experiments. Each method is made of several *segments*, or individual instructions (*e.g.*, Equilibrate, Ramp), that control the state of the instrument.

Setting Up Accessories

If your experiment requires external accessories, ensure that they are turned on, and make any necessary adjustments before you start your experiment. Make sure that the system can achieve the conditions of all segments in the method.

This section describes how to use the following accessories with the TGA 2950:

- Air cool option
- Purge gas
- TA Instruments Gas Switching Accessory
- Evolved Gas Analysis (EGA) Furnace (see also Appendix D).

For the TGA HiRes™ Option, see Appendix B, and for the TGA Autosampler, see Appendix C.

The TGA can also be used with other accessories, such as vacuum, FTIR, gas chromatographs, mass spectrometers, and evolved gas analyzers. Consult the appropriate local instrument manufacturer for further information.

Using the Air Cool Option

You can program the system to air-cool the furnace automatically at the end of the experiment by selecting the air cool method-end condition as one of your instrument parameters (see the online help and documentation for further information.). After the air cool is activated, it will continue to run for the desired time.

Before you start an experiment that uses the air cool option, ensure that the supply valve from the air source is open and that the pressure is regulated to between 25 and 120 psig. Nitrogen can also be used as a cooling gas.

NOTE:

If you are using the EGA furnace, Air Cool can be used with the furnace closed. However, if the temperature is above 500°C, the EGA furnace will cool naturally until it is 500°C or less, then Air Cool will begin.

Using a Purge Gas

You can control the sample atmosphere during TGA experiments by connecting a purge gas to the system. Purge gas is distributed separately to two parts of the TGA: the furnace and the balance chamber.

The balance purge maintains a positive pressure in the balance chamber to prevent decomposition products from contaminating the sensitive balance mechanism. The balance purge flows from the balance chamber via two routes: down the thermocouple tube and through an outlet in the balance chamber to the right of the thermocouple tube. It then exits across the sample pan along with the furnace purge.

The purge flow through the furnace is horizontal to the sample (see Figure 3.4), permitting rapid removal of decomposition products from the sample environment.




Figure 3.4
Furnace Purge

You can choose nitrogen, oxygen*, helium, air, or argon for your purge environment. Do not use any other gases in the TGA 2950. See the “Chemical Safety” section on page xv for further details.

! WARNING

Do not use hydrogen or any other explosive gas in the TGA 2950 furnace or the TGA 2950 EGA furnace. *Oxygen may be used as a purge gas, but the furnace must be kept clean of volatile hydrocarbons to prevent combustion.

! WARNING

Do not use any liquid in the purge lines.

Purge gas can be obtained from a pressurized cylinder or an in-house supply source. Gas supplied from an in-house source should be passed through a sieve dryer to remove any trace of moisture before it enters the TGA.

It is important to maintain the proper ratio of flow rates between the balance chamber and the furnace housing. Having the separate balance chamber purge prevents decomposition gases from entering the balance chamber. The recommended setting for the purge rate is 100 mL per minute or less. When you use the standard TGA 2950 furnace, the flow distribution should be 40 percent to the balance chamber and 60 percent to the furnace. When you use the TGA 2950 EGA furnace, the flow distribution should be 10 percent to the balance chamber and 90 percent to the furnace.

To maintain this flow distribution, you will need to connect a flowmeter to each of the purge fittings on the back of the TGA instrument. Set the purge gas flow rates by adjusting these meters. The PURGE port goes to the TGA furnace, and the BALANCE PURGE goes to the balance chamber.

Before you start the purge gas, make sure that the desired gas is connected to the purge ports, that all lines are clear, and that your supply of purge gas is sufficient for the experiment.

Always maintain constant purge flow rates and distribution throughout your experiment; changing the purge during an experiment can affect the data.

Using the Gas Switching Accessory

You can use the Gas Switching Accessory to turn the purge gas on and off or to switch between two different purge gases during a TGA experiment. Before starting an experiment that uses the Gas Switching Accessory, make sure its power switch is on, and make sure the necessary gas sources are properly connected.

The Gas Switching Accessory can be controlled by the Gas segment in the method (see the online help and documentation for further information).

Connect the Gas Switching Accessory to the purge port only, when switching between gases during an experiment. Attach the inert gas to GAS 1 and the other gas to GAS 2.

Consult your Gas Switching Accessory operator's manual for further instructions.

Starting an Experiment

Before you start the experiment, ensure that the TGA 2950 is online with the controller and you have entered all necessary experimental parameters.

NOTE:

Once the experiment is started, operations are best performed at the controller keyboard. The TGA 2950 is very sensitive to motion and might pick up the vibration caused by pressing a key on the instrument keypad.

Start the experiment by pressing the START key on the instrument keypad, or selecting start on the TGA Instrument Control program (see the online help and documentation for further information). When you press the START key, the system automatically loads the sample pan and closes the furnace if necessary, and then runs the loaded method to completion.

Forced Start

If you wish to start collecting data during instrument setup, you can use the *forced start* feature. This is most useful for samples that lose a significant amount of weight during the set-up period (*i.e.*, samples with volatile solvents). When a forced start is initiated, the current sample weight is stored as the initial weight, data collection is started immediately, and the instrument status changes from “Set Up” to “Started.” The methods begin when the normal set-up procedures are completed.

To begin a *forced start*, press START on the instrument keypad while the status line displays “Set Up.”

Stopping an Experiment

If for some reason you need to discontinue the experiment, you can stop it at any point by pressing either the STOP key on the TGA 2950 keypad or **Stop** on the on the TGA Instrument Control program (see the online help and documentation for further information). Another function that stops the experiment is **Reject**. However, the **Reject** function discards all of the data from the experiment; the **Stop** function saves any data collected up to the point at which the experiment was stopped.

NOTE:

The Heat Exchanger will continue to run as long as the Air Cool option is activated or until the indicated temperature is below 50°C.

◆ **CAUTION:**

The **REJECT** function discards all experiment data.

Unloading the Sample

If you select the method-end option that enables the furnace to open and unload at the end of the experiment, the TGA will automatically unload the sample at the end of the run. (See the online help and documentation for further information.) If you need to unload the sample manually, wait until the run and all method-end operations are complete, and then press the UNLOAD key. The sample pan may not line up with the sample platform groove at method end.

Use in an Oxygen-Free Atmosphere

If you choose to perform TGA experiments in an oxygen-free atmosphere, a few extra precautions to the purge gas system and instrument setup are necessary to ensure an oxygen-free environment.

Purge Gas System

When performing TGA experiments in an oxygen-free atmosphere, take the following precautions:

- Use a high purity, inert gas of grade 5 or better. It may be necessary to use an oxygen trap in-line, depending on the purity grade.
- Choose a 2-stage regulator of diaphragm construction for high purity applications.
- Use copper or stainless steel tubing from the gas regulator, to the flowmeters, and to the TGA purge inlet ports.
- Allow the TGA to prepurge (under closed conditions) for at least 30 minutes after first turning on your purge gas.
- Increase the standard purge rate during this time to 100 mL/minute flow into the balance chamber and 100 mL/minute into the furnace.

Instrument Set-Up

When performing TGA experiments in an oxygen-free atmosphere, take the following steps to set up your instrument:

- Readjust your flowmeter(s) for standard operating flow rates: (a) for the standard furnace, 40 mL/minute flow into balance chamber and 60 mL/minute into the furnace or (b) for the EGA furnace, 10 mL/minute into the balance chamber and 90 mL/minute into the furnace.
- Tare sample pan (as needed).
- Load sample and close system
- Purge 20 minutes, if possible, before starting a run.
- When run is complete, allow the furnace to cool in the closed position until the temperature is less than 350°C. This can be automatically programmed into your method; after last segment add an equilibrate at 350°C step. Another option is to change method end conditions to leave furnace closed at method end.
- When using air or oxygen during an experiment, introduce new gas through furnace purge port only and switch back to nitrogen before cooling down.

NOTE:

When the TGA is idle, leave the system closed and continue purging with nitrogen.

Running Experiments

CHAPTER 4: Technical Reference

Description of the TGA 2950	4-3
Components	4-5
Balance	4-6
Sample Loading Assembly	4-8
Furnace	4-9
TGA Standard Furnace	4-9
EGA Furnace	4-11
Cabinet	4-13
Heat Exchanger	4-15
Theory of Operation	4-16
Status Codes	4-17

Description of the TGA 2950

The TGA 2950 operates on a null balance principle. Physically attached to a taut-band meter movement, the balance arm is maintained in a horizontal reference position by an optically actuated servo loop. When the balance is in a null position, a flag located on top of the balance arm blocks an equal amount of light to each of the photodiodes (the light is supplied by a constant current infrared LED). As sample weight is lost or gained, the beam becomes unbalanced, causing an unequal amount of light to strike the photodiodes. The unbalanced signal, called the error signal, is acted upon by the control circuitry and reduced to zero, or nulled. This is accomplished by an increase or decrease in the current to the meter movement, causing it to rotate back to its original position (null position). The change in current necessary to accomplish this task is directly proportional to the change in mass of the sample. This current is converted to the weight signal.

The TGA 2950 has two weight ranges: 1 gm and 100 mg. Both ranges are continuous over their weight loss operating range, which means that the entire weight loss range can be viewed without any loss of information. The weight loss operating ranges are:

- 1 gm to 0 μ g for the 1 gm range
- 100 mg to 0.0 μ g for the 100 mg range.

Negative weight (tare imbalance) is limited to 150 mg in the 1 gm range, and to 15 mg in the 100 mg range. Range control is automatic.

During normal operation of the TGA, the sample may evolve gases. To prevent back diffusion of these liberated gases to the balance chamber, the balance chamber is purged with an inert gas (standard furnace = 40 mL/min, EGA furnace = 10 mL/min). An inert gas must be used to prevent contamination or corrosion of the balance.

Heating rate and sample temperature are measured by the thermocouple located above the sample.

Components

The TGA 2950 has five major components, illustrated in Figure 4.1: the balance, the sample loading assembly, the furnace, the cabinet, and the heat exchanger.




Figure 4.1
TGA 2950 Components

The **balance**, the most important part of the TGA system, provides precise measurement of the sample weight. The **sample loading assembly** automatically loads and unloads samples from the TGA balance. The **furnace** controls the sample atmosphere and temperature. The **cabinet** contains the system electronics and mechanics. The **heat exchanger** dissipates heat from the furnace.

Balance

The TGA balance assembly (Figure 4.2) consists of the balance meter movement, the balance arm, the balance arm sensor, the hang-down wire assemblies, the sample pan, and the tare pan.




Figure 4.2
TGA Balance Assembly

The **balance meter movement** is a taut-band meter movement to which the balance arm is attached.

The **balance arm** is a rhombic piece of aluminum attached to the meter movement. It is in a null balance system. A hang-down loop is attached to each end to hold the hang-down wires.

The **balance arm sensor** is a printed circuit board assembly that detects the null position of the meter movement. The balance beam sensor is mounted above the balance arm. It is used in conjunction with the analog circuitry to maintain a null position.

The TGA has two **hang-down wire assemblies**: one for the tare pan and one for the sample pan. Each assembly consists of a hang-down wire and loop. The hang-down wire has hooks at each end and connects the pan to the loop. The loop has eyelets at each end; it is used to connect the hang-down wire to the balance arm. The longer hang-down wire (4 inches) is for the sample pan.

Sample pans are available in platinum in 50, and 100 μL sizes, alumina ceramic in 100, 250, and 500 μL sizes, and aluminum in 100 μL . All pans are 0.4 inches in diameter.

The **tare pan** holds the counter-balance weight that mechanically subtracts out the weight of the sample pan and hang-down wire.

Sample Loading Assembly

The sample loading assembly (Figure 4.3) is a platform that pivots the sample pan to the furnace area, where the pan engages the hang-down wire from the balance assembly. It also pivots the platform away from the furnace area for easy sample loading and unloading.




Figure 4.3
Sample Loading Assembly

Furnace

Two types of furnaces may be used with the TGA 2950 instrument, the standard TGA furnace, or the Evolved Gas Analysis (EGA) furnace. The EGA Furnace (see Appendix D) is an optional accessory that allows you to connect a spectrometer to the instrument so that the gases evolved by sample decomposition can be analyzed.

TGA Standard Furnace

The TGA standard furnace (Figure 4.4) consists of a furnace housing, a heater, and a furnace base that moves these parts up (to closed position) and down (to open position).




Figure 4.4
TGA Furnace

The stainless steel **furnace housing** has a purge opening on either side to allow gas flow or evacuation. Purge gas enters through a purge fitting on the right side of the furnace housing. It then passes through an opening in the heater, flows around the sample pan, and exits through openings on the opposite side of the furnace housing and heater. From here, the evolved gas can be sent through appropriate connections to an external analyzer, if desired.

The **heater** is a resistance-wound unit of low thermal mass alumina material that can be heated and cooled rapidly. Controlled heating rates of up to 100°C per minute can be obtained, to an upper limit of 1000°C.

Cooling air enters through the holes in the base of the furnace assembly at the completion of test runs, if desired.

A Platinel II* **thermocouple** extends through the bottom of the balance chamber, down along the hang-down wire, and is positioned just above the sample pan, where it monitors the sample environment temperature.

The **furnace base** moves the furnace assembly up around the sample pan to the closed position, or down away from the sample pan to the open position.

EGA Furnace

The (Evolved Gas Analysis) EGA furnace consists of a quartz glass sample tube surrounded by an electric resistance heater, both of which are contained within a water-cooled furnace housing. The housing is mounted to a furnace base that raises and lowers the furnace for sample loading and unloading. See Figure 4.5




Figure 4.5
EGA Furnace

The **sample tube** has a purge gas inlet that passes through the right side of the furnace housing. A fitting on the left side of the housing allows connection of a transfer line to carry exhaust gas to a spectrometer such as a FTIR. Because the heater is external to the sample tube, evolved gases from sample decomposition

within the sample tube do not come in contact with the resistance elements or the furnace ceramic refractory.

Cooling air enters through the furnace base and passes upward between the outside of the sample tube and the inside of the furnace, completely separating the cooling air from the sample and the sample zone.

The **furnace** is a resistance heater wound on alumina ceramic, which allows sample zone temperatures as high as 1000°C with heating rates up to 50°C/min.

The **thermocouple** and **furnace base** are the same as those described above for the standard TGA furnace.

*Platinel II is a registered trademark of Engelhard Industries.

Cabinet

The TGA cabinet (Figure 4.6) consists of the cabinet housing, the instrument display and keypad, the electronics compartment, the purge and cooling gas fittings, and the rear panel.




Figure 4.6
TGA Cabinet Components

The TGA **cabinet housing** consists of a base casting, a rear cover, and a sample preparation tray. The base casting is one-piece casting of heavy-weight aluminum, designed to provide a stable platform for the TGA instrument parts. The rear cover is injection molded using a heavy-gauge thermoplastic material, designed for easy cleaning. The removable sample preparation tray is located on the right side of the instrument. The tray is designed to keep liquids from spilling into the instrument.

The **instrument display and keypad** are described in Chapter 1.

The **electronics compartment** contains the electronics that control the instrument functions.

Two **purge fittings** are located on the left side of the instrument back, one for the balance chamber and one for the furnace housing.

The **cooling gas fitting** is located on the right side of the instrument back and is for furnace cool-down air.

The **rear panel** (Figure 4.7) has the signal and power connections for the instrument, the fuses, the ready light, the address switches, and the Reset button.




Figure 4.7
Rear Panel

Heat Exchanger

The heat exchanger (Figure 4.8) consists of a fan, a radiator, a water reservoir, a pump, a temperature switch, and a flow switch.

The **fan** blows cool air through the radiator.

The **radiator** exchanges heat between the water and air.

The **water reservoir** holds additional water that may be required by the system.

The **pump** pushes the water through the system.

The **temperature switch** detects over-temperature conditions, which could be caused by failure of the fan.

The **flow switch** detects lack of flow, which could be caused by failure of the pump or a leak in the system.




Figure 4.8
Heat Exchanger

Theory of Operation

Thermogravimetric analysis (TGA) is a thermal analysis technique for measuring the amount and rate of change in sample mass as a function of temperature and time. It is used to characterize any material that exhibits weight loss or phase changes as a result of decomposition, dehydration, and oxidation. Two modes are commonly used for investigating thermal stability behavior in controlled atmospheres: (1) dynamic, in which the temperature is increased at a linear rate, and (2) isothermal, in which the temperature is kept constant.

Status Codes

Status codes are character strings that are continuously displayed at the top left of the TGA instrument display. These codes tell you what segment in the method is currently being performed by the instrument.

Table 4.1
Status Codes

Code	Meaning
Air Cool	The furnace air cool line has been opened to cool the furnace.
Calib	The TGA is in calibration mode.
Closing	The furnace assembly is closing.
Cold	The instrument heater cannot supply heat fast enough to keep up with the thermal program. This may be caused by a large ballistic jump in the program, a faulty heater, or a faulty control thermocouple signal.
Complete	The thermal method has finished.

(table continued)

Table 4.1
(continued)

Code	Meaning
Cooling	The heater is cooling, as specified by a Ramp segment.
Ending	The method is complete and the EGA furnace is cooling until it can Air Cool or open and unload.
Equilb	The temperature is being equilibrated to the desired set point.
Err <i>n</i>	An error has occurred. The instrument display will give the error code number (<i>n</i> , a two-or-three-digit code); the controller screen will also show the complete error message.
Heating	The heater temperature is increasing, as specified by a Ramp segment.
Holding	Thermal experiment conditions are holding; the program is suspended. Choose Start to continue the run.

(table continued)

Table 4.1
(continued)

Code	Meaning
Hot	The temperature is beyond the set point, and the instrument cannot remove heat fast enough to follow the thermal program. This is usually caused by a large ballistic jump to a lower temperature.
Initial	The temperature is being equilibrated to the desired set point. When the temperature has reached equilibrium, the status will change to Ready .
Iso	The thermal program is holding the current temperature isothermally.
Jumping	The heater is jumping ballistically to the set point temperature.
Load	The TGA is loading a sample pan onto the balance.
No Power	No power is being applied to the heater. Check the heater switch and fuse.

(table continued)

Table 4.1
(continued)

Code	Meaning
Opening	The furnace assembly is opening.
Ready	The system has equilibrated at the initial temperature and is ready to begin the next segment. Choose Start to continue the method.
Reject	The experiment has been terminated and the data erased.
Repeat	The method is executing a repeat loop that does not involve temperature control segments.
Set Up	The system is loading the sample, closing the furnace, and letting the weight stabilize before beginning the first segment.
Stand by	The method and method-end operations are complete.

(table continued)

Table 4.1
(continued)

Code	Meaning
Started	The TGA is still setting up to start the experiment (see Set Up above), but the initial weight has been measured and data collection has begun. The thermal method will start when the normal setup process has been completed.
Tare	The TGA is measuring the weight difference between an empty sample pan and the tare pan. The measured weight is used as an offset so that the displayed weight value indicates the weight of the sample only.
Temp °C	The heater is in stand-by mode, and the experiment has been terminated.
Temp*	Temperature calibration is in effect. The heater is in stand-by mode, and the experiment has been terminated.
Unload	The TGA is unloading a sample from the balance.
Weight #	The weight reading is not stable.

CHAPTER 5: Maintenance and Diagnostics

Overview	5-3
Routine Maintenance	5-4
Inspection	5-4
Cleaning the Instrument	5-4
Cleaning the Furnace Housing	5-5
TGA Standard Furnace Only	5-5
Heat Exchanger	5-6
Maintaining Heat Exchanger Coolant	5-6
Draining and Refilling the Water Reservoir	5-6
Replacing the Thermocouple	5-9
Removing and Reinstalling the Furnace	5-11
Furnace Removal	5-11
Furnace Replacement	5-14
Diagnosing Power Problems	5-16
Fuses	5-16
Power Failures	5-18

Maintenance and Diagnostics

TGA Test Functions 5-19

The Confidence Test 5-19

Replacement Parts 5-21

Overview

The procedures described in this section are the customer's responsibility. Any further maintenance should be performed by a representative of TA Instruments or other qualified service personnel.



Because of the high voltages in this instrument, untrained personnel must not attempt to test or repair any electrical circuits.

Routine Maintenance

Inspection

Examine the instrument periodically to keep it free of dust, debris, and moisture. Keep the furnace area clean. Any sample spillage of residue should be removed before the next experiment.

Cleaning the Instrument

You can clean the TGA instrument keypad as often as you like. The keypad is covered with a silk-screen Mylar* overlay that is reasonably water resistant but not waterproof or resistant to strong solvents or abrasives.

A household liquid glass cleaner and paper towel are best for cleaning the instrument keypad. Wet the towel, not the keypad, with the glass cleaner, and then wipe off the keypad and display.

Cleaning the Furnace Housing



Do not touch the EGA furnace sample tube with your bare fingers. Skin oils may cause devitrification of the quartz glass, resulting in severely reduced sample tube life. Do not insert instruments inside the sample tube to scrape or chip contaminants from the sample tube as breakage may result.

TGA Standard Furnace Only

To ensure long furnace life, we recommend that you clean the furnace housing at least once a month to remove condensation materials. Follow steps 1-6 to remove furnace housing (page 5-8 and 5-9). Disconnect the purge line and invert the furnace housing over paper towels. Clean the inside with a solvent and cotton squabs. Be sure to dry the housing and purge ports with air to remove any solvent traces before replacing the furnace housing. Follow the instructions on page 5-10 and 5-11 for replacing the standard furnace housing. After replacing the furnace, heat the TGA to 900°C to remove any remaining solvent.



If you are routinely evaluating materials in the TGA that lose a large amount of volatile hydrocarbons (e.g., lubricating oils), you need to clean the standard furnace more frequently to prevent dangerous buildup of debris in the furnace. This is particularly important if you are using oxygen as a purge gas.

Heat Exchanger

The heat exchanger does not require any maintenance other than to maintain the level and quality of the liquid coolant. If the level drops too low, or the coolant becomes contaminated, this could result in problems with your instrument.



Do not put any liquid other than distilled water and TA Conditioner in the heat exchanger reservoir.

Maintaining Heat Exchanger Coolant

You should check the level and condition of the heat exchanger coolant periodically. We recommend routine checks every three to six months, depending on use of the instrument.

Add distilled water to the reservoir, if necessary, to keep the reservoir at least 2/3 full. If algae growth is visible, drain the reservoir bottle, refill it with distilled water, and add TA Instruments TGA Conditioner, as described in the next section.

Draining and Refilling the Water Reservoir

Drain and refill the heat exchanger water reservoir as follows:

1. Turn off the POWER switch and disconnect the heat exchanger cable and water lines from the instrument cabinet (see Chapter 2 for instructions).

2. Unscrew and remove the water reservoir cap.
3. Drain the coolant and flush out the system as follows:
 - a. Lift the heat exchanger and dump out the contents of the water reservoir bottle.
 - b. Fill the bottle to 2/3 full with distilled water only and replace the cap.
 - c. Reconnect the heat exchanger cable and water lines to the instrument cabinet.
 - d. Turn on the POWER switch.
 - e. Turn on the pump by activating Air Cool and allow the water to circulate for several minutes.
 - f. Turn off the pump by deactivating Air Cool, and check the clarity of the water in the reservoir bottle.
 - g. If the water clarity is still unacceptable, disconnect the heat exchanger cable and water lines from the instrument cabinet, and repeat steps a through f.
 - h. Continue repeating this procedure until you are satisfied with the clarity of the water in the bottle after it has circulated.
4. Dispose of the water and fill the bottle with TGA Conditioner (PN 952377.901) and fresh distilled water as directed in Chapter 2.

5. Turn on the pump again by activating Air Cool, and circulate the water until the air bubbles disappear from the water lines. (You may see “Err 119” on the instrument display until all the air has been removed.)
6. Replace and tighten the water reservoir cap.

Replacing the Thermocouple

1. Unload the sample pan and open the furnace.
2. Using the ball driver supplied in your TGA accessory kit, loosen and remove the six screws securing the balance chamber faceplate to the instrument.
3. Take off the faceplate.
4. Push the thermocouple up from the bottom, to feed it back into the balance chamber (Figure 5.1).




Figure 5.1
Removing the
Thermocouple

5. Unplug the thermocouple from its connector and remove the thermocouple from the balance chamber.

6. Plug the new TGA thermocouple into the connector.
7. Thread the new thermocouple down through the hole next to the thermocouple tube.
8. Thread the end of the thermocouple just through the ceramic disk at the end of the thermocouple tube.
9. Load a sample pan to make sure that the end of the thermocouple does not touch it (Figure 5.2).




Figure 5.2
Checking Position
of Thermocouple

10. Make sure that the hang-down wire does not touch the top of the thermocouple inside the balance chamber.
11. Replace the balance chamber faceplate and screws.

Removing and Reinstalling the Furnace

To remove or reinstall the furnace, you will have to remove the furnace arm from its connection inside the slot on the front of the instrument cabinet.

Furnace Removal

To remove the furnace use the following procedure:

1. Press the FURNACE key to open the furnace completely.
2. With the ball driver supplied in your TGA Accessory Kit, loosen the two screws on each side of the furnace arm connection, within the slot on the front of the instrument cabinet.

In Figure 5.3 on the next page, the furnace arm/base has been removed to better show the location of the screws and alignment of the cleats.

NOTE:

To obtain access to the upper left screw and cleat, loosen the three hold-down screws on the furnace base. Rotate the furnace housing gently counter-clockwise to move the coolant connections away from the front of the cabinet slot.




Figure 5.3
Aligning Cleats for
Removal of
Furnace Arm/Base

3. Rotate each screw to turn the D-shaped cleat so that the flat edge of the cleat is aligned vertically and parallel with the groove in the furnace arm (Figure 5.3).
4. While holding the furnace base in one hand, press the FURNACE key to raise the furnace about 1/4 of the way up, and press STOP.
5. Unplug and remove the furnace arm/base from the instrument cabinet (Figure 5.4).
6. Loosen the hold-down screws, if necessary, and remove the furnace housing from the furnace base, being careful not to disturb the coolant connections. Lay the furnace housing on the front ledge of the cabinet.




Figure 5.4
Removing Furnace
and Housing
from Cabinet

7. Unplug the Air Cool line from the bottom of the furnace arm/base (Figure 5.4). The furnace arm/base is now completely free of the instrument.

NOTE:

When you remove the Air Cool line, do not let it slip back into the instrument cabinet.

Furnace Replacement

To replace or reinstall the furnace:

1. Plug the Air Cool line into the bottom of the furnace arm/base.
2. Slip the furnace housing over the furnace arm/base and tighten the three hold-down screws only enough to keep the housing and base together while you reinstall them.

NOTE:

Before trying to reconnect the furnace arm to the plug inside the cabinet slot, it may help to back each of the four screws inside the slot out one more complete turn, again aligning the flat edge of each cleat so that it is parallel with the groove on the furnace arm. (Loosening the screws one more turn each will give you a bit more room to maneuver the arm into its connection.) In order to loosen the screws, press the FURNACE key to lower the connection so that you can reach it through the wide part of the slot. When you have aligned the cleats, press the FURNACE key again to raise the furnace about 1/4 of the way up.

3. Plug in the furnace arm.
4. Continuing to hold the furnace base, press the FURNACE key to lower the furnace completely.
5. Using the ball driver, tighten the two screws on either side of the furnace arm connection, making sure that the curved edges of all four cleats engage the groove on the furnace arm.

NOTE:

To obtain access to the upper left screw and cleat, rotate the furnace housing gently counterclockwise. This will move the purge and coolant connections away from the front of the cabinet slot.

7. Rotate the furnace housing clockwise until it is aligned correctly, and tighten the three hold-down screws completely.

Diagnosing Power Problems

Fuses

The TGA contains internal fuses that are not user serviceable. If any of the internal fuses blows, a hazard may exist. Call your TA Instruments service representative.

The only fuses that you should service yourself are the external fuses, located on the TGA's rear panel. Both are housed in safety-approved fuse carriers, labeled F1 and F2 (Figure 5.5).



Always unplug the instrument before you examine or replace the fuses.




Figure 5.5
Fuse Locations

Fuse 1 is in the circuit between the main electrical input and the POWER switch. All power for internal operations and instrument functions, except heater power, passes through this fuse. If this fuse blows, you will get no response from the instrument when you attempt to turn it on.

Fuse 2 protects the heater coils in the furnace. Because fuse 2 does not power the internal logic, you may not know that this fuse is blown until you try to heat a sample; the TGA passes the confidence test with fuse 2 open.

Fuse 2 is always checked at the beginning of a method. Power supplied by this circuit is switched by a computer-controlled relay as well as by the HEATER switch located on the instrument's front panel. When both devices are active, the light in the HEATER switch will glow.

Power Failures

A power failure caused by a temporary drop in line voltage results in one of two responses by the TGA instrument:

If the drop is fairly large and of long duration (20 milliseconds or more), the system will reset and go into its power-up sequence when power resumes.

If the drop is small or of short duration, the system may halt, and you may see “Err F02” on the display. This message means that the system has detected a power failure and has shut down. The instrument will not restart until it is reset. To reset, press the Reset button on the TGA’s back panel.

If “Err F02” appears at start-up and remains even after you have tried to restart the instrument, the detection circuitry itself is probably at fault. Do not try to repair it yourself; call your TA Instruments service representative.

The TGA is designed for a nominal line voltage of 115 volts AC (+ 10%), 50 or 60 Hz. It should not be operated outside this range. Low line voltage may result in poor instrument operation; high line voltage may damage the instrument.

TGA 2950

Test Functions

The TGA 2950 has three levels of test and diagnostic functions:

- The confidence test that is run every time the instrument is started.
- Cycling test functions that continuously test specific.
- A manufacturing verifier test mode that coordinates and logs the results of a sequence of confidence test and drift runs.

These test functions are always present in the instrument. They are designed to aid manufacturing and service in checking and repairing the instrument.

The Confidence Test

The TGA confidence test is run each time the instrument is turned on or reset. The confidence test checks most of the computer and interface components in the system.

When the confidence test is running, the number of the test currently being performed is shown on the display. The test number appears as a two-digit hex number on the lower right of the display. This number is changed as each new test is started. Most of the tests are very brief, so their test numbers may not be apparent.

A standard TGA system takes about 12 seconds to run the confidence test. The longest tests are the RAM tests, which take about 6 seconds.

After the tests are completed, a sign-on message is displayed for 3 seconds. The system then starts running, and the Ready light on the back of the instrument glows.

If an error is detected, an error message is posted on the bottom line of the display. Nonfatal errors are displayed for 3 seconds, and then the confidence test continues. A fatal error occurs when a circuit essential to the operation of the instrument has failed the confidence test; the instrument cannot reliably perform any further functions. The system stops when the fatal error is posted, and the Ready light remains off.

Table 5.1 summarizes the primary confidence tests for the TGA. If any errors occur during the confidence test, call your TA Instruments service representative.

Table 5.1
TGA Confidence Test

Test Number	Area Being Tested
—	CPU logic
30	CMOS RAM
4 <i>n</i>	Program memory
5 <i>n</i>	CPU board I/O functions
6 <i>n</i>	DRAM data storage memory
70	GPIB test
82	Keypad test
A <i>n</i>	Analog board tests
B <i>n</i>	Drive board tests
D0	Saved memory checksum

Replacement Parts

Replacement parts for the TGA 2950 are available from TA Instruments and are listed in Table 5.2.

Table 5.2
List of TGA
2950 Parts

Part Number	Description
952018.906	100 μ L platinum sample pan kit
952018.907	100 μ L ceramic sample pan kit
952040.901	Sample hang-down wire
952040.902	Tare hang-down wire
952011.906	Class M calibration weight kit (100 mg and 1 gm)
269931.001	Class M cal. wt. 100 mg
269931.002	Class M cal wt. 1 gm
952018.908	50 μ L platinum sample pan kit
952323.902	100 μ L aluminum sample pan kit
952018.909	250 μ L ceramic sample pan kit
952018.910	500 μ L ceramic sample pan kit
952384.901	TGA Temperature Calibration kit
952385.901	TGA nickel reference material
952013.901	Furnace assembly
952014.901	Balance assembly
952017.001	Tare tube
	(table continued)

Table 5.2
(continued)

Part Number	Description
952310.901	Motor drive PCB
952060.901	Analog PCB
952068.901	Sample thermocouple assembly
952080.901	Sample motor assembly
952081.902	Furnace ∇ sensor
952081.903	Furnace Δ sensor
952081.904	Sample position sensor
952082.901	Furnace motor assembly
952121.001	Work surface tray
952183.901	Aluminum temperature calibration standard
900905.901	Calcium oxalate sample
990806.901	Air purge valve assembly
952324.001	TGA 2950 keypad assembly
990820.010	Instrument keypad
990828.901	Power supply assembly
990850.901	Central processor PCB
990880.901	Communications PCB
990870.901	Triac drive PCB
259508.000	Brass tweezers
259509.000	Spatula, curved, 165 mm long
265303.001	Instrument display
265749.001	O-ring, bottom of furnace housing
200063.029	O-ring, bottom cap plate of furnace
269845.001	O-ring, furnace housing to balance chamber
269920.002	Balldriver, 0.050-inch
269920.026	Balldriver, 7/64-inch
	(table continued)

Table 5.2
(continued)

Part Number	Description
269930.001	Class C calibration weight kit (1 mg to 500 mg)
952160.901	TGA 2950 Cooling Accessory
952160.903	Heat exchanger fan/ assembly
952162.901	Cooling Accessory tubing
952166.901	Cooling Accessory water reservoir bottle
952172.901	Cooling Accessory pump assembly
952161.901	Flow switch assembly
269932.001	Solid state relay
952377.901	Conditioner Kit

Maintenance and Diagnostics

Appendix A: Ordering Information

Email address: <http://www.tainst.com>; click on
“Answerman” icon.

TA Instruments, Inc.
109 Lukens Drive
New Castle, DE 19720
Telephone: 1-302-427-4000 or 1-302-427-4040
Fax: 1-302-427-4001

HELPLINE—U.S.A.
For technical assistance with current or
potential thermal analysis applications,
please call the Thermal Analysis Help Desk
at 1-302-427-4070.

SERVICE—U.S.A.
For instrument service and repairs,
please call 1-302-427-4050.

TA Instruments Ltd.
Europe House, Bilton Centre
Cleeve Road
Leatherhead, Surrey KT22 7UQ
England
Telephone: 44-1372-360363
Fax: 44-1372-360135

TA Instruments GmbH
Max-Planck-Strasse 11
D-63755 Alzenau
Germany
Telephone: 49-6023-9647-0
Fax: 49-6023-9647-77

TA Instruments Belgium
A Division of Waters s.a./n.v.
Raketstraat 60
B-1130 Brussels
Telephone 32-2- 706 00 80
Fax 32-2- 706 00 81

Appendix A

TA Instruments The Netherlands
A Division of Waters Chromatography B.V.
Florijnstraat 19
4879 AH Etten-Leur
Telephone 31-76-508 72 70
Fax 31-76-508 72 80

TA Instruments Japan
No. 5 Koike Bldg.
1-3-12 Kitashinagawa
Shinagawa-Ku, Tokyo 140
Japan
Telephone: 813/3450-0981
Fax: 813/3450-1322

TA Instruments France
B.P. 608
78056 Saint-Quentin-Yvelines
Cedex
France
Telephone: 33-1-30-48 94 60
Fax: 33-1-30-48 94 51

TA Instruments Spain
Waters Cromatografia, S.A.
División TA Instruments
Avda. Europa, 21. Pta. Baja
28108 Alcobendas
Madrid, Spain
Telephone: 34-91-203-9100
Fax: 34-91-661-0855

TA Instruments Australia
Unit 3
38-46 South Street
Rydalmere NSW 2116
Australia
Telephone: 61-29-9331-705
Fax: 61-29-8981-455

TA Instruments Italy
Division of Waters SpA
via Achille Grandi 27
20090 Vimodrone (MI), Italy
Telephone: 39-02-27421-1
Fax: 39-02-250-1827

Printed in U.S.A.

Appendix B: High Resolution™ TGA Option

Overview	B-5
Option Installation	B-6
Using Hi-Res™ TGA	B-7
Background	B-7
The TA Instruments Hi-Res™ Technique	B-8
The Hi-Res™ Ramp Segment	B-9
Calcium Oxalate Example Scans	B-13
Advanced Hi-Res™ Techniques	B-18
What Can Be Resolved and What Cannot?	B-18
Unresolved Transitions	B-19
Selecting a Hi-Res™ Technique	B-22
Dynamic Rate Hi-Res™ Ramp	B-22
Constant Reaction Rate Hi-Res™ Ramp	B-23
Weight Gain Experiments	B-26
Signature Analysis	B-27

Sample Quantity and Orientation	B-28
Exposed Surface Area	B-28
Bubble Formation	B-29
Thermocouple Placement	B-32
Data Analysis Effects	B-32
Derivative Plots	B-33
Adjusting Heating Rate	B-34
Adjusting Resolution Setting	B-37
Useful Resolutions Settings	B-39
Temperature Calibration	B-41
Hi-Res™ Transition Temperatures ...	B-42
Hi-Res™ Sensitivity Segment	B-43
Understanding Sensitivity Setting	B-44
Adjusting Sensitivity in Dynamic Rate Mode	B-45
Adjusting Sensitivity in Constant Reaction Rate Mode	B-47
Abort Segment	B-51
Stepwise Isothermal Heating	B-53

Hi-Res™ TGA Examples	B-58
Mixture of Bicarbonates	B-58
Dynamic Rate Scans	B-59
Varying Resolution Setting	B-61
Varying Sensitivity Setting	B-62
Constant Reaction Rate Scans	B-63
Stepwise Isothermal Scans	B-65
Monosodium Glutamate	B-68
Banana Taffy	B-69
Plastic Laboratory Tubing	B-71
References	B-75

Appendix B

Overview

This appendix describes how to use the High Resolution option for the TGA 2950 instrument.

Some of the benefits provided by the Hi-Res™ option are:

- Improved Transition Resolution
- Faster Survey Scans
- Enhanced Signature Analysis Capability
- Transition Temperatures Closer to Isothermal Values
- Increased Method Programming Versatility

The Hi-Res™ option provides three method programming steps (segments) which are used for the high resolution TGA capability. The new method segments are:

- High Resolution Ramp
- High Resolution Sensitivity
- Abort Next Segment

With the addition of these segments, method programming becomes more versatile and powerful than ever before. The Hi-Res™ ramp can be used alone as a simple single-segment method, or the method segments can be combined with more traditional segments, such as constant heating rate ramps and timed isothermal periods, for maximum programming flexibility.

What these segments do and how to use them to control TGA experiments and improve transition resolution is described in the following sections. For further details regarding creating methods, refer to the online help and documentation.

Option Installation

The Hi-Res™ TGA option is field installable by qualified service personnel (see separate installation procedure included with the Hi-Res™ option kit). The following items are required:

- Version 2.0 or higher TGA 2950 Instrument Software (Included in Kit)
- Version 8.2 or higher controller operating system
- Instrument “Software Option” circuit board (Included in Kit)
- Hi-Res™ TGA software option key (Included in Kit)

A TGA instrument with Hi-Res™ capability properly installed can be identified by the “Hi-Res TGA Installed” message on the instrument display screen following the confidence test, and by the letters “HR” in the instrument identification string on the configuration screen of the controller (*e.g.*, “TGA 2950 HR V2.0A”).

Using Hi-Res™ TGA

Background

TGA is particularly useful for observing the thermal decomposition of compounds. When individual thermal decompositions occur at well separated temperatures, quantitative information about sample composition can be obtained from the percent weight change at each transition. However, many TGA decomposition transitions overlap or appear drawn out in temperature due to the time-dependent nature of the reactions taking place. This overlap substantially reduces the ability to obtain an accurate measurement of weight change and reaction temperature.

It has long been known that using very slow heating rates will improve the separation of some overlapping transitions and, thus, increase the resolution of the TGA scan. Another technique to increase resolution is to increase furnace temperature until the onset of decomposition and then hold temperature isothermally until the decomposition is complete. After which, the temperature is raised again until the next decomposition begins, and so on until the final temperature of interest is reached. A third technique to increase resolution is controlling the furnace temperature in such a way as to maintain a preselected constant reaction rate (%/minute). This results in slower or even negative heating rates during a transition which gives the reaction more time to reach completion before the next transition is encountered.

The major drawback to these techniques is that they increase substantially the total time required for a measurement, thereby reducing laboratory productivity. Moreover, increasing measurement time often reduces the accuracy and reliability of the analysis. This is due to exposing the sample to high temperatures for long periods of time which may cause slow time-dependent changes such as oxidation and absorption, or exposure to changing ambient conditions such as humidity and pressure.

The TA Instruments Hi-Res™ Technique

The TA Instruments Hi-Res™ technique, dynamic rate TGA (DRTGA), differs from previous control techniques in that the heating rate of the sample material is dynamically and continuously modified in response to changes in the rate of decomposition of the sample so as to maximize weight change resolution. This technique allows the use of very high maximum heating rates during Hi-Res™ ramp segments while avoiding transition temperature overshoot. Typical Hi-Res™ ramps often take the same or less time to complete than a comparable constant heating rate experiment run at a lower heating rate, while providing improved resolution.

The Hi-Res™ Ramp Segment

The new Hi-Res™ ramp segment varies the heating/cooling rate of the furnace in response to changes in the rate of decomposition of the sample so as to improve weight change resolution. The new segment has the following format:

Ramp <rate>°C/min res <res_setting> to <temp>°C

where:

<rate> is the maximum ramp heating rate
(0.01 to 200°C/minute)

<res_setting> is the resolution setting
(-8.0 to +8.0)

<temp> is the ramp final temperature
(-200 to 1000°C)

example:

Ramp 50.00°C/min res 4.0 to 800.00°C

The Hi-Res™ ramp segment operates similarly to the traditional constant heating rate ramp segment, except that the heating rate is varied dynamically during the ramp in response to the derivative of weight change (%/minute). As percent/minute increases, heating rate is decreased. As percent/minute decreases, heating rate is increased. The heating rate is constrained to the range 0.001°C/minute (minimum) to the

maximum specified in the ramp segment. The resolution setting is a unitless number used to select the most useful band of percent/minute values for proportional heating rate control.

Higher resolution settings select lower percent/minute values, and generally result in increased resolution and longer experiment times. Lower settings have the opposite effect.

Resolution settings may be selected anywhere in the range of -8.0 to +8.0. Positive settings indicate that dynamic rate mode is to be used during the ramp. Negative settings indicate that constant reaction rate mode is to be used. More details on using each mode is covered in the Advanced Hi-Res™ Techniques section.

Positive resolution settings are the most universally useful settings and the least likely to have undesirable side effects. Although there are no hard and fast rules about which resolution setting to use for a given experiment, there are some general guidelines which will be helpful.

It has already been stated that higher resolution settings usually provide better resolution results and lower settings the reverse. The closer the setting is to zero the larger the derivative of weight change (%/minute) must be for a reduction in heating rate to occur. In fact, a resolution setting of exactly zero completely disables the application of the Hi-Res™ technique, resulting in a normal constant heating rate ramp at the maximum rate specified.

A resolution setting of +1.0 will produce a TGA scan at 50°C/minute that roughly approximates the resolution obtained by a constant heating rate scan at 20°C/minute. In other words, you

can zip through baseline sections of your scan at a higher heating rate while slowing down only for transitions and still get the resolution of the slower heating rate scan. For scans which contain a large amount of baseline this can result in very significant overall time savings with no loss of resolution. The same speed/resolution relationship applies to other heating rates as well.

Higher resolution settings apply the Hi-Res™ technique more aggressively by reacting to smaller percent/minute values. Some general rules-of-thumb for selecting resolution settings are as follows:

1. If you are uncertain what the resolution setting and heating rate should be, then try setting +3.0 and 50°C/minute. (Negative resolution settings are covered in the Advanced Hi-Res™ Techniques section.)
2. When trying to obtain better resolution, try progressively higher resolution settings, 1.0 at a time, while leaving the heating rate fixed.
3. The most useful resolution settings are 3.0 to 5.0 because they cover the most commonly encountered bands of percent/minute during typical decompositions. If your decompositions tend to be explosive in nature, with large percent/minute peaks (greater than 50 %/minute), or if you wish to minimize experiment time, then use settings less than 3.0. If your decompositions are very gradual (less than 0.5 %/minute peaks), or if it is important to limit the rate of decomposition, try settings greater than 5.0.

The most useful maximum heating rates are 10 to 50°C/minute for positive resolution settings. However, other values are perfectly acceptable when required by the needs of the experiment.

Use lower heating rates when transitions are very closely spaced in temperature, or if the sample material reacts very rapidly. A higher heating rate conventional ramp can be used to skip over baseline sections in order to shorten experiment time. Generally speaking, you no longer need to use very slow heating rates, less than 5°C/minute, because the dynamic rate Hi-Res™ technique automatically reduces heating rate with the accompanying improvement in resolution.

There are no special method programming constraints on Hi-Res™ ramp segments. They may appear anywhere in a method that a normal ramp could appear. The maximum ramp rate, resolution setting and final temperature of an executing Hi-Res™ ramp may be changed via the Modify Segment feature on the controller.

Selecting the resolution setting is covered in greater detail in the Advanced Hi-Res™ Techniques section.

Calcium Oxalate Example Scans

In this example, five TGA scans of calcium oxalate monohydrate ($\text{CaC}_2\text{O}_4 \cdot \text{H}_2\text{O}$) were run in nitrogen to compare the results using conventional TGA and Hi-Res™ TGA. In all cases a single ramp or Hi-Res™ ramp segment from ambient to 800°C was used for the method.

Figures B.1 and B.2 show the results of constant heating rate scans at 20°C/minute and 1°C/minute. Figures B.3, B.4 and B.5 show the results of Hi-Res™ scans at 50°C/minute with resolution settings 3.0, 4.0 and 5.0 respectively. Figure B.6 shows a composite plot of the derivative of weight loss for each of the five scans. The derivative smoothing window for all plots was set to 5°C.




Figure B.1




Figure B.2




Figure B.3




Figure B.4





Figure B.5

**Figure B.6**

As can be seen in these examples, calcium oxalate has three well resolved transitions corresponding to the loss of water (1st weight loss), carbon monoxide (2nd weight loss) and carbon dioxide (3rd weight loss). Comparing the 20°C/minute and the 1°C/minute scans (Figures B.1 and B.2), we see some improvement in resolution of the water loss (1st transition), but little improvement in the other two transitions.

Note that the transition temperatures in the slower scan are shifted to lower temperatures as expected. The 20°C/minute scan took 39 minutes and the 1°C/minute scan took nearly 13 hours to complete.

Comparing the results of the two constant heating rate scans to the Hi-Res™ scans in Figures B.3, B.4 and B.5 we can see that the resolution 3.0 scan (Figure B.3) gave comparable results in two thirds the time of the 20°C/minute scan. The resolution 4.0 scan (Figure B.4) gives much improved resolution in about the same time as the 20°C/minute scan, and the resolution 5.0 scan gives a dramatic improvement in only twice the time.

Figure B.6 shows a plot of the weight percent derivatives from each of the calcium oxalate scans overlaid on the same scale. Note how much taller and narrower the Hi-Res™ peaks are compared to the conventional scans. We can also clearly see that transition temperature is reduced by increasing the resolution setting. This effect is normal because with progressively higher resolution settings, transitions are constrained to lower decomposition rates which can only be maintained at lower temperatures.

As can be seen in these simple examples, the Hi-Res™ ramp segment is easy to use and provides significant resolution improvement, in the same time frame as conventional constant heating rate TGA. The following section on Advanced Hi-Res™ Techniques contains valuable information about additional Hi-Res™ methods and adjustments which will help you obtain maximum performance from this powerful analysis tool.

Advanced Hi-Res™ Techniques

This section discusses in greater detail how to use the Hi-Res™ features and gives specific advice which may be helpful in setting up your own experiments. As with any new analytical tool, there is a learning period required during which the user becomes familiar with the options and adjustments and gains knowledge about what can and cannot be expected from the analysis.

What Can Be Resolved and What Cannot?

“Will Hi-Res™ TGA improve the resolution of my transitions?” This is one of the first questions asked by most people when they are introduced to Hi-Res™ TGA. It would be wonderful, if the answer were always an unqualified “yes”. Unfortunately, there are some applications where resolution enhancement will be minimal. Therefore, some criteria for selecting likely candidates for resolution improvement is needed. Outlined below are some guidelines which can be employed whenever a new material is considered, or when a sample is tried and no resolution improvement is obtained. (It is worth noting that substantial productivity gains are still possible, even if transition resolution is not improved.)

Unresolvable Transitions

The Hi-Res™ techniques provide useful tools to improve transition resolution of many sample materials, but some materials will show little or no resolution improvement. This is because these materials have transitions which cannot be separated by time and temperature alone. The transitions are usually overlapped such that the components of interest decompose at or very near the same temperature and at approximately the same rate of reaction. For these materials it may be necessary to employ other techniques separately or in conjunction with TGA, such as vacuum, switching purge gases, semi-pressurized sample containment, or evolved gas analysis.

There are some general rules-of-thumb which will help you decide if the material you are working with will be a successful candidate for resolution improvement. First, do the components of the sample material decompose at sufficiently different temperatures? This can often be determined by running a slow heating rate survey scan of the material at 1°C/minute and comparing the result to a 20°C/minute scan of the same material. Generally, the temperatures of transition will be lower in the slower scan, but if there is no observable improvement in separation of the components, then it is likely that Hi-Res™ TGA will not produce a significant separation of the components either.

Always consider purge gas to be a factor when running survey scans or heating rate trials. Generally, nitrogen and air are the most common choices for purge. In some cases, transitions which appear to be one large weight loss in

nitrogen will separate in air due to reaction with oxygen in the purge stream. Switching from air to nitrogen may help to eliminate oxidation which tends to counteract a simultaneous weight loss. A few materials are reactive with nitrogen and are better run in argon. Also, consider the purity and moisture content of the purge gas. Adding or removing moisture may change the rate or nature of the reactions taking place. When the nature of the sample is relatively unknown, it is wise to repeat scans using both air and nitrogen to see if any transitions appear or disappear, or shift in temperature or weight loss.

Another test is to determine if the components of the sample material have significantly different rates of decomposition. If so, it may be possible to improve resolution using stepwise isothermal heating, or using the constant reaction rate Hi-Res™ mode (negative resolution settings). To find out, run a conventional constant heating rate ramp up to the transition temperature and then hold isothermally. After the decomposition is complete, plot the derivative of weight percent versus time and look at the shape of the curve. If both components of the transition are decomposing at about the same rate, then the curve will be a continuous exponential decay. However, if the curve appears to be a rather rapid exponential decay, followed by a very gradual and somewhat constant rate of weight change, then the components have different reaction rates and improvement in separation is probably achievable.

A third technique, which may help to improve resolution, is to control the atmospheric pressure surrounding the sample while in the furnace. This can be done by placing the sample in a semi-sealed container such as a hermetic DSC pan with a very small pin hole (0.1 mm diameter or less) or in a sample cup with a lid. When the sample starts to react, the evolved gas will build a slight vapor pressure inside the sample container. This pressure may reduce or stop one or more of the overlapped reactions and, thereby, allow completion at a higher temperature. To make this test, simply try running the sample using a Hi-Res™ ramp with and without the sample containment to see if a difference in separation can be observed. Generally, the temperature of reaction will shift to a higher temperature, even if no improvement in separation occurs. The pressure containment technique is particularly helpful when using constant reaction rate Hi-res mode (negative resolution settings).

Selecting a Hi-Res™ Technique

The Hi-Res™ TGA option consists of not one, but several techniques, which all have the common characteristic of controlling the thermal experiment based on changes in the sample weight. Both absolute weight change and the rate of weight change can be utilized. These capabilities add a whole new dimension to TGA experiments.

Some general guidelines have already been presented in the previous sections on how to set some of the parameters involved in using Hi-Res™ TGA. In the next section we consider more closely why one technique may give better results than another, and present additional guidelines for selecting techniques and setting parameters.

Dynamic Rate Hi-Res™ Ramp

In dynamic rate Hi-Res™ mode, one or more Hi-Res™ ramp segments are used with positive resolution settings. In this mode, the furnace heating rate is varied between a fixed minimum and the maximum specified in the ramp segment, but is never reduced to zero (isothermal). A mathematical function is used to relate the rate of weight change (%/minute) to the sample heating rate (°C/minute). Dependant variables to this function are resolution setting, sensitivity setting and maximum heating rate. Independent variables are both short and long-term rates of weight-change (%/minute), time and temperature. The result is the ability to directly compute

the appropriate heating rate for the current weight change conditions.

Because the dynamic rate Hi-Res™ mode reduces heating rate smoothly and only when necessary, it is the fastest and most reliable of the various techniques. This mode gives good results with most temperature separable transitions. It is preferred for fast survey scans of unknown materials over wide temperature ranges. If no other criteria exists to select a Hi-Res™ technique, then dynamic rate is the preferred choice.

Constant Reaction Rate Hi-Res™ Ramp

In constant reaction rate Hi-Res™ mode, one or more Hi-Res™ ramp segments are used with negative resolution settings. In this mode the heater control system varies the temperature of the furnace as required to maintain a constant preselected rate of weight change (%/minute). Whenever the rate of weight change exceeds the percent/minute threshold, the heating rate of the furnace is reduced, even to the point of cooling if necessary. When percent/minute falls below the threshold the heating rate is increased up to the maximum specified for the ramp segment. Transition resolution is improved because sample heating is reduced or reversed during transitions allowing them to complete at the selected reaction rate before moving on to the next transition. Figure B.10 on page B-49 shows a good example of a material (sodium bicarbonate) which was analyzed using a constant reaction rate ramp at 10°C/minute with resolution setting -4.0.

When used with semi-pressurized sample containment, constant reaction rate mode becomes even more powerful. The vapor pressure which builds up inside the sample container limits the rate of reaction of the sample. This allows the reaction to complete at a nearly constant rate and temperature. The reaction progresses more uniformly throughout the sample because vapor pressure gradients in and around the sample material are substantially reduced. The onset of higher temperature reactions is effectively suppressed until the completion of lower temperature reactions.

Constant reaction rate mode is preferred for any sample where it is important to limit or control the rate of reaction. These may include pyrotechnics, self-heating reactions, auto-catalyzing reactions and gas diffusion reactions. Constant reaction rate mode is also a good choice when it is important to accurately determine the transition temperature at a given reaction rate.

Another area where constant reaction rate heating can be helpful, is when the sample material exhibits a relatively large and somewhat constant background weight change, onto which is superimposed a relatively small transition. If the decomposition rate threshold is chosen to be close to the background percent/minute at the maximum heating rate, then the heating rate will only be changed significantly when the smaller transition occurs.

Derivative of weight change curves that are plotted versus temperature may appear cyclic and have negative peaks as well as positive ones. This effect is caused by the automatic application of cooling whenever the rate of

weight change (%/minute) exceeds the specified set point. In many cases the appearance of the derivative curve can be improved by increasing the derivative smoothing window in the data analysis program.

Constant reaction rate mode works best at lower heating rates (1 to 10°C/minute), where significant reaction rate and transition temperature overshoot can be avoided. Usually several scans of the same material are required to determine the best reaction rate threshold to use. If the sample material is very reactive, or it is important not to overshoot the selected reaction rate, then even lower maximum heating rates may be required.

This is particularly true for very small reaction rates (less than 0.1 %/minute)

Since the heater control system concentrates on a very narrow band of reaction rates, transitions with slightly different reaction rates in the same scan are often given very different treatment. For example, a transition which falls short of the %/minute threshold may be passed at a fairly high heating rate, with results similar to conventional constant heating rate TGA.

Whereas, a transition which just crosses the %/minute threshold may cause a significant reduction in heating rate or even reversal of the heating process. The two transitions, although similar in nature, may appear quite different on a plot of weight change versus sample temperature. This effect can be observed in the weight loss curve of figure B.10 (page B-49). Note that

the surface water loss at 85°C and the bicarbonate transition at 100°C are treated quite differently. This effect is most noticeable at low sensitivity settings. The effect can be reduced or eliminated by using multiple ramp segments in the method, each tailored to the needs of specific transitions. Increasing sensitivity setting may also be helpful. (See the section entitled “Adjusting Sensitivity Setting in Constant Reaction Rate Mode”).

Weight Gain Experiments

It is important to note that, while most TGA work involves decomposition analysis, some applications involve weight gain such as in oxidation studies. The Hi-Res™ heating control techniques apply equally well to weight gains as to weight losses. In this case, the absolute value of the weight change signal is used for control. Weight gains of up to 200% can be accommodated. The rates of weight gain (%/minute) and their relationship to heating rate, resolution setting and sensitivity setting are exactly the same as for weight loss. No special parameters or controls are needed for Hi-Res™ weight gain analysis. Combinations of weight gain and weight loss in the same TGA scan are handled automatically. It is important to recognize, however, that when weight gain and weight loss transitions overlap, the resultant weight change is additive and may not be separable.

Signature Analysis

For many materials it will not be possible to separate overlapped transitions sufficiently to allow quantitative analysis of weight change. However, this does not mean that no useful information can be gained from the TGA scan of the material. It is frequently the case, particularly in quality control work, that an exact determination of sample composition is not needed. Instead, the requirement is only to identify which material of a group of known standards the unknown sample most closely resembles. Another usage is to identify lot-to-lot variation from an acceptable standard. In both of these cases the location, size, and shape of the derivative of weight change peaks of the TGA scan, or the weight curve itself, is used to create a unique pattern or “signature” of the sample material. Signature scans of standards and the unknown material are then compared to make the identification or “accept/don’t accept” decision.

Since the various components of a sample, when run separately, usually decompose or evolve at unique and reproducible rates and temperatures, it is often felt that it should be possible to determine exactly what is in an unknown mixture by comparison to a library of known TGA scans of the individual components. Unfortunately, this usually does not work because the various components of a mixture typically interact with one another, so that the resultant scan is unlike either material run separately. The interactions can be chemical or physical. Some examples are evolved gases from one decomposition which slow or accelerate the decomposition or evolution of another

component (*e.g.*, CO₂ from oxidizing carbon). Molecular attractions prevent a more volatile component from evolving when expected, and at the same time hasten the evolution of another less volatile component (*e.g.*, chain linkage in polymer blends). The physical matrix of the mixture may retard the evolution of a more volatile component, which then evolves at a higher than normal temperature and at a slower than normal rate (*e.g.*, oil evaporating from rubber). Or the components may react chemically at elevated temperatures and produce new compounds which decompose at different temperatures than the individual components.

Sample Quantity and Orientation

As with conventional constant heating rate TGA, sample quantity and orientation in the sample holder can be important during Hi-Res experiments. This is particularly true if the sample is not homogeneous (*e.g.*, a laminated sheet or coated surface). With these types of samples it is wise to try scans with different surfaces exposed. Be aware that chopping or grinding the material may produce an entirely different result since physical or chemical properties may be altered.

Exposed Surface Area

Exposed surface area is often important. When samples melt they usually spread out over the bottom surface of the sample container. This will expose more or less surface area, depending on whether the original configuration was a single block (area typically increases) or a

powder (area typically decreases). Usually with open sample pans it is best to try to maximize exposed surface area at all times so that evolved gases escape quickly and reactions proceed uniformly. This suggests the use of small powdered or thin samples which are uniformly distributed in the sample vessel. When semi-pressurized sample vessels are used the issue of exposed surface area is far less important.

Generally, sample sizes in the range of 5 to 15 milligrams are recommended. If the material is self heating or auto-catalytic, then smaller sample quantities may help with heating control. (This is particularly important for constant reaction rate Hi-Res™ mode.) On the other hand, larger sample quantities (50 to 100 mg) are recommended for reactions in which a very small weight change (less than 1 percent) is being measured. For maximum weight resolution it is advisable to keep sample weight below the TGA 2950 weight range change at 100 mg.

A problem with very large samples, which decompose rapidly and almost completely, is that the furnace purge may not be able to remove all of the evolved components and some contamination of the furnace wall and cooling jacket may result. This may affect the remainder of the experiment or future scans.

Bubble Formation

Whenever medium to large sample quantities are being run, use caution in thermocouple placement. Some materials, particularly polymers, will form a “skin” on the outer surface of the sample as it is heated which inhibits mass transport of the more volatile components.

These samples form bubbles which can rise up and touch the end of the thermocouple, ruining the experiment and possibly contaminating the thermocouple. Just as the largest chewing gum bubbles are produced by blowing slow prolonged breaths, the largest gas bubbles form in samples which are gradually heated. For this reason bubble formation is more of a concern for Hi-Res™ than conventional TGA.

Another effect of bubble formation is a sudden small unexpected change in weight and accompanying spike in the percent/minute curve as the bubble bursts. An excellent example of bubble noise can be seen in the scan of ethylene-vinyl acetate in figures B.7 and B.8. In figure B.7 the effect of bubble formation and bursting can be seen as a sudden drop in weight at about 400°C during the second weight loss. This appears in the derivative curve as a peak shoulder. In Figure B.8 we can clearly see the formation and bursting of several large bubbles in the percent/minute curve between 65 and 85 minutes into the run. Again, these effects tend to be more noticeable in Hi-Res™ TGA because larger bubbles form due to the slow heating process. The best solution to bubble noise is to decrease resolution setting and/or increase maximum heating rate. Reducing sample size may also help.




Figure B.7




Figure B.8

Thermocouple Placement

Generally, the thermocouple should be placed 2 to 5 mm above the bottom of the sample container (1 to 4 mm above the edge). Placement 1 mm above the edge is recommended for most applications. Longer distances will not degrade heater control and are sometimes helpful in reducing any adverse effects of sample self-heating or bubble formation.

Data Analysis Effects

Hi-Res™ TGA, due to its technique of changing heating rates dynamically, causes very significant non-linearity of data points across transition regions when plotted versus temperature. This is not a problem when making plots, but when automatic limit selection is used during data analysis report generation the analysis program can sometimes become confused about where to place the limits. This situation can be identified by curve tick marks and tangent lines which are poorly placed on the weight curve or may even be plotted off of the curve.

The difficulty is caused by the fixed size of the transition onset and endset windows (the first and last 12.5% of the analysis region). The data analysis program assumes that data points are uniformly distributed throughout this region. If the number of data points in the baseline portion of a window is very small compared to the number of points in the sloped portion of the window (the usual case with Hi-Res™ scans) then the tangent lines will be excessively

weighted toward this cluster of points and may be nearly vertical or even fall off the curve entirely.

Improved results can be obtained by moving the transition start and stop limits farther away from the transition so that the onset and endset windows remain completely on the baseline portion of the curve. If this fails to help, the transition tick marks can be placed manually on the curve.

Derivative Plots

Another data analysis problem caused by unequally spaced temperature data points is unusually shaped or flattened derivative plots versus temperature. This is due to the assumption that data points are equally distributed over a moving window (the smoothing window) used to compute the derivative. If the derivative smoothing window is too wide, derivative peaks will be flattened and apparent resolution will be reduced. If the window is too narrow, derivative peaks will be needle sharp and noisy.

The default smoothing window is 0.2 minutes for derivatives with respect to time (%/min and %/min/min) and 10°C for derivatives with respect to temperature (°C/min and °C/min/min). For typical Hi-Res™ TGA scans, the quality of derivative plots with respect to temperature can be improved by decreasing the temperature window to 5°C. If extremely sharp results are desired, try a smaller window. Less than 1°C is not recommended. The smoothing window for time does not normally need adjustment. If the time derivative seems particularly noisy, try increasing the window to 1 minute.

In conventional constant heating rate TGA, the plot of percent/minute and the plot of percent/°C are essentially identical because there is a direct linear relationship between time and temperature. Percent/minute is conventionally chosen for transition peak analysis and plotting. With Hi-Res™ TGA there is no linear relationship between time and temperature because the heating rate is constantly changing. When plotting the derivative of weight change versus temperature for a Hi-Res™ TGA scan, you should use percent/°C. A plot of percent/minute versus temperature can still be useful for locating minor transitions and determining the rate of reactions.

Adjusting Heating Rate

Heating rate has long been used to control transition resolution. The typical thermal analyst makes most runs at 20°C/minute and drops down to 5 or 1°C/minute for hard-to-resolve transitions, or just to see if there is anything else interesting going on that might have been missed in the faster scan. 1°C/minute scans are generally not routine due to the enormous amount of time required to run them.

With Hi-Res™ TGA the heating rate is varied automatically to give the benefit of slow heating rates during transitions and fast rates during baseline. However, it is still necessary to specify a maximum heating rate. This is because in Hi-Res™ TGA, just as in constant heating rate TGA, the maximum heating rate affects the results of the experiment and, as such, is an important adjustment to consider.

Heating rate adjustment is particularly critical if the transitions of an overlapped group are very closely spaced in temperature. At 50°C/minute the normal temperature lags in the TGA furnace can be enough to overshoot the first transition, which may reduce separation from the other transitions in the group. This will be most noticeable when the percent/minute baseline preceding a rapid transition is relatively constant and several orders of magnitude below the peak rate of the transition. Running at 20°C/minute instead of 50°C/minute is recommended for most routine work using positive resolution settings particularly if only one scan can be made of each unknown material. Less than 10°C/minute is usually not required.

For negative resolution settings the selection of heating rates is usually quite different because reaction rate overshoot must be minimized. Generally, 1, 2, 5 and 10°C/minute are the most useful rates to use with negative resolution settings. Try 5°C/minute to start.

An important difference between the dynamic rate (positive resolution setting) and constant reaction rate (negative resolution setting) Hi-Res™ mode, is that the maximum heating rate selected is only an upper limit in negative mode, but it is a gain factor in positive mode. In other words, in negative mode the maximum heating rate is irrelevant to heating control, except when the reaction rate drops to such a low level that heating rates higher than this maximum limit are required to maintain the selected percent/minute.

In negative mode, as long as the maximum heating rate is high enough to allow completion of the transition at the selected percent/minute

set point, and not so high as to cause percent/minute overshoot, transitions will appear largely the same from one heating rate to the next. A word of caution, however: *always keep in mind that the background (baseline) rate of weight change will be accelerated at higher maximum heating rates.* This will move the background percent/minute closer to the control point you have selected with the resolution setting. If noise is present in the background percent/minute, it may fool the control algorithm into thinking a transition is starting and result in a premature reduction in heating rate to avoid predicted set point overshoot. In extreme cases, heating rate cycling can occur, particularly at high sensitivity settings. (See Figure B.13 on page B-61) To cure this problem reduce maximum heating rate and/or sensitivity setting.

Adjusting Resolution Setting

The purpose of the resolution parameter is to select the range of percent/minute values over which the heater control system will vary heating rate in response to changes in the rate of weight change. In dynamic rate mode (positive resolution settings) the range of percent/minute values selected by each resolution setting is fairly wide (about two orders of magnitude). This width is adjustable using the sensitivity parameter.

In constant reaction rate mode (negative resolution settings) the resolution setting specifies the percent/minute value that will be used as the control set point for furnace heating. In this case the system will adjust heating rate as required to maintain this constant rate of weight change. Table B.1 (on the next page) shows the negative resolution settings and their associated percent/minute values.

Table B.1

Res	%/min	Res	%/min	Res	%min	Res	%/mi
-0.1	28.2	-2.1	2.82	-4.1	0.282	-6.1	0.0282
-0.2	25.1	-2.2	2.51	-4.2	0.251	-6.2	0.0251
-0.3	22.4	-2.3	2.24	-4.3	0.224	-6.3	0.0224
-0.4	20.0	-2.4	2.00	-4.4	0.200	-6.4	0.0200
-0.5	17.8	-2.5	1.78	-4.5	0.178	-6.5	0.0178
-0.6	15.8	-2.6	1.58	-4.6	0.158	-6.6	0.0158
-0.7	14.1	-2.7	1.41	-4.7	0.141	-6.7	0.0141
-0.8	12.6	-2.8	1.26	-4.8	0.126	-6.8	0.0126
-0.9	11.2	-2.9	1.12	-4.9	0.112	-6.9	0.0112
-1.0	10.0	-3.0	1.00	-5.0	0.100	-7.0	0.0100
-1.1	8.91	-3.1	0.891	-5.1	0.089	-7.1	0.0089
-1.2	7.94	-3.2	0.794	-5.2	0.079	-7.2	0.0079
-1.3	7.08	-3.3	0.708	-5.3	0.071	-7.3	0.0071
-1.4	6.31	-3.4	0.631	-5.4	0.063	-7.4	0.0063
-1.5	5.62	-3.5	0.562	-5.5	0.056	-7.5	0.0056
-1.6	5.01	-3.6	0.501	-5.6	0.050	-7.6	0.0050
-1.7	4.47	-3.7	0.447	-5.7	0.045	-7.7	0.0045
-1.8	3.98	-3.8	0.398	-5.8	0.040	-7.8	0.0040
-1.9	3.55	-3.9	0.355	-5.9	0.036	-7.9	0.0036
-2.0	3.16	-4.0	0.316	-6.0	0.032	-8.0	0.0032

The process of initially picking, and then adjusting resolution setting, is not exacting or calculated. It is based largely on experience and some general guidelines. This is because no single resolution setting will give dramatically different results from all the others. The change from one number to another is rather gradual. Another reason you might wish to experiment with more than one setting is that some materials react differently from others to increasing or decreasing the resolution setting. This is due to the time as well as the temperature-dependant nature of transitions, and to the interaction

between the rate of weight change (%/minute) and heating rate (°C/minute). As percent/minute increases heating rate is reduced by the control algorithm, but the reduction in heating rate usually causes an accompanying reduction in percent/minute, and vice versa. Therefore, attempting to directly compute optimal resolution settings from percent/minute information gathered from previous runs becomes a very questionable and usually frustrating experience.

Let us then consider what guidelines and rules-of-thumb we can use to help make the selection process easier. As stated earlier in the section on Hi-Res™ ramps, if you do not know what resolution setting to start with, try resolution +3.0 and 50°C/minute heating rate. This will give a rapid scan with moderate application of the Hi-Res™ heating technique. Results should be at least as good as a 20°C/minute conventional scan of the same material, and will usually be better. If time permits, it is often helpful to have a constant heating rate 20°C/minute scan of the material available for comparison.

Useful Resolution Settings

After some experience with Hi-Res™ TGA, you will find that the most useful resolution settings fall within the range +3.0 to +5.0 for the positive numbers and -3.0 to -5.0 for the negative numbers, and that adjustment by +/-0.5 is usually adequate. This is similar to the situation with heating rates. You can adjust heating rate to any value from 0.01 to 200.0°C/minute in steps of 0.01°C but most people use 1, 5, 10, 20 and 50°C/minute exclusively because a finer adjustment does not produce significantly different results. With this guideline alone we have reduced the number of resolution settings to deal

with from 80 to only 5 for each Hi-Res™ mode while covering the majority of materials of interest.

The next step is how to proceed after the first Hi-Res™ run is complete. If the first run was made at resolution setting 3.0, try the second one at 4.0. Generally, increasing resolution setting by a whole number will increase the time to complete the TGA scan by a factor of 2 to 5 times. Therefore, you must consider whether you can afford the added run time as well as the extra setup time. This balance between increased run time and adequate resolution is usually the determining factor on what resolution setting to live with.

The benefit to having more range and “resolution” to the resolution setting than seems to be necessary, is that on a rare occasion a material requires a very fine adjustment or an extreme treatment. Don’t forget that the maximum heating rate is a factor in determining resolution as well as the resolution setting.

Lower resolution settings allow materials which already have well-separated transitions to be analyzed at super high heating rates such as 200°C/minute with excellent resolution in a fraction of the time required at a constant rate of 20°C/minute. Resolution settings above 5.0 are useful when exact decomposition temperatures are needed, or when the decompositions are explosive in nature, or when overlapped transitions are extremely close but highly temperature selective.

The effect of adjusting the resolution setting in dynamic rate mode while holding other experimental factors constant can be seen in Figure B.13 (page B-61).

Temperature Calibration

TGA temperature calibration is useful if accurate transition temperatures are required. The major causes of temperature inaccuracy in a TGA are thermal gradients between the sample thermocouple and the sample being studied. The magnitude of these gradients is proportional to heating rate. The Hi-Res™ TGA techniques inherently reduce thermal gradients by slowing down the heating rate during transitions.

Another way to reduce the effect of thermal gradients is to temperature calibrate the TGA. The general procedure for temperature calibration is found in the online help and documentation. Temperature calibration involves analyzing a magnetic standard to determine its curie temperature. The curie temperature corresponds to the extrapolated endpoint on the “S” shaped thermal curve.

However, when the calibration is intended for use with Hi-Res™ TGA experiments (*i.e.*, dynamic rate, constant reaction rate or stepwise isothermal), then a slow heating rate conventional ramp of 5°C/minute or less should be used for calibration. A faster ramp rate is only used when calibrating for constant heating rate experiments. The reason for this is that the Hi-Res™ heating control system reduces the heating rate during transitions.

Hi-Res™ Transition Temperatures

The TGA provides precise weight measurements coupled with relative temperature information. The resolution setting of a Hi-Res™ ramp controls the reaction rate of sample transitions. It is reaction rate, more than anything else, which will determine the apparent transition temperature of a decomposition reaction.

The shift in measured transition temperature caused by changing resolution setting can easily be an order of magnitude larger than the thermal gradients you are trying to correct with calibration. This effect can be clearly observed in the mixture of bicarbonates example (Figure B.13, page B-61). In light of this fact, it is acceptable in many cases to simply not use temperature calibration when employing the Hi-Res™ TGA techniques for decomposition analysis.

Hi-Res™ Sensitivity Segment

The Hi-Res™ sensitivity segment sets an additional parameter associated with Hi-Res™ ramp segments which can be used to adjust the response of the Hi-Res™ temperature control algorithm. This is sometimes necessary because of the wide variation in decomposition mechanisms of typical sample materials. This segment has the following format:

Hi-Res™ sensitivity <sens_setting>

where:

<sens_setting> is the Hi-Res™ sensitivity setting (1.0 to 8.0)

example:

Hi-Res™ sensitivity 2.0

Hi-Res™ sensitivity segments execute immediately when encountered in a method and simply set the sensitivity setting to the new value provided. The last value set is used for all subsequent Hi-Res™ ramps until a new value is set. If no Hi-Res™ sensitivity segment has been encountered in the method before the execution of a Hi-Res™ ramp segment, then the default sensitivity (1.0) is used.

The sensitivity setting is a unitless number, ranging from 1.0 (lowest sensitivity) to 8.0 (highest sensitivity). The setting is used by both the dynamic rate (positive resolution setting)

and the constant reaction rate (negative resolution setting) modes of the Hi-Res™ ramp segment. There is no limit to how many times the setting can be changed during a method. Increasing sensitivity setting tends to increase experiment time.

Understanding Sensitivity Setting

The TGA 2950 Hi-Res™ control algorithms have been pretuned to respond correctly to most transition situations with the default sensitivity setting of 1.0. This means that in most cases it will not be necessary to adjust sensitivity setting at all. The key is knowing when and how to make the adjustment.

It is easy to confuse resolution setting and sensitivity setting since both values can affect the resolution of the TGA scan. However, there is a simple way to think of the difference between the two parameters.

Resolution setting controls the temperature at which the transition will occur (*i.e.*, how far from the theoretical isothermal decomposition temperature) by selecting the reaction rate (%/minute) at which heating rate is reduced. The closer the reaction is to the isothermal decomposition temperature the lower the reaction rate and the longer the reaction will take. You will find that you can use the resolution setting to literally move the measurement of transitions on the temperature axis. (See Figure B.13, page B-61.)

Sensitivity setting controls the response of the Hi-Res™ system to changes in the rate of reaction (%/minute). Larger sensitivity settings cause the system to be more reactive or “sensitive” to small changes in the rate of the reaction. Lower sensitivity settings dampen this response. Generally, it is best to adjust resolution setting first with sensitivity set to a low value, and then after a good result is obtained, try increasing sensitivity to see if any resolution improvement can be obtained.

Use caution when adjusting sensitivity since over adjustment can cause oscillation or anomalies in the weight versus temperature curve.

Adjusting Sensitivity in Dynamic Rate Mode

In dynamic rate Hi-Res™ mode (positive resolution settings), the sensitivity setting is used to further increase the resolution of some transitions once an appropriate resolution setting has been determined. This is accomplished by narrowing the range of percent/minute values over which the heating rate is proportionally varied. Higher sensitivity settings result in progressively narrower percent/minute ranges and generally increased resolution.

In this mode the resolution setting selects the general neighborhood of percent/minute values that will drive changes in furnace heating rate. For example, resolution setting 3.0 selects the neighborhood of approximately 1.0 to 20.0%/minute for most of the variation in heating rate. Whereas, setting 4.5 selects 0.1 to 2.0%/minute. Sensitivity setting controls the relative width of

this range. Setting 1.0 allows use of the full range. Setting 2.0 reduces the range to about one half the full range. Setting 3.0 to about a third and so on up to 8.0. In general, higher sensitivity settings bring the furnace to the transition temperature more quickly but then tend to remain at that temperature longer. In other words, higher sensitivity settings bring the control closer to stepwise isothermal heating. (See “Stepwise Isothermal Heating” in the section on using Abort Segments for more information.)

The recommended procedure for adjusting sensitivity setting for dynamic rate mode is to start with a sensitivity setting of 1.0 and adjust resolution setting to get the best separation possible in the desired time frame. Then increase sensitivity to 2.0, 4.0 and 8.0 to see if a useful improvement in resolution results.

It is possible that no improvement in resolution will result. This is usually caused by overlapped transitions which are weakly temperature-dependant and strongly time-dependant. In this case, it doesn't matter how precisely we control at a specific temperature, the individual components of the sample material are all going to decompose more or less together in the neighborhood of the decomposition temperatures we have selected via the resolution setting. About all we can do is lengthen or shorten the total time of decomposition by selecting larger or smaller resolution settings.

The effect of changing sensitivity settings in dynamic rate mode can be seen in Figure B.14 on page B-62.

Adjusting Sensitivity in Constant Reaction Rate Mode

In constant reaction rate Hi-Res™ mode (negative resolution settings) the sensitivity setting is used to adjust the heater control system to minimize transition temperature overshoot and heating control fluctuation. Higher sensitivity settings result in decreased percent/minute overshoot and tighter control at the beginning of transitions. Lower settings have the opposite effect.

In this mode sensitivity setting is used to adjust the response of the heater control system to changes in the rate of weight change (%/minute). For materials that react gradually, low sensitivity settings are generally preferred because they help dampen noise and greatly reduce the possibility of control cycling. However, when it is very important to avoid percent/minute overshoot, or if the sample is highly reactive then higher settings will be required. The problem with too high a sensitivity setting is that control cycling or heating rate “ringing” may occur (see Figure B.11).

The recommended procedure to adjust sensitivity setting for constant reaction rate mode is to start with a sensitivity setting of 1.0 and observe the transition for percent/minute overshoot and control cycling. If overshoot is acceptable, no further adjustment is needed. If overshoot is excessive, increase sensitivity setting by 1.0 and recheck overshoot and cycling. Continue increasing sensitivity until the results are acceptable or until control cycling becomes excessive.

If no satisfactory sensitivity setting can be found, the problem may be too low a resolution setting or too high a heating rate. Try a higher resolution setting (larger negative number) to reduce the percent/minute set point and rerun the experiment. Set points in the range of 0.1 to 1.0%/minute (resolution settings -5.0 to -3.0) generally give the best results.

If the percent/minute overshoot is primarily associated with the first transition of a group of overlapped transitions, the problem may be too high a maximum heating rate. Try reducing the heating rate of the Hi-Res™ ramp segment by one half and rerunning the experiment. Heating rates in the range of 1.0 to 5.0°C/minute generally give the best results.

When heating rates higher than 5.0°C/minute are used, proper adjustment of sensitivity setting becomes critical to maintaining smooth heating control. The default sensitivity of 1.0 is usually too low at these higher heating rates and typically results in significant transition temperature overshoot and heating control “ringing” as shown in Figure B.9. At high heating rates a setting of 3.0 or 4.0 will give better results for most materials. The improvement can be seen in Figure B.10. If sensitivity setting is adjusted too high, then continuous control cycling may result as shown in Figure B.11. With some experimentation, an optimal setting can usually be found. When very high heating rates are used (greater than 10°C/minute) it may be impossible to completely eliminate control ringing. However this should not affect the quantitative measurement of weight loss for the transition.




Figure B.9
Sensitivity Setting Too Low




Figure B.10
Correct Sensitivity Adjustment




Figure B.11
Sensitivity Setting Too High

Abort Segment

The abort segment provides a mechanism to skip over or terminate other method segments when specific weight change conditions are met. This segment has the following format:

*Abort next seg if <signal> <condition>
<value>*

where:

<signal> is the real-time measurement used to decide whether to abort or not (weight % or weight %/minute)

<condition> is the limit condition for aborting (“<” or “>”)

<value> is the abort limit which is compared with the real-time signal

example:

Abort next seg if % < 20.0

Abort segments execute immediately when encountered in a method and simply establish conditions for testing the next segment. The specified limit, designated by the <signal>, <condition> and <value> parameters, is tested before and during the execution of the next method segment. If the limit is reached at the beginning of a segment, that segment is skipped and method execution continues with the next segment. If the segment following the abort segment is equilibrate, initial temperature, ramp, Hi-Res™ ramp, isothermal or step, and the limit has not been reached yet, the limit will be tested

at the rate of 2 times per second until the segment terminates normally or the limit is reached. If the limit is reached during the execution of a segment, the remaining portion of the segment is skipped. Method execution then continues with the next segment in the method.

The conditions for determining whether limits have been reached for each type of limit are as follows:

% -If the condition operator is “<” and the sample weight percent is less than or equal to the limit percent, or if the condition operator is “>” and the weight percent is greater than or equal to the limit percent, then the limit is reached. Values greater than 100% are permitted to accommodate weight gains.

%/min -If the condition operator is “<” and the derivative of sample weight percent is less than or equal to the limit percent/minute, or if the condition operator is “>” and the derivative of weight percent is greater than or equal to the limit percent/minute, then the limit is reached. Negative values indicate weight gain and positive values indicate weight loss.

Abort segments are very versatile because they can be used in front of any method segment (including other abort segments) to dynamically change the execution of a method. For example, an abort segment can be used to control the switching of a purge gas or activation of an event relay based on the rate of reaction or the amount of weight loss. An abort segment in front of a repeat segment provides a mechanism to terminate a loop prematurely. Abort segments

can be used to customize a method for a particular material by allowing different ramp and/or isothermal segments to be used for each transition region without regard to specific temperature limits. Abort segments can be used to control data storage so that file size is minimized by turning off storage or increasing sampling interval during baseline sections of a scan.

Abort segments provide a convenient mechanism for shortening experiments after the data of interest has been collected. For example, if a heating ramp is used to analyze a material which has two weight losses and the only data of interest is the percent weight loss during the first transition, then an abort segment can be used to terminate the ramp after the beginning of the second weight loss by specifying a “%” limit condition. This is particularly useful for Hi-Res™ ramps because the reduced heating rate during a weight loss causes the majority of the transition to occur during a very narrow temperature range, making termination by final temperature difficult to predict.

Stepwise Isothermal Heating

Another useful TGA technique which can be implemented with abort segments is transition-controlled *stepwise isothermal heating*. This process consists of heating a sample via a ramp segment until a certain rate of weight change is detected, and then switching to an isothermal segment to hold constant temperature until the transition has completed. Then the sample heating is continued until the next transition is detected, whereupon, isothermal holding is

again initiated, and so on until the final temperature is reached.

The stepwise heating technique can be easily implemented by placing a ramp segment followed by an isothermal segment into a “repeat to final temperature” loop, and preceding the ramp and isothermal segments with abort segments. The abort segment preceding the ramp is setup to terminate the ramp when the “%/minute” is greater than a specified limit. The abort segment preceding the isothermal segment terminates the holding period when the “%/minute” is less than a second limit. An example of this type of method is shown below.

- 1: Abort next seg if %/min > 0.5*
- 2: Ramp 10 °C/min to 700 °C*
- 3: Abort next seg if %/min < 0.05*
- 4: Isothermal for 500 minutes*
- 5: Repeat segment 1 til 700 °C*

The “%/min” limit for the isothermal segment is selected to be equal to the baseline rate of weight change encountered during the onset of the transition of interest via a normal constant heating rate scan of this material. (Be sure to use the same ramp rate as that selected for stepwise heating.) The “%/min” limit for the ramp segment is selected to be about an order of magnitude greater than that selected for the isothermal segment (but not more than the maximum rate of weight change encountered during the transition of interest). The ramp final temperature and the repeat final temperature are set to the final experiment temperature. The isothermal time is set to an arbitrary time, which is sufficiently large such that the segment will not terminate until the percent/minute limit has been reached.

Stepwise heating often improves transition resolution because transitions are time dependent as well as temperature dependent. Stepwise heating gives transitions more time to complete, thereby, reducing overlap with neighboring transitions.

To get the maximum benefit from stepwise heating it will be necessary to run several TGA scans to properly “tune” the reaction rate thresholds used to start and stop heating. Relatively slow heating rates are generally required to prevent transition overshoot. A rough rule-of-thumb is to use a heating rate which is about one tenth the transition temperature difference between the transitions being resolved. For example, if the transitions are separated by 10°C then use 1°C/min as the heating rate prior to, between and following the transitions. If precise reaction temperatures are important, then slow heating rates should always be used prior to encountering transitions of interest even though they may be well separated in temperature, and the “%/min” limit for the ramp segment should be set closer to the limit for the isothermal segment. To avoid excessively long experiments, higher heating rate ramps or equilibrate segments can be used to skip over baseline portions of the scan.

A disadvantage to stepwise heating is that most experiments take much longer in total time to complete than by a conventional constant heating rate scan. Another disadvantage of stepwise heating is that the decision to leave isothermal mode and continue heating is somewhat arbitrary and may lead to incorrect assumptions about the number and size of transitions.

This is particularly true for materials which have transitions that are overlapped even at very slow heating rates (such as the sodium/potassium bicarbonate mixture in the Examples section). As a general rule, stepwise isothermal heating cannot be used to reliably separate transitions which cannot be separated by conventional TGA at very slow heating rates.

Another problem to watch out for with stepwise heating is the creation of anomalies in the weight loss versus temperature curve. These appear as small unexpected secondary weight losses following a larger transition. These anomalies can be caused by two factors.

The first cause is using too high a heating rate in the ramp which follows the transition. Any small amount of sample material which did not finish decomposing during the transition will now quickly decompose due to the rapid elevation of the furnace temperature. This will cause the decomposition rate (%/minute) to rise substantially. If the rising decomposition rate crosses the abort threshold for the ramp segment then a second isothermal period will be introduced which will appear on the weight loss versus temperature plot as a small unexplained transition. Generally it is best to use the same heating rate on both sides of a transition. This rate can then be accelerated if desired after the furnace temperature is some distance from the transition.

The second cause of anomalies is leaving the isothermal holding period prematurely during a transition because the percent/minute threshold for aborting the isothermal segment was set too high. This situation leaves a significant amount

of undecomposed sample material which now accelerates its rate of decomposition. As in the case of too high a ramp rate, the increasing temperature causes the remaining sample to quickly decompose which raises the rate of weight loss (%/minute) to a high level and either triggers another isothermal hold period or shows up as a backside shoulder on the weight loss curve.

An example of using stepwise isothermal heating is shown in the following “Hi-Res™ TGA Examples” section.

Hi-Res™ TGA Examples

Included in this section of the manual are example TGA scans of common materials. These examples can be used to compare the results of using Hi-Res™ TGA heating control with the results of conventional constant heating rate TGA. Where possible the effects of using the different Hi-Res™ modes and parameters are shown.

Mixture of Bicarbonates

A mixture of potassium bicarbonate (potassium hydrogencarbonate) and sodium bicarbonate (sodium hydrogencarbonate) was chosen to demonstrate the effects of different Hi-Res™ techniques and parameter settings. The individual bicarbonates decompose to carbonates between 100°C and 200°C with the simultaneous release of CO₂ and H₂O. Potassium bicarbonate decomposes at approximately 50°C higher temperature than sodium bicarbonate. When mixed together the decompositions of the two bicarbonates are overlapped in temperature and are very difficult to resolve. You can easily make this sample yourself by thoroughly mixing equal parts (by weight) of finely powdered potassium bicarbonate (KHCO₃) and sodium bicarbonate (NaHCO₃). Inadequate mixing or large granule size will reduce weight loss reproducibility. Note that potassium bicarbonate is very hygroscopic. The mixture must not be exposed to ambient humidity for long or a

significant surface water transition will become evident between 70°C and 100°C which will affect the overall weight loss percentages. When mixing and using this sample be sure to keep the sample supply containers tightly capped. Load the TGA quickly and use a dry purge gas (air, nitrogen or argon). Dry air purge at 100 mL/min was used for all example scans shown. Sample sizes varied from 20 to 40 mg. Small variations observable in weight loss reproducibility in the example scans are largely due to the non-homogeneity of the sample mixture and variations in ambient humidity from run to run.

Dynamic Rate Scans

In Figure B.12 (on the next page) we have overlaid the individual bicarbonate decompositions (curves b and d) along with the decomposition of the mixture by conventional TGA (curves a and e) and by dynamic rate Hi-res™ TGA (curve c). Reducing the heating rate of the mixture from 20°C/minute (curve e) to 1°C/minute (curve a) gave a very slight improvement in resolution. Comparing these results to the 50°C/minute Hi-res™ scan we observe a significant improvement in resolution with Hi-res™ TGA in about twice the time of the 20°C/minute scan, and about one tenth the time of the 1°C/minute scan. The method used for all scans was a single ramp segment.




Figure B.12

Varying Resolution Setting

In Figure B.13 we have overlaid the conventional constant heating rate decompositions of the bicarbonate mixture (curves a and b) with dynamic rate Hi-Res™ scans at eight different resolution settings (curves 1 through 8). All of the Hi-Res™ scans were run at 50°C/minute with the default sensitivity setting of 1.0. Note that increasing resolution setting increases the resolution of each transition and simultaneously reduces the transition temperature. The initial weight losses of approximately 1% on each curve are due to the evaporation of surface water absorbed by the mixture.




Figure B.13

Varying Sensitivity Setting

In Figure B.14 we have overlaid the 1°C/minute conventional constant heating rate decomposition of the mixture (curve a) with a dynamic rate Hi-Res™ scan at four different sensitivity settings (curves b through e). All of the Hi-Res™ scans were run at 20°C/minute with a resolution setting of 5.0. Note that increasing sensitivity setting increases the sharpness of each transition, but does not substantially change the transition temperature.




Figure B.14

Constant Reaction Rate Scans

In Figure B.15 we have overlaid the conventional constant heating rate decompositions of the bicarbonate mixture (curves c and e) with constant reaction rate Hi-Res™ scans at different resolution and sensitivity settings. Curve b shows the Hi-Res™ scan run in an open sample pan at resolution setting -4.0 and sensitivity setting 1.0. For curve d (resolution -4.0) and curve a (resolution -5.0) the sample was contained in a hermetic aluminum DSC sample pan with a 0.1mm pin hole in the top. All of the Hi-Res™ scans were run at 5°C/min.

Comparing curve b to curves a and d, we can see a significant improvement in resolution due to the vapor pressure build up in the semi-hermetic sample container. Because the sample pan was open in curve b, there was no opportunity for a vapor pressure/reaction rate equilibrium to occur as the sample decomposed resulting in only partial separation of the transitions. The pressure buildup in the closed container (Curves a & d) retarded the potassium bicarbonate decomposition until the sodium bicarbonate decomposition had completed. As with dynamic rate mode (Figure B.13), we see that higher resolution setting (larger negative number) reduces transition temperature.




Figure B.15

Stepwise Isothermal Scans

In Figures B.16 and B.17 we see the result of stepwise isothermal scans of the bicarbonate mixture using a conventional 1°C/minute ramp and the abort segment. The following method was used for the scan in Figure B.16:

- 1: Abort next seg if %/min > 0.15
- 2: Ramp 1 °C/min to 300 °C
- 3: Abort next seg if %/min < 0.015
- 4: Isothermal for 500 minutes
- 5: Repeat segment 1 til 300 °C

Although this method gives apparently excellent separation results, the quantitative value of the weight loss plateau between the two bicarbonate decompositions is in question because there is no inflection point in the plateau and the rate of weight loss immediately increases after the isothermal segment is aborted. This indicates that the two decompositions are still overlapped and that holding for a longer isothermal time period during the first transition would have resulted in a lower weight loss plateau between transitions.

In Figure B.17 the same stepwise isothermal method is repeated with the %/minute limits for the abort segments set to smaller values (0.05%/minute for the ramp abort and 0.005%/minute for the isothermal abort). The following method was used for the scan in Figure B.17:

- 1: Abort next seg if %/min > 0.05
- 2: Ramp 1 °C/min to 300 °C
- 3: Abort next seg if %/min < 0.005
- 4: Isothermal for 500 minutes
- 5: Repeat segment 1 til 300 °C

The scan in Figure B.17 is a definite improvement over the result in Figure B.16. This is because the abort limit for the isothermal segment (0.005%/minute) was chosen to be equal to the baseline %/minute immediately preceding the sodium bicarbonate transition observed in a conventional constant heating rate scan of the mixture at 1°C/minute. The theory supporting this decision is that if the two transitions are separable then the rate of weight loss should return to baseline between the transitions. The %/minute limit for the ramp segment was then chosen to be ten times larger than that for the isothermal segment.

Although the weight loss result in Figure B.17 seems more reasonable, we are suspicious that the decomposition of the potassium bicarbonate (2nd transition) has already started because the rate of weight loss immediately increases as heating is resumed at 88°C. Another problem is the torturous 1300 minute time frame of the experiment.

As can be seen by these results, caution must always be employed when interpreting results from stepwise isothermal heating experiments. It is usually wise to run confirming experiments by other TGA techniques, particularly if the sample material is of relatively unknown composition.




Figure B.16




Figure B.17

Monosodium Glutamate

In Figure B.18 we analyzed Accent® brand monosodium glutamate (MSG), a common salt used for seasoning foods, which has three well resolved transitions below 500°C. Curves a and c show the result of conventional constant heating rate scans of MSG at 1 and 20°C/minute. Curve b shows the result of a dynamic rate Hi-Res™ scan at resolution setting 4.0 and sensitivity setting 1.0. As can be seen by the derivative of weight loss curves, the Hi-Res™ scan gives resolution comparable to the 1°C/minute scan in a fraction of the time.




Figure B.18

If you should decide to run this sample, be aware that MSG foams significantly at these temperatures. If a sample size greater than

about 10 mg is used, the sample material may rise up in the pan and touch the sample temperature thermocouple. If MSG is heated to temperatures well above 500°C, it will leave a residue which is difficult to remove from the sample pan.

Banana Taffy

In Figures B.19 and B.20 we analyzed a sample of artificial banana taffy, a common confectionery product composed primarily of water and sugar, which has a number of overlapped transitions between 100°C and 500°C. Figure B.19 shows the result of a conventional constant heating rate scan of taffy at 10°C/minute. Figure B.20 shows the result of a 50°C/minute dynamic rate Hi-Res™ scan at resolution setting 4.0 and sensitivity setting 1.0.




Figure B.19




Figure B.20

As can be seen by comparing the derivative of weight loss curves in Figures B.19 and B.20, the Hi-Res™ scan gives improved resolution in the same time as the conventional scan. This is possible because the Hi-Res™ TGA ramp heats the taffy sample rapidly during baseline portions of the scan and slowly during transitions, resulting in an average heating rate of about 10°C/minute. A 200°C/minute Hi-Res™ scan of this sample takes about the same time as a 20°C/minute conventional scan while providing resolution improvement similar to that observed with the 50°C/minute Hi-Res™ scan.

Plastic Laboratory Tubing

In Figures B.21 through B.25 we analyzed a sample of Tygon® R-3603 plastic tubing, a polyvinyl chloride (PVC) based clear flexible tubing commonly used in laboratories and industry. A number of overlapped transitions are evident between 100°C and 350°C, followed by two well resolved transitions at approximately 400°C and 500°C. Of interest are the changes in resolution of the various transitions as maximum heating rate changes in both the conventional and Hi-Res™ scans.

Observing the two conventional scans (Figures B.21 and B.22) we see that the resolution of the initial overlapped transitions is good in the 1°C/minute scan, but poor in the 20°C/minute scan. In contrast, the resolution of the two transitions at 400°C and 500°C are best in the 20°C/minute scan and reduced in the 1°C/minute scan. In both cases, however, the small backside transition at 25% to 35% weight loss is barely discernable. Comparing the conventional TGA results in figures B.21 and B.22 to the 50°C/minute Hi-Res™ scan in Figure B.23, we see that all of these transitions are better resolved in the Hi-Res™ TGA scan in a timely fashion.

Using Figures B.23, B.24, and B.25 we can compare the result of changing the maximum heating rate of a dynamic rate Hi-Res™ scan. Maximum heating rates of 50°C/minute (Figure B.23), 20°C/minute (Figure B.24) and 10°C/minute (Figure B.25) were used with a method consisting of a single Hi-Res™ ramp segment.

Sample size (approximately 10.45mg), sensitivity setting (1.0), purge gas (100 mL/minute dry air) and sample pan (open platinum) were all held constant from run to run.

As can be seen in the examples, the results are similar in each scan except that the measurement of transitions is moved to slightly lower temperatures as heating rate is reduced, and each run takes about 50% longer to complete when heating rate is reduced by 50%. Most noticeable, however, is that transition resolution is best in the fastest scan. This is because higher maximum heating rates allow the TGA furnace temperature to change more quickly between transitions which reduces transition overlap and flattens weight loss baseline. As an added benefit, experiment time is reduced compared to traditional resolution enhancement techniques. Here we can see the real beauty of dynamic rate Hi-Res™ TGA: better results in the same or less time as traditional constant heating rate TGA.




Figure B.21




Figure B.22




Figure B.23

Appendix B




Figure B.24




Figure B.25

References

Listed below are two excellent additional references for specific materials and techniques. These have been chosen because they give a broad review of the techniques involved and address the physical and chemical processes taking place, as opposed to a specific analysis of any one sample material. Each paper contains an extensive list of additional references for specific materials and techniques.

1. Thermoanalytical Examinations Under Quasi-Isothermal - Quasi-Isobaric Conditions, F. Paulik & J. Paulik, *Thermochimia Acta.*, 100 (1986) 23-59.
2. Controlled Transformation Rate Thermal Analysis: The Hidden Face of Thermal Analysis, J. Rouquerol, *Thermochimica Acta.*, 144 (1989) 209-224.

Hi-Res™ is a trademark of TA Instruments, Inc.
TYGON® is a registered trademark of NORTON Co.
ACCENT® is a registered trademark of PET, Inc.

Appendix B

Appendix C: TGA 2950 Autosampler Option

Introducing the Auto TGA	C-3
Getting Started	C-4
The TGA 2950 Instrument	C-6
Calibrating the Auto TGA	C-10
Running Experiments	C-11
Preparing the Samples	C-12
Selecting Sample and Tare Pans	C-12
Taring the Sample Pans	C-13
Loading the Samples	C-16
Setting Up an Experiment	C-18
Manual Operation	C-19
Tracking the TGA Autosampler Status	C-19
Interrupting a Run	C-20

Appendix C

Introducing the Auto TGA

The TGA Autosampler, known as the Auto TGA, is an accessory to the TA Instruments Thermogravimetric Analyzer (TGA) 2950 (see Figure C.1 on the next page). It allows you to place up to 16 samples at one time on the TGA instrument to measure the amount and rate of weight change in a material. Experiments are performed as they normally would be using the TGA—but now you can run samples on a continual basis and keep a log of the results using the Autosampler screens. The six (6) standard TGA pans listed below are used with the Auto TGA:

- 100 μL aluminum
- 50 and 100 μL platinum pans and
- 100, 250, and 500 μL alumina ceramic pans.

This appendix provides information on the set up of the Auto TGA. For instructions on the use of the Auto TGA through the Instrument Control software, please refer to the online help and documentation for further information.




Figure C.1
Auto TGA

Getting Started

The Auto TGA, as an accessory to the TGA, does not alter the procedures used to start up and shut down the TGA instrument and the controller; refer to the procedures found in Chapter 3 of this manual when starting your instrument.

The TGA 2950 Instrument

When you receive your Auto TGA, you will notice that a key has been added to the TGA instrument keypad—the AUTO SELECT key, shown in the figure below.




Figure C.2
TGA Instrument Display
and Keypad with Auto
TGA Accessory

Table C.1 on page C-5 explains the functions of the instrument keys when used in conjunction with the Auto TGA accessory.

Table C.1
Auto TGA
Instrument Keys

Key/Function	Explanation
SCROLL	Scrolls the realtime signals shown on the bottom line of the display. For more information on the progress of the experiment, refer to the online help and documentation. <u>Auto TGA function:</u> This key has an added function, to act as a shift key for the AUTO SELECT, TARE, and LOAD keys.
AUTO SELECT	<u>Auto TGA function:</u> Increments the next pan to be loaded when pressed alone.
SCROLL & AUTO SELECT	<u>Auto TGA function:</u> Decrements the next pan to be loaded when you hold down the SCROLL key while pressing AUTO SELECT. The selected sample is loaded or tared when the appropriate key is pressed. <i>(table continued)</i>

Table C.1
(continued)

Key/Function	Explanation
TARE	<u>Auto TGA functions:</u> Zeros the displayed weight of an empty sample pan: automatically loads the pan from the sample platform, raises the furnace to protect the pan from air currents, weighs the pan, stores the weight as an offset, and then unloads the pan. Performs a tare on the <i>selected</i> pan when pressed alone.
SCROLL/TARE	<u>Auto TGA function:</u> Tares <i>all</i> of the pans on the sample platform when you hold down the SCROLL key while pressing TARE.
LOAD	Loads the selected pan from the sample platform onto the balance. <u>Auto TGA function:</u> Loads the <i>selected</i> sample pan when pressed alone. (table continued)

Table C.1
(continued)

Key/Function	Explanation
SCROLL/LOAD	Continuously loads and unloads <i>all</i> of the pans on the sample platform when you hold down the SCROLL key while pressing LOAD.
NOTE:	LOAD functions can be stopped by pushing the STOP key on the instrument keypad.
UNLOAD	Functions of these remaining keys do not change when used with the Auto TGA accessory; see Chapter 1.
Δ FURNACE ▽	
START	
STOP	
REJECT (SCROLL/STOP)	

Calibrating the Auto TGA

When using the Auto TGA accessory, the need for calibration remains the same—weight calibration and temperature calibration are recommended at least once a month or when replacing the thermocouple; and the platform adjustment procedure needs to be performed if the sample hang-down wire fails to pick up a sample pan. See the online help and documentation for further information.

Running Experiments

Performing experiments with the Auto TGA is similar in many respects to the regular operation of the TGA instrument, the exceptions are explained in this section.

All of your TGA experiments will have the following general outline. In some cases, not all of these steps will be performed.

- Selecting the pan type and material
- Loading the pans
- Taring the empty sample pans
- Loading the samples into the pans
- Entering experiment information through the TA controller (sample and instrument information).
- Creating and selecting the thermal method on the controller.
- Attaching and setting up external accessories as required (*e.g.*, purge gas, Gas Switching Accessory).
- Starting the experiment.

Preparing the Samples

Selecting Sample and Tare Pans

The Auto TGA can utilize the same types of pans that are available for the TGA 2950 instrument.

When you are using the Auto TGA, it is possible to prepare from 1 to 64 different samples. Each sample platform can hold 16 numbered samples. The four sample platforms are numbered in the following ranges:

- 1-16 samples
- 17-32 samples
- 33-48 samples
- 49-64 samples.

Once you have selected the type of pan that you wish to use, you must use the *same type of pan* for all of the samples on the sample platform disk. The *same type of pan* that you use for experiments must also be used as a tare pan.

Tare Pan

1. Obtain a pan of the same type and size to be used for your experiments.
2. Remove the tare tube.
3. Use brass tweezers to hang the tare pan from the tare hook.
4. Mechanically tare the balance using the first step of the weight calibration procedure.

5. Replace the tare tube.

Before running any experiments on the TGA you must tare the sample pans to ensure that the weight measured by the balance reflects the weight of the sample only. You should tare the sample pans before each experiment, even if you use the same set of pans in consecutive experiments.

Taring the Sample Pans

When you tare the sample pans, the TGA reads the weight of the empty pans in their numbered order and then stores these weights as a set of offsets. These offsets are subtracted from subsequent weight measurements for each numbered sample. You can tare the pans manually or automatically as explained here.

NOTE:

View the Autosampler log, found on the Instrument Control software, after taring to determine which pan, if any, does not tare during the procedure (Error 118). If a tare error occurs, replace the pan and manually retare it.

Automatic Tare

Because the TGA 2950 has two weight ranges, taring is done for both ranges. The tare weight is stored by the instrument for the appropriate weight range.

1. Place the platform on the sample arm as shown in Figure C.3. Make sure that the small pin is inserted in the hole in the platform. This will seat the platform correctly.
2. Using brass tweezers, place each sample pan in a numbered place on the platform, making sure it is stable.




Figure C.3
Loading the Auto
TGA Sample Platform

3. Hold down the SCROLL key, then press the TARE key on the instrument keypad. The TGA will perform the following tare functions on each pan automatically:
 - Load the pan
 - Raise the furnace (to protect the pan from air currents)
 - Weigh the pan
 - Store the weight as the offset for each weight range
 - Unload the pan.

The Auto TGA will automatically tare each of the remaining pans in sequence. It should take approximately 45 minutes to tare all 16 pans on one sample platform. When the Auto TGA has completed taring each pan on the sample platform, the tare status message will disappear.

NOTE:

The advantage to using the Auto TGA is to save time and effort; therefore, it would not be efficient to perform manual tare operations on each sample, as that would defeat the purpose of an automatic operation. But, if manual taring is desired, see Chapter 3 for the procedure.

Loading the Samples

After you have done the taring procedures for all of the empty sample pans, load the samples as follows:

1. Place each sample in the correct sample pan making sure that you do not switch around the pans on the sample platform (they have already been tared in the numbered position).
2. Place the sample pans on the sample platform in their original order. Make sure that the wire on the bottom of the sample pans align with the groove in the panhole, so that the sample pan can be picked up by the sample hang-down wire.

NOTE:

Always use the brass tweezers to handle the sample pans.

◆ CAUTION:

Manually loading the sample pan onto the hang-down wire may damage the balance mechanism.

3. Press the LOAD key. The TGA will automatically load the selected sample.
4. Position the thermocouple at the edge of the sample pan, rather than in the middle for the best results (see Figure C.4).

NOTE:

The position of the thermocouple should be the same as it was during temperature calibration.

5. Press the UNLOAD key. The TGA will unload the sample and be ready for automatic sequencing.




Figure C.4
Adjusting the
Thermocouple

Setting Up an Experiment

Once you have prepared the sample, the next step in your experiment is to enter the needed information in the TA controller. All of the controller functions described in this section are accessed through the Instrument Control screen. Refer to the online help and documentation to learn how to perform the following steps.

1. **Select the Instrument.**
2. **Select the Instrument Mode.**
3. **Access the Autosampler Sequence.**
4. **Enter Sample Information.**
5. **Enter Instrument Information.**

NOTE:

If you are planning to run the Autosampler, be sure to specify furnace open at method end, as well as air cool, to cool down the furnace between samples.

6. **Create and Select Thermal Methods.**

The first time you use your TGA you will need to create at least one thermal *method* to control experiments. Each method is made of several *segments*, or individual instructions (*e.g.*, Equilibrate, Ramp), that control the state of the instrument. A different method may be selected for each sample.

Manual Operation

Use the AUTO SELECT and/or SCROLL/AUTO SELECT to manually select the pan that you want to run, then follow the procedures found on in Chapter 3 to start, stop, and monitor an experiment.

For details setting up the methods consult the online help and documentation.

Tracking the TGA Autosampler Status

To monitor the current run, observe the status of the Auto TGA , etc. you will need to access the TA Instruments Instrument Control software. Refer to the online help and documentation for further information.

Interrupting a Run

If you want to stop an Auto TGA run that is in progress, you can use one of the following keys on the TGA instrument keypad:

- **Stop** Terminates the current method and run and causes the Auto TGA to unload the pan and start the next run. However, the data file for the interrupted run is saved.
- **Reject** Terminates the current method and AS run and causes the Auto TGA to unload the cell and start the next run. The data file for the interrupted run is discarded.
- Other options are available through the controller functions, see the online help and documentation for further information.

Appendix D: TGA 2950 EGA Furnace Option

Introducing the EGA Furnace	D-3
EGA Furnace Specifications	D-5
Installing the EGA Furnace	D-6
First-Time Installation	D-6
Removing and Reinstalling the EGA Furnace	D-14
EGA Furnae Removal	D-14
EGA Furnace Installation	D-16
Connecting the Spectrometer	D-18
Using the EGA Furnace	D-21
Cleaning the Quartz Furnace Tube	D-22

Appendix D

Introducing the EGA Furnace

The Evolved Gas Analysis (EGA) Furnace (Figure D.1) is an optional accessory that allows you to connect a spectrometer to the instrument so that the gases evolved by sample decomposition can be analyzed. The EGA furnace and the standard TGA furnace can be exchanged easily on the instrument after initial hardware installation. Version 3.3B (or higher) of the TGA 2950 instrument software is required for the EGA furnace.




Figure D.1
TGA 2950 EGA Furnace

NOTE:

The following description of the EGA furnace may also be found in Chapter 4 of this manual.

The EGA furnace consists of a quartz glass sample tube surrounded by an electric resistance heater, both of which are contained within a water-cooled furnace housing. The housing is mounted to a furnace base that raises and lowers the furnace for sample loading and unloading.

The **sample tube** has a purge gas inlet that passes through the right side of the furnace housing. A fitting on the left side of the housing allows connection of a transfer line to carry exhaust gas to a spectrometer such as a mass spectrometer. Because the heater is external to the sample tube, evolved gases from sample decomposition within the sample tube do not come in contact with the resistance elements or the furnace ceramic refractory.

Cooling air enters through the furnace base and passes upward between the outside of the sample tube and the inside of the furnace, completely separating the cooling air from the sample and the sample zone.

The **furnace** is a resistance heater wound on alumina ceramic, which allows sample zone temperatures as high as 1000°C with heating rates up to 50°C/min.

A Platinel II* **thermocouple** is positioned in the furnace, just above the sample pan, where it monitors the sample environment temperature.

*Platinel II is a registered trademark of Engelhard Industries.

The **furnace base** moves the furnace assembly up around the sample pan to the closed position, or down away from the sample pan to the open position.

EGA Furnace Specifications

Refer to Table D.1 below for the specifications of the EGA furnace. Some of these may be similar to the standard furnace for the TGA 2950.

Table D.1
EGA Furnace
Specifications

Temperature range	25°C to 1000°C
Thermocouple	Platinel II*
Heating rate	0.1 to 50°C/min
*Platinel II is a registered trademark of Engelhard Industries.	

Installing the EGA Furnace



Make sure that you place the heater power switch in the off position before removing or installing the furnace on the TGA 2950 instrument.

Version 3.3B (or higher) of the TGA 2950 instrument software is required for the EGA furnace.

First-Time Installation

The first time that you install the EGA furnace involves the removal of the standard TGA furnace and replacement of the four furnace mounting screws and cleats with those supplied with the EGA furnace.

After this has been completed according to the following directions, you may interchange the two types of furnaces without changing the mounting hardware. Refer to page D-13 “Removing and Reinstalling the EGA Furnace” for subsequent furnace exchanges.

1. Place a drip pan to the left of the instrument to catch the coolant that will leak from the hose connections when the standard furnace is removed.
2. Remove the standard TGA furnace using steps 1 through 5 found in Chapter 5 “Removing and Reinstalling the Furnace.”

Make sure that you do not remove the furnace from its housing.

3. Unplug the AIR COOL line from the bottom of the furnace arm/base.

NOTE:

When you remove the AIR COOL line, do not let it slip back into the instrument cabinet.

4. Lay the furnace assembly down on the left side of the instrument so that the hose connections are positioned over the drip pan. Visually note the left (AAA) and right (CCC) orientation of the water lines so that you do not cross them later when they are reconnected. (See Figure D.2.) Then carefully snip the wire ties and disconnect the cooling water lines from the housing. (A small amount of cooling water will drain out into the pan when the hoses are disconnected.)
5. Press the FURNACE key to lower the furnace carriage completely.
6. Remove the four mounting screws and cleats from the furnace carriage (see Figure D.2).




Figure D.2
Removing the Furnace
Carriage Mounting Screws

7. Use the 3/32-inch ball driver supplied with the EGA installation kit to replace the screws with the new screws, spacers and cleats provided with the kit. Refer to Figure D.3. Replace the upper screws using the long screws with spacers and replace the lower screws using the short screws without spacers.




Figure D.3
Installing the EGA Mounting
Screws, Spacers, and Cleats

8. Tighten the upper two mounting screws fully, then loosen each of them no more than one full turn, so that the flat sides of the cleats are aligned vertically as shown in Figure D.4.

Now tighten the lower two mounting screws fully, then loosen each of them two full turns, plus a fraction of a turn, so that the flat sides of the cleats are aligned vertically as shown in Figure D.4.




Figure D.4
Aligning the Cleats

! WARNING

Loosening the upper two mounting screws any more than one full turn may cause the screws to interfere with the inside of the instrument cabinet, causing damage to the instrument.

9. Plug the AIR COOL line into the base of the EGA furnace, then connect the two water-cooling lines, making sure that the lines are not crossed. See Figure D.5 for the correct placement of the lines. Be sure to install the wire ties (supplied in the kit) around the cooling lines to prevent water leakage.




Figure D.5
Plugging in the
AIR COOL Line

10. Press the FURNACE key to raise the furnace carriage until the lower mounting screws are below the top edge of the enlarged cutout in the instrument faceplate. Then press the STOP key.
11. Plug the EGA furnace arm into the connector on the carriage and tighten the lower two mounting screws using the 3/32-inch ball driver supplied with the TGA 2950 EGA. See Figure D.6 on the next page.




Figure D.6
Tightening the
Lower Mounting Screws

12. Press the FURNACE key to completely lower the furnace. Use the 3/32-inch ball driver to tighten the upper furnace mounting screws. To reach the upper left mounting screw, insert the ball driver between the water connections on the left side of the furnace housing as shown in Figure D.7.




Figure D.7
Tightening the
Upper Mounting Screws

13. Connect the purge hose to the gas purge inlet on the right side of furnace.

◆ **CAUTION:**

Hold onto the glass purge tube with one hand while you install the purge hose to avoid breaking the glass.

14. Check the Heat Exchanger reservoir water level and add water if needed. See Chapter 2 “Filling the Heat Exchanger” for instructions.

Removing and Reinstalling the EGA Furnace



Make sure that you place the heater power switch in the off position before removing or installing the furnace on the TGA 2950 instrument.

To remove or reinstall the furnace, you will have to remove the furnace arm from its connection inside the slot on the front of the instrument cabinet.

EGA Furnace Removal

To remove the EGA furnace use the following procedure:

1. Press the FURNACE key to open the furnace completely.
2. Locate the top two mounting screws on each side of the furnace arm connection, found in the slot on the front of the instrument. Using the 3/32-inch ball driver supplied with the EGA furnace, loosen the two screws no more than one full turn. To reach the upper left mounting screw, insert the ball driver between the water connections on the left side of the furnace housing. Refer to Figure D.7 for the location of these screws.
3. Press the FURNACE key to raise the furnace about one (1) inch and press STOP.

4. Loosen the bottom two mounting screws using the 3/32-inch ball driver supplied with the TGA 2950 EGA. See Figure D.6.
5. Unplug and remove the EGA furnace arm/base from the instrument cabinet.
6. Unplug the AIR COOL line from the bottom of the furnace arm/base.

NOTE:

When you remove the AIR COOL line, do not let it slip back into the instrument cabinet.

7. Place a drip pan to the left of the instrument to catch the coolant that will leak from the hose connections when the EGA furnace is removed.
8. Lay the furnace assembly down on the left side of the instrument so that the hose connections are positioned over the drip pan. Then carefully snip the wire ties and disconnect the cooling water lines from the housing. (A small amount of cooling water will drain out into the pan when the hoses are disconnected.)

The furnace is now completely free from the instrument.

EGA Furnace Installation

To replace or reinstall the EGA furnace:

1. Plug the AIR COOL line into the bottom of the furnace arm/base as shown in Figure D.5 on page D-11.
2. Slip the water cooling hoses over the EGA furnace water connections and secure them with the wire ties.
3. Tighten the upper two mounting screws fully, then loosen each of them no more than one full turn, so that the flat sides of the cleats are aligned vertically as shown in Figure D.4 on page D-10.

Now tighten the lower two mounting screws fully, then loosen each of them two full turns, plus a fraction of a turn, so that the flat sides of the cleats are aligned vertically as shown in Figure D.4 on page D-10.



Loosening the upper two mounting screws any more than one full turn may cause the screws to interfere with the inside of the instrument cabinet, causing damage to the instrument.

4. Press the FURNACE key to raise the furnace carriage until the lower mounting screws are just below the top edge of the enlarged cutout in the instrument faceplate. Then press the STOP key.

5. Plug the EGA furnace arm into the connector on the carriage and tighten the lower two mounting screws using the 3/32-inch ball driver supplied with the EGA. See Figure D.6 on page D-12.
6. Press the FURNACE key to completely lower the furnace. Use the 3/32-inch ball driver to tighten the upper furnace mounting screws. To reach the upper left mounting screw, insert the ball driver between the water connections on the left side of the furnace housing as shown in Figure D.7 on page D-13.
7. Connect the purge hose to the gas purge inlet on the right side of furnace.

◆ **CAUTION:**

Hold onto the glass purge tube with one hand while you install the purge hose to avoid breaking the glass.

8. Check the Heat Exchanger reservoir water level and add water if needed. See “Filling the Heat Exchanger” for instructions.

Connecting the Spectrometer

The TGA 2950 EGA furnace allows you to connect a spectrometer, such as a FTIR spectrometer, to the instrument. To connect any spectrometer, you will need to use a transfer line (supplied by the spectrometer manufacturer) to transport the gas evolved from the sample on the TGA to the spectrometer.

- The transfer line should be 1/8 inch in diameter to connect with a 1/8-inch Swagelok™ fitting on the exhaust gas connection.
- The transfer line should be made of heat-resistant alloy capable of resisting corrosion by the evolved gas and oxidation at temperatures up to 1000°C.
- The transfer line must pass through the exhaust gas fitting and a glass branch tube in the sample tube. It should end at a point just short of the inside diameter of the sample tube to ensure that the evolved gases do not condense before entering the transfer line.
- The transfer line must be long enough to allow flexible movement. It must accommodate movement of the EGA furnace up and down 3 5/8 inches to open and close for sample loading and unloading. (If the transfer line is not long enough, it must be disconnected and reconnected each time the furnace is opened and closed.)

To connect your spectrometer to the EGA furnace follow these steps:

1. Install a Swagelok™ nut, drive ring, and ferrule on the correct length of transfer line, leaving more than two inches of transfer line projecting beyond the ferrule.
2. Swage the ferrule, then cut the end of the transfer tube off so that two inches of the tube projects beyond the ferrule. See Figure D.8 below

NOTE:

Extending the transfer line more than two inches beyond the Swagelok™ fitting may cause the TGA 2950 to operate improperly.




Figure D.8
EGA Transfer Line
with Ferrule

3. Make sure that the end of the transfer line is straight and free of oxide deposits before you insert it into the exhaust gas connection.

4. Insert the transfer line and tighten the Swagelok™ nut to seal the connector. When you tighten the Swagelok™ nut, use a 3/8-inch wrench on the exhaust fitting flats to prevent them from turning.



If the transfer line is not straight, or has heavy oxide deposits on it, the sample tube may be broken as the line is inserted.

Using the EGA Furnace

After proper installation, the EGA furnace can be used as you would normally use the TGA standard furnace. No special procedures must be followed when preparing TGA samples, setting up methods, or running experiments. Refer to the appropriate sections in the main manual for information. Follow the manufacturer's instructions for the use of your spectrometer when connected to the EGA furnace.

Cleaning the Quartz Furnace Tube



If the TGA is used to evaluate materials using an oxygen purge, the furnace should be cleaned routinely to prevent build-up of volatile hydrocarbon residues that could combust.



Do not touch the furnace sample tube with your bare fingers. Skin oils may cause devitrification of the quartz glass, resulting in severely reduced sample tube life. Do not insert metallic instruments inside the sample tube to scrape or chip contaminants from the sample tube as breakage may result.

To clean the furnace quartz sample tube, use the following procedure:



Do not disturb the hangdown wire and furnace thermocouple located directly above the furnace when cleaning the furnace, as damage may result.

1. Press the FURNACE key to open the furnace completely.
2. Remove any sample pans.
3. Remove the rubber cap located on the underside of the furnace base.
4. Place a small cup under the furnace tube. Rinse the furnace tube using a solvent (such as alcohol) to remove debris. The solvent will drain out of the bottom of the tube into the cup.

5. Using a soft bristle brush (we recommend a flexible bottle brush), gently slide the brush up and down to clean out the inside of the furnace tube, allowing the handle to bend freely (see Figure D.9).




Figure D.9
Cleaning the
Inside of the
Furnace Tube

6. Rinse the furnace tube with the solvent again.
7. Replace the rubber cap on the quartz tube stem when you have completed the cleaning procedure.
8. Purge the system with nitrogen for one hour.
9. Heat the furnace to 900°C to remove any remaining solvent.

Appendix D

Index

A

accessories 1-11
 setting up 3-15

Air Cool (status code) 4-17

air-cool
 using 3-16

Autosampler C-1 to C-20

Auto Zero 3-10

B

balance 1-4, 3-16, 4-6
 arm 4-6
 arm sensor 4-6
 unpacking 2-20

binary address switches 2-16

C

cabinet 1-4, 4-13

cable
 GPIB 2-15

Calib (status code) 4-17

calibration 3-5
 sample platform 3-6
 temperature 3-5
 weight 3-6

cleaning 3-9

 EGA furnace quartz tube D-22

Closing (status code) 4-17

Cold (status code) 4-17

Complete (status code) 4-17

components 1-4, 4-5

confidence test 5-19

controller
 description 1-3

Cooling (status code) 4-18

cooling gas line 2-18

D

description
 TGA 2950 4-3

display 1-5

E

EGA furnace 1-11, 4-11

Ending (status code) 4-18

Equilb (status code) 4-18

Evolved Gas Analysis (EGA)

Furnace D-3 to D-24. *See also*

EGA furnace

cleaning D-22

description D-3

installation D-6, D-16

reinstallation D-14

specifications D-5

usage D-21

experiment

oxygen-free atmosphere 3-22

set up 3-23

procedure 3-7

rejecting 3-21

setting up 3-14

starting 3-20

stopping 3-21

F

furnace 1-4

cleaning 5-5

EGA 4-9, 4-11, D-1

reinstalling 5-11

removing 5-11

standard 4-9

furnace base 4-10

furnace housing 4-10

furnace tube

cleaning D-22

FURNACE up key 1-7

fuses 5-16

G

gas

purge 3-16

types 3-17

Gas Switching Accessory 1-11,
3-19

gases

recommended 2-17

GPIO cable 2-15

H

hang-down wires 4-7

aligning 2-25

installing 2-22

heat exchanger 1-4, 4-15, 5-6

cables 2-12

coolant 5-6

filling 2-10

lines 2-12

water reservoir 5-6

heater 4-10, 4-12

HEATER switch 1-10

Heating (status code) 4-18

High Resolution option B-5 to
B-12

Holding (status code) 4-18

Hot (status code) 4-19

I

Initial (status code) 4-19

instrument. *See* TGA 2950

Iso (status code) 4-19

J

Jumping (status code) 4-19

K

keypad 1-6

L

lines
connecting 2-12

Load (status code) 4-19

LOAD key 1-7

M

meter movement 4-6

N

No Power (status code) 4-19

O

Opening (status code) 4-20

ordering information A-1

oxygen-free atmosphere 3-22

P

pans 3-8, 4-7
cleaning 3-9
unloading 3-21

parts 5-21

power problems
failures 5-18

POWER switch 1-10

purge gas 2-18, 3-16

purge lines 2-17

R

Ready (status code) 4-20

Reject (status code) 4-20

REJECT function 1-8

Repeat (status code) 4-20

routine maintenance 5-4

S

sample
loading 3-11, 4-8
preparing 3-8
unloading 3-21

Index

- sample pans 3-8
- sample platform 1-4
 - adjusting 2-29
 - calibration 3-6
- SCROLL key 1-6
- SCROLL-STOP keys 1-8
- Set Up (status code) 4-20
- specifications 1-12
- spectrometer
 - connecting to EGA furnace D-18
- Stand by (status code) 4-20
- START key 1-7
- Started (status code) 4-21
- status codes 4-17. *See also* name of code
- STOP key 1-8
- T**
- Tare (status code) 4-21
- TARE key 1-6
- taring 3-9
 - automatic 3-10
 - manual 3-10
- Temp °C (status code) 4-21
- Temp* (status code) 4-21
- test functions 5-19
- TGA 2950
 - accessories 1-11 to 1-12
 - address switches 2-15
 - air cool 3-16
 - autosampler C-3 to C-20
 - cables
 - GPIB 2-15
 - power 2-19
 - calibrating 3-5
 - temperature 3-5
 - weight 3-6
 - cleaning 5-4
 - furnace 5-5
 - components 1-4, 4-5
 - balance 4-6
 - cabinet 4-13
 - hang-down wire 4-7
 - confidence test 5-19
 - description 1-3, 4-3
 - display 1-5
 - EGA furnace D-3 to D-23
 - furnace
 - installing 5-14
 - removing 5-11
 - hang-down wires
 - aligning 2-25
 - heat exchanger 4-15, 5-6
 - filling 2-10
 - water lines 2-12
 - High Resolution option B-5
 - inspecting 2-8
 - installing 2-7–2-19
 - hang-down wires 2-22

TGA 2950 (*cont'd*)
 instrument characteristics 1-12
 keypad 1-6
 lines
 cooling gas 2-18
 purge 2-17
 loading samples 3-11
 location 2-9
 operating parameters 1-12
 power problems 5-18
 power switch 1-10
 purge gas 2-18, 3-16
 purge gas types 3-17
 repacking 2-6–2-7
 replacement parts 5-21
 routine maintenance 5-4
 sample pans 4-7
 loading 4-8
 sample platform
 adjusting 2-29
 sampling system 1-13
 setting up an experiment 3-14
 specifications 1-12
 starting 2-30
 status codes 4-17
 stopping 2-32
 tare pans 4-7
 test functions 5-19
 theory of operation 4-16
 thermocouple
 replacing 5-9
 unpacking 2-3–2-6
 weight ranges 4-3
thermocouple 4-10, 4-12
 replacing 5-9

Thermogravimetric analysis 4-16

Thermogravimetric Analyzer. *See*
 TGA 2950

U

Unload (status code) 4-21

UNLOAD key 1-7

W

Weight # (status code) 4-21

Index

Artisan Technology Group is an independent supplier of quality pre-owned equipment

Gold-standard solutions

Extend the life of your critical industrial, commercial, and military systems with our superior service and support.

We buy equipment

Planning to upgrade your current equipment? Have surplus equipment taking up shelf space? We'll give it a new home.

Learn more!

Visit us at [artisanth.com](https://www.artisanth.com) for more info on price quotes, drivers, technical specifications, manuals, and documentation.

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

We're here to make your life easier. How can we help you today?

(217) 352-9330 | sales@artisanth.com | [artisanth.com](https://www.artisanth.com)

