ARTISAN®
TECHNOLOGY GROUP

Your **definitive** source for quality pre-owned equipment.

Artisan Technology Group
(217) 352-9330 | sales@artisantg.com | artisantg.com

Full-service, independent repair center
with experienced engineers and technicians on staff.

We buy your excess, underutilized, and idle equipment
along with credit for buybacks and trade-ins.

Custom engineering
so your equipment works exactly as you specify.

- Critical and expedited services
- In stock / Ready-to-ship
- Leasing / Rentals / Demos
- ITAR-certified secure asset solutions

Expert team | Trust guarantee | 100% satisfaction

All trademarks, brand names, and brands appearing herein are the property of their respective owners.

Find the *Alphi Technology cPCI-AD320* at our website: ***Click HERE***

CPCI-AD320

16-32 Channel High Performance Analog Data Acquisition Card for 6U *CompactPCI*™ systems

REFERENCE MANUAL

767-13-000-4000
Version 1.3
March 2001

ALPHI TECHNOLOGY CORPORATION

6202 S. Maple Avenue #120
Tempe, AZ 85283 USA
Tel: (480) 838-2428
Fax: (480) 838-4477

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

NOTICE

The information in this document has been carefully checked and is believed to be entirely reliable. While all reasonable efforts to ensure accuracy have been taken in the preparation of this manual, ALPHI TECHNOLOGY assumes no responsibility resulting from omissions or errors in this manual, or from the use of information contain herein.

ALPHI TECHNOLOGY reserves the right to make any changes, without notice, to this or any of ALPHI TECHNOLOGY's products to improve reliability, performance, function or design.

ALPHI TECHNOLOGY does not assume any liability arising out of the application or use of any product or circuit described herein; nor does ALPHI TECHNOLOGY convey any license under its patent rights or the rights of others.

ALPHI TECHNOLOGY CORPORATION

All Rights Reserved

This document shall not be duplicated, nor its contents used
for any purpose, unless express permission has been granted in advance.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

TABLE OF CONTENTS

1. GENERAL DESCRIPTION 1
1.1 INTRODUCTION 1
1.2 FUNCTIONAL DESCRIPTION 2
1.3 TRIGGERING DESCRIPTION 2
1.3.1 CLOCK / TRIGGER SOURCES 3
1.3.2 CLOCK / TRIGGER TARGETS 4
1.3.3 PXI TRIGGER OUTPUTS 4
1.4 SOFTWARE SUPPORT 5
1.5 REFERENCE MATERIALS LIST 6

2. CPCI BRIDGE TO LOCAL PCI BUS 7
2.1 OFFLOADING OF CPCI BUS 7
2.2 DSP HANDLING OF PMC INTERRUPT 7
2.3 NO BRIDGE PROGRAMMING REQUIRED 7
2.4 STAND ALONE OPERATION 7

3. LOCAL PCI BRIDGE TO LOCAL DSP BUS 8
3.1 INTERFACE TO HOST (CPCI) 8
3.2 CPCI CONFIGURATION SPACE 8
3.3 CPCI BASE ADDRESS REGIONS 10

4. LOCAL DSP BUS 11
4.1 LOCAL SRAM 12
4.2 CARD CONTROL / STATUS REGISTERS 12
4.2.1 9080 CONTROL / STATUS REGISTER 13
4.2.2 HOST INTERRUPT ACKNOWLEDGE 14
4.2.3 BOARD CONTROL / STATUS REGISTER 15
4.2.4 RESET THE HARDWARE FIFO 16
4.2.5 READ THE HARDWARE FIFO 16
4.2.6 WRITE THE HARDWARE FIFO 16
4.2.7 RESET WATCHDOG TIMER 17
4.2.8 PGA GAIN REGISTERS 17
4.2.9 SOFTWARE STROBE FOR MANUAL CLOCKING BY DSP 18
4.2.10 TRIGGER / GATE CONFIGURATION REGISTERS 18
4.2.11 8530 SERIAL PORTS 20
4.2.12 SOFTWARE STROBE FOR MANUAL CLOCKING BY HOST 21
4.2.13 RPIO SERIAL BUS AND DIGITAL PORT CONFIGURATION 21
4.2.14 DIGITAL PORTS A AND B 23
4.2.15 DSP INTERRUPT STATUS REGISTER 23
4.2.16 DSP INTERRUPT ENABLE REGISTER 25
4.2.17 DSP BOOTMODE REGISTER 26
4.2.18 NULL PACE / LIMIT REGISTER 27
4.2.19 DEMAND DMA CONTROL REGISTER 28
4.2.20 DEMAND DMA 0 TRANSFER COUNT REGISTER 29
4.2.21 DEMAND DMA 1 TRANSFER COUNT REGISTER 30
4.3 LOCAL FLASH 30
4.4 PLX OPERATION REGISTERS 30

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.4.1 LOCAL CONFIGURATION REGISTERS 31
4.4.2 RUNTIME REGISTERS 32
4.4.3 DMA REGISTERS 33
4.5 CPCI PASS-THROUGH REGION 34

5. HARDWARE DETAILS 35
5.1 BOARD PERIPHERALS 35
5.1.1 A/D CONVERTERS 35
5.1.2 DIGITAL I/O 36
5.1.3 SERIAL PERIPHERAL CHAIN 36
5.2 DMA 37
5.2.1 DSP DMA 37
5.2.2 DSP TO PCI BRIDGE DMA 37
5.2.3 DEMAND MODE DMA 38
5.3 PXI 38
5.4 INTERRUPTS 39
5.4.1 DSP INTERRUPTS 39
5.4.2 HOST INTERRUPTS 40
5.5 RESET 41
5.5.1 PCI HARDWARE RESET 41
5.5.2 SOFTWARE RESET FROM A PCI DEVICE 41
5.5.3 HOLDING DSP IN RESET 41
5.5.4 WATCHDOG RESET 42
5.5.5 DSP BOOT MODES 42
5.6 CONNECTORS, JUMPERS, AND LEDS 43
5.6.1 JUMPER DESCRIPTIONS 44
5.6.2 FRONT PANEL DESCRIPTION 45
5.6.3 REAR PANEL DESCRIPTION 51
5.6.4 CONNECTORS ON COMPONENT SIDE DESCRIPTION 54

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

1. GENERAL DESCRIPTION

1.1 INTRODUCTION

The **CPCI-AD320** is a DSP based data acquisition and processing system, in a 6U CPCI form factor, targeted for general purpose analog data acquisition systems. The **CPCI-AD320** provides from 16 to 32 differential analog inputs (depending upon option ordered), with programmable gain, to individual A/D converters, operating at a maximum conversion rate of 200 kHz.

The **CPCI-AD320** is also a fully compatible PMC carrier for a single PMC module. Additional processing horsepower is available by adding an additional PMC module with DSP processing to the **CPCI-AD320**. Software control of this card is fully available to the on board DSP, any DSP or processor on the PMC module, as well as the HOST processor on the CPCI bus.

Connections to the input signals are available at the front panel and at the rear panel I/O connectors, and are fully supported by ALPHI Technology's line of Rear Panel I/O modules targeted towards data acquisition applications. In particular, the following modules are applicable:

- **RPIO-SGEX**: 8 channels of DC Bridge/Load Cell input.
- **RPIO-SGEX-M**: Additional 8 channels of DC Bridge/Load Cell input.
- **RPIO-CAL-M**: Adds bridge calibration, bridge completion, and monitoring of the excitation current for the 8 channels in the **RPIO-SGEX**.
- **RPIO-LVDT-SGEX**: 4 channels of LVDT signal conditioning and 4 channels of DC Bridge/Load Cell input.

Additionally, the **CPCI-AD320** is fully PXI compatible, and can make use of the PXI triggering signals to support synchronizing multiple boards.

The primary features of the **CPCI-AD320** are as follows:

- 16-32 channels of simultaneous 16 bit A/D acquisition at 200 kHz.
- Integrated DSP (TI TMS320C32) at 60 MHz available to process the data, if desired, or to offload I/O operation from PMC processors or the HOST.
- 128K x 32 bit zero wait state SRAM for the DSP.
- 512K by 8 bit FLASH device for bootloader and customer applications.
- Programmable PGAs allow for gains of 1, 2, 4, or 8, or 1, 2, 5, or 10 on a per channel basis (determined at time of order).
- Extremely flexible triggering, including internal, external, and PXI trigger/gate sources.
- Two channels of Bus Master DMA with Scatter/Gather support to offload the PMC and the HOST processor.
- Fully shared access to SRAM, card peripherals by HOST and PMC Processor.
- DSP is fully capable of accessing PMC, HOST RAM, and any other devices present on the CPCI bus
- 16 bits of digital input and output.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

- 4 bits of digital output to control peripherals on the RPIO modules.
- One RS232 port and one RS422 port.
- Full software support including DSP code and Drivers/DLL for WinNT.

1.2 FUNCTIONAL DESCRIPTION

An overview block diagram of the **CPCI-AD320** is presented in Figure 1-1.




Figure 1.1: Overview Block Diagram

The CPCI to local PCI bridge completely isolates local PCI accesses between the PMC and this card from the main CPCI bus. This will allow high speed, low latency transfers to the PMC module.

There is full bi-directional access through the PCI to Local Bus bridge. The DSP can directly address and access memory on the HOST processor in addition to DMA. The HOST can directly access resources on the Local DSP bus.

1.3 TRIGGERING DESCRIPTION

The **CPCI-AD320** has extremely flexible triggering. The triggering clock / gate block diagram is presented in Figure 1-2.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Figure 1.2: Trigger Clock / Gate Signal Block Diagram

It is easier to describe the sources and targets separately. Any source can be connected to any target, except for the PXI Outputs which are described as a special case.

1.3.1 CLOCK / TRIGGER SOURCES

DSP Timers

There are two timers internal to the DSP which are capable of generating clock outputs. The outputs can be selected between positive or negative pulses, and square wave. Periods are multiples of 66.7 nS.

External TTL inputs

There are two front panel SMB connectors labeled Clock and Gate. The inputs are TTL compatible.

PXI Trigger Inputs

There are 8 PXI trigger lines, which may be used to synchronize multiple boards, provided that the other boards and the CPCI backplane are PXI compatible.

Trigger by DSP

The DSP can trigger a pulse by writing to a location in the registers.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Trigger by HOST
The HOST can trigger a pulse by writing to a location in the registers.

Discrete Logic Level
A forced 0 or 1 can be used as a source. This may be useful for the A/D Gate if the Gate functionality is not desired.

1.3.2 CLOCK / TRIGGER TARGETS

These targets can be connected to any of the above sources.

A/D Converters
The A/D converters are controlled by two signals: Clock and Gate. A rising edge on Clock will trigger the start of a conversion. Gate serves as a Clock qualifier, and must be high for Clock to be recognized.

Demand DMA Engine for DSP to PCI Transfers
One of the two DSP to PCI bridge DMA engines can be operated in demand mode, transferring a fixed number of DWORDs based on a selected trigger signal.

External Clock Output
There is a front panel SMB connector labeled Clock Output. The output is TTL compatible.

HOST Interrupt
The HOST can be interrupted on a rising edge of the source.

DSP Interrupt
The DSP can be interrupted on a rising edge of the source.

1.3.3 PXI TRIGGER OUTPUTS

The card can source signals on the PXI Trigger lines from the following sources.

- DSP Timers
- External TTL inputs
- Trigger by DSP
- Trigger by HOST
- Discrete Logic Level

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

1.4 SOFTWARE SUPPORT

The **CPCI-AD320** is supported under *Windows NT* by a **Board Support Package** which is supplied with the card.

The card is also supported by a **DSP Support Library** providing the following features:

- Common access routines for communications with the HOST. These routines also allow for Direct and DMA access to HOST memory.
- Identify the applicable card resources and parameters.
- Allow for serial communication through the provided serial port for debugging DSP code and configuring the card.

A bootloader provided on the card allows for control by the HOST, including downloading custom DSP code. User code can also be downloaded to FLASH memory and booted automatically on reset.

All these are provided in a manner consistent across ALPHI Technology platforms. Other documentation supplied with the card will describe the support in full detail.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

1.5 REFERENCE MATERIALS LIST

PCI Local Bus Specification:

PCI Special Interest Group
Tel: (800) 433-5177
Tel: (503) 797-4207
Fax: (503) 234-6762

PXI Specification: PCI Extensions for Instrumentation:

PXI System Alliance
http://www.pxisa.org

PCI 9080 PCI Controller data book:

PLX Technology Corporation
Tel: (408) 744-9060
http://www.plxtech.com
apps@plxtech.com

DSP (TMS320C32) C Compiler, Assembler, Linker:

Part Number TMDU3243855-02

Texas Instruments
Tel: (410) 312-7900
http://www.ti.com

DSP Debugger and Integrated Development Environment:

Code Composer

GO-DSP Corporation
Tel: (416) 599-6868
http://www.go-dsp.com

DSP Emulators suitable for use with *Code Composer* or TI debugger:

White Mountain DSP
(603) 833-2430
http://www.wmdsp.com

WindowsNT and Windows95 Programming Tools:

BlueWater Systems
Tel: (206) 771-3610
Fax: (206) 771-2742
info@bluewatersystems.com
http://www.bluewatersystems.com

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

2. CPCI BRIDGE TO LOCAL PCI BUS

A Standard PCI to PCI bridge chip is used to connect the CPCI Bus to the local PCI bus present on the **CPCI-AD320**. There are two PCI type devices present on the local PCI bus.

- PLX 9080 bridge between the Local PCI bus and the DSP bus.
- Installed PMC Module.

The bridge is necessary to meet the CPCI bus loading requirements, and it also affords a few added benefits.

2.1 OFFLOADING OF CPCI BUS

Local PCI Bus traffic does not affect the CPCI bus. The PLX9080 and the PMC module can communicate at high rates and low latency without affecting the HOST CPCI bus. Both devices can request the Local PCI bus for direct bus mastered transfers, both to each other, as well as the rest of the CPCI bus through the CPCI bridge.

This bridge allows the HOST CPCI bus to independently access both the PMC and the **CPCI-AD320** registers on the DSP bus. Standard HOST drivers for the PMC module will work without any modification.

Because of buffering within the CPCI to Local PCI bridge, overall bus utilization can be maintained at a high percentage.

2.2 DSP HANDLING OF PMC INTERRUPT

The hardware on the **CPCI-AD320** makes it possible to intercept the PMC interrupt line from reaching the CPCI bus, and can route the interrupt to the DSP. This may be helpful for applications involving processor PMC modules or data communication PMC modules. Custom software on the DSP can then offload the HOST processor.

2.3 NO BRIDGE PROGRAMMING REQUIRED

Many HOST operating systems will automatically configure the bridge with no intervention required. The CPCI to Local PCI bridge is fully compatible with the PCI bridge specifications.

2.4 STAND ALONE OPERATION

If the **CPCI-AD320** is used in a stand-alone configuration without a CPCI bus, the CPCI to Local PCI bridge will not be populated, and hardware on the **CPCI-AD320** will take over the local PCI clocking and bus request and grant. The DSP and the PLX9080 can configure the PMC PCI configuration registers.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3. LOCAL PCI BRIDGE TO LOCAL DSP BUS

A PLX PCI9080 bridge chip is used to interface the Local DSP Bus to the local PCI Bus on the **CPCI-AD320**. This interface provides the card with the following features.

- Completely transparent access by HOST and PMC to the Local DSP Bus.
- Completely transparent access by DSP to the CPCI bus and PMC.
- Eight mailbox registers and two doorbell registers for communication between the HOST/PMC and the DSP.
- Two Bus Master DMA channels with scatter / gather support for transfers between the local PCI or CPCI bus and the DSP bus.
- Full interrupt support to HOST and DSP.
- Ability to intercept PCI interrupt from PMC to handle by DSP.

3.1 INTERFACE TO HOST (CPCI)

All PCI devices contain a set of registers in Configuration Space which allow for determining the manufacturer and model of the device, determining the resources necessary for the correct operation of the device, and other configuration information. Configuration space is decoded on a per slot / device basis via a mechanism described in the PCI specification.

All PCI devices can be relocated in physical memory by means of several Base Address Registers. These registers contain the high address bits, which must be matched for any access to the card to be successful. Part of the protocol of programming the Base Address Registers allows for the transfer of information regarding the size of the regions. The actual Base Address Registers are located in Configuration Space.

Normally, configuring the card is performed by the system controller and some form of firmware or BIOS. Unfortunately, determining Base Addresses and other configuration information in Configuration Space is operating system dependent. One of the primary functions of the device driver under Windows NT is to map these resources to the user application.

The card is actually accessed through the decoded base address registers.

3.2 CPCI CONFIGURATION SPACE

DSP Address: *0x880000 – 0x88000F*
PCI Address: *CONFIG:0x00 – 0x3C*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset*

The card has the following registers available to PCI Configuration Space. They are implemented in the PLX chip. The registers are also accessible from the DSP Local bus.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Offset Into PCI CFG	DSP Address	31 – 24	23 – 16	15 – 8	7 – 0
0x00	0x880000	Device ID		Vendor ID	
0x04	0x880001	Status		Command	
0x08	0x880002	Class Code			Revision ID
0x0C	0x880003	BIST	Header Type	PCI Latency Timer	Cache Line Size
0x10	0x880004	PCI Base Address 0 (Memory Access to PLX Registers)			
0x14	0x880005	PCI Base Address 1 (I/O Access to PLX Registers)			
0x18	0x880006	PCI Base Address 2 (Memory Access to DSP SRAM and card registers)			
0x1C	0x880007	PCI Base Address 3 (Not Used for this card)			
0x20	0x880008	Unused PCI Base Address 4			
0x24	0x880009	Unused PCI Base Address 5			
0x28	0x88000A	Cardbus CIS Pointer (Not Supported)			
0x2C	0x88000B	Subsystem ID		Subsystem Vendor ID	
0x30	0x88000C	PCI Base Address for Expansion ROM			
0x34	0x88000D	Reserved			
0x38	0x88000E	Reserved			
0x3C	0x88000F	Max Latency	Min Grant	Interrupt Pin	Interrupt Line

Table 3.1: CPCI Configuration Space

The card presents the following initial configuration values to the CPCI system, based on the values stored in the NVRAM device read by the PLX PCI9080 interface chip.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Register	Value (Meaning)
Vendor ID	0x13C5 (ALPHI Technology)
Device ID	0x0205
Revision ID	0x00
Class Code	0xff0000 (Device does not fit into defined class codes)
Interrupt Line	0xff
Interrupt Pin	A
Multifunction Device	No
Build In Self Test	No
Latency Timer	0x00
Minimum Grant	0x00
Maximum Latency	0x00
Base Address 0 Size	0x100 Bytes (Memory Access to PLX Registers)
Base Address 1 Size	0x100 Bytes (I/O Access to PLX Registers)
Base Address 2 Size	0x200000 Bytes (Memory Access to DSP SRAM and card registers)
Base Address 3 Size	None (Not applicable to this card)
Expansion ROM Size	None

Table 3.2: PCI Configuration Register Default Values

3.3 CPCI BASE ADDRESS REGIONS

There are 4 base address regions available on the PLX PCI9080. The **CPCI-AD320** uses 3 of these regions.

- BAR0: Memory access the PLX operation registers.
- BAR1: I/O access the PLX operation registers.
- BAR2: Passthru to the DSP bus and is used to access the DSP SRAM and the card control / status registers. The passthru is established at PCI reset via the NVRAM.
- BAR3: Not used on this design.

NOTE: Regions, which are accessed by memory cycles, will reside at physical addresses above 1MB. Therefore, non-protected (real) mode operating systems running on a x86 class processor, such as DOS, will not be able to access these regions. Any applications, which must run under DOS, must be compiled with a DOS Protected Mode Extender, such as PharLap or will require making DPMI calls to switch to protected mode. ALPHI Technology does not provide any DOS software support, but we will be happy to help with hardware technical issues.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4. LOCAL DSP BUS

The following devices are present on the Local DSP Bus at the addresses specified. Many of these devices are also accessible by the HOST directly through the Base Address Regions. The PLX PCI9080 accesses the Local DSP Bus by placing the DSP in a HOLD state, performing any cycles necessary, and releasing the DSP from HOLD.

HOST Address	DSP Address	Data	R/W	Description
BAR2: 0x000000 – 0x07FFFF	0x000000 – 0x01FFFF	D31-D00	R/W	Local SRAM
BAR2: 0x080000 – 0x0FFFFF	0x020000 – 0x03FFFF	D31-D00	R/W	Local SRAM repeated
BAR2: 0x100000 – 0x17FFFF	0x040000 – 0x05FFFF	D31-D00	R/W	Card Control / Status Registers
BAR0: 0x00 – 0xFF	0x880000 – 0x88003F	D31-D00	R/W	PLX Operation Registers
Not Accessible	0x900000 – 0x97FFFF	D07-D00	RO	Local FLASH
Not Accessible	0xC00000 – 0xFFFFFF	D31-D00	R/W	Pass through to CPCI bus

Table 4.1: Local DSP Bus Overview

In certain applications, it may be necessary to calculate the impact of accessing the Local DSP Bus from the HOST. The following table of events should help to explain the timing of a SRAM read. A single 32 bit read will stall the DSP for 10 clocks, or 333 nS. Burst access will add a single clock for each additional word. Of course, the DSP can continue running from the DSP cache and using internal RAM inside the DSP. Additionally, access to the Local Bus by the PLX can be turned off by the DSP for short periods of time during critical code sections.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Cycle	Action by DSP	Action by PLX
0	Normal Access	Signals request for bus
1	Idles bus prior to relinquish	Waits for acknowledgement
2	Relinquishes bus and acknowledges	Seizes bus, driving idle
3	Waits for release	Drives Address Phase for Read
3+1	Waits for release	Performs Data Phase for Read
3+2	Waits for release	Optionally continue burst Read
…	…	…
3+n	Waits for release	Last burst read
4+n	Waits for release	Internal Synchronization
5+n	Waits for release	Internal Synchronization
6+n	Waits for release	Releases bus. Driven idle by card.
7+n	One cycle lost to synchronization	Idle
8+n	DSP recognizes release	Idle
9+n	DSP performs idle cycle	Idle

Table 4.2: HOST Read of Local DSP SRAM

4.1 LOCAL SRAM

DSP Address: *0x000000 - 0x03FFFF, Bits 31-00*
PCI Address: *BAR2:0x000000 - 0x0FFFFF, Bits 03-00*
Mode of Access: *Read/Write*

There is 128k x 32 of Local SRAM for storing DSP programs and buffering data. The SRAM is accessible to both the DSP and the HOST.

The SRAM operates at zero wait states to the DSP, and one wait state on the first read by the PLX PCI9080 due to the separate address and data phase. Additional access in burst mode by the PLX will occur at zero wait state.

The decoding of the SRAM occurs twice, at 0x000000 and 0x020000 in order to make use of the 4k section at 0x000000 which is normally lost to internal DSP access. This memory can be accessed at 0x020000 – 0x020FFF.

4.2 CARD CONTROL / STATUS REGISTERS

The following locations are the Control / Status registers for the card. These registers are accessible to both the DSP and the HOST.

These registers operate with two wait states to the DSP.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

HOST Address BAR2:	DSP Address	Data	R/W	Description
0x100000	0x040000	D03-D00	R/W	9080 Control / Status Register
0x100020	0x040008	N/A	WS	HOST Interrupt Acknowledge
0x100040	0x040010	D07-D00	R/W	Board Control / Status Register
0x100060	0x040018	N/A	WS	Reset the hardware FIFO
0x100080	0x040020	D15-D00	R	Read the hardware FIFO
0x100080	0x040020	D15-D00	W	Write the hardware FIFO for testing and configuration
0x1000A4	0x040029	N/A	WS	Reset Watchdog Timer
0x1000C0	0x040030	D31-D00	WO	PGA Gain Registers
0x1000E0	0x040038	N/A	WS	Software Strobe for manual clocking by DSP
0x100100	0x040040	D07-D00	R/W	Trigger / Gate Configuration Registers
0x100120	0x040048	D07-D00	R/W	8530 Serial Ports
0x100140	0x040050	N/A	WS	Software Strobe for manual clocking by HOST
0x100160	0x040058	D07-D00	R/W	RPIO Serial Bus and Digital Port Configuration
0x100180	0x040060	D15-D00	R/W	Digital Ports A and B
0x100200	0x040080	D07-D00	R/W	DSP Interrupt Status Register
0x100224	0x040081	D07-D00	R/W	DSP Interrupt Enable Register
0x100228	0x040082	D07-D00	R/W	DSP Bootmode Register
0x10022C	0x040083	D07-D00	R/W	Null Pace / Limit Register
0x100234	0x040085	D07-D00	R/W	Demand DMA Control Register
0x100238	0x040086	D07-D00	R/W	Demand DMA 0 Transfer Count Register
0x10023C	0x040087	D07-D00	R/W	Demand DMA 1 Transfer Count Register

Table 4.3: Control / Status Registers

4.2.1 9080 CONTROL / STATUS REGISTER

DSP Address: *0x040000, Bits 03-00*
PCI Address: *BAR2:0x100000, Bits 03-00*
Mode of Access: *Mixed*

This register allows for simplified determination of the cause of a DSP interrupt from the PLX, as well as providing a way to temporarily lockout access by the HOST.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

BIT 03	BIT 02	BIT 01	BIT 00
LSERR	LINTO	DMPAF	LOCKOUT9080

LSERR
DSP Address: *0x040000, Bit 03*
PCI Address: *BAR2:0x100000, Bit 03*
Mode of Access: *Read Only*
Reset By *See PLX Documentation*
This bit tracks the current state of the LSERR line from the PLX PCI9080. When this is low, (0), it implies that the DSP interrupt was caused by this hardware signal. See the PLX documentation for more details and for a list of causes.

LINTO
DSP Address: *0x040000, Bit 02*
PCI Address: *BAR2:0x100000, Bit 02*
Mode of Access: *Read Only*
Reset By *See PLX Documentation*
This bit tracks the current state of the LINTO line from the PLX PCI9080. This line is the normal way in which the PLX interrupts the DSP. When this is low, (0), it implies that the DSP interrupt was caused by this hardware signal. See the PLX documentation for more details and for the enable bits.

DMPAF
DSP Address: *0x040000, Bit 01*
PCI Address: *BAR2:0x100000, Bit 01*
Mode of Access: *Read Only*
Reset By *See PLX Documentation*
This bit tracks the current state of the DMPAF line from the PLX PCI9080. This line can be used to throttle direct writes by the DSP to the HOST PCI bus. When this is low, (0), it implies that the FIFO inside the PLX has a programmable number of accesses queued, but not yet completed for the HOST PCI bus. See the PLX documentation for more details and for the enable bits.

LOCKOUT9080
DSP Address: *0x040000, Bit 00*
PCI Address: *BAR2:0x100000, Bit 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*
This bit, when set (1) will prevent the PLX PCI9080 from seizing the Local Bus. This can allow for time critical DSP code to complete in a timely manner. This bit should only be set for short intervals of time, since it may potentially lock up the HOST PCI bus until it has been cleared (when an access to the card is performed).

4.2.2 HOST INTERRUPT ACKNOWLEDGE

DSP Address: *0x040008*
PCI Address: *BAR2:0x100020*
Mode of Access: *Write Strobe*
The HOST can write to this register to turn off the HOST interrupt generated by LINTI. It is intended for use by the HOST device driver.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.2.3 BOARD CONTROL / STATUS REGISTER

DSP Address: *0x040010, Bits 07-00*
PCI Address: *BAR2:0x100040, Bits 07-00*
Mode of Access: *Mixed*

This register allows for querying the current state of the hardware FIFO flags and a means to program the FIFO.

BIT 07	BIT 06	BIT 05	BIT 04	BIT 03	BIT 02	BIT 01	BIT 00
FIFO FF	FIFO PAF	FIFO PAE	FIFO EF	N/A	N/A	DISABLE PMC INT	FIFO REG

FIFO FF

DSP Address: *0x040010, Bit 07*
PCI Address: *BAR2:0x100040, Bit 07*
Mode of Access: *Read Only*
Set By *Write to **Reset the Hardware FIFO***

This bit, when low (0), indicates that there are 8192 samples stored in the hardware FIFO.

FIFO PAF

DSP Address: *0x040010, Bit 06*
PCI Address: *BAR2:0x100040, Bit 06*
Mode of Access: *Read Only*
Set By *Write to **Reset the Hardware FIFO***

This bit, when low (0), indicates that there is a programmable number of samples stored in the hardware FIFO.

FIFO PAE

DSP Address: *0x040010, Bit 05*
PCI Address: *BAR2:0x100040, Bit 05*
Mode of Access: *Read Only*
Reset By *Write to **Reset the Hardware FIFO***

This bit, when low (0), indicates that there is less than a programmable number of samples stored in the hardware FIFO.

FIFO EF

DSP Address: *0x040010, Bit 04*
PCI Address: *BAR2:0x100040, Bit 04*
Mode of Access: *Read Only*
Reset By *Write to **Reset the Hardware FIFO***

This bit, when low (0), indicates that there are no samples stored in the hardware FIFO.

DISABLE PMC INT

DSP Address: *0x040010, Bit 01*
PCI Address: *BAR2:0x100040, Bit 01*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is high (1), the PMC interrupt is blocked from causing an interrupt to the HOST computer. The DSP can catch the interrupt on its INT2 line, and perform its own handling of the PMC, if desired. When this bit is low (0), the PMC interrupt is passed to the HOST for normal processing.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

FIFO REG

DSP Address: *0x040010, Bit 00*
PCI Address: *BAR2:0x100040, Bit 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This bit, when cleared to low (0), causes writes to **Write the Hardware FIFO** to be placed into the FIFO for testing purposes. Reads of **Read the Hardware FIFO** will access the contents of the FIFO. When the bit is set high (1), writes to and reads of the FIFO will access the FIFO programming registers.

4.2.4 RESET THE HARDWARE FIFO

DSP Address: *0x040018*
PCI Address: *BAR2:0x100060*
Mode of Access: *Write Strobe*

A write to this location will reset the hardware FIFO to empty. It is required to initialize the FIFO before any access is made.

4.2.5 READ THE HARDWARE FIFO

DSP Address: *0x040020, Bits 31-00*
PCI Address: *BAR2:0x100080, Bits 31-00*
Mode of Access: *Read Access*
Reset By *Write to **Reset the Hardware FIFO***

If the **FIFO REG** bit is clear (0), then a read of this location will respond with the oldest pair of samples in the FIFO. If the **FIFO REG** bit is set (1), then a read of this location will read the internal configuration registers of the FIFO.

Data is stored as pairs of 16 bit WORDs in 2's complement format, but must be read as 32 bit DWORDs. The even numbered channels are in the high WORD, and the odd channels in the low WORD.

When the state machine writes to the FIFOs, it writes all 32 conversion results as 16 pairs in channel order (ch2|ch1, ch4|ch3, ch6|ch5, ch8|ch7, ch10|ch9, ch12|ch11, ch14|ch13, ch16|ch15, ch18|ch17, ch20|ch19, ch22|ch21, ch24|ch23, ch26|ch25, ch28|ch27, ch30|ch29, ch32|ch31).

4.2.6 WRITE THE HARDWARE FIFO

DSP Address: *0x040020, Bits 31-00*
PCI Address: *BAR2:0x100080, Bits 31-00*
Mode of Access: *Write Access*
Reset By *Write to **Reset the Hardware FIFO***

If the **FIFO REG** bit is clear (0), then a write to this location will directly add the value to the FIFO for testing purposes. If the **FIFO REG** bit is set (1), then a write to this location will program the internal configuration registers of the FIFO.

The FIFO is actually four 8k x 9 FIFOs from Cypress, model CY7C5251.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.2.7 RESET WATCHDOG TIMER

DSP Address: *0x040029*
PCI Address: *BAR2:0x1000A4*
Mode of Access: *Write Strobe*

The **CPCI-AD320** incorporates a watchdog timer to help ensure that the DSP will always be operating correctly.

Use of the watchdog is optional. By default, and as a result of a CPCI Hardware Reset, a Software Reset, and the Watchdog Reset, the hardware will automatically strobe the watchdog timer to prevent inadvertent RESET. The first time that this location is written, the hardware will enter a mode where automatic strobes stop, and the write to this location resets the watchdog timer.

If this location is then not repeatedly written within a period of approximately 250 – 1000 mS, the watchdog timer will reset the entire **CPCI-AD320**. Since the watchdog RESET will also clear the **Bootmode Register**, the DSP will be forced to boot from the FLASH device.

4.2.8 PGA GAIN REGISTERS

DSP Address: *0x040030 – 0x40033, Bits 07-00*
PCI Address: *BAR2:0x1000C0 – 0x1000CC, Bits 07-00*
Mode of Access: *Write Only*
Reset By *Not Reset by any signal*

There are four to eight PGA gain registers as follows. Channels 17-32 only exist if the optional additional 16 channels (on the daughter card) is present.

HOST Address BAR2:	DSP Address	BITS 07-06	BITS 05-04	BITS 03-02	BITS 01-00
0x1000C0	0x040030	Chan 4	Chan 3	Chan 2	Chan 1
0x1000C4	0x040031	Chan 8	Chan 7	Chan 6	Chan 5
0x1000C8	0x040032	Chan 12	Chan 11	Chan 10	Chan 9
0x1000CC	0x040033	Chan 16	Chan 15	Chan 14	Chan 13
0x1000D0	0x040034	Chan 20	Chan 19	Chan 18	Chan 17
0x1000D4	0x040035	Chan 24	Chan 23	Chan 22	Chan 21
0x1000D8	0x040036	Chan 28	Chan 27	Chan 26	Chan 25
0x1000DC	0x040037	Chan 32	Chan 31	Chan 30	Chan 29

The gains can be selected from the following:

BITS 01-00	Factory Default	Special Order	Special Order
00	x1	x1	x1
01	x2	X2	x10
10	x4	X5	x100
11	x8	X10	X1000

Table 4.4: Programmable Gain Selections

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.2.9 SOFTWARE STROBE FOR MANUAL CLOCKING BY DSP

DSP Address: *0x040038*
PCI Address: *BAR2:0x1000E0*
Mode of Access: *Write Strobe*

A write to this location can potentially trigger a conversion or other event if selected in the **Trigger / Gate Configuration Registers**.

4.2.10 TRIGGER / GATE CONFIGURATION REGISTERS

DSP Address: *0x040040 – 0x40047, Bits 07 – 00*
PCI Address: *BAR2:0x100100 – 0x10011C, Bits 07 – 00*
Mode of Access: *Mixed*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

These registers control the trigger and gate signal routing inside the card, as well as allow for the control and monitoring of the PXI trigger lines. All of the registers are read/write, except for the PXI monitoring bits, which are read only.

HOST Address BAR2:	DSP Address	BITS 07-04	BITS 03-00
0x100100	0x040040	A/D Clock Source	A/D Gate Source
0x100104	0x040041	HOST Interrupt Source	Front Panel /Rear Clock Output Source
0x100108	0x040042	N/A	DSP Interrupt Source

Table 4.5: Trigger / Gate Source Registers

HOST Address BAR2:	DSP Address	BITS 07-00
0x10010C	0x040043	PXI Trigger Direct Output

Table 4.6: PXI Programmable Output Register

HOST Address BAR2:	DSP Address	BIT 07	BITS 06-04	BIT 03	BITS 02-00
0x100110	0x040044	PXI1 Mon	PXI1 Source	PXI0 Mon	PXI0 Source
0x100114	0x040045	PXI3 Mon	PXI3 Source	PXI2 Mon	PXI3 Source
0x100118	0x040046	PXI5 Mon	PXI5 Source	PXI4 Mon	PXI4 Source
0x10011C	0x040047	PXI7 Mon	PXI7 Source	PXI6 Mon	PXI6 Source

Table 4.7: PXI Trigger Source Registers

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Trigger / Gate Sources

DSP Address: *0x040040 – 0x40042, Bits 07 – 00*
PCI Address: *BAR2:0x100100 – 0x100108, Bits 07 – 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

The following signals can be selected from several sources.

- A/D Clock
- A/D Gate
- Front Panel / Rear Clock Output
- HOST Interrupt
- DSP Interrupt

BITS 03-00 BITS 07-04	Source
0x0	Low (0)
0x1	High (1)
0x2	DSP Timer 0
0x3	DSP Timer 1
0x4	Trigger Input (Front Panel / Rear Connector)
0x5	Gate Input (Front Panel/ Rear Connector)
0x6	Strobe Register from HOST
0x7	Strobe Register from DSP
0x8	PXI Trigger 0
0x9	PXI Trigger 1
0xA	PXI Trigger 2
0xB	PXI Trigger 3
0xC	PXI Trigger 4
0xD	PXI Trigger 5
0xE	PXI Trigger 6
0xF	PXI Trigger 7

Table 4.8: Trigger / Gate Sources

PXI Trigger Direct Output

DSP Address: *0x040043, Bits 07 – 00*
PCI Address: *BAR2:0x10010C, Bits 07 – 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

The PXI Trigger can be driven to a specific logic state for gating or other applications.

If the **PXI Trigger Source** is set to PXI Trigger Direct Output, then the associated bit is output to the PXI Trigger line.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

PXI Trigger Monitor

DSP Address: *0x040044 – 0x40047, Bits 07and 03*
PCI Address: *BAR2:0x100110 – 0x10011C, Bits 07and 03*
Mode of Access: *Read Only*

These bits reflect the current state of the PXI Trigger lines.

PXI Trigger Source

DSP Address: *0x040044 – 0x40047, Bits 06-04 and 02-00*
PCI Address: *BAR2:0x100110 – 0x10011C, Bits 06-04 and 02-00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

The PXI Trigger lines can be sourced from several sources, if desired.

BITS 02-00 BITS 06-04	Source
0x0	Line used as input or unused
0x1	PXI Trigger Direct Output
0x2	DSP Timer 0
0x3	DSP Timer 1
0x4	Trigger Input (Front Panel / Rear Connector)
0x5	Gate Input (Front Panel/ Rear Connector)
0x6	Strobe Register from HOST
0x7	Strobe Register from DSP

Table 4.9: PXI Trigger Sources

4.2.11 8530 SERIAL PORTS

The **CPCI-AD320** includes an Z85C30 serial communications controller chip from ZILOG. There are two communications ports on this chip. Port 1 is connected to RS-232 drivers and is available at the front panel and at the rear panel I//O connectors. Port 2 is connected to RS-422 drivers, and is also available at the front panel and the rear panel.

The Z85C30 is clocked at 7.3728 MHz, and allows for a multitude of common frequencies, including 115,200, 57,600, 38,400, and 19,200 baud.

The bootloader and hardware support libraries supplied with the **CPCI-AD320** utilizes the RS-232 port for a console for standard input and output by the DSP. The customer may alternatively wish to write his own software to use this port. Examples are provided in the **Board Support Package**.

There is hardware support for many RS422 and RS485 applications including HDLC, SDLC, and multidrop configurations. Clocking can be provided externally or internally.

The 8530 can interrupt the DSP for servicing. The 8530 is routed to INT2 via the interrupt control logic. See the DSP Interrupt Status Register and the DSP Interrupt Enable Register for details.

See documentation from ZILOG for more details about programming this device.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

HOST Address BAR2:	DSP Address	BITS 07-00
0x100120	0x040048	Control Port for RS-422
0x100124	0x040049	Data Port for RS-422
0x100128	0x04004A	Control Port for RS-232
0x10012C	0x04004B	Data Port for RS-232

Table 4.10: Z85C30 Serial Port

4.2.12 SOFTWARE STROBE FOR MANUAL CLOCKING BY HOST

DSP Address: *0x040050*
PCI Address: *BAR2:0x100140*
Mode of Access: *Write Strobe*

A write to this location can potentially trigger a conversion or other event if selected in the **Trigger / Gate Configuration Registers**.

4.2.13 RPIO SERIAL BUS AND DIGITAL PORT CONFIGURATION

DSP Address: *0x040058*
PCI Address: *BAR2:0x100160*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This register controls the direction of the 16 digital I/O lines, and allows control of the RPIO Serial Peripheral Bus.

BIT 07-06	BIT 05	BIT 04	BIT 03	BIT 02	BIT 01	BIT 00
N/A	PB DIR	PA DIR	SRST	SLD	SCK	SDI

PB DIR

DSP Address: *0x040058, Bit 05*
PCI Address: *BAR2:0x100160, Bit 05*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is clear (0), digital port B is configured as an input. When this bit is set (1), digital port B is configured as an output. There are 10k pull resistors on the digital lines, so the RESET state will be HIGH on the digital lines.

PA DIR

DSP Address: *0x040058, Bit 04*
PCI Address: *BAR2:0x100160, Bit 04*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is clear (0), digital port A is configured as an input. When this bit is set (1), digital port A is configured as an output. There are 10k pull resistors on the digital lines, so the RESET state will be HIGH on the digital lines.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SRST

DSP Address: *0x040058, Bit 03*
PCI Address: *BAR2:0x100160, Bit 03*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This bit directly controls the SRST line to the RPIO devices on the Serial Peripheral Bus.

When this bit is clear (0), the RPIO devices on the Serial Peripheral Bus will be reset. Since this bit is cleared at RESET, the RPIO devices will also be cleared at RESET. When this bit is set (1), the RPIO devices on the Serial Peripheral Bus will be in their normal operating mode.

SLD

DSP Address: *0x040058, Bit 02*
PCI Address: *BAR2:0x100160, Bit 02*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This bit directly controls the SLD line to the RPIO devices on the Serial Peripheral Bus. Note that the SLD line is driven through an inverter.

When this bit is set and then cleared, an active low pulse will appear on the SLD line to the RPIO devices on the Serial Peripheral Bus. Since this bit is cleared on a RESET, and the line is driven through an inverter, the hardware line will be HIGH on RESET.

SCK

DSP Address: *0x040058, Bit 01*
PCI Address: *BAR2:0x100160, Bit 01*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This bit directly controls the SCK line to the RPIO devices on the Serial Peripheral Bus.

When this bit is set and then cleared, an active high pulse will appear on the SCK line to the RPIO devices on the Serial Peripheral Bus. This will cause the first device to latch the data on SDI, and the remaining devices to move the configuration data to the next device.

SDI

DSP Address: *0x040058, Bit 00*
PCI Address: *BAR2:0x100160, Bit 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This bit directly controls the SDI line to the RPIO devices on the Serial Peripheral Bus.

This bit represents the next data bit to be clocked along the Serial Peripheral Bus on the next clock edge.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.2.14 DIGITAL PORTS A AND B

DSP Address: *0x040060, Bits 15-00*
PCI Address: *BAR2:0x100180, Bits 15-00*
Mode of Access: *Read/Write*
Reset By *Not Reset by any signal*

When written, this register sets the output state for the two digital ports for when they are in output mode.

When read, this register will report the current state of the digital lines.

When initializing the digital ports for output, be sure to program any values to be output before setting the direction bits in **RPIO Serial Bus and Digital Port Configuration** register.

There are 10k pull resistors on the digital lines, so the RESET state will be HIGH on the digital lines.

BITS 15 – 08	BITS 07 – 00
Digital Port B	Digital Port A

4.2.15 DSP INTERRUPT STATUS REGISTER

DSP Address: *0x040080, Bits 07-00*
PCI Address: *BAR2:0x100200, Bits 07-00*
Mode of Access: *Mixed Access*

This register allows the DSP interrupt routine for INT2 to determine which of several sources are causing an interrupt. Additionally, it allows for determination of the current state of the HOST and DSP trigger interrupts.

BIT 07	BIT 06	BIT 05	BIT 04	BIT 03	BIT 02	BIT 01	BIT 00
N/A	HOST TRIGGER	DSP TRIGGER	INT2 ACTIVE	EXT INT	PMC INT	N/A	8530 INT

HOST TRIGGER

DSP Address: *0x040080, Bit 06*
PCI Address: *BAR2:0x100200, Bit 06*
Mode of Access: *Read Only*

This bit monitors the current state of the HOST TRIGGER selected in the **Trigger / Gate Configuration Registers**. A rising edge on HOST TRIGGER will generate an interrupt to the HOST processor.

DSP TRIGGER

DSP Address: *0x040080, Bit 05*
PCI Address: *BAR2:0x100200, Bit 05*
Mode of Access: *Read Only*

This bit monitors the current state of the DSP TRIGGER selected in the **Trigger / Gate Configuration Registers**. A rising edge on DSP TRIGGER will generate INT1 to the DSP processor.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

INT2 ACTIVE
DSP Address: *0x040080, Bit 04*
PCI Address: *BAR2:0x100200, Bit 04*
Mode of Access: *Read Only*
When this bit is high (1), it indicates that one of the enabled sources for INT2 is currently generating an interrupt. Since the interrupt logic on the DSP is edge triggered, and a falling edge on this combined interrupt is required to interrupt the DSP, this line makes it possible in the DSP interrupt handler to ensure that all the causes have been handled, and that the line is de-asserted before returning. The line is de-asserted by writing a 1 to the active interrupt source in the lower 4 bits of this register, or by clearing the cause in the appropriate device.

EXT INT
DSP Address: *0x040080, Bit 03*
PCI Address: *BAR2:0x100200, Bit 03*
Mode of Access: *Read / Write 1 to set*
When this bit is read as low (0), it indicates that a falling edge has been detected on the external interrupt line located on the rear panel I/O connector. It is cleared by writing a 1 to this bit. This will reset the bit to set (1), and prepare the edge detector for the next interrupt.

If this interrupt is enabled, and this bit is low, then **INT2 ACTIVE** will also be low. Be sure at the end of the interrupt routine that all the active interrupts have been acknowledged by writing 1's to this status register, or by clearing the interrupt in the device.

PMC INT
DSP Address: *0x040080, Bit 02*
PCI Address: *BAR2:0x100200, Bit 02*
Mode of Access: *Read Only*
When this bit is low (0), it indicates that the PMC module is currently asserting its interrupt line. Since the interrupt logic on the DSP is edge triggered, and a falling edge on the PMC interrupt is required to interrupt the DSP, this bit makes it possible in the DSP interrupt handler to ensure that the line is de-asserted before returning.

If this interrupt is enabled, and this bit is low, then **INT2 ACTIVE** will also be low. Be sure at the end of the interrupt routine that all the active interrupts have been acknowledged by writing 1's to this status register, or by clearing the interrupt in the device.

8530 INT
DSP Address: *0x040080, Bit 00*
PCI Address: *BAR2:0x100200, Bit 00*
Mode of Access: *Read Only*
When this bit is read as low (0), it indicates that the 8530 serial controller chip is asserting its interrupt line. It is cleared by performing the appropriate action to the 8530 controller chip.

If this interrupt is enabled, and this bit is low, then **INT2 ACTIVE** will also be low. Be sure at the end of the interrupt routine that all the active interrupts have been

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

acknowledged by writing 1's to this status register, or by clearing the interrupt in the device.

4.2.16 DSP INTERRUPT ENABLE REGISTER

DSP Address: *0x040081, Bits 07-00*
PCI Address: *BAR2:0x100204, Bits 07-00*
Mode of Access: *Mixed*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This register allows the DSP to select which interrupts should be allowed to generate INT2 on the DSP. Additionally, it allows for determination of the type of option board installed.

BIT 07	BIT 06-04	BIT 03	BIT 02	BIT 01	BIT 00
N/A	OPTION SENSE	EXT INT EN	PMC INT EN	N/A	8530 INT EN

OPTION SENSE

DSP Address: *0x040081, Bits 06-04*
PCI Address: *BAR2:0x100204, Bits 06-04*
Mode of Access: *Read Only*

These three bits allow a determination of the type of option daughter-board installed on the main card. As of this time, the following options are defined.

BITS 06-04	OPTION
0x7	No option present (total of 16 analog channels).
0x6	Additional 16 channel option for CPCI-AD320 (total of 32 analog channels).
0x5	Option for CPCI-SERVO8 (total of 16 analog input channels, 8 analog output channels, other special logic)

EXT INT EN

DSP Address: *0x040081, Bit 03*
PCI Address: *BAR2:0x100204, Bit 03*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), a falling edge on the external interrupt line located on the rear panel I/O connector will cause the DSP to see INT2. When this bit is clear (0), the external interrupt line is ignored, and any pending interrupt is cleared.

PMC INT EN

DSP Address: *0x040081, Bit 02*
PCI Address: *BAR2:0x100204, Bit 02*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), a falling edge on the PMC interrupt line will cause the DSP to see INT2. When this bit is clear (0), the PMC interrupt line is ignored.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

8530 INT EN

DSP Address: *0x040081, Bit 00*
PCI Address: *BAR2:0x100204, Bit 00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), a falling edge on the 8530 serial controller chip interrupt line will cause the DSP to see INT2. When this bit is clear (0), the 8530 interrupt line is ignored.

4.2.17 DSP BOOTMODE REGISTER

DSP Address: *0x040082, Bits 03-00*
PCI Address: *BAR2:0x100208, Bits 03-00*
Mode of Access: *Read/Write*
Reset By *Watchdog Reset*

This register allows for selecting several sources for booting the DSP. The choices are as follows.

- Booting a ROM image from FLASH device at 0x900000.
- Booting a ROM image from SRAM at 0x1000. This requires that the DSP be held in RESET, using the mechanism described in the RESET section, and an external PCI device copy the ROM image to the SRAM.
- Directly executing an executable image in SRAM at the reset vector located at 0x0. This requires that the DSP be held in RESET, using the mechanism described in the RESET section, and an external PCI device copy the executable image to the SRAM.

BIT 03	BIT 02	BIT 01	BIT 00
N/A	NB	BOOT RAM	MC_MP

NB

DSP Address: *0x040082, Bit 02*
PCI Address: *BAR2:0x100208, Bit 02*
Mode of Access: *Read/Write*
Reset By *Watchdog Reset*

This bit directly controls the /NB line to the PLX, through an inverter. When it is LOW, after a PCI reset the PCI devices will be issued a retry until the DSP sets the Local Init Status bit in register CNTRL in the PLX. If this bit is HIGH, PCI accesses will be permitted immediately.

BOOT RAM

DSP Address: *0x040082, Bit 01*
PCI Address: *BAR2:0x100208, Bit 01*
Mode of Access: *Read/Write*
Reset By *Watchdog Reset*

When this bit is set (1), and the **MC_MP** bit is clear (0), the DSP will boot from a RESET into a ROM image placed in SRAM by an external PCI agent.

When this bit is clear (0), and the **MC_MP** bit is clear (0), the DSP will boot from a RESET into a ROM image placed in FLASH.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

MC_MP
DSP Address: *0x040082, Bit 00*
PCI Address: *BAR2:0x100208, Bit 00*
Mode of Access: *Read/Write*
Reset By *Watchdog Reset*
When this bit is set (1), the DSP will immediately begin executing an executable image in SRAM at the reset vector located at 0x0, once it is taken from RESET. The executable image must be placed in RAM by an external PCI agent.

When this bit is clear (0), the DSP will boot a ROM image from either the FLASH or the RAM.

Note that this bit is reversed from the meaning in the DSP processor manuals.

4.2.18 NULL PACE / LIMIT REGISTER

DSP Address: *0x040083, Bits 07-00*
PCI Address: *BAR2:0x10020C, Bits 07-00*
Mode of Access: *Mixed*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*
The Clock In and Gate In lines on the front panel have a second function on other configurations of this hardware, which are not applicable to the **CPCI-AD320**. Therefore, to preserve functionality, some of these bits must remain 0.

This register allows for control and monitoring of the NULL PACE and LIMIT lines, and control of the USER LED.

BIT 07-06	BIT 05	BIT 04	BIT 03	BIT 02	BIT 01	BIT 00
N/A	CLOCK IN	GATE IN	N/A	USER LED	KEEP AS 0	KEEP AS 0

CLOCK IN
DSP Address: *0x040083, Bit 05*
PCI Address: *BAR2:0x10020C, Bit 05*
Mode of Access: *Read Only*
This bit represents the current state of the CLOCK IN line from the front panel or the rear panel I/O. It will read the current state of the line.

GATE IN
DSP Address: *0x040083, Bit 04*
PCI Address: *BAR2:0x10020C, Bit 04*
Mode of Access: *Read Only*
This bit represents the current state of the GATE IN line from the front panel or the rear panel I/O. It will read the current state of the line.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

USER LED

DSP Address: *0x040083, Bit 02*
PCI Address: *BAR2:0x10020C, Bit 02*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), the USER LED will be illuminated. When this bit is clear (0), the USER LED will extinguish.

KEEP AS 0

DSP Address: *0x040083, Bits 01-00*
PCI Address: *BAR2:0x10020C, Bits 01-00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

Keep these bits set to 0 to maintain full functionality.

4.2.19 DEMAND DMA CONTROL REGISTER

DSP Address: *0x040085, Bits 07-00*
PCI Address: *BAR2:0x100214, Bits 07-00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

The two DMA engines in the PLX can operate in a mode where they are throttled by external hardware signals. There is a state machine on the card to transfer a fixed number of DWORDs based upon a trigger signal.

DMA engine 0 is triggered by a programmable clock source which includes the PXI trigger lines and the DSP timers.

DMA engine 1 is triggered by the PAE flag from the hardware FIFO indicating that the programmed number of samples are available.

BITS 07-06	BIT 05	BIT 04	BITS 03-00
N/A	DMA 1 Trigger Enable	DMA 0 Trigger Enable	DMA 0 Trigger Source

DMA 1 Trigger Enable

DSP Address: *0x040085, Bit 05*
PCI Address: *BAR2:0x100214, Bit 05*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), the DMA hardware will be enabled. Whenever PAE is asserted indicating that there are more than the programmed number of samples stored in the hardware FIFO, this state machine will request that the number of DWORDS specified in **Demand DMA 1 Transfer Count Register** will be read by the PLX DMA engine 1. Presumably, the PLX will have been set up for demand mode.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

DMA 0 Trigger Enable

DSP Address: *0x040085, Bit 04*
PCI Address: *BAR2:0x100214, Bit 04*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

When this bit is set (1), the DMA hardware will be enabled. Whenever a rising edge of the selected trigger source in **DMA 0 Trigger Source** is detected, this state machine will request that the number of DWORDS specified in **Demand DMA 0 Transfer Count Register** will be read by the PLX DMA engine 0. Presumably, the PLX will have been set up for demand mode.

DMA 0 Trigger Source

DSP Address: *0x040085, Bits 03-00*
PCI Address: *BAR2:0x100214, Bits 03-00*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This register selects the trigger source for the PLX DMA engine 0.

BITS 03-00	Source
0x0	Low (0)
0x1	High (1)
0x2	DSP Timer 0
0x3	DSP Timer 1
0x4	Trigger Input (Front Panel / Rear Connector)
0x5	Gate Input (Front Panel / Rear Connector)
0x6	Strobe Register from HOST
0x7	Strobe Register from DSP
0x8	PXI Trigger 0
0x9	PXI Trigger 1
0xA	PXI Trigger 2
0xB	PXI Trigger 3
0xC	PXI Trigger 4
0xD	PXI Trigger 5
0xE	PXI Trigger 6
0xF	PXI Trigger 7

Table 4.11: Trigger / Gate Sources

4.2.20 DEMAND DMA 0 TRANSFER COUNT REGISTER

DSP Address: *0x040086*
PCI Address: *BAR2:0x100218*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This 8 bit register sets the number of DWORDs to be transferred by the PLX DMA engine 0 when it is operating in demand mode.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

The value programmed is one less than the desired number to transfer. Use 0 to transfer 1 DWORD, and 7 to transfer 8.

4.2.21 DEMAND DMA 1 TRANSFER COUNT REGISTER

DSP Address: *0x040087*
PCI Address: *BAR2:0x10021C*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

This 8 bit register sets the number of DWORDs to be transferred by the PLX DMA engine 1 when it is operating in demand mode. It should be set to the same value as the PAE flag in the hardware FIFO.

The value programmed is one less than the desired number to transfer. Use 0 to transfer 1 DWORD, and 7 to transfer 8.

4.3 LOCAL FLASH

DSP Address: *0x900000 - 0x97FFFF, Bits 07-00*
PCI Address: *Not Accessible*
Mode of Access: *Read/Write in page mode*

There is 512k x 8 of FLASH memory available to the DSP on the Local Bus. The first 64k is reserved for the use of the DSP Bootloader with the remainder available for customer applications. Access occurs with about 10 wait states due to the slow speed of the device.

Code examples of writing to this device is located in the Board Support Package.

4.4 PLX OPERATION REGISTERS

DSP Address: *0x880000 – 0x88005E*
PCI Address: *BAR0/1:0x0 - 0xF8*
Mode of Access: *Read/Write*
Reset By *Depends upon register*

The HOST processor can access the internal PLX registers through BAR0 and BAR1. BAR0 is accessed via normal memory operations and BAR1 is accessed via I/O operations.

The DSP can access these registers through normal memory cycles at the addresses specified.

The registers can be divided into several categories.

- Local Configuration Registers
- Runtime Registers
- DMA Registers

There are insufficient resources on the card to support I_2O message queues, so these registers are not described.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.4.1 LOCAL CONFIGURATION REGISTERS

DSP Address: *0x880020 – 0x88005E*
PCI Address: *BAR0/1:0x0 – 0xF8*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

See the PLX PCI9080 Documentation for more details about these registers.

HOST Address BAR0/1:	DSP Address	31 – 24	23 – 16	15 – 8	7 – 0
0x00	0x880020	Range for PCI to Local Address Space 0			
0x04	0x880021	Local Base Address for PCI to Local Address Space 0			
0x08	0x880022	Mode / Arbitration Register			
0x0C	0x880023	Big / Little Endian Register			
0x10	0x880024	Range for PCI to Local Expansion ROM			
0x14	0x880025	Local Base Address for PCI to Local Expansion ROM			
0x18	0x880026	Local Bus Region Descriptors			
0x1C	0x880027	Range for Direct Master to PCI			
0x20	0x880028	Local Base Address for Direct Master to PCI Memory			
0x24	0x880029	Local Base Address for Direct Master to PCI IO/CFG			
0x28	0x88002A	PCI Base Address for Direct Master to PCI			
0x2C	0x88002B	PCI Configuration Address for Direct Master to PCI IO/CFG			
0xF0	0x88005C	Range for PCI to Local Address Space 1			
0xF4	0x88005D	Local Base Address for PCI to Local Address Space 1			
0xF8	0x88005E	Local Bus Region Descriptor for PCI to Local			

Table 4.12: PLX Local Configuration Registers

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.4.2 RUNTIME REGISTERS

DSP Address: *0x880030 – 0x88003F*
PCI Address: *BAR0/1:0x40 – 0x7C*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset*

See the PLX PCI9080 Documentation for more details about these registers.

HOST Address BAR0/1:	DSP Address	31 – 24	23 – 16	15 – 8	7 – 0
0x40	0x880030	Mailbox Register 0			
0x44	0x880031	Mailbox Register 1			
0x48	0x880032	Mailbox Register 2			
0x4C	0x880033	Mailbox Register 3			
0x50	0x880034	Mailbox Register 4			
0x54	0x880035	Mailbox Register 5			
0x58	0x880036	Mailbox Register 6			
0x5C	0x880037	Mailbox Register 7			
0x60	0x880038	PCI to Local Doorbell Register			
0x64	0x880039	Local to PCI Doorbell Register			
0x68	0x88003A	Interrupt Control / Status			
0x6C	0x88003B	Serial EEPROM, PCI Command Codes, User I/O, Init			
0x70	0x88003C	Device ID		Vendor ID	
0x74	0x88003D	Unused		Revision ID	
0x78	0x88003E	Mailbox Register 0			
0x7C	0x88003F	Mailbox Register 1			

Table 4.13: PLX Runtime Registers

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.4.3 DMA REGISTERS

DSP Address: *0x880040 – 0x88004C*
PCI Address: *BAR0/1:0x80 – 0xB0*
Mode of Access: *Read/Write*
Reset By *CPCI Hardware Reset, Software Reset, Watchdog Reset*

See the PLX PCI9080 Documentation for more details about these registers.

HOST Address BAR0/1:	DSP Address	31 – 24	23 – 16	15 – 8	7 – 0
0x80	0x880040	DMA Ch0 Mode			
0x84	0x880041	DMA Ch0 PCI Address			
0x88	0x880042	DMA Ch0 Local Address			
0x8C	0x880043	DMA Ch0 Transfer Byte Count			
0x90	0x880044	DMA Ch0 Descriptor Pointer			
0x94	0x880045	DMA Ch1 Mode			
0x98	0x880046	DMA Ch1 PCI Address			
0x9C	0x880047	DMA Ch1 Local Address			
0xA0	0x880048	DMA Ch1 Transfer Byte Count			
0xA4	0x880049	DMA Ch1 Descriptor Pointer			
0xA8	0x88004A	Reserved		DMA Ch1 CSR	DMA Ch0 CSR
0xAC	0x88004B	Mode / Arbitration Register			
0xB0	0x88004C	DMA Threshold Register			

Table 4.14: PLX DMA Registers

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

4.5 CPCI PASS-THROUGH REGION

DSP Address: *0xC00000 – 0xFFFFFF*
Mode of Access: *Read/Write*

The **CPCI-AD320** can directly access HOST memory and I/O ports on the CPCI system and the PMC Module through correctly programming the PLX interface. Up to 16 Mbytes of PCI address space can be mapped at a time.

NOTE: Windows NT/2000 and Windows 95/98 supports virtual memory and separate flat address spaces for each task. Physical memory pages are not contiguous.

There is support in the Board Support Package and alphi_io DSP library to allow the DSP to directly access HOST memory on a per task basis.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5. HARDWARE DETAILS

5.1 BOARD PERIPHERALS

5.1.1 A/D CONVERTERS




Figure 5.1: Main A/D Converters

Each of the 16-32 main input channels (AIN01 – AIN32) has its own Programmable Gain Amplifier (PGA). Inputs are differential mode, and the gains can be selected under DSP control. Depending on factory input option, the gains are x1, x2, x4, and x8, x1, x2, x5, and x10, or optionally x1, x10, x100, and x1000. Individual channels can have different input options. Note that the X10, X100, and X1000 option will greatly reduce the input bandwidth, depending upon the gain selected. Contact the factory for more details.

16 channels reside on the main board, and an additional 16 channels reside on a daughterboard. The daughterboard is optional.

Upon receipt of a rising edge on the trigger when gate is high, the A/D converters simultaneously sample the output voltages from the PGAs and start a conversion. When the conversions are complete, a hardware state machine copies the conversion results to a hardware FIFO.

When a programmable number of samples are available in the FIFO, either the DSP can transfer them to the on board SRAM for use by the control loop, or the PLX CPCI Interface can DMA the conversion results immediately to the processor(s) on the PMC slot, or send them directly to HOST memory.

The DSP code and NT driver, provided in the Board Support Package, supports full scatter / gather type DMA to the HOST.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5.1.2 DIGITAL I/O

There are two 8 bit bi-directional digital ports. Each port can be set for input or output. There are 10k Ohm resistors serving as pull-ups to ensure that the lines will default to high on reset.

5.1.3 SERIAL PERIPHERAL CHAIN

The **CPCI-AD320** is designed to interface with a series of rear panel I/O modules which incorporate peripherals which are programmed using a serial bus. This bus is a chained serial bus which connect all the modules and daughter-boards in sequence. Each module and daughter may have different peripherals and the documentation for each module and daughter-board should be consulted for the overall programming format.

The software for this card will include functions to simplify the programming of this and any other RPIO Modules. See the software manuals for more details.

Details about the programming the devices on the RPIO modules is provided for reference.




Figure 5.2: Serial Peripherals

Each peripheral on the serial bus consists of four signals and five connections. These signals originate using a digital port at the A/D Acquisition card. The four signals (five connections) are as follows:

SDI Serial Data In

Serial stream of program information. Originated by A/D acquisition card, and possibly clocked through earlier devices in the chain.

SDO Serial Data Out

Serial stream of data which is clocked through the device to the next device in the chain. The number of bits between SDI and SDO is device dependent, and documented below.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SCK Serial Clock
This signal clocks the data along the SDI/SDO chain of shift registers. It is suggested that both a rising edge and falling edge be present on SCK before any changes to the SDI for the next bit. The SCK is not toggled if peripherals are not being programmed.

SLD Serial Load
A falling edge followed by a rising edge on this signal will transfer the contents of the shift registers on all devices to the holding registers on the devices, thus programming the device. All devices on the chain must be programmed at the same time.

SCLR Serial Clear
A falling edge followed by a rising edge on this signal will clear the holding registers to a preset value, depending upon the device. The preset value depends upon the device, and some devices may not support it.

5.2 DMA

There are a total of four DMA channels available on the **CPCI-AD320**. Two are located in the TMS320C32 DSP and two are located in the PLX 9080 PCI to DSP bridge.

5.2.1 DSP DMA

There are two fully independent DMA channels in the DSP which can be used for such purposes as:

- Transferring the main A/D conversion results from hardware FIFO to SRAM memory.
- Performing transfers between the DSP local bus and the PCI, with a slight performance penalty, if the PCI window is set up in the PLX.

These DMA engines can be free-running, or they can be triggered from several DSP interrupt sources, such as:

- FIFO has a programmable number of samples ready. (DSP INT3)
- External trigger event, such as PXI trigger, external PCI device writing to a strobe register. (DSP INT1, source selected in the **Trigger / Gate Configuration Registers**)
- DSP timer causing a trigger. (Internal to DSP)

See the DSP Processor documentation for more details.

5.2.2 DSP TO PCI BRIDGE DMA

The PLX 9080 DSP to PCI bridge contains two DMA engines for transferring between the DSP local bus and the PCI bus, either to the PMC or to the CPCI bus. These DMA engines are the most efficient means of transferring data between the two busses.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

The DMA engines are capable of processing a chained list of transfer requests, and can generate interrupts to the DSP or to the HOST processor with each transfer. Scatter/Gather is supported with these chains of requests. The list of requests can reside in PCI memory or in SRAM on the card.

5.2.3 DEMAND MODE DMA

Additional logic on the card allow for control of the PLX DMA engines in demand mode. A programmed number of transfers can occur for each hardware request.

Demand mode is independent of the chains of requests. It is of use when scatter/gather modes are being used.

The DMA hardware is pre-assigned for certain tasks:

PLX DMA Engine 1 is intended to transfer the contents of the hardware FIFO to the PCI bus with a minimum of latency. When the PAE flag in the hardware FIFO indicates that more than a certain number of samples are available for transfer, the DMA will trigger, transferring that certain number. Profiling indicates that the time between data available in the FIFO to the request of the PCI bus for access will occur in 1 μS.

PLX DMA Engine 0 is intended to be triggered by an external hardware event. This event could be a PXI trigger line, an external PCI device writing to a strobe register, or a DSP timer. A programmed number of samples is transferred for each trigger.

Of course, the demand modes do not have to be used, such as for programmed DMA transfers.

5.3 PXI

In 1997, National Instruments Corporation introduced an extension of the CPCI specification called PXI. The intent was to port some of the features of the VXI triggering bus to the CPCI realm.

One of the features of the PXI specification is a bus of 8 trigger lines which are common to all of the slots on a PXI compatible backplane.

By design or by coincidence, 6 of the 8 trigger lines defined by the PXI specification were already reserved and specified to be tied in common on the backplane, as per the CPCI specification.

Therefore, on a correct implementation of a CPCI compatible backplane, but non PXI compatible backplane, at least the first 6 trigger lines should be available for triggering between cards.

The **CPCI-AD320** is designed to make exhaustive use of the PXI trigger lines for communication and synchronization between other **CPCI-AD320** boards and other PXI compatible boards.

The **CPCI-AD320** is capable of driving the following signals to a PXI line:

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

- A constant 0 or 1, or a signal created by the DSP or PCI device toggling the bit which is output to the PXI trigger line.
- A timing source generated by either of the DSP timers.
- Either the Clock Input or the Gate Input from the front panel connector or through the rear panel I/O connector.
- A pulse generated by a write to a register by either the DSP or another PCI device.

Of course, the trigger output can be tri-stated when it is not used or is for input.

The **CPCI-AD320** is capable of performing the following actions based on a PXI trigger:

- Provide a trigger and a gate to the A/D conversion hardware.
- Outputting the signal to the Clock Output at the front panel connector or through the rear panel I/O connector.
- Provide an interrupt to the HOST.
- Provide an interrupt to the DSP.
- Provide a trigger of the DSP DMA engines.
- Provide a trigger of one of the PLX DMA engines.
- Can be read by the DSP or PCI processor to determine its current state.

5.4 INTERRUPTS

5.4.1 DSP INTERRUPTS

The TMS320C32 DSP processor has four external interrupt sources. They are called INT0 – INT3. The sources of these interrupts are described below.

INT0
INT0 is dedicated to handling PLX9080 issues. This assertion of either LINTO# or LSERR# by the PLX will generate INT0. The DSP can query the **9080 Control / Status Register** bits **LINTO** and **LSERR** to determine the cause of the interrupt, and determine that both signals are HIGH before returning from the interrupt handler.

Individual bits in the PLX will enable the following sources:

- DMA Completion. (**LINTO**)
- DMA Terminal Count Reached. (**LINTO**)
- One or more bits in Doorbell register are set. (**LINTO**)
- PCI write to a Mailbox. (**LINTO**)
- Master Abort on PCI bus. (**LSERR**)
- Target Abort on PCI bus. (**LSERR**)
- Retry count exceeded on PCI bus. (**LSERR**)
- Parity Error on PCI bus. (**LSERR**)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

INT1
INT1 is dedicated to handling an interrupt from the source selected in the **Trigger / Gate Configuration Registers**. The following sources can be connected to this interrupt:

- Rising edge on the Clock Input or the Gate Input from the front panel connector or through the rear panel I/O connector.
- Rising edge on a PXI trigger line.
- External PCI device writing to one of the strobe registers.

This interrupt line can be used to trigger a DSP DMA engine, or be handled as an interrupt, of both.

INT2
INT2 is a catchall for the other available interrupt sources. Interrupts can be generated by the following:

- 8530 Serial Interface.
- PMC Module's PCI Interrupt.
- A falling edge on the EXT INT line from the rear panel I/O.

Interrupt enables and status bits are available in the **DSP Interrupt Status Register** and the **DSP Interrupt Enable Register**.

Note that the PMC interrupt can be turned off from interrupting the HOST by setting the **DISABLE PMC INT** bit in the **Board Control / Status Register**. Either the HOST or the DSP should handle the PMC interrupt.

INT3
INT3 is dedicated to handling the A/D hardware FIFO. An interrupt is generated when the PAE line from the FIFO indicates that there are a programmable number of samples available. The FIFO will need to be emptied to reset this interrupt.

This interrupt line can be used to trigger a DSP DMA engine, or be handled as an interrupt, of both.

5.4.2 HOST INTERRUPTS

The PLX can interrupt the CPCI HOST from any of the following causes.

- DMA Completion.
- DMA Terminal Count Reached.
- One or more bits in Doorbell register are set.
- Master Abort on PCI bus.
- Target Abort on PCI bus.
- Retry count exceeded on PCI bus.
- Parity Error on PCI bus.

Individual bits in the PLX will enable the sources. Additionally, the following hardware source can generate a CPCI HOST interrupt.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

- HOST interrupt source selected in the Trigger / Gate Configuration Registers.

5.5 RESET

The **CPCI-AD320** is reset by several different means.

- PCI Hardware RESET
- Software RESET from a PCI device
- Watchdog RESET

Additionally, it is possible to hold the DSP in reset after a software reset, to allow for loading DSP code, as well as performing any other operation to the card hardware.

5.5.1 PCI HARDWARE RESET

The entire card is reset when the PCI reset line is held low by the CPCI bus. All of the card's registers are reset as described in this manual, and the D/A converters will output 0 Volts. The PLX is fully reset, and the NVRAM image is automatically read by the PLX.

All access by the CPCI bus is locked out until either the DSP sets the **Local Init Status** bit in the **CNTRL** register of the PLX, or the DSP sets the **NB** bit in the **DSP Bootmode Register**.

The DSP will boot into the mode selected in the **DSP Bootmode Register**.

5.5.2 SOFTWARE RESET FROM A PCI DEVICE

A PCI device can reset the DSP, the card registers, and partially reset the PLX. Additionally, the DSP can be held in RESET, after the software is complete.

The reset is accomplished by setting the **PCI Adapter Software Reset** bit in the PLX **CNTRL** register to a 1. After 200 mS, this bit needs to be set to a 0.

The DSP will be forced into reset, the card registers will reset, the D/A converters will output 0 Volts, and the PLX Local Configuration and PLX DMA registers will be reset. This will stop any active DMA, and will eliminate the passthru region BAR2. Any access to BAR2 by the PCI bus will not be acknowledged. Behavior of accessing a region which is not acknowledged is machine dependent.

After the reset is complete, one of two things needs to occur. One is that the PCI device toggle the **Reload Configuration Registers** bit in the PLX **CNTRL** register. This forces the PLX to reload from the NVRAM, thus re-enabling the passthru region. The other possibility is that the DSP is free to run, which will reload the PLX registers from the NVRAM.

5.5.3 HOLDING DSP IN RESET

If it is desired to hold the DSP in reset, the **General Purpose Output** bit in the **CNTRL** register of the PLX should be set to 0 at the same time as the **PCI Adapter Software Reset** is set high. Leave the **General Purpose Output** bit low for as long as it is desired to hold the DSP.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

The DSP boot mode can be set in the **DSP Bootmode Register**.

5.5.4 WATCHDOG RESET

The **CPCI-AD320** incorporates a watchdog timer, which can optionally be used.

At any RESET cause, the hardware will automatically generate reset strobes to the watchdog timer, to prevent the board from being reset.

If the DSP software desires to make use of the watchdog timer, it makes periodic writes to the **Reset Watchdog Timer** register. The first write will set the hardware into a mode which requires periodic reset of the timer with further writes to **Reset Watchdog Timer**.

In the event that the reset has not occurred within between 600 and 1000 mS, the Watchdog timer will force a reset.

The DSP will be forced into reset, the card registers will reset, the D/A converters (if present) will output 0 Volts, digital lines will tri-state and pull high, and the PLX Local Configuration and PLX DMA registers will be reset. This will stop any active DMA, and will eliminate the passthru region BAR2. Any access to BAR2 by the PCI bus will not be acknowledged. Behavior of accessing a region which is not acknowledged is machine dependent.

This reset will also reset the **DSP Bootmode Register**, which will force the DSP to boot from the FLASH device.

After the reset is complete, the DSP is free to run, which will reload the PLX registers from the NVRAM. This will re-enable the passthru region at BAR2.

5.5.5 DSP BOOT MODES

The DSP can be booted from several sources, once it is allowed to run. The choices are as follows.

- Booting a ROM image from FLASH device at 0x900000.
- Booting a ROM image from SRAM at 0x1000. This requires that the DSP be held in RESET, using the mechanism described in the RESET section, and an external PCI device copy the ROM image to the SRAM.
- Directly executing an executable image in SRAM at the reset vector located at 0x0. This requires that the DSP be held in RESET, using the mechanism described in the RESET section, and an external PCI device copy the executable image to the SRAM.

Determination of the boot source is made depending upon the settings in the **DSP Bootmode Register**.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5.6 CONNECTORS, JUMPERS, AND LEDS

The jumper and connector placement is depicted below.




Figure 5.1: Jumper and Connector Locations

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5.6.1 JUMPER DESCRIPTIONS

JUMPER	FACTORY SETTING	DESCRIPTION
W1	None	Does not exist.
W2	1-2	Factory use. Always connect 1-2.
W3	None	Factory use. Never connect.
W4	None	Not populated in this configuration.
W5	None	Not populated in this configuration.
W6	2-3	Select output to RS422 pair 6 and 7. 1-2 to output TXCLK generated by 8530. 2-3 to output RTS from the 8530. RTS is default.
W7	2-3	Select input from RS422 pair 8 and 9. 1-2 to connect to receive clock of 8530; 2-3 to connect to CTS of 8530. CTS is default.
W8	2-3	Factory use. Always connect 2-3.
W9	None	RS422 Driver Mode. If present, drive both outputs on RTS; otherwise drive always. Drive always is default.
W10	2-3	Provides an identical RESET to that of the watchdog timer.
W11	1-2	Connect 1-2 in this configuration.
W12	None	W12 is used at the factory to supply 3.3 volts to certain components when a CPCI bus is not present. It should not be installed when the card is in a CPCI backplane.

Table 5.15 Jumper Descriptions

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5.6.2 FRONT PANEL DESCRIPTION

The following picture shows the front panel.




Figure 5.2: Front Panel

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Analog Input Connector (P7)

A pair of 50 pin ribbon-style connectors are used to connect the analog input signals to this card. ALPHI Technology can supply transition modules and cable assemblies to meet any special requirements. Contact the factory for additional information.

Pins are numbered according to the following diagram. To simplify the description of the connections, P7 is split into two parts, P7LEFT, and P7RIGHT to represent the left and right sides of the connector.




Figure 5.3: Analog Input Connector (P7)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Pin	Connection	Pin	Connection
50	AGND	49	AGND
48	AIN01+	47	AIN01-
46	AIN02+	45	AIN02-
44	AGND	43	AGND
42	AIN03+	41	AIN03-
40	AIN04+	39	AIN04-
38	AGND	37	AGND
36	AIN05+	35	AIN05-
34	AIN06+	33	AIN06-
32	AGND	31	AGND
30	AIN07+	29	AIN07-
28	AIN08+	27	AIN08-
26	AGND	25	AGND
24	AIN09+	23	AIN09-
22	AIN10+	21	AIN10-
20	AGND	19	AGND
18	AIN11+	17	AIN11-
16	AIN12+	15	AIN12-
14	AGND	13	AGND
12	AIN13+	11	AIN13-
10	AIN14+	9	AIN14-
8	AGND	7	AGND
6	AIN15+	5	AIN15-
4	AIN16+	3	AIN16-
2	AGND	1	AGND

Table 5.16: Analog Input Connector (P7LEFT)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Pin	Connection	Pin	Connection
50	AGND	49	AGND
48	AIN17+	47	AIN17-
46	AIN18+	45	AIN18-
44	AGND	43	AGND
42	AIN19+	41	AIN19-
40	AIN20+	39	AIN20-
38	AGND	37	AGND
36	AIN21+	35	AIN21-
34	AIN22+	33	AIN22-
32	AGND	31	AGND
30	AIN23+	29	AIN23-
28	AIN24+	27	AIN24-
26	AGND	25	AGND
24	AIN25+	23	AIN25-
22	AIN26+	21	AIN26-
20	AGND	19	AGND
18	AIN27+	17	AIN27-
16	AIN28+	15	AIN28-
14	AGND	13	AGND
12	AIN29+	11	AIN29-
10	AIN30+	9	AIN30-
8	AGND	7	AGND
6	AIN31+	5	AIN31-
4	AIN32+	3	AIN32-
2	AGND	1	AGND

Table 5.17: Analog Input Connector (P7RIGHT)

CUTOUT FOR PMC MODULE CONNECTOR

This cutout allows the front panel from the PMC module to show through the front panel. If no PMC module is present, then a filler panel will be installed.

RS-232 INTERFACE CONNECTOR

A 9 pin subminiature D shelled connector is used to route the RS-232 signals off the card. Port A of the 8530 is configured as the RS-232 port, and it serves as the console output for the bootloader and any DSP programs linked with the ALPHI_IO.LIB library discussed in the **DSP Support Library** documentation.

Connectors are manufactured by ITT Cannon.

Use	Model
On PC Board	MDSM-9PE-C10
Suggested Plug	MDSM-9SC-Z11

Table 5.2: Serial Connector Model Numbers

The pinout is described in the table below.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Pin	Description	Pin	Description
1	No Connection	2	Transmit Data
3	Receive Data	4	Clear To Send
5	Ground	6	Request To Send
7	Ground	8	No Connection
9	No Connection		

Table 5.3: Serial RS232 Port (P4)




Figure 5.2: Serial RS232 Port (P4)

RS-422 INTERFACE CONNECTOR

A 9 pin subminiature D shelled connector is used to route the RS-422 signals off the card. Port B of the 8530 is configured as the RS-422 port.

Connectors are manufactured by ITT Cannon.

Use	Model
On PC Board	MDSM-9PE-C10
Suggested Plug	MDSM-9SC-Z11

Table 5.2: Serial Connector Model Numbers

The pinout is described in the table below.

Pin	Description	Pin	Description
1	Transmit Data +	2	Transmit Data-
3	Receive Data +	4	Receive Data -
5	Ground	6	Request To Send +
7	Request To Send -	8	Clear To Send +
9	Clear To Send -		

Table 5.3: Serial RS232 Port (P4)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Figure 5.2: Serial RS232 Port (P4)

CLOCK INPUT
This SMB connector can provide the CLOCK INPUT signal to the card. The input signal is a TTL compatible signal with pull-up resistors. It also can be read by the DSP or a PCI processor. The connector is labeled “IC” on the front panel.

This signal is also available at the rear panel.

GATE INPUT
This SMB connector can provide the GATE INPUT signal to the card. The input signal is a TTL compatible signal with pull-up resistors. It also can be read by the DSP or a PCI processor. The connector is labeled “IC” on the front panel.

This signal is also available at the rear panel.

CLOCK OUTPUT
This SMB connector routes the CLOCK OUTPUT signal off of the card. The output signal is an open drain TTL compatible signal with pull-up resistors. It cannot be read back. The connector is labeled “OC” on the front panel.

This signal is also available at the rear panel.

A/D CONVERT LED
When this LED is glowing, the A/D converters are being triggered. The LED is labeled “CV” on the front panel.

PLX ACCESS BY DSP LED
When this LED is glowing, the DSP is accessing registers on the PLX. The LED is labeled “AC” on the front panel.

ACCESS BY PCI LED
When this LED is glowing, the PLX has been granted access to the Local Bus, usually for access by the PCI bus. The LED is labeled “PC” on the front panel.

USER LED
This LED will glow as a result of setting the **USER LED** bit in the **Null Pace / Limit Register**. The LED is labeled “US” on the front panel.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

5.6.3 REAR PANEL DESCRIPTION

32 Bit CPCI Bus Connection (J1)

This connector plugs into the backplane and provides the standard CPCI signals for all CPCI systems. See the CPCI Bus Documentation for details.

PXI Trigger Extension Bus (J2)

This connector plugs into the backplane and provides the PXI triggering connections for PXI systems. It is compatible with both a PXI backplane and a standard 64 bit CPCI backplane.

PIN	A	B	C	D	E	F
22						GND
21		GND				GND
20						GND
19		GND				GND
18	PXITRG3	PXITRG4	PXITRG5		PXITRG6	GND
17	PXITRG2	GND				GND
16	PXITRG1	PXITRG0			PXITRG7	GND
15		GND				GND
14						GND
13		GND				GND
12						GND
11		GND				GND
10						GND
9		GND				GND
8						GND
7		GND				GND
6						GND
5		GND				GND
4						GND
3		GND				GND
2						GND
1		GND				GND

Table 5.18: PXI Trigger Extension Bus (J2)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

PMC Module Rear Panel I/O Connection (J3)

This connector brings the rear panel I/O connector of the PMC module to the rear panel. The pinout specified is the standard one for a PMC module installed into slot 2 on a 6U carrier board.

PIN	A	B	C	D	E	F
19						GND
18						GND
17						GND
16						GND
15						GND
14	+5	+5	+3.3	+3.3	+3.3	GND
13	PMCIO00	PMCIO01	PMCIO02	PMCIO03	PMCIO04	GND
12	PMCIO05	PMCIO06	PMCIO07	PMCIO08	PMCIO09	GND
11	PMCIO10	PMCIO11	PMCIO12	PMCIO13	PMCIO14	GND
10	PMCIO15	PMCIO16	PMCIO17	PMCIO18	PMCIO19	GND
9	PMCIO20	PMCIO21	PMCIO22	PMCIO23	PMCIO24	GND
8	PMCIO25	PMCIO26	PMCIO27	PMCIO28	PMCIO29	GND
7	PMCIO30	PMCIO31	PMCIO32	PMCIO33	PMCIO34	GND
6	PMCIO35	PMCIO36	PMCIO37	PMCIO38	PMCIO39	GND
5	PMCIO40	PMCIO41	PMCIO42	PMCIO43	PMCIO44	GND
4	PMCIO45	PMCIO46	PMCIO47	PMCIO48	PMCIO49	GND
3	PMCIO50	PMCIO51	PMCIO52	PMCIO53	PMCIO54	GND
2	PMCIO55	PMCIO56	PMCIO57	PMCIO58	PMCIO59	GND
1	PMCIO60	PMCIO61	PMCIO62	PMCIO63	+5	GND

Table 5.19: PMC Module Rear Panel I/O Connection (J3)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Analog Outputs and Digital Connection to Rear Panel I/O Module (J4)

PIN	A	B	C	D	E	F
25			AOUT01	AGND	AOUT02	GND
24	AGND			AOUT03	AGND	GND
23	AOUT04	AGND			AOUT05	GND
22	AGND	AOUT06	AGND			GND
21	AOUT07	AGND	AOUT08	AGND		GND
20						GND
19						GND
18						GND
17						GND
16						GND
15	+3.3	+3.3	+3.3	+5	+5	GND
12 - 14	KEY AREA					
11	RXD0	TXD0	RTS0	CTS0	PA_04	GND
10	PA_06		PA_02	PB_02		GND
9	PB_04	PB_06	PB_00	SCK	SLD	GND
8	EXT INT	FAULT				GND
7	LIMIT	NULL PACE				GND
6		PA_01	PA_03		PA_05	GND
5	PA_07		PB_01	PB_03		GND
4	PB_05	PB_07			SRST	GND
3	TXD1-	RTS1+	RTS1-	CTS1+	CTS1-	GND
2	PA_00	RXD1+	RXD1-	SDI	TXD1+	GND
1	+3.3	+3.3	+3.3	+5	+5	GND

Table 5.20: Analog Outputs and Digital Connection to Rear Panel I/O Module (J4)

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Analog Inputs Connection to Rear Panel I/O Module (J5)
Note that channels 09-16 and channels 17-24 are swapped from the RPIO documentation, in order to match correctly with the RPIO modules.

Channels 01-08 will be located on the Primary RPIO Module.

Channels 09-16 will be located on the Secondary RPIO Module.

Channels 17-24 will be located on the Primary RPIO Module's Daughter Board.

Channels 25-32 will be located on the Secondary RPIO Module's Daughter Board.

PIN	A	B	C	D	E	F
22			AIN17+	AIN01+	AIN18+	GND
21	AIN02+			AIN19+	AIN03+	GND
20	AIN20+	AIN04+			AIN21+	GND
19	AIN05+	AIN22+	AIN06+			GND
18	AIN23+	AIN07+	AIN24+	AIN08+		GND
17		AIN17-	AIN01-	AIN18-	AIN02-	GND
16			AIN19-	AIN03-	AIN20-	GND
15	AIN04-			AIN21-	AIN05-	GND
14	AIN22-	AIN06-			AIN23-	GND
13	AIN07-	AIN24-	AIN08-			GND
12	+3.3	+3.3	+3.3	+5	+5	GND
11			AIN25+	AIN09+	AIN26+	GND
10	AIN10+			AIN27+	AIN11+	GND
9	AIN28+	AIN12+			AIN29+	GND
8	AIN13+	AIN30+	AIN14+			GND
7	AIN31+	AIN15+	AIN32+	AIN16+		GND
6		AIN25-	AIN09-	AIN26-	AIN10-	GND
5			AIN27-	AIN11-	AIN28-	GND
4	AIN12-			AIN29-	AIN13-	GND
3	AIN30-	AIN14-			AIN31-	GND
2	AIN15-	AIN32-	AIN16-			GND
1	+3.3	+3.3	+3.3	+5	+5	GND

Table 5.21: Analog Inputs Connection to Rear Panel I/O Module (J5)

5.6.4 CONNECTORS ON COMPONENT SIDE DESCRIPTION

Emulator Connector (P9)
This connector is used to connect the emulator to the C32 DSP.

Factory Use (P8)
This connector is used at the factory for programming the FPGA.

 Copyright © 1999, ALPHI Technology Corporation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Artisan Technology Group is an independent supplier of quality pre-owned equipment

Gold-standard solutions

Extend the life of your critical industrial, commercial, and military systems with our superior service and support.

We buy equipment

Planning to upgrade your current equipment? Have surplus equipment taking up shelf space? We'll give it a new home.

Learn more!

Visit us at **artisantg.com** for more info on price quotes, drivers, technical specifications, manuals, and documentation.

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

We're here to make your life easier. How can we help you today?

(217) 352-9330 | sales@artisantg.com | artisantg.com

ARTISAN® TECHNOLOGY GROUP