ARTISAN®
TECHNOLOGY GROUP

Your **definitive** source for quality pre-owned equipment.

Artisan Technology Group
(217) 352-9330 | sales@artisantg.com | artisantg.com

Full-service, independent repair center
with experienced engineers and technicians on staff.

We buy your excess, underutilized, and idle equipment
along with credit for buybacks and trade-ins.

Custom engineering
so your equipment works exactly as you specify.

- Critical and expedited services
- In stock / Ready-to-ship
- Leasing / Rentals / Demos
- ITAR-certified secure asset solutions

Expert team | Trust guarantee | 100% satisfaction

All trademarks, brand names, and brands appearing herein are the property of their respective owners.

Find the *Yokogawa / Ando AQ-6310B* at our website: *Click HERE*

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

AS-61712
INSTRUCTION MANUAL
FOR
AQ-6310 OPTICAL
SPECTRUM ANALYZER

ANDO ELECTRIC CO., LTD.
Tokyo, Japan


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

The apparatus to which this manual is attached is designed to operate with any of AC 100V to 120V, 220V to 240V.

This apparatus has been set before shipment from the factory, to the operating voltage to be used. Before use, confirm that the operating voltage indication in the window of (AC LINE) connector at the rear conforms to the voltage of the AC power socket-outlet.


The line voltage AC ***V referred to in this manual should be read as the voltage of the operating voltage indication in the window described above.

If the operating voltage indication, however, does not conform to the voltage of the AC power socket-outlet, change the operating voltage by the procedure illustrated below.



Operating voltage indication

How to change the operating voltage:



1. Open cover door. Pull [FUSE PULL] lever and rotate it to left. Remove the fuse.
2. Remove the operating voltage selection card. Position this card

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

so that the same value as the voltage of the AC power socket-outlet appears at top-left corner, and put it firmly into the slot.

3. Rotage [FUSE PULL] lever to normal position. Insert a fuse conforming to the operating voltage-fuse current chart provided above (AC LINE) connector in the holder. Close the cover.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

CONTENTS

PAGE

SECTION 1 GENERAL

1.1 Introduction 1-1
1.2 General Description 1-1
1.3 Specifications 1-1
1.4 Options 1-2
1.5 Standard Accessories 1-2
1.6 Precautionary Concerns 1-6

SECTION 2 PREPARATIONS FOR USE

2.1 Introduction 2-1
2.2 Unpacking and Acceptance Inspection 2-1
2.2.1 Mechanical Inspection 2-1
2.2.2 Performance Test 2-1
2.3 Damage or Fault 2-1
2.4 Repacking 2-2

SECTION 3 OPERATION

3.1 Introduction 3-1
3.2 Operating Controls, Indicators, and Connectors 3-1
3.3 Soft Key Descriptions 3-8
3.4 Preoperational Preparations 3-19
3.5 Operating Procedures 3-19
3.5.1 Initialization 3-19
3.5.2 FUNCTION and MARKER Setup Procedure 3-20
3.5.3 CRT Display Screen Contents Descriptions 3-21
3.5.4 AUTO Measurement 3-22
3.5.5 Manual Measurement 3-25
3.5.6 Measurement Condition Setup 3-28
3.5.7 Sweep Mode Setup 3-38
3.5.8 Display Method Setup 3-43
3.5.9 Marker Usage 3-54
3.5.10 Level Difference Measurement 3-58
3.5.11 Wavelength Difference Measurement 3-60
3.5.12 Half-Width Measurement 3-63
3.5.13 Half-Width Algorithms 3-66

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.14 MEMORY and DATA MEMORY 3-72
3.5.15 Waveform Storage and Recall by MEMORY Functions ... 3-73
3.5.16 Data Overlap Display 3-76
3.5.17 Data Subtraction Display by DATA MEMORY Function ... 3-78
3.5.18 Printer Operating Procedure 3-85
3.5.19 Automatic Calibration by CALIBRATION Function 3-90
3.5.20 Comment Display by LABEL Function 3-91
3.5.21 Shutdown Procedure 3-93
3.6 GP-IB Control 3-93
3.6.1 Address Setup 3-93
3.6.2 Function Setup 3-94
3.6.3 Measurement Start 3-94
3.6.4 Measured Data Output 3-95
3.6.5 Other Remote Messages 3-95
3.6.6 SRQ Function 3-96
3.6.7 DC Function 3-96
3.6.8 DT Function 3-96
3.6.9 GP-IB Usage Precautions 3-96
3.7 ATTENTION Display Function 3-109
3.8 Video Interface 3-111
3.9 Abbreviations 3-115
3.10 Measuring Precautions 3-117
SECTION 4 CIRCUITRY AND CONSTRUCTION
4.1 Introduction 4-1
4.2 Circuitry 4-1
4.2.1 Circuitry Overview 4-1
4.2.2 Individual Circuit Descriptions 4-1
4.3 Construction 4-2
SECTION 5 MAINTENANCE
5.1 Introduction 5-1
5.2 Pre-operational Checkout 5-1
5.3 Mechanical Inspection 5-1

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

LIST OF DRAWINGS AND TABLES

Fig. 3-1 Front Panel 3-2
Fig. 3-2 Rear Panel 3-2
Fig. 3-3 CRT Display Screen on Initialization 3-20
Fig. 3-4 CRT Screen Example 3-21
Fig. 3-5 Soft Key Menu Screen (1) 3-21
Fig. 3-6 Soft Key Menu Screen (2) 3-22
Fig. 3-7 AUTO Measurement Example 3-23
Fig. 3-8 LD Light Source First Sweep Waveform 3-26
Fig. 3-9 LD Light Source Last Sweep Waveform 3-27
Fig. 3-10 LED Light Source First Sweep Waveform 3-27
Fig. 3-11 LED Light Source Last Sweep Waveform 3-28
Fig. 3-12 State Prevailing after **[CENTER WAVELENGTH]** Switch Activation 3-29
Fig. 3-13 State Prevailing after **[SWEEP WIDTH]** Switch Activation 3-31
Fig. 3-14 State Prevailing after **<SWEEP TIME>** Switch Activation 3-31
Fig. 3-15 State Prevailing after **[△W→SPAN]** Switch Activation 3-32
Fig. 3-16 State Prevailing after **[HIGH SENSITIVITY]** Switch Activation 3-33
Fig. 3-17 State Prevailing after **[RESOLUTION]** Switch Activation 3-34
Fig. 3-18 State Prevailing after **[REFERENCE LEVEL]** Switch Activation 3-36
Fig. 3-19 State Prevailing after **[AVERAGE TIMES]** Switch Activation 3-37
Fig. 3-20 State Prevailing upon **[REPEAT]** Switch Activation 3-39
Fig. 3-21 State Prevailing upon **<WAIT>** Key Activation 3-40
Fig. 3-22 State Prevailing upon **[SINGLE]** Switch Activation 3-41
Fig. 3-23 State Prevailing upon **[STOP]** Switch Activation 3-42
Fig. 3-24 LOG Scale Screen 3-43
Fig. 3-25 LIN Scale Screen 3-44
Fig. 3-26 NORMALIZE Function Execution Examples **(LOG Scale Mode)** 3-46
Fig. 3-27 NORMALIZE Function Execution Examples **(LIN Scale Mode)** 3-47
Fig. 3-28 DOMINANT Function Execution Examples 3-49
Fig. 3-29 3-D Mode Initial Screen 3-50
Fig. 3-30 3-D Display Example 3-52
Fig. 3-31 Peak Level Measurement Example 3-55
Fig. 3-32 Peak Wavelength Measurement Example 3-57
Fig. 3-33 Level Difference Measurement Examples 3-60

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Fig. 3-34 Wavelength Difference Measurement Examples 3-62
Fig. 3-35 Half-Width Measurement Examples 3-65
Fig. 3-36 Half-Width Determination Algorithm (ENV Method) 3-66
Fig. 3-37 Mode Determination 3-67
Fig. 3-38 3-67
Fig. 3-39 3-68
Fig. 3-40 3-68
Fig. 3-41 3-69
Fig. 3-42 3-70
Fig. 3-43 3-71
Fig. 3-44 MEMORY Function Execution Examples 3-75
Fig. 3-45 Overlap Display Function Execution Examples 3-77
Fig. 3-46 Loss Wavelength Characteristic Measurement Setup 3-79
Fig. 3-47 Loss Wavelength Characteristic Measurement Examples 3-82
Fig. 3-48 Optical Filter Transmittance Measurement Setup 3-83
Fig. 3-49 Optical Filter Transmittance Measurement Examples 3-84
Fig. 3-50 [COPY ON] Switch Activated 3-87
Fig. 3-51 <PAUSE> Key Pressed after [COPY ON] Switch Activation 3-88
Fig. 3-52 Recording Paper Loading 3-88
Fig. 3-53 [COPY ON] Switch Activated with Head Lifted Up 3-89
Fig. 3-54 [COPY ON] Switch Activated with No Recording Paper Loaded ... 3-89
Fig. 3-55 LABEL Function Execution Example 3-92
Fig. 3-56 Bus Line Description and Connector Difference 3-108
Fig. 3-57 Output Video Signal 3-112
Fig. 3-58 Video Plotter Adjustment 3-113
Fig. 3-59 Fine Adjustment 3-114
Fig. 3-60 3-117
Fig. 3-61 3-119
Fig. 3-62 3-119
Fig. 3-63 3-120
Fig. 3-64 Principles of Czerny Turner Monochromator 3-121
Fig. 3-65 Grating 3-122

Table 1-1 Specifications 1-3
Table 1-2 Option List 1-5
Table 1-3 Standard Accessory List 1-6

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-1 Switch Descriptions 3-5
Table 3-2 Function Codes 3-97
Table 3-3 Talker Formats 3-106
Table 3-4 STATUS Contents 3-107
Table 3-5 AQ-6310 Interface Construction 3-107
Table 3-6 ATTN Indications No. 0 to 10 3-109
Table 3-7 ATTN Indications No. 21 to 27 3-110
Table 3-8 ATTN Indications No. 31 to 37 3-111
Table 3-9 Abbreviation List 3-115

LIST OF ATTACHED DRAWINGS AND TABLES

AQ-6310 Optical Spectrum Analyzer External Drawing ASD-61712-1-1/3
AQ-6310 Optical Spectrum Analyzer External Drawing (Front View) ASD-61712-1-2/3
AQ-6310 Optical Spectrum Analyzer External Drawing (Rear View) ASD-61712-1-3/3

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SECTION 1
GENERAL

1.1 Introduction

This manual describes the operating, handling, and maintenance procedures for the AQ-6310 Optical Spectrum Analyzer.

1.2 General Description

The AQ-6310 is a general-purpose wavelength characteristic measuring instrument. It measures not only the LD, LED, and other light source spectrums, but also the loss wavelength and transmission characteristics of optical fiber cables, filters, and other devices.

The AQ-6310 has a wide wavelength measurement range of 400 to 1750 nm, a high sensitivity to permit measurement of light as weak as -70 dBm, a wavelength resolution of ±1 nm, and a high sweep rate, thereby covering a wide scope of applications.

In addition to the basic functions such as the center wavelength, sweep width, resolution, and reference level setup functions, the AQ-6310 performs various other functions including the averaging, data memory, data subtraction, marker, peak search, half-width search, label, normalize, dominant, and 3-D display functions.

One major advantage offered by the AQ-6310 is the automatic function setup feature. This feature is provided for enhanced simplicity of operating steps; that is, the center wavelength, sweep width, resolution, and reference level are automatically set to the optimum values when a light source is connected via optical fiber to the AQ-6310.

The AQ-6310 incorporates a printer so that on-screen image hard-copy printout is obtained with ease. Further, a GP-IB is standard on the AQ-6310 to permit automatic measurement.

1.3 Specifications

The product specifications for the AQ-6310 are presented in Table 1-1.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

1.4 Options

The options available for use with the AQ-6310 are listed in Table 1-2.

1.5 Standard Accessories

The AQ-6310 is supplied with the standard accessories listed in Table 1-3.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


Table 1-1 Specifications

Measurement wavelength range	400 to 1750 nm
Measurement level range	-70 to + 10 dBm/resolution (600 to 1600 nm) -60 to + 10 dBm/resolution (400 to 1750 nm) (See NOTE 1) Wavelength sensitivity automatic correction function provided.
Linearity	± 1 dB/60 dB ± 0.5dB/20dB ± 0.2dB/10dB
Level scale	0.5 to 10 dB/div and LINEAR
Wavelength accuracy	± 1 nm (5° to 40°C), ± 0.5 nm (25°C)
Wavelength sweep width	0.0 to 150 mm/div (see NOTE 2)
Resolution	0.1 to 10 nm (600 to 1750 nm) 0.1 to 5nm (400 to 600nm)
Sweep time	1 sec or less (span: 100 nm or less) 2 sec or less (span: 500 nm or less) 5 sec or less (span: 1500 nm or less) (See NOTE 3)
Functions	Automatic measurement, peak search, half-width search, data subtraction, averaging, labeling, normalization, dominant waveform display, three-dimensional display, wavelength calibration (optional calibration light source available), data/measurement condition memory (nonvolatile), etc.
CRT	System : electromagnetic deflection Size : 7-inch Resolution : 640 x 474 dots
Optical input connector	FC type (standard), D4 type (NEC Specification), OF2 type (Fujitsu Specification), W/E type, diamond type, etc.
Printer function	High-speed printer incorporated

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Video output	Approx. 1 Vp-p, composite video signal, 75 Ω load, BNC connector
Analog output	0 to 5 V Load: 1 kΩ or more; BNC connector
GP-IB	Provided as a standard item
Power requirements	100/115-120/220/230-240 VAC, 50/60 Hz, approx. 150 VA
Dimensions and weight	Approx. 266 H x 426 W x 450 mm D Approx. 30 kg

NOTE 1: The resolution is expressed by △λ. Therefore, the power per resolution or the smallest measurable power approximates to the shaded area below.



NOTE 2: The 0.0 nm sweep function is exercised to observe how the level changes with time at a certain wavelength. This function is useful for optical axis alignment.

NOTE 3: These values are obtained when the HIGH SENSITIVITY switch is OFF, the AVERAGE TIMES setting is 1, and the center wavelength setting is 1050 nm.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 1-2 Option List

AQ-4302 He-Ne Laser

(For automatic calibration)

Output light wavelength : 632.8 nm

Optical output level : -8 dBm or higher (For GI50/125 µm fiber)

-15 dBm or higher (For SM10/125µm fiber)

AQ-4303B White Light Source

(For wavelength loss characteristic measurement)

Output light wavelength : 400 to 1800 nm

Optical output level : -45 dBm or higher

(wavelength: 850 nm, 1300 nm; wavelength band: 10 nm; CW light)

Output waveform : CW light and 270 Hz chopped light

AQ-9313 Device Adapter

(For LD/LED device characteristic measurement)

Output form : FC connector attached SM10/125 µm fiber

Device DC power supply incorporated

AQ-9314 Parallel Beam Mount

(For transmission and loss characteristic measurements of optical filters and other devices)

Fiber used : core diameter, 800 µm; GI fiber

Large-diameter Optical Fiber Cable

Core diameter : 800 µm, 400 µm

Plotter

Produces the hard-copy printout of all CRT screen contents.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 1-3 Standard Accessory List

Name	Qty	Remarks
Power cord	1	Approx. 3 m (attached)
Fuse	1	* A (installed) NOTE
Instruction manual	1	
Recording paper	2 rolls	For built-in printer, TF50KS- E2 (Jujo Paper)

NOTE: The ampere is maked on the supply voltage rating indication above the rear panel AC LINE connector.

1.6 Precautionary Concerns

1.6.1 Operating Temperature Range

The operating temperature range within specifications is from + 5 to + 40°C.

1.6.2 Environmental Conditions

The AQ-6310 incorporates a precision spectroscope, therefore, when storing or transporting the AQ-6310, use extreme caution to avoid undue temperature, shock, and vibration.

If the following conditions are not met, the AQ-6310 may fail to recover its expected performance.

(1) Storage temperature range: -20 to + 50°C

(2) Vibration

Frequency	: 10 Hz
Complex amplitude	: 2 ± 0.5 mm
Direction of vibration	: vibration in all three planes
Vibration time	: 10 minutes in each plane

(3) Shock

The shock caused when one side of the apparatus resting still on a hard wooden floor is raised 2.5 cm and then let fall.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SECTION 2
PREPARATIONS FOR USE

2.1 Introduction

This section describes the procedures for unpacking, acceptance inspection, and repacking.

2.2 Unpacking and Acceptance Inspection

The AQ-6310 has been factory inspected, mechanically and electrically, prior to shipment to ensure that it gives satisfactory performance. When your order is received, promptly unpack it and check its contents for damage that may have been sustained in transit.

When unpacking the apparatus, save the wooden box, corrugated fiberboard box, cushions, and other packing materials so that they may be reused when the apparatus is to be packed again for shipment.

2.2.1 Mechanical Inspection

Visually check the apparatus for damage or faults. Also check its switches and knobs for faults or damage that may have been sustained in transit. Further, check to ensure that the specified number of correct accessories and spare parts are supplied.

2.2.2 Performance Test

If the apparatus is found by the mechanical inspection to be in good order, check its performance for compliance with the product specifications indicated in Table 1-1.

2.3 Damage or Fault

If the apparatus is found damaged or faulty in the acceptance inspection, immediately report the damage or fault to your nearest dealer.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

2.4 Repacking

When repacking the apparatus for shipment, use the packing materials, if saved for later use. If they have not been saved, repack the apparatus using the procedure outlined below.

(1) Wrap the apparatus in strong paper or vinyl sheeting. Protect all protrusions with cushions against damage.

(2) Place the wrapped apparatus in a wooden or corrugated fiberboard box which is larger by about 5 to 10 cm than the apparatus on all sides.

(3) Fill all open spaces between the apparatus and the box with polyurethane foam or any other suitable cushioning material. The apparatus may rattle and be damaged in transit if cushioning is insufficient.

(4) Cover the box and brace it up with steel bands. If a corrugated fiberboard box is used, seal it with adhesive tape or the like.

(5) Indicate the contents, shipping marks, and other relevant shipping information on the box in a legible and durable way.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SECTION 3
OPERATION

3.1 Introduction

This section describes the handling and operating procedures for the AQ-6310.

3.2 Operating Controls, Indicators, and Connectors

The names of the operating controls, indicators, and connectors are indicated in Figs. 3-1 and 3-2, and their functions are described in Table 3-1.

The numbers enclosed in circles in the figures correlate to those indicated in the Remarks column of the tables.

The names presented in this section should be correctly remembered because they are used in the subsequent explanations of operating procedures.

The names enclosed within square brackets ([ ]) are the same as the enclosure panel markings (panel control name lettering, setting class marks, etc.).

The switches marked "Provided" in the Soft Key Option column have optional functions. For the information on the soft key functions, see section 3.3.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-1 Front Panel




Fig. 3-2 Rear Panel

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

<Front Panel>

(1) CRT Display

Shows the measurement conditions, measured waveforms and values, and others.

(2) Measurement Switches

Used to set up the measurement method and conditions or start and stop measurements.

(3) Display Switches

Used to change the screen Y scale, perform data subtraction, or manipulate the markers.

(4) Memory Switches

Used to store waveforms and measurement conditions in the data memory or read such stored data.

(5) Copying Switches

Used when, for instance, the built-in printer is to be set to produce the hard-copy printout of screen contents.

(6) GP-IB Switches

Used to perform equipment address setup or return the apparatus to the local mode.

(7) Label Switch

Used to enter a comment or the like onto the upper part of the CRT screen.

(8) Calibration Switch

Used when automatic wavelength calibration is to be effected with He-Ne laser light entered.

(9) Data Entry Switch and Knob

Used for the setup of measurement conditions and other data.

(10) Soft Key Switches

Used when the optional functions provided for the above switches are to be exercised.

(11) Optical Input Connector

(12) Dustproof Cap

(13) Power Switch

(14) Brightness Adjustment Control

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

<Rear Panel>

(15) Composite Video Signal Output Terminal

(16) Analog Output Terminal

Used for measurement wavelength level external monitoring.

(17) GP-IB Connector

(18) AC Power Connector

Comprised of an AC power connector and a fuse unit.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-1 Switch Descriptions

Switch Marking		Function	Soft Key Option	Remarks	Ref. Page
MEASUREMENT				(2)	
AUTO	ON/OFF	Used to choose between automatic and manual function measurements. The lamp comes on when the switch is turned ON. This switch is effective during sweep stop periods only.			3-22
SWEEP	REPEAT	Repetitive sweep. The lamp lights during sweep.	Provided		3-38
	SINGLE	Single sweep. The lamp lights during sweep.	Provided		3-41
	STOP	Stops a sweep operation.	Provided		3-42
HIGH SENSITIVITY		Measurement sensitivity selector switch.	Provided		3-33
CENTER WAVELENGTH		Used for center wavelength selection.	Provided		3-28
SWEEP WIDTH		Used for sweep width selection.	Provided		3-29
RESOLUTION		Used for resolution setup.			3-34
REFERENCE LEVEL		Used for reference level setup.			3-35
AVERAGE TIMES		Used to select the number of averaging operations to be performed.			3-36
PEAK→ CENTER		Searches for the peak wavelength and sets its value as the central wavelength.			3-29
MARKER→ CENTER		Causes the wavelength value of the wavelength marker position to be set as the central wavelength.			3-29
PEAK→ REF LEVEL		Searches for the peak level and sets its value as the reference level.			3-35
MARKER→ REF LEVEL		Causes the level value of the level marker position to be set as the reference level.			3-35
△W→ SPAN		Searches for the half-width and sets its value as the measurement wavelength range.	Provided		3-30

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-1 Switch Descriptions (Continued)

Switch Marking		Function	Soft Key Option	Remarks	Ref. Page
DISPLAY				(3)	
LOG/LIN		Places the display vertical axis scale in the LOG or LIN mode. When the LOG scale is selected, the lamp lights.			3-43
NORMALIZE		Normalizes a displayed waveform. This switch is effective during sweep stop periods only.			3-45
3D		Provides a 3-D display of measured waveforms. Switching can be effected during sweep stop periods only.	Provided		3-50
DOMINANT		Subjects a displayed waveform to human visual sensitivity correction to provide a normalized waveform display. This switch is effective during sweep stop periods only.			3-48
DATA MEM-ORY	Acc A	Nonvolatile memory for storing waveforms and measurement conditions. This switch is effective during sweep stop periods only.			3-76
	Acc B	Nonvolatile memory for storing waveforms and measurement conditions. This switch is effective during sweep stop periods only.			3-76
	A-B A/B	Displays A-B (in the LOG mode) or A/B (in the LIN mode) after scaling. This switch is effective during sweep stop periods only.	Provided		3-78
	B-A B/A	Displays B-A (in the LOG mode) or B/A (in the LIN mode) after scaling. This switch is effective during sweep stop periods only.	Provided		3-82
	RECALL	Used for A and B memory recalling. This switch is effective during sweep stop periods only.			3-76
	GRAPH CLEAR	Clears a displayed waveform.			3-74
MARK-ER	PEAK LEVEL	Searches for the peak value power and places the movable level marker at the peak value power position.			3-54
	PEAK WAVE LENGTH	Searches for the peak value power and places the movable wavelength marker at the peak value power position.	Provided		3-55
	HALF-WIDTH SEARCH	Searches for the half-width points and places markers 1 and 2 at the half-width points.	Provided		3-63
	1	Sets up marker 1 at the current marker position.			3-58
	2	Sets up marker 2 at the current marker position.			3-58
	CLEAR	Clears the marker.			3-54

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-1 Switch Descriptions (Continued)

Switch Marking	Function	Soft Key Option	Remarks	Ref. Page
MEMORY			(4)	
SAVE	When any one of label keys 0 through 9 or A through F is pressed after SAVE key activation, the associated waveform and measurement conditions are stored into memory. Only the measurement conditions are to be stored into nonvolatile memory. This switch is effective during sweep stop periods only.			3-73
RECALL	When any one of label keys 0 through 9 or A through F is pressed after RECALL key activation, the associated waveform and measurement conditions are recalled. This switch is effective during sweep stop periods only.			3-73
COPY			(5)	
ON	Starts screen contents printing. This switch is effective during sweep stop periods only.	Provided		3-85
FEED	While this switch is pressed, the paper is fed.			3-86
GP-IB			(6)	
ADDRESS	Performs equipment address setup.			3-93
LOCAL	Used to exit the REMOTE mode when the GP-IB is used. However, this switch is inoperative in the LLO state.			3-95
LABEL			(7)	
ON/ OFF	Used when labeling is conducted. When the LABEL switch lamp is lit, the key switch group green light indications take effect. This switch is effective during sweep stop periods only.	Provided		3-91
CALIBRATION			(8)	
(632.8 mm)	Wavelength automatic calibration function. (The option AQ-4302 is required.) This switch is effective during sweep stop periods only.	Provided		3-90
DATA ENTRY			(9)	
FAST	Places the rotary knob in the fast-forward mode.			3-20
(Rotary knob)	Used for function setup.			3-20

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-1 Switch Descriptions (Continued)

Switch Marking	Function	Soft Key Option	Remarks	Ref. Page
— SOFT KEY —			(10)	
ON/ OFF	Turns on and off the soft key screen display feature. The switch lamp comes on when the switch is turned ON.			3-8
(Soft key switch)	Pressed when the associated optional switch function is to be exercised.			3-9

3.3 Soft Key Descriptions


When the switches indicated in Table 3-1 in section 3.2 are marked "Provided" in the Soft Key Option column, soft keys are displayed at the bottom of the CRT screen upon switch activation. The displayed soft key functions can be exercised by pressing the associated switches located at the lower side of the CRT body.

The soft keys are displayed only when the **[SOFT KEY ON/OFF]** switch is ON (with the LED illuminated). However, even if this **[ON/OFF]** switch is OFF, the optional functions can be performed by pressing the soft keys although they are not displayed on the CRT screen.

The following describes the soft keys displayed upon switch activation. Note that the name of the hard switches (the panel mounted pushbutton switches except the soft keys) are enclosed by square brackets ([***]) while the soft key names are enclosed by angle brackets (<***>) for ease of explanation. Further, if the soft key name is marked "(Latch)", the press of such a soft key causes the display to appear in reverse field, the associated function to be executed, and the key to be stored. This ensures that when the same switch is pressed, such a stored soft key function is immediately executed. This condition will be retained until another latch function incorporated soft key is pressed. It should also be noted that latched information is nonvolatile and therefore does not change upon power turn-off.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(1) [REPEAT] Switch



(a) <SWEEP NORMAL > (Latch)

Repeatedly makes measurement over the range from the left- to the right-end of the screen at the intervals preset by <WAIT NORMAL> or <WAIT>.

(b) <SWEEP MKR 1-2>(Latch)

Repeatedly conducts sweep between wavelength markers 1 and 2 only when such markers are set up on the screen.

However, there are some limitations on the sweep between markers 1 and 2. For the information on such limitations, see section 3.5.7-(1)-(a).

(c) <WAIT NORMAL> (Latch)

Selects the minimum WAIT time. When the key is pressed, the soft key contents return to the states indicated under (A), causing the measurement to start.

(d) <WAIT> (Latch)

This key is pressed to determine the measurement time intervals for repetitive measurement. When the key is pressed, the current setting appears in the interrupt display area and the soft key contents change as indicated under (B).

(e) <s> <m> <h>


Used to specify the unit for WAIT time selection. When the key is pressed, the setting is reset to 1. When the rotary knob is turned in this state, the interrupt display area indication changes from 1 to 59 for "s" (second) and "m" (minute) or from 1 to 24 for "h" (hour).

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(f) **<ENTER>**

When this key is pressed after the WAIT time values are set up with the rotary knob with the units specified as indicated in paragraph (e), the WAIT time is set up. Further, the soft key contents return to the state indicated under (A), causing measurement to start.

(2) **[SINGLE]** Switch




(a) **<SWEEP NORMAL>** (Latch)

Makes single measurement over the range from the left- to the right-end of the screen.

(b) **<SWEEP MKR 1-2>** (Latch)

Conducts single sweep between wavelength markers 1 and 2 only when such markers are set up on the screen.
However, there are some limitations on the sweep between markers 1 and 2. For the information on such limitations, see section 3.5.7-(1)-(a).

(3) **[STOP]** Switch




Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(a) <MANUAL SWEEP>

This key is pressed when the rotary knob is turned to manually start measurement from the point at which sweep was stopped by [STOP] switch activation.

(4) [HIGH SENSITIVITY] Switch



(a) <H-SENS FAST>(Latch; (a), (b), or (c))

Makes high-sensitivity measurement. The maximum reference level selectable is -60 dB. The high- sensitivity measurement is suitable for the light sources having -60 dB or higher output levels. Note that the measurement speed is lower than that in the <H-SENS OFF> mode.

(b) <H-SENS SLOW> (Latch)


Makes superhigh-sensitivity measurement. The maximum reference level selectable is -70 dB. This mode makes it possible to measure extremely weak light below -60 dBm at enhanced S/N ratios because signal amplification is limited to a narrow band. However, the measurement speed is considerably lowered when -60 dBm or lower light levels are measured.

(c) <H-SENS OFF> (Latch)

Provides regular-sensitivity measurement, turning off the high-sensitivity measurement feature. The maximum reference level selectable is -42 dB. The measurement speed is higher than those in the above two modes.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(5) [CENTER WAVELENGTH] Switch



(a) <START WL>

This key is used when the measurement start wavelength is to be used as the measurement wavelength setting instead of the center wavelength. When the key is pressed, the soft key contents change so that the current measurement start wavelength appears in the interrupt display area. This displayed wavelength can be changed with the rotary knob.

(b) <STOP WL>

When this key is pressed after measurement start wavelength setup, the current stop wavelength appears in the interrupt display area. This displayed wavelength can be changed with the rotary knob. However, in contrast to start wavelength setup, the stop wavelength setting cannot be changed continuously. It can be changed stepwise in accordance with the selected sweep intervals (1, 2, or 5 nm).

(c) <RETURN>

Returns the soft key contents to the state indicated under (A).

(d) <CTR WL **** .*>


Sets the center wavelength to **** .* nm.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(e) <ETC>

Each time this key is pressed, the soft key contents alternate between (A) and (C).

(6) [SWEEP WIDTH] Switch



(a) </div.> (Latch; Either (a) or (b))

Sets up the sweep width in terms of "/div".

(b) <SPAN> (Latch)

Sets up the sweep width by expressing its full span (measurement stop wavelength - start wavelength).

(c) <INTVL NORMAL> (Latch; Either (c) or (d))

Selects the standard data sampling interval [(measurement stop wavelength -measurement start wavelength) /580] for measurement.

(d) <MEAS WL INTVL> (Latch)

Sets the data sampling interval to 0.1, 0.2, 0.5, 1, 2, 5, 10, or 20 nm. However, the maximum acceptable setting is half the full span.

(e) <****–****nm>

Sets up the center wavelength and sweep width so that the measurement wavelength range is from **** to **** nm.


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(f) <SWEEP TIME>

This key is operative only when the sweep width is 0 nm with the [HIGH SENSITIVITY] switch OFF, and selects 1, 2, 5, 10, 20, or 50 seconds as the time for left- to right- end screen sweep. Note, however, that this sweep time setting applies to an averaging times setting of 1.

(7) [△W→ SPAN] Switch

[△W→SPAN] Switch Activation




(a) <× *> (*: 1 to 5; One of Them Is to Be Latched)

Conducts half-width search by the currently defined method, multiplies the obtained value by *, and sets the sweep width to the smallest of acceptable step full span values greater than the multiplied value.

(8) [3D] Switch

[3D] Switch Activation



(a) <ANGLE CHANGE>

Changes the three-dimensional waveform display angle. When this key is pressed, the current angle setting appears in the interrupt display area. The rotary knob varies the angle setting in 10 degree steps from + 50 to - 50 degrees.


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(b) <EXECUTE>

When this key is pressed with the angle setting changed with the rotary knob after <ANGLE CHANGE> key activation, the displayed waveform angle changes accordingly.

(9) [A-B A/B], [B-A B/A] Switch

[A-B A/B], [B-A B/A] Switch Activation




(a) <MANUAL>

When the [A-B A/B] , [B-A B/A] switch is pressed, the memory A contents are subtracted from the memory contents B or vice versa and the results are displayed after auto scaling. When the vertical axis scale is to be manually changed at this time, the <MANUAL> key is pressed. When the current scale appears in the interrupt display area after <MANUAL> key activation, turn the rotary knob to change the displayed scale value. This causes the vertical axis scale and displayed waveform to be updated accordingly.

(10) <PEAK WAVELENGTH>Switch

[PEAK WAVELENGTH] Switch Activation



(a) <AUTO ON> (Latch; Either (a) or (b))

When this key is activated, the peak waveform is searched for to display a marker after each REPEAT sweep. This is performed without respect to [PEAK WAVELENGTH] switch activation.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(b) <AUTO ON> (Latch)

When this key is activated, the displayed peak waveform is searched for to display a marker only when the [PEAK WAVELENGTH] switch is pressed.

(c) <CROSS ON>(Latch; Either (c) or (d))


When this key is activated, the marker functions as a cross marker (the intersection of a wavelength marker and a level marker).

(d) <CROSS OFF> (Latch)

When this key is activated, only the wavelength marker function is exercised.

(11) [HALF-WIDTH SEARCH] Switch

[HALF-WIDTH SEARCH] Switch Activation



(a) <ENV> (Latch; Any One of (a) through (e))

Determines the half-width by the envelope method.

(b) <RMS> (Latch)

Determines the half-width by the RMS method.

(c) <RMS X*.**> (Latch)

The half-width obtained by the RMS method is multiplied by *.**, and the resultant value is used as the half-width.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


(d) <THRESH **dB> (Latch)

Determines the half-width considering that the threshold level is ** dB.

(e) <THRESH 1/e 2> (Latch)

Determines the half-width considering that the threshold level is $1/e^2$ the displayed waveform peak level (therefore the threshold level is approx. 0.135).

(12) [ON] Switch (COPY)



(a) <PAUSE>

When this key is pressed within 2 seconds after [ON] switch activation, date and time setting update is made possible. Upon <PAUSE> key activation, the soft key contents change as indicated under (B).

(b) <CONT>

When this key is pressed after date and time setup, printing starts.

(c) <→>

This key is used to move the cursor for date and time setup.

(d) <HORIZE> (Latch either (d) or (e))

Produces a horizontal hard-copy printout of the screen contents.

(e) <VERT> (Latch)

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Produces a vertical hard-copy printout of the screen contents.


(f) **<DATE & TIME P ON>** (Either One Is to Be Latched)

When the "P ON" key is pressed, the date and time are printed onto the printer.

When the "P OFF" key is pressed, the date and time are not printed onto the printer.

(13)[LABEL] Switch

[LABEL] Switch Activation



(a) **<SHIFT>**


When this key is pressed in message input into the screen label area, the display appears in reverse field, indicating that the right-hand side of the label character is effective. When the label character is pressed, the character located to the right of the pressed label character is entered and displayed back in normal field. This character also appears in normal field when the **<SHIFT>** key is pressed again.

(b) **<CLEAR>**

Clears the label area and places the cursor at the left end of the label area.

(14)[CALIBRATION] Switch

[CALIBRATION] Switch Activation



(a) **<START>**

When this key is pressed after the He-Ne laser light source (632.8 nm) is connected to the input connector, automatic waveform calibration starts.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.4 Preoperational Preparations

Perform the following checks before connecting the AQ-6310 power cord to a commercial power source wall outlet.

(1) Ensure that the **[POWER]** switch (13) is OFF.

(2) Check whether the voltage supplied to the wall outlet agrees with the AQ-6310 operating voltage.

3.5 Operating Procedures

In the following subsections, the name of the hard switches (the panel mounted pushbutton switches except the soft keys) are enclosed by square brackets ([***]) while the soft key names are enclosed within a square (<***>). Further, the following explanations are given on the assumption that the **[SOFT KEY ON/OFF]** switch is ON.

3.5.1 Initialization

(1) Turn ON the **[POWER]** switch.

(2) The CRT screen then looks like Fig. 3-3 while the initial conditions are automatically set up with the automatic offset function executed. When initialization is entirely completed, the CRT screen display contents change.
The screen contents appearing first time after initialization are those which prevailed immediately before the last **[POWER]** switch turn-OFF. Not only the measurement conditions but also the waveform displayed at that time are reproduced on the screen.

(3) As stated above, the AQ-6310 stores the measurement conditions, soft key latch status, displayed waveform, and other relevant data in nonvolatile memory. All these data stored in nonvolatile memory are retained even after power turn-off.
However, when all the stored conditions or soft key latch status are to be initialized, turn ON the **[POWER]** switch while the **[AUTO ON/OFF]** switch

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

is held down. When this procedure is completed, all the stored conditions are initialized to the states prevailing just after apparatus shipment.





Fig. 3-3 CRT Display Screen on Initialization

3.5.2 FUNCTION and MARKER Setup Procedure

When the FUNCTION to be set up is pressed, the current setting is displayed at the upper right corner of the CRT screen. When the MARKER is pressed, the marker value is displayed at the bottom of the CRT screen. To change the displayed values, use the [DATA ENTRY] rotary knob.

When the [FAST] switch is pressed to ON (so that its LED lights), the fast forward mode prevails.



Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.3 CRT Display Screen Contents Descriptions

(1) State Where Soft Key Menu Is Not Displayed




Fig. 3-4 CRT Screen Example

(2) State Where Soft Key Menu (MARKER Excluded) Is Displayed

The lower-end display area changes as shown below.




Fig. 3-5 Soft Key Menu Screen (1)

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(3) State Where Soft Key Menu (MARKER) Is Displayed

The lower-end display are changes as shown below.




Fig. 3-6 Soft Key Menu Screen (2)

3.5.4 AUTO Measurement

(1) Connect to the **[INPUT]** connector the light source under test.

(2) While sweep is not being effected, press the **[AUTO ON/OFF]** switch so that its switch lamp lights.


(3) The appropriate center wavelength, sweep width, reference level, and resolution are automatically selected to suit the input signal spectrum so that sweep is repeatedly conducted under the optimum conditions.

(4) When the peak level is below -40 dBm in the **[AUTO]** mode, sweep is repeatedly conducted under the initial conditions without optimizing the conditions. When such a condition exists, the **[AUTO ON/OFF]** switch should be pressed to turn OFF the **[AUTO]** function so that manual measurement is made.

NOTE: The **[AUTO]** measurement wavelength range is 600 to 1750 nm.


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

The [AUTO] measurement examples are presented below.




First Sweep

Fig. 3-7 [AUTO] Measurement Example



Second Sweep

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



Sweep after Convergence

NOTE

In the [AUTO] mode measurement, convergence may not be achieved under the optimum conditions due to input signal (waveform or level) or other condition variations so that the measurement parameters change each sweep. When such a condition occurs, turn off the [AUTO] function and make manual measurement.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.5 Manual Measurement

Typical LD light source (wavelength: approx. 1300 nm; level: approx. -10 dBm) and LED light source (wavelength: approx. 850 nm; level: approx. -30 dBm) measurement procedures are described below.

(1) LD Light Source Measurement

(a) Connect the light source under test, to the optical input connector with optical fiber.

(b) Turn OFF the **[AUTO ON /OFF]** switch (its lamp goes off).

(c) Set the **[LOG/LIN]** switch to Log (its lamp comes on) and turn the rotary knob to 5 dB/div.

(d) Press the **[CENTER WAVELENGTH]** switch and turn the rotary knob to set the center wavelength to 1300.0 nm.

(e) Press the **[RESOLUTION]** switch and turn the rotary knob to set the resolution to 10 nm.

(f) Press the **[REFERENCE LEVEL]** switch and turn the rotary knob to select a reference level of –10 dB.

(g) Press the **[SWEEP WIDTH]** and turn the rotary knob to select a sweep width of 10 nm/div.

(h) Press the **[AVERAGE TIMES]** switch and turn the rotary knob to set the averaging times to 1.

(i) Press the **[REPEAT]** switch to start repetitive sweep.

(j) While observing the measurement results (displayed waveform), adjust the measurement conditions set up in steps (d) through (h).

(k) Press the **[STOP]** switch as needed to stop sweep operation, press the **[LABEL]** switch to carry out labeling, and press the **[COPY ON]** switch to produce a hard-copy printout.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(2) LED Light Source Measurement

(a)-(c) Perform the same steps as indicated in paragraph (1).

(d) Press the [CENTER WAVELENGTH] switch and turn the rotary knob to set the center wavelength to 850.0 nm.

(e) Perform the same step as indicated in paragraph (1).

(f) Press the [CENTER WAVELENGTH] switch and turn the rotary knob to set the reference level to -30 dBm.

(g) Press the [SWEEP WIDTH] switch and turn the rotary knob to select a sweep width of 50 nm/div.

(h)-(k) Perform the same steps as indicated in paragraph (1).

Typical LD and LED light source measurement examples are presented in Figs. 3-8 through 3-11.




Fig. 3-8 LD Light Source First Sweep Waveform

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-9 LD Light Source Last Sweep Waveform




Fig. 3-10 LED Light Source First Sweep Waveform

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-11 LED Light Source Last Sweep Waveform

3.5.6 Measurement Condition Setup

(1) Center Wavelength

Center wavelength setup is performed by either the **[CENTER WAVELENGTH]** switch or **[PEAK → CENTER/MARKER→ CENTER]** switch.

A) Center Wavelength Setup with **[CENTER WAVELENGTH]** Switch

(a) Press the **[CENTER WAVELENGTH]** switch. The current center wavelength setting appears in the interrupt display area and the soft key menu appears on the CRT. While noting the CRT display, turn the rotary knob as required to select a desired value. (Refer to Fig. 3-12.)

(b) When the desired value is equal to the value of soft key **<CTR WL ****.*>**, press that soft key. If the desired value is close to ****.*, press the soft key and then turn the rotary knob until the desired setting is obtained.

(c) When performing setup on the basis of the measurement start and measurement stop wavelengths, press soft keys **<START WL>** and **<STOP WL>** and then turn the rotary knob as needed.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

B) Center Wavelength Setup with [PEAK → CENTER/ MARKER→ CENTER] Switch

(a) When the displayed waveform peak wavelength is to be employed as the center wavelength setting, press the [PEAK→CENTER] switch.

(b) When the value indicated by the displayed wavelength marker is to be employed as the center wavelength setting, press the [MARKER→CENTER] switch.




Fig. 3-12 State Prevailing after [CENTER WAVELENGTH] Switch Activation

(2) Sweep Width

Press the [SWEEP WIDTH] switch. The soft key menu then appears on the CRT. Further, the current sweep width setting appears in the interrupt display area. Perform any one of the following procedures to select a setting of 0.5, 1, 2, 5, 10, 20, 50, 100, or 150 nm/div. (Refer to Fig. 3-13.)

(a) When performing setup on the per-div. basis, press soft key For setup on the full-span basis, press soft key . After such soft key activation, turn the rotary knob as needed to obtain a desired setting.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(b) Press soft key <ETC> so that soft key <****-****nm> appears. When the sweep range indicated by the displayed soft key is to be employed, press that soft key. The sweep width and center wavelength are then set up to suit the displayed sweep range.

(c) Press the [△W→SPAN] switch so that the soft key menu appears. The half-width is then determined by the currently defined method. The determined value is multiplied by the latched magnification power in the soft key menu. The value obtained is set up as the full span value (see Fig. 3-15).

(d) When the sweep width is 0.0 nm with the[HIGH SENSITTIVITY] switch OFF, press soft key <SWEEP TIME>. The time for left- to right-end screen sweep can then be set by rotary knob rotation to a value between 1 and 50 seconds. The selectable sweep time settings are 1, 2, 5, 10, 20, and 50 seconds. Note, however, that this sweep time setting applies to an averaging times setting of 1. The sweep time is indicated in the interrupt display area during REPEAT sweep (see Fig. 3-14).
(In the case the sweep time is approx.)

(e) Press soft key <MEAS WL INTVL> and then turn the rotary knob to set the sweep state data sampling interval to a value between 0.1 and 20 nm. The selectable settings are 0.1, 0.2, 0.5, 1, 2, 5, 10, and 20 nm. Note, however, that the setting must be smaller than half the full span (measurement stop wavelength - measurement start wavelength) and greater than (full span / 580) nm. The standard data sampling interval is (full span / 580) nm, and this setting is selected when soft key <INTVL NORMAL> is pressed.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-13 State Prevailing after **[SWEEP WIDTH]** Switch Activation




Fig. 3-14 State Prevailing after **<SWEEP TIME>** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-15 State Prevailing after [△W→ SPAN] Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(3) High-Sensitivity Measurement Mode Selection

When high-sensitivity measurement is to be made, press the **[HIGH SENSITIVITY]** switch. This causes the soft key menu to appear on the CRT. Press the **<H-SENS FAST>** key for relatively high level (higher than -60 dBm) light source measurement, or the **<H-SENS SLOW>** key for extremely low level (lower than -60 dBm) light source measurement. (Refer to Fig. 3-16.)




Fig. 3-16 State Prevailing after **[HIGH SENSITIVITY]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(4) Resolution

Press the **[RESOLUTION]** switch. The current resolution setting then appears in the interrupt display area. Turn the rotary knob as required to set the resolution to 0.1, 0.2, 0.5, 1, 2, 5, or 10 nm. (However, if the measurement wavelength range includes a wavelength 560 nm or shorter, the selectable resolution setting is between 0.1 and 5 nm.) (Refer to Fig. 3-17.)




Fig. 3-17 State Prevailing after **[RESOLUTION]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(5) Reference Level

Reference level setup is performed by either the [REFERENCE LEVEL] switch or [PEAK →REF LEVEL/ MARKER → REF LEVEL] switch.

A) Reference Level Setup with [REFERENCE LEVEL] Switch

(a) Press the [REFERENCE LEVEL] switch. The current reference level setting then appears in the interrupt display area. Turn the rotary knob to change the setting as desired. (Refer to Fig. 3-18.)

B) Reference Level Setup with [PEAK →REF LEVEL/ MARKER → REF LEVEL] Switch

(a) When the displayed waveform peak level is to be employed as the reference level, press the [PEAK →REF LEVEL] switch.

(b) When the value indicated by the level marker set up on the screen is to be used as the reference level, press the [MARKER →REF LEVEL] switch.

In either of the above cases, the acceptable reference level settings are as follows.

	HIGH-SENSITIVITY		
	OFF	ON	
		H-SENS FAST	H-SENS SLOW
REFERENCE LEVEL (dB)	0 to -42	0 to -60	0 to -70

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-18 State Prevailing after [**REFERENCE LEVEL**] Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(6) Averaging Times

Press the **[AVERAGE TIME]** switch. The current averaging times setting then appears in the interrupt display area. Vary the displayed setting as desired with the rotary knob. The selectable averaging times settings are 1, 5, 10, 20, 50, 100, 200, 500, and 1000 for the cases where the **[HIGH SENSITIVITY]** switch is OFF, or 1, 2, 5, 10, and 20 for the cases where the **[HIGH SENSITIVITY]** switch is ON. (Refer to Fig. 3-19.)




Fig. 3-19 State Prevailing after **[AVERAGE TIMES]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.7 Sweep Mode Setup

(1) REPEAT Sweep

To conduct repetitive sweep, press the [REPEAT] switch. This causes the switch lamp to come on and the soft key menu to appear on the CRT. Further, either soft key <SWEEP NORMAL> or <SWEEP MKR1-2> is latched and displayed in reverse field (white on black). Sweep is then conducted over the wavelength range indicated by the latched soft key (sweep takes places between the left and right ends of the screen when <SWEEP NORMAL> is latched or between markers 1 and 2 when <SWEEP MKR1-2>is latched). (Refer to Fig. 3-20.)

(a) Sweep between Markers 1 and 2

As long as wavelength markers 1 and 2 are set up on the screen, <SWEEP MKR1-2> key activation causes the system to sweep between markers 1 and 2. While sweep is being conducted between markers 1 and 2, the waveforms prevailing before marker 1-to-marker 2 sweep are displayed for the portions outside the sweep area.

While soft key <SWEEP MKR1-2> is latched, all switches except the following are inoperative. If an attempt is made to press such inoperative switches, the interrupt display area reads "ATTN 5".

[REPEAT] , [SINGLE] , [STOP] , [HIGH SENSITIVITY] , [RESOLUTION] , [AVERAGE TIMES] , [COPY ON] , [COPY FEED] , [LABEL ON/OFF] , [ADDRESS] , [LOCAL], [FAST].

When soft key <AUTO ON>is latched by the[PEAK WAVELENGTH] switch, the auto peak search function is not executed during sweep between markers 1 and 2. However, this does not change the <AUTO ON> latch status; therefore, the auto peak search operation is resumed when sweep is effected after the latched soft key has been changed from <SWEEP MKR1-2> to <SWEEP NORMAL>.

The above-mentioned limitations on marker 1-to- marker 2 sweep also apply to the SINGLE sweep modes.

(b) WAIT Time Setup

The inter-sweep pause time (WAIT time) can be set up with soft key <WAIT>.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

When the <WAIT> key is pressed, the soft key menu changes and the interrupt display area shows the current WAIT time. When no units of time are to be changed, just turn the rotary knob to obtain a desired value. When a certain unit is to be changed, press soft key <s> , <m>, or <h> and then turn the rotary knob to obtain a desired value.

The values selectable range from 1 to 59 for s (second) and m (minute) or from 1 to 24 for h (hour).

When soft key <ENTER> is pressed after the WAIT time has been selected with the rotary knob, WAIT time setup takes place and measurement starts immediately.

To minimize the WAIT time setting, press soft key <WAIT NORMAL>.




Fig. 3-20 State Prevailing upon [REPEAT] Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-21 State Prevailing upon <WAIT> Key Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(2) SINGLE Sweep

To conduct single sweep, press the **[SINGLE]** switch. This causes the switch lamp to come on and the soft key menu to appear on the CRT. Further, either soft key **<SWEEP NORMAL>** or **<SWEEP MKR1-2>** is latched and displayed in reverse field. Sweep is then conducted over the wavelength range indicated by the latched soft key. The **[SINGLE]** switch lamp goes off upon completion of sweep. (Refer to Fig. 3-22.)

NOTE: There are some limitations on the sweep between markers 1 and 2. Therefore, reference should be made to paragraph (1)-(a).




Fig. 3-22 State Prevailing upon **[SINGLE]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(3) MANUAL Sweep

The term "MANUAL sweep" refers to the manual sweep operation that is performed only enough to cover the range indicated by rotary knob rotation. This function is provided as a soft key option for the [STOP] switch.

When the [STOP] switch is pressed while sweep is conducted or stopped, soft key **<MANUAL SWEEP>** appears on the CRT. To effect MANUAL sweep, press that soft key. (Refer to Fig. 3-23.)

When the rotary knob is turned after soft key activation, sweep starts. If the [STOP] switch is pressed during sweep and then the system is placed in the MANUAL sweep mode, sweep starts from the point at which the last sweep was terminated by [STOP] switch activation.




Fig. 3-23 State Prevailing upon [STOP] Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.8 Display Method Setup

(1) LOG/LIN Scale

Whenever the [LOG/LIN] switch is pressed, the on- screen Y-axis scale alternates between the LOG and LIN types.

When the LOG scale is employed, the switch lamp is lit. Further, the current Y-axis scale is indicated in the interrupt display area. In the HIGH-SENSITIVITY-OFF state, scale change can be effected up to 0.5, 1, 2, and 5 dB/div with the rotary knob. In the HIGH- SENSITIVITY-ON state, scale change can be effected further up to 10 dB/div. Also, the displayed waveform is updated in accordance with the changed scale. (Refer to Figs. 3-24 and 3-25.)




Fig. 3-24 LOG Scale Screen

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-25 LIN Scale Screen


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(2) NORMALIZE

When the **[NORMALIZE]** switch is pressed, the indication "NORM" appears at the upper right corner of the screen. Further, normalization is effected with reference to the displayed waveform peak level value, and then the waveform is changed with the peak value considered to be 0 dB in the LOG scale mode or 1 in the LIN scale mode. (Refer to Figs. 3-26 and 3-27.)

However, if the waveform peak level value is extremely small, the associated ATTN number is displayed and waveform updating does not take place.


The NORMALIZE function can be exercised only while sweep is stopped with the [3D] switch OFF (see subsection (4)).



Before Execution

61


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




After Execution

Fig. 3-26 NORMALIZE Function Execution Examples (LOG Scale Mode)

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



Before Execution



After Execution

Fig. 3-27 NORMALIZE Function Execution Examples (LIN Scale Mode)


Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(3) DOMINANT

When the **[DOMINANT]** switch is pressed, the indication "DOM" appears at the upper right corner of the screen. Further, the displayed waveform is subjected to human visual sensitivity correction and normalized. (Refer to Fig. 3-28.)


However, if the waveform wavelength range is not within the scope of human visual sensitivity, the associated ATTN number is displayed and waveform updating does not take place.

The DOMINANT function can be exercised only while sweep is stopped with the **[3D]** switch OFF (see subsection (4)).



Before Execution

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



After Execution

Fig. 3-28 DOMINANT Function Execution Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(4) 3-D (Three-dimensional) Display

The 3-D display function is useful for observing measured waveform changes with time. Up to 16 waveforms can be measured at certain time intervals and displayed altogether on the CRT.

(a) When the **[3D]** switch is pressed while sweep is stopped, the switch lamp comes on and the system goes into the 3-D display mode, causing the CRT to show as indicated in Fig. 3-29.




Fig. 3-29 3-D Mode Initial Screen

(b) Sweep starts when the **[REPEAT]** or **[SINGLE]** switch is pressed. In **[REPEAT]** sweep, the first swept waveform is displayed in the forefront row of the screen, with the number "00" attached to the Z-axis. The second swept waveform then appears in the forefront row of the screen, with the number "01" attached to the Z-axis. At this time, the first waveform and its number attached to the Z-axis are scrolled backward.

Whenever sweep is effected subsequently, the last- swept waveform is displayed in the forefront row of the screen, incrementing the Z-axis

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


number by 1 and scrolling the preceding waveform backward. The Z-axis number indicates the number of sweeps minus 1. This Z- axis count increases in hexadecimal notation from 00 to FF (up to the decimal equivalent of 255) and then resets itself back to 00. Also note that up to 16 waveforms are displayed on the screen at a time; therefore, when 17th and subsequent sweeps are performed, the old displayed waveforms go off one by one.

In **[SINGLE]** sweep, on the other hand, single sweep takes place. The swept waveform is displayed in the same manner as in **[REPEAT]** sweep. Whenever the switch is pressed, the measured waveform appears in the forefront row of the screen, scrolling older waveforms backward.


When sweep is resumed after the **[STOP]** switch has been pressed to halt an ongoing **[REPEAT]** or **[SINGLE]** sweep operation, newly swept waveforms are displayed subsequently to the previously swept ones. If new waveforms are to be displayed with previously swept waveforms deleted, clear such old waveforms by pressing the **[GRAPH CLEAR]** switch while sweep is stopped.

(c) Soft key **<ANGLE CHANGE>** is used to change the waveform display angle in 10 degree steps from + 50 to -50 degrees. When the **<ANGLE CHANGE>** key is pressed, the current display angle setting appears in the interrupt display area and the soft key menu changes. Turn the rotary knob to select a desired angle, and then press the **<EXECUTE>** key to update the displayed waveform (see Fig. 3-30). Note that the soft key is displayed only when sweep operations are not performed.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



Angle: -30°



Angle: +30°

Fig. 3-30 3-D Display Example

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(d) The 16 waveforms displayed on the screen are stored in memories 0 through F. The memory number correlates to the rightmost digit of the number attached to the Z- axis of each waveform. For instance, when the Z-axis number is 01, the associated waveform is stored in memory 1, or when the Z-axis number is 7F, the waveform is stored in memory F. When one of the 16 displayed waveforms is to be observed, press the [MEMORY], [RECALL], and [*] (*: any number between 0 and F) switches in sequence.

(e) In contrast to the other display modes, the following restrictions are imposed on 3-D display.

1) In the 3-D display mode, it is impossible to change the center wavelength, sweep width, reference level, or other measurement conditions. Therefore, it is necessary that such conditions be optimized before the 3-D display mode is selected. If the measurement conditions are found in the 3-D mode to be in need of some changes, exit the 3-D display mode to return to a regular screen mode, change the measurement conditions, and then re-enter the 3- D display mode.
2) The markers cannot be used.
3) If the [REPEAT] or [SINGLE] switch is pressed during sweep, the system responds to such switch activation after completion of on-going sweep operation (in a regular screen mode, the system restarts a sweep operation immediately after switch activation). It should also be noted that the soft key options provided for the [REPEAT] or [SINGLE] switch cannot be used. Sweep is conducted in the NORMAL mode (sweep is conducted from the left- to the right-end of the screen). However, the WAIT time setting for REPEAT sweep is as specified with soft key <WAIT> in a regular screen mode.
4) Only the following switches are operative in the 3-D display mode.
[REPEAT], [SINGLE], [STOP], [ADDRESS], [LOCAL], [COPY ON] [COPY FEED], [FAST], [LABEL ON/OFF], [LOG/LIN], [3D] , [GRAPH CLEAR] rotary knob, soft keys.
All the above switches except [COPY ON], [LABEL ON/OFF], [LOG LIN], [ADDRESS], [LOCAL], [3D] and [GRAPH CLEAR] are operative even during sweep.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.9 Marker Usage

Two types of markers are provided: level (horizontal line) markers and wavelength (vertical line) markers. Two fixed markers and one movable marker are provided each for the two marker types.

A marker appearing after **[PEAK LEVEL]** or **[PEAK WAVELENGTH]** switch activation is called the "movable marker." This marker moves in coordination with the lower-end display area marker value as the rotary knob is turned.

When the **[MARKER 1]** or **[MARKER 2]** switch is pressed with the movable marker placed in any desired position, the marker is fixed in that position. Such a marker is then called the "fixed marker." Both the level and wavelength markers can be designated as markers 1 and 2.

All these markers are used for wavelength and level difference measurements and peak wavelength, peak level, or half-width search.

(1) Level Marker

Press the **[PEAK LEVEL]** switch when the displayed waveform peak level is to be searched for or a level marker is to be displayed.

(a) When the **[PEAK LEVEL]** switch is pressed while a waveform is displayed, the system searches for the displayed waveform peak level value, places a level marker at the peak level value position, and displays the marker value in the LMARK column (see Fig. 3-31).

(b) If no waveform is displayed when the **[PEAK LEVEL]** switch is pressed, the system places a level marker at the bottom of the screen and displays the marker value in the LMARK column.

(c) When the rotary knob is turned clockwise after marker display, the marker moves upward. Counterclockwise rotary knob rotation causes the marker to move downward. The marker value indicated in the LMARK column varies with the marker movement in real time.

(d) When the **[MARKER CLEAR]** switch is pressed, all displayed markers and the marker values in the WMARK and LMARK columns are cleared.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-31 Peak Level Measurement Example

(2) Wavelength Marker

Press the **[PEAK WAVELENGTH]** switch when the displayed waveform peak wavelength is to be searched for or a wavelength marker is to be displayed on the screen.

(a) When the **[PEAK WAVELENGTH]** switch is pressed, the system displays the soft key menu on the CRT, searches for the displayed waveform peak wavelength, places a marker at the peak wavelength position, and displays the marker value in the WMARK column (see Fig. 3-32).


(b) If no waveform is displayed when the **[PEAK WAVELENGTH]** switch is pressed, the system places the marker at the right-hand end of the screen and displays the marker value in the WMARK column.

(c) When soft key **<CROSS ON>** is activated, the displayed marker functions as the cross marker. When the key **<CROSS OFF>** is activated, on the other hand, the displayed marker serves as the wavelength marker (see Fig. 3-32).

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


(d) When soft key **<AUTO ON>** is activated, the system searches for the peak wavelength in each REPEAT sweep and displays the marker. When the **<AUTO OFF>** key is activated, on the other hand, the system searches for the peak wavelength upon **[PEAK WAVELENGTH]** switch activation only, and not in every sweep. Note, however, that even if the **< AUTO ON>** key is activated, marker display does not take place while sweep is being conducted between markers 1 and 2.

(e) When the rotary knob is turned clockwise after marker display, the marker moves to the right. Counterclockwise rotary knob rotation causes the marker to move to the left. The marker value indicated in the WMARK column (plus the value indicated in the LMARK column when the cross marker is used) varies with the marker movement in real time.



(Wavelength Marker)

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



(Cross Marker)

Fig. 3-32 Peak Wavelength Measurement Example

(3) Center Wavelength/Reference Level Setup by Markers

The displayed wavelength and level markers can be used to set up the center wavelength or reference level.

(a) When the [MARKER→ CENTER] switch is pressed with a movable wavelength marker displayed on the CRT, the displayed marker value is selected as the center wavelength setting. If the marker value cannot be selected as the center wavelength, the associated ATTN number appears in the interrupt display area.

(b) When the [MARKER→ REF LEVEL] switch is pressed with a movable level marker displayed on the CRT, the displayed marker value is selected as the reference level. If the marker value cannot be selected as the reference level, the associated ATTN number appears in the interrupt display area.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.10 Level Difference Measurement

A marker appearing upon [**PEAK LEVEL**] or [**PEAK WAVELENGTH**] switch activation is called the "movable marker." This marker moves in coordination with the lower-end display area marker value as the rotary knob is turned.

When the [**MARKER 1**] or [**MARKER 2**] switch is pressed with the movable marker placed in any desired position, the marker is fixed in that position. Such a marker is then called the "fixed marker." Both the level and wavelength markers can be designated as markers 1 and 2.

The fixed marker function is exercised for level difference and wavelength difference (see the next section) measurements.

(1) Press the [**PEAK LEVEL**] switch to display level markers.


(2) When the [**MARKER 1**] (or [**MARKER 2**]) switch is pressed after appropriate rotary knob rotation, the system fixes marker 1 (or 2) in the current position and displays the fixed marker 1 (or 2) value in the LMARK column. The movable marker value is then assigned to the unfixed marker 2 (or 1), and the right- hand end of the LMARK column shows the level difference between the fixed and movable markers.

(3) When the rotary knob is turned, the level difference indicated in the LMARK column varies in real time with the movable marker movement.


(4) When the [**MARKER 2**] (or [**MARKER 1**]) switch is pressed to fix the marker that was not fixed in step (2), the level difference indicated in the LMARK column is also fixed.

(5) The indicated level difference refers to the value "marker 2 - marker 1" when the Y-axis is in the LOG scale mode or the value "marker 2 / marker 1" when the Y-axis is in the LIN scale mode.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




State Prevailing upon **[PEAK LEVEL]** Switch Activation



State Prevailing upon **[MARKER 1]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



State Prevailing upon **[MARKER 2]** Switch Activation

Fig. 3-33 Level Difference Measurement Examples

3.5.11 Wavelength Difference Measurement


(1) When the wavelength markers are used after **[PEAK WAVELENGTH]** switch and **<CROSS OFF>** key activation, the same display indications are given as described in section 3.5.10 except that the level markers are replaced by the wavelength markers.

(2) When the **<CROSS ON>** key is activated to use a cross marker, **[MARKER 1]** (or **[MARKER 2]**) switch activation fixes both the wavelength marker and the level marker that is perpendicular to the wavelength marker. The movable marker value is then indicated in the unfixed marker value area of the MARK column while the right-hand end of the same column shows the level and wavelength differences between the fixed and movable markers.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


(3) When the rotary knob is turned, the level and wavelength differences indicated in the MARK column vary in real time with the movable marker movement.

(4) When the **[MARKER 2]** (or **[MARKER 1]**) switch is pressed to fix the marker that was not fixed in step (2), the level and wavelength differences indicated in the MARK column are also fixed. (Refer to Fig. 3-34.)




State Prevailing upon **[PEAK WAVELENGTH]** Switch Activation

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



State Prevailing upon **[MARKER 1]** Switch Activation



State Prevailing upon **[MARKER 2]** Switch Activation

Fig. 3-34 Wavelength Difference Measurement Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.12 Half-Width Measurement

There are several half-width definitions. The AQ- 6310 makes half-width measurement by the envelope method, RMS (root-mean-square) method, or threshold method.

(1) To measure the half-width of a waveform displayed on the CRT, press the **[HALF-WIDTH SEARCH]** switch.


(2) The soft key menu then appears on the CRT, and the system calculates the half-width according to the latched algorithm and displays a wavelength marker.

(3) The wavelength difference between markers 1 and 2 is indicated in the WMARK column. This indicated value represents the half-width.


(4) When determining the half-width according to another algorithm, press the associated soft key. The system then performs calculations immediately, displays a wavelength marker, and indicates the half-width in the WMARK column.

(5) Fig. 3-35 presents the examples of measurements made by various methods.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




<ENV> Method



<RMS> Method

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



<THRESH 15 dB> Method

Fig. 3-35 Half-Width Measurement Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.13 Half-Width Algorithms

(1) Envelope Method

The half-width is determined according to the following algorithm after **[HALF-WIDTH SEARCH]** switch activation.




Fig. 3-36 Half-Width Determination Algorithm (ENV Method)

1) Mode Determination

The level 10 dB lower than the peak level is referred to as mode judgment level I and the level 7 dB lower than the peak level is called mode judgment level II (see Fig. 3-37). The portions above mode judgment level I are considered mode candidates, and when the peak level of an individual mode candidate is higher than mode judgment level II, such a mode candidate is treated as one mode.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com





Fig. 3-37 Mode Determination

2) Half-Width Determination

(a) One Mode




The half-width must be determined at a level 3 dB lower than the peak.

$\Delta\lambda = \lambda_2 - \lambda_1$

Fig. 3-38

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(b) Two Modes Not More than 3 dB away from Peak Level




The distance between the peak positions of the two modes must be regarded as the half-width.

$\Delta\lambda = \lambda_2 - \lambda_1$

Fig. 3-39

(c) Two Modes Whose Levels Are More than 3 dB apart from Each Other



As indicated in Fig. 3-40, the distance between the peak position (λ_1) and 3-dB difference position (λ_3) must be regarded as the half-width.

Fig. 3-40

$$\Delta\lambda = \lambda_3 - \lambda_1 = -3^* \frac{\lambda_2 - \lambda_1}{P_2 - P_1}$$

$$\lambda_3 = \lambda_1 - 3^* \frac{\lambda_2 - \lambda_1}{P_2 - P_1}$$

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(d) More than Two Modes




Fig. 3-41

(i) Locate the rightmost and leftmost modes.

(ii) When the rightmost and leftmost mode levels are less than 3 dB lower than the peak level, they are regarded as the half-width points (λ_3 and λ'_3).

(iii) When the rightmost and leftmost mode levels are more than 3 dB lower than the peak level, check the inner modes until you encounter the mode levels less than 3 dB lower than the peak level.

(iv) Draw straight lines so as to connect the modes more than 3 dB lower than the peak level, to the modes less than 3 dB lower than the peak level.

(v) The intersections of the drawn straight lines and 3-dB-lower-than-peak-level lines must be regarded as the half-width points (λ_3 and λ'_3), as indicated in Fig. 3-41.

(vi) When all mode levels except the peak level are more than 3 dB lower than the peak level, draw straight lines so as to connect the right- and left-hand side highest modes to the peak level point. In this case, the intersections of the drawn straight lines and 3-dB- lower-than-peak-level lines must be regarded as the half-width points.

$$\Delta\lambda = \lambda_3 - \lambda'_3$$

$$\lambda_3 = \frac{\lambda_2 - \lambda_1}{P_2 - P_1}(P_0 - 3 - P_1) + \lambda_1$$

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(2) RMS (Root-Mean-Square) Method




Fig. 3-42

(a) As indicated in the above figure, the values greater than 1/100 of the peak level value are used for half-width calculation. The calculation procedure is described below.

(b) The 0- to 580-dot resolutions are expressed along the on-screen X-axis. When the wavelength related to each dot position is expressed as λi and its level as Pi, the center wavelength (λo) is determined by the following formula.

$$\lambda o = \frac{\sum_{i=0}^{580} Pi \cdot \lambda i}{\sum_{i=o}^{580} Pi}$$

(c) Calculate the half-width (Δλ) by substituting the determined center wavelength (λo) into the following equation.

$$\Delta\lambda = \sqrt{\frac{\sum_{i=0}^{580} Pi \cdot (\lambda i)^2}{\sum_{i=o}^{580} Pi} - \lambda o^2}$$

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(d) Set up wavelength marker 1 at the position expressed by ($\lambda_o - \Delta\lambda/2$) and wavelength marker 2 at the position prescribed by ($\lambda_o + \Delta\lambda/2$). Further, place a movable marker at the position of center wavelength (λ_o). The half-width is then expressed as (marker 2 - marker 1).

(e) When the **<RMS ×2.0>** or **<RMS ×2.35>** is activated, the value determined in step (d) is multiplied by 2.0 or 2.35, and the product is displayed as the half-width.

(3) Threshold Method

(a) Displayed spectrum waveforms that meet all the following conditions are treated as the modes.

(1) Level	$Pj \geqq Pmax - K_1$	(K_1 = **dB)
(2) Difference between peak and trough	$Pj - Lj \geqq K_2$	(K_2 = 3 dB)
Difference between peak and trough	$Pj - Lj\text{-}1 \geqq K_2$	(K_2 = 3 dB)
(3) Spectral width	$Lj - Lj - 1 \geqq K_3$	(K_3 = 0.05 nm)




Fig. 3-43

(b) Locate the intersections of the rightmost and leftmost modes and threshold value (** dB) line. When these intersections are expressed as λ_1 and λ_2, the half-width = λ_2-λ_1.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(c) When the <THRESH 1/e^2> key is activated, the value e^2 (about 7.389) is substituted for K_1 in the above equations to find a solution in the same manner.

(d) When the threshold method is employed, the wavelength markers are placed in the positions of λ_1, λ_2, and $(\lambda_1 + \lambda_2) / 2$. Further, the number of displayed waveform modes is determined in accordance with the conditions indicated in paragraph (a) and displayed in the interrupt display area.

3.5.14 MEMORY and DATA MEMORY

The AQ-6310 has 18 memories which are used to store waveforms, measurement conditions, and other relevant data. As regards 16 out of the 18 memories, label character switch ([0] through [9] and [A] through [F]) activation after **[MEMORY SAVE]** or **[MEMORY RECALL]** switch depression effects data storage or recall. The measurement condition memories are nonvolatile and therefore retain their stored data even after power turn-off. However, the waveform data memories lose their stored data upon power off. For the remaining two memories, data storage takes place upon **[DATA MEMORY Acc A]** or **[DATA MEMORY Acc B]** switch activation, and data recall occurs when the **[RECALL]** and **[Acc A]** or **[Acc B]** switches are pressed in sequence. The **[Acc A]** and **[Acc B]** memories have an accumulator function. Therefore, it is possible to perform stored waveform data subtractions and display the results. These memories are nonvolatile and therefore do not lose their stored waveforms and measurement conditions upon power turn-off.

Subtraction computations cannot be performed with **[MEMORY 0]** through **[MEMORY F]** contents alone. However, the **[RECALL]** and **[*]** (*: number between 0 and F) switches can be pressed to call up desired data onto the screen and then the **[DATA MEMORY Acc A]** or **[DATA MEMORY Acc B]** switch pressed so that the data is stored in the accumulator to carry out subtraction.

NOTE

The **[MEMORY 0]** through **[MEMORY F]** memories (16 in total) are used as the 3-D display buffers. Therefore, when measurement is made with 3-D display feature activated, the memories lose all previously saved waveforms.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.15 Waveform Storage and Recall by MEMORY Functions

When a displayed waveform is to be stored into memory or a waveform stored in memory is to be displayed on the screen, perform the following procedure.

(1) Make measurement so that a waveform to be stored is displayed on the screen, and be sure that sweep is stopped.

(2) Press the [MEMORY SAVE] switch and then any one of label character switches [0] through [9] or [A] through [F] (numbers or alphabetical letters are printed in green). The waveform and measurement conditions are then stored into memory. Upon [MEMORY SAVE] switch activation, the upper right-hand corner of the screen reads "SAVE". Further, when any one of [0] through [F] switches is pressed subsequently to [MEMORY SAVE] switch activation, the screen reads "MEM *" (*: number between 0 and F) to indicate the name of the memory into which the data have been saved.

(3) To recall a waveform stored in memory, press the [MEMORY RECALL] switch and then any one of label character switches [0] through [9] or [A] through [F]. Upon [MEMORY RECALL] switch activation, the upper right-hand corner of the screen reads "RCL". Further, when any one of the [0] through [F] switches is pressed subsequently to [MEMORY RECALL] switch activation, the screen reads "MEM *" (*: number between 0 and F) to indicate the name of the memory being recalled, and then shows the recalled waveform. (Refer to Fig. 3-44.)

(4) If a switch other than label character switches [0] through [9] and [A] through [F] is pressed in step (2) or (3), the keying operation is aborted. When such a condition occurs, press the [SAVE] or [RECALL] switch again and proceed to the next keying operation.
If the specified memory stores no waveforms or measurement conditions when the [RECALL] operation is performed, the operation is aborted and the associated ATTN number appears on the screen.


(5) In the [0] through [9] and [A] through [F] memories, the waveform storage areas are volatile and the measurement conditions storage areas, nonvolatile. Therefore, when the power is turned off, the waveform data

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


are lost. However, the measurement condition data will be retained so that they can be recalled by pressing the [RECALL] and [*] (*: number between 0 and F) keys in sequence after power restoration.

(6) When the [GRAPH CLEAR] switch is pressed, the displayed waveform is cleared. Note that only the waveform is cleared at this time, and that the scale marks, lower-end display area, or memory contents are not cleared.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



State Prevailing upon **[MEMORY SAVE]-[1]** Switch Activation



State Prevailing upon **[MEMORY RECALL]-[1]** Switch Activation

Fig. 3-44 MEMORY Function Execution Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.16 Data Overlap Display

The following procedure is performed to display overlapping waveform data after they have been derived from accumulators A and B and memories 0 through F. The following explanation applies to the cases where accumulators A and B are used.

(1) Make measurement so that a waveform is displayed on the screen. Be sure that sweep is stopped.

(2) Press the **[Acc A]** or **[Acc B]** switch to store the displayed waveform into accumulator A **(Acc A)** or accumulator B **(Acc B)**. At this time, the upper right- hand corner of the screen reads "Acc A" or "Acc B" to indicate which accumulator has stored the waveform.

(3) Repeat steps (1) and (2) to store a waveform into the other accumulator.


(4) Press the **[DATA MEMORY RECALL]** and **[Acc A]** switches in sequence so that the Acc A waveform is recalled and displayed again on the screen. At this time, the upper right-hand corner of the screen reads "Acc A" as is the case with step (2).

(5) Press the **[DATA MEMORY RECALL]** and **[Acc B]** switches in sequence. The Acc B waveform is then recalled and appears on the screen with the Acc A waveform continuously displayed.


(6) To provide overlap display, it is necessary that the Acc A and Acc B waveform measurement conditions coincide in center wavelength, sweep width, and resolution. However, if the measurement conditions differ in any one of them, overlap display does not take place so that the screen shows only the waveform recalled later.
If the Acc A and Acc B waveforms coincide in the above-mentioned measurement conditions but differ in other conditions, overlap display is provided with the "*" mark displayed along the Y-axis or in the associated lower-end display area column. In such a case, the waveform recalled later is displayed after being adjusted to suit the measurement conditions for the previously recalled waveform.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(7) The displayed waveforms are cleared when the [GRAPH CLEAR] switch is pressed.



State Prevailing upon [DATA MEMORY RECALL]-[Acc A] Switch Activation



State Prevailing upon [DATA MEMORY RECALL]-[Acc B] Switch Activation

Fig. 3-45 Overlap Display Function Execution Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.17 Data Subtraction Display by DATA MEMORY Function

The data stored in the Acc A and Acc B can be subjected to subtraction (A - B when the Y-axis is in the LOG scale mode or A / B when the Y-axis is in the LIN scale mode) and displayed, using the [DATA MEMORY A-B A/B] or [DATA MEMORY B-A B/A] switch.

The following subsections present the examples of optical fiber loss wavelength characteristic and optical filter transmittance measurements that are made using the DATA MEMORY data subtraction function.

(1) Optical Fiber Loss Wavelength Characteristic Measurement

(a) Set the AQ-6310 measurement conditions as follows.
[CENTER WAVELENGTH] <START WL>: 700nm, [CENTER WAVELENGTH] <STOP WL>: 1700 nm, [SWEEP WIDTH] <MEAS WL INTVL>: 10 nm, [RESOLUTION]: 10nm, [REFERENCE LEVEL]: – 20 dB, [AVERAGE TIMES]: 1, [HIGH SENSITIVITY] <H-SENS SLOW>, [LOG/LIN]: 10 dB/div

(b) After set the wavelength 700 to 1000 nm with AQ-4303B White light source. (See Note 1)
(700 to 1750 nm, it is not possible to measure 700 nm or less.)
As indicated in Fig. 3-46, connect the AQ-6310 to a white light source with the reference (short) fiber. Press the [SWEEP SINGLE] swith to make reference spectrum measurement.

(c) Press the [Acc A] switch to store the reference spectrum into the Acc A.

(d) As shown in Fig. 3-46, replace the reference fiber with the fiber under test. Press the [SWEEP SINGLE] switch to make comparison spectrum measurement.

(e) Press the [Acc B] switch to store the comparison spectrum into the Acc B.

(f) To subtract the comparison value from the reference value, press the [A-B A/B] switch. The soft key menu then appears on the screen, and the value obtained by subtracting the Acc B data from the Acc A data is subjected to automatic scaling and displayed. At this time, the upper

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

right-hand corner of the screen reads "A-B" to indicate that the (A - B) data is displayed. The resultant displayed waveform shows the loss wavelength characteristic.

(g) The automatic scaling display function exercised upon [A-B A/B] or [B-A B/A] switch activation is useful because the optimum scale is obtained at all times. However, when two or more data are to be compared against the reference data, the scale unexpected varies causing inconvenience to the measurement personnel. To avoid such inconvenience, the Y-axis scale should be manually changed after automatic scaling.


To effect such Y-axis scale change, press soft key **MANUAL** after automatic scaling display. The waveform is then updated with the Y-axis bottom considered to be 0 dB (or 1 in the LIN scale mode) when the [A-B A/B] key is activated or with the Y-axis top considered to be 0 dB when the [B-A B/A] key is activated. At this time, the interrupt display area shows the prevailing Y scale. When the rotary knob is turned to vary the value in this state, the Y-axis scale and waveform are updated at the same time.

(h) The above data subtraction processing takes place only when the measurement conditions for the waveforms stored in the Acc A and Acc B coincide in center wavelength, sweep width, and resolution. If the Acc A and Acc B waveforms differ in any one of them, the system displays the associated ATTN number on the screen and does not execute the data subtraction display function.

If the Acc A and Acc B data coincide in the above- mentioned three conditions but differ in averaging times setting and HIGH-SENSITIVITY ON/OFF status, subtraction display is provided with the "*" mark displayed in the associated lower-end display area column and the Acc B averaging times setting adjusted to agree with the Acc A value.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(Reference Value Measurement)



(Comparison Value Measurement)





Fig. 3-46 Loss Wavelength Characteristic Measurement Setup

Note 1


This unit provides for measurement throughout a range of 700 to 1700 nm, using white light source wavelength over a 700 to 1000 nm range. This is because the AQ-4303B cuts the wavelength of 700 nm or less but passes the light of 1000 nm or longer wavelength. Conventional spectroscopes pass light of integral multiples of a certain wavelength so that when they are used over a wavelength range of 1000 nm or more, they must be set in the range of 1000 to 1800 nm. This unit, however, eliminates the need of such settings because it contains the filter to cut high-order light on the longer-wavelength side.

When 1700 nm or longer wavelength measurements are to be made, set the AQ-4303B for a wavelength range of 1000 to 1800 nm. When this wavelength range setting is employed, wavelengths of 1000 to 1750 nm can be covered.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




[Acc A] Switch Activated after Reference Fiber Measurement



[Acc B] Switch Activated after Fiber-under-Test Measurement

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



[A-B A/B] Switch Activated

Fig. 3-47 Loss Wavelength Characteristic Measurement Examples

(2) Optical Filter Transmittance Measurement (600 to 1700 nm)

(a) After set the wavelength 700 to 1000 nm with AQ-4303B White light source.
(700 to 1750 nm, it is not possible to measure 700 nm or less.)
As indicated in Fig. 3-48, measure the reference spectrum and store it into the Acc A.

(b) Insert the optical filter, as indicated in Fig. 3- 48, and measure the comparison spectrum and store it into the Acc B.

(c) Press the [B-A B/A] switch to subtract the reference value from the comparison value. The soft key menu then appears on the screen, and the value obtained by subtracting the Acc A data from the Acc B data is subjected to automatic scaling and displayed. At this time, the upper right-hand corner of the screen reads "B-A" to indicate that the (B - A) data is displayed. The resultant displayed waveform shows the transmittance.

(d) When the Y-axis scale is to be changed, perform the same procedure as indicated in step (1)-(f).

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com






Fig. 3-48 Optical Filter Transmittance Measurement Setup




[Acc A] Switch Activated after Unfiltered Data Measurement

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



[Acc B] Switch Activated after Filtered Data Measurement



[B-A B/A] Switch Activated

Fig. 3-49 Optical Filter Transmittance Measurement Examples

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.18 Printer Operating Procedure

The AQ-6310 incorporates a printer that produces hard- copy printout of screen contents. Print mode selection (vertical or horizontal screen contents printing) and date/time printing feature ON/OFF can be controlled with soft keys.

(1) When the **[COPY ON]** switch is pressed during sweep stop periods, the currently displayed screen contents are printed onto the printer. That is, the screen contents prevailing upon switch activation are stored and printed out.

(2) When the **[COPY ON]** switch is pressed, the screen display changes as indicated in Fig. 3-50. If the displayed date or time is to be changed at this time, press soft key **<PAUSE>** within about 2 seconds after **[COPY ON]** switch activation. When **<PAUSE>** is not pressed, the system concludes that there is no need of changing the data, starts printing about 2 seconds later, and, upon completion of printing, returns to the normal screen mode.

(3) When the **<PAUSE>** key is pressed, the cursor appears on the screen as shown in Fig. 3-51 with the soft key menu changed. Move the cursor with the **<→>** key and turn the rotary knob to change the settings as desired.

(4) When selecting the direction of printing or turning ON or OFF the date/time printing feature, press the **<ETC>** key while the soft keys are conditioned as indicated in step (3). After the soft key menu is changed, press the **<HORIZE>** key for horizontal printing, the **<VERT>** key for vertical printing, the **<DATE & TIME P ON>** key for date/time printing feature ON, or the **<DATE & TIME P OFF>** key for date/time printing feature OFF. Note that either one of the two printing direction keys and either one of the two date/time printing feature keys are latched and kept effective until their conditions are changed.
Fig. 3-49 presents a date/time printout example.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(5) When date/time change is completed in step (3), press soft key <CONT>. The system then starts printing, and returns to the regular screen mode upon completion of printing.

(6) To halt an ongoing printing operation, press the [COPY ON] switch again.

(7) The recording paper is fed while the [COPY FEED] switch is pressed.

(8) When loading the recording paper, proceed as directed below.
   (a) Open the cover located on the top panel of the AQ-6310. Set the recording paper on the paper holder. Be sure to position the recording paper so that when its leading end is pulled outside, its outer side faces up.
   (b) Raise the head lift-up lever into the [FREE] position. Remove any residual paper from the head section. Thread the newly loaded (in step (a)) paper leading end between the paper guide and platen, and then advance the paper about 5 cm (see Fig. 3-52).
   (c) Lower the head lift-up lever into the [LOCK] position. Press the [FEED] switch until the recording paper leading end comes out of the top plate paper outlet.
   (d) Close the top panel cover.

(9) If the [COPY ON] switch is pressed with the head lift-up lever raised (it must be lowered after paper loading), the screen looks like Fig. 3-53. When such a condition occurs, open the top panel cover, lower the head lift-up lever, and press soft key <READY>. The system then returns to the state indicated in step (1). To return to the regular screen mode without lowering the head lift-up lever after the screen looks like Fig. 3-53, press the [COPY ON] switch again.

(10) If the [COPY ON] switch is pressed with no recording paper remaining or the recording paper runs out during printing, the screen looks like Fig. 3-54 to prompt for recording paper loading. When such a condition exists, load new recording paper as directed in step (8) and press soft key <READY>. The system then returns to the state indicated in step (1). To

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

return to the regular screen mode without performing recording paper loading, press the **[COPY ON]** switch again.

NOTE: When the **[COPY ON]** switch is activated, battery backup is not provided for date/time display. Therefore, when the power is turned off, the date/time display system is deactivated, and when the power is turned back on, the date/time display system restarts. When accurate time information is needed, it is necessary to correct the displayed time after each power turn-on. It should also be noted that the date display operation is not controlled in perfect accordance with the calendar. Therefore, the displayed date needs to be corrected at the end of some months.




Fig. 3-50 **[COPY ON]** Switch Activated

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-51 <PAUSE> Key Pressed after [COPY ON] Switch Activation




Fig. 3-52 Recording Paper Loading

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-53 **[COPY ON]** Switch Activated with Head Lifted Up




Fig. 3-54 **[COPY ON]** Switch Activated with No Recording Paper Loaded

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.19 Automatic Calibration by CALIBRATION Function

(1) Connect the AQ-4302 He-Ne laser (option) to the AQ- 6310 with a short fiber.


(2) When the **[CALIBRATION]** (632.8 nm) switch is pressed during sweep stop periods, the switch lamp comes on and the following message and soft key menu appear on the CRT.

(3) When soft key **<START>** is pressed, calibration is automatically effected to 632.8 nm. If the peak level is below -30 dBm, the associated ATTN number appears in the interrupt display area to indicate that the level is too low for calibration.

(4) To stop an ongoing calibration, press the **[CALIBRATION]** (632.8 nm) switch again.


(5) The calibration range is ±5 nm. If the deviation is outside this range, calibration does not take place and the associated ATTN number appears on the CRT.

(6) The CALIBRATION function can be executed only when no sweep operations are performed.



Automatic Calibration System

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



Screen Display Prevailing upon **[CALIBRATION]** (632.8 nm) Switch Activation

3.5.20 Comment Display by LABEL Function

(1) When the **[LABEL ON/OFF]** switch is pressed during sweep stop periods, the switch lamp comes on, with the cursor and soft key menu appearing on the CRT.

(2) Turn the rotary knob to move the cursor into a desired position within the label area.

(3) Upon each switch activation, the green letter marking provided to the left of each switch appears in the uppermost area of the CRT. This letter marking indication position moves from left to right as the switches are pressed one by one.

(4) When soft key **<SHIFT>** is pressed, the green letter marking indicated to the right of a switch takes effect only for the first switch pressed after **<SHIFT>** key activation.

(5) For label deletion, press soft key **<CLEAR>**

(6) When labeling is completed, press the **[LABEL ON/OFF]** switch, causing the switch lamp to go off. The system then returns to the normal mode.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-55 LABEL Function Execution Example

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.5.21 Shutdown Procedure

(1) Turn OFF the **[POWER]** switch.

(2) Disconnect the **[INPUT]** connector and measurement cords from the apparatus.

(3) Disconnect the power cord plug from the power source (wall outlet). Wind the power cord around the take-up foot on the rear side of the apparatus.

3.6 GP-IB Control

The AQ-6310 can be externally controlled using the GP- IB (conforming to the IEEE-488 bus standard). To exercise external control of the AQ-6310, a controller is required. Connect a controller to the AQ-6310 with a GP-IB cord, and then turn ON the AQ-6310 **[POWER]** switch (13).

> CAUTION
>
> When connecting the GP-IB cord to the **[GP- IB]** connector (17), ensure that the **[POWER]** switch (13) is OFF, and that the bus system has completed handshaking.

3.6.1 Address Setup

(1) For both the listen and talk address setup operations for the AQ-6310, use the **[GP-IB ADDRESS]** switch. When this switch is pressed, the current address appears in the interrupt display area. Turn the rotary knob as required to obtain a desired address setting. Address settings of 0 through 30 are selectable.

(2) When a listen address is tranmitted from the controller to the AQ-6310 with REN and ATN set for "True", the AQ-6310 **[GP-IB REMOTE]** LED comes on.

(3) When a talk address is transmitted to the AQ-6310 in the same manner, the AQ-6310 becomes a talker and transmits data. The latest measured data, marker values, and AQ-6310 current function information are to be transmitted. The data to be transmitted must be specified using the function codes listed in Table 3-2.

The talker formats are indicated in Table 3-3.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(4) The address setting is cleared when the following conditions exit.

Listener ..... IFC is "True".
"Unlisten" command
My talk address

Talker ..... IFC is "True".
"Untalk" command
My listen command
Other talk address

3.6.2 Function Setup

(1) When data is transmitted in the formats indicated in Table 3-2 after the AQ-6310 has been designated as a listener by the controller, function or data setup can be achieved.

Example) To select a center wavelength of 1300 nm and a sweep width of 10 nm/div, the following procedure is performed.
CTR△WL1300, SWP△WD10 CR LF or EOI

(2) The function codes can be joined together using the symbol ",". The maximum acceptable number of characters except blanks is 512 (the CR, LF, and other control codes included). If this limit is exceeded, all transmitted codes are ignored.

(3) Blanks are all ignored so that they can be inserted as appropriate.

(4) Numerical value setup takes place when the values consist of no more than 8 digits (the decimal point excluded).

Example) RESOLN10 = RESOLN010 = RESOLN10. = RESOLN10.000

(5) The delimiter is either CR LF or EOI.

3.6.3 Measurement Start

Measurement starts when data transmission is effected in the formats indicated in Table 3-2 after the controller has designated the AQ-6310 as a listener.

Example) To make measurement by a single sweep, proceed as follows.
SGL CR LF or EOI

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.6.4 Measured Data Output

When the AQ-6310 is designated by the controller as a talker, measured data output takes place in the formats indicated in Table 3-3 immediately after completion of measurement.

Example) When the LOG scale is employed, proceed as follows.

DB△581, – 12.34, – 23.11,...., – 65.71 CR LF

(△: space)

3.6.5 Other Remote Messages

The following remote messages can be used with the AQ- 6310 in addition to those presented in sections 3.6.1 through 3.6.4.

(1) GO TO LOCAL

When this command is received, the AQ-6310 can no longer be externally controlled, but can be controlled by the local front panel controls. When such a condition occurs, the AQ-6310 **[GP-IB REMOTE]** LED goes off.

Such a control status change can also be effected by setting REN to be "False".

This GO TO LOCAL function can also be executed with the **[LOCAL]** switch on the AQ-6310.

NOTE

When the AQ-6310 is to be used in the remote mode after the [LOCAL] switch has been pressed to enter the local mode, designate the AQ-6310 as a listener and then follow the specified procedure.

(2) LOCAL LOCKOUT

When the **[LOCAL]** switch is pressed while the AQ- 6310 is in the remote mode, the apparatus goes into the local mode. However, the **[LOCAL]** switch becomes inoperative after the apparatus has received the LOCAL LOCKOUT command. Once this command is received, it remains effective until REN is rendered "False" (the **[LOCAL]** switch does not take effect until REN becomes "False").

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.6.6 SRQ Function

The SRQ signal is transmitted when measurement is terminated or printer paper has run out. The STATUS byte is subsequently transmitted by controller serial polling. The STATUS byte contents are indicated in Table 3-4.

3.6.7 DC Function

When DCL or SDC is receieved, the AQ-6310 goes into the state that prevails immediately after power turn-on.

3.6.8 DT Function

Upon receipt of GET, the AQ-6310 performs the same operation as specified by function code SGL.

3.6.9 GP-IB Usage Precautions

(1) When connecting the GP-IB cord to the **[GP-IB]** connector (17), be sure that the AQ-6310 is turned off, and that the bus system has completed handshaking.

(2) After the system setup has been changed, clear the entire interface system before resuming operations.

(3) When the AQ-6310 is connected to the GP-IB, the power must not be turned on or off unless handshaking is completed. This precaution must be observed whether the AQ-6310 is designated as a listener or talker. If this precaution is not followed, erroneous operations may occur because of noise generation, rendering the entire system inoperative.

(4) The employed **[GP-IB]** connector (17) has 24 pins and conforms to the IEEE-488 bus standard. If an IEC- IB interface bus is to be employed, remember that it complies with the IEEE-488 bus standard, but requires the use of 25-pin connectors. The difference between the IEEE-488 and IEC-IB connectors is indicated in Fig. 3-56.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes

No.	Category	Description	Data Format	Unit	Remarks
1	Center wavelength setup	400 to 1750	CENTER △WAVELENGTH ****.* or CTR△WL****.*	mm	
		Soft key <START WL> 400 to 1750	START △WAVELENGTH ****.* or STA△WL****.*		
		Soft key <STOP WL> 400 to 1750	STOP △WAVELENGTH ****.* or STP△WL****.*		(See note 1)
2	Sweep width setup	Setup on the /div basis 0.5, 1, 2, 5, 10, 20, 50, 100, 150	SWEEP △WIDTH***.* or SWP△WD***.*	mm	
		Setup on the SPAN basis 5, 10, 20, 50, 100, 200, 500, 1000, 1500	SWEEP :SPAN ****.* or SWP:SPAN****.*		
		Soft key <INTVL NORMAL>	SWEEP:INTVL△NORMAL or SWP:INTVL△NORM		
		Soft key <MEAS WL INTVL> 0.1, 0.2, 0.5, 1, 2, 5, 10, 20	SWEEP:INTVL*** or SWP:INTVL***		
		Soft key <SWEEP TIME> 1, 2, 5, 10, 20, 50	SWEEP :TIME ** or SWP:T**	sec	
3	Resolution setup	0.1, 0.2, 0.5, 1, 2, 5, 10	RESOLUTION **.* or RESOLN**.*	nm	

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

<table>
<tr><th>No.</th><th>Category</th><th colspan="2">Description</th><th>Data Format</th><th>Unit</th><th>Remarks</th></tr>
<tr><td>4</td><td>Reference level setup</td><td colspan="2"><table><tr><td colspan="3">HIGH SENSITIVITY</td></tr><tr><td rowspan="2">OFF</td><td colspan="2">ON</td></tr><tr><td>H-SENS FAST</td><td>H-SENS SLOW</td></tr><tr><td>0-42</td><td>0-60</td><td>0-70</td></tr></table></td><td>REFERENCE△LEVEL-**
or
REF△LEV-**</td><td>dB</td><td></td></tr>
<tr><td>5</td><td>Averaging times setup</td><td colspan="2">1, 2, 5, 10, 20, 50, 100, 200, 500, 1000</td><td>AVERAGE****
or
AVR****</td><td>Number of averaging operations</td><td></td></tr>
<tr><td rowspan="3">6</td><td rowspan="3">HIGH SENSITIVITY ON/OFF</td><td rowspan="2">ON</td><td>Soft key
<H-SENS SLOW>
The portion within square brackets ([ ]) can be omitted.</td><td>H-SENS[:SLOW]
or HIGH</td><td rowspan="3"></td><td rowspan="3"></td></tr>
<tr><td>Soft key
<H-SENS FAST></td><td>H-SENS:FAST</td></tr>
<tr><td>OFF</td><td>Soft key
<H-SENS OFF></td><td>H-SENS:OFF
or LOW</td></tr>
<tr><td rowspan="2">7</td><td rowspan="2">AUTO sweep</td><td colspan="2">AUTO ON

The portion within square brackets ([ ]) can be omitted.</td><td>AUTO [:NORMAL]</td><td rowspan="2"></td><td rowspan="2"></td></tr>
<tr><td colspan="2">AUTO OFF</td><td>AUTO:OFF</td></tr>
<tr><td rowspan="2">8</td><td rowspan="2">REPEAT sweep</td><td colspan="2">Soft key
<SWEEP NORMAL>

The portion within square brackets ([ ]) can be omitted.</td><td>REPEAT[:NORMAL]
or
RPT[:NORM]</td><td rowspan="2"></td><td rowspan="2"></td></tr>
<tr><td colspan="2">Soft key
<SWEEP MKR 1-2></td><td>REPEAT:MARKER
or
RPT:MKR</td></tr>
</table>

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
8	REPEAT sweep	Soft key **<WAIT>** 1 to 59S, 1 to 59M 1 to 24H (Lower-case letters may be used for unit entry.)	REPEAT:WAIT△ *** or RPT:WAIT△ ***	sec, min, hr	
		Soft key **<WAIT NORMAL>**	REPEAT: WAIT△NORMAL or RPT:WAIT△NORM		
9	SINGLE sweep	Soft key **<SWEEP NORMAL>** The portion within square brackets ([ ]) can be omitted.	**SINGLE[:NORMAL]** or **SGL[:NORM]**		
		Soft key **<SWEEP MKR1-2>**	SINGLE:MARKER or SGL:MKR		
10	Sweep stop		STOP or STP		
11	Y-SCALE LOG setup	0.5, 1, 2, 5, 10 (The value "10" is acceptable in the HIGH-SENSITIVITY-ON state only.)	YSCL**.*	dB	
12	Y-SCALE LIN setup		YSCL△LIN		
13	NORMALIZE setup		NORMALIZE or NORM		
14	DOMINANT setup		DOMINANT or DOMI		

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
15	3D setup	-50, -40, -30, -20, -10, 0, + 10, + 20, + 30, + 40, + 50, OFF When the portion within square brackets is omitted, a setting of -20 is automatically selected. If OFF is selected, the system returns to the normal screen mode.	3D[***]	deg	
16	Peak level search		PEAK△LEVEL or LPK△SRCH		
17	Peak wavelength search	*** = ON or OFF When the portion within square brackets is omitted, both the AUTO and CROSS operations are performed in the OFF state.	PEAK△WAVELENGTH (:AUTO***:CROSS***) or WPK△SRCH[:AUTO***:CROSS***]		
18	Half-width search	ENV method The portion within square brackets ([ ]) can be omitted.	HALF-WIDTH△SEARCH [:ENV] or HALF△SRCH[:ENV]		
		RMS method Soft key <RMS x*.**> *.** = 2.00, 2.35 Soft key <RMS> if the portion within square brackets is omitted	HALF-WIDTH△SEARCH :RMS[*.**] or HALF△SRCH:RMS[*.**]	Number of times	
		Soft key <THRESH **dB> ** = 3, 15, 20	HALF-WIDTH△SEARCH :THRESH** or HALF△SRCH:THRESH**	dB	
		Soft key <THRESH 1/e 2>	HALF-WIDTH△SEARCH :THRESH1/E2 or HALF△SRCH:THRESH1/E2		

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
19	Marker clear		MARKER△CLEAR or MKR△CLR		
20	PEAK → CENTER setup		CENTER = PEAK or CTR = PK		
21	MARKER → CENTER setup		CENTER = MARKER or CTR = MKR		
22	PEAK → REF LEVEL setup		REF△LEVEL = PEAK or REF△LEV = PK		
23	MARKER → REF LEVEL setup		REF△LEVEL = MARKER or REF△LEV = MKR		
24	△W→ SPAN setup	* = 1, 2, 3, 4, 5 Soft key <X*> * = 1 if the portion within square brackets is omitted	SPAN = W[*]	Number of times	
25	DATA MEMORY selection	Acc A selection	MEMORY:A or MEM:A or REF		
		Acc B selection	MEMORY:B or MEM:B or MEAS		
26	DATA MEMORY recall	Acc A recall	RECALL:ACCA or RCL:ACCA or RECALL△REF		
		Acc B recall	RECALL:ACCB or RCL:ACCB or RECALL△MEAS		

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
27	Data subtraction by DATA MEMORY	Acc A - Acc B in LOG scale mode	MEMORY:A-B or MEM:A-B or LOSS		
		Acc A / Acc B in LIN scale mode	MEMORY:A/B or MEM:A/B		
		Acc B - Acc A in LOG scale mode	MEMORY:B-A or MEM:B-A or TRANS.M		
		Acc B / Acc A in LIN scale mode	MEMORY:B/A or MEM:B/A or TRANS.M		
28	Waveform clear		GRAPH△CLEAR or CLR		
29	MEMORY selection	* = 0 to 9, A to F	SAVEM:*		
30	MEMORY recall	* = 0 to 9, A to F	RECALL:* or RCL:*		
31	Printer operation	Printing start	PRINTER△ON or PR△ON		
		Ongoing printing stop	PRINTER△OFF or PR△OFF		
		About 40 mm recording paper feed	PRINTER△FEED or PR△FEED		

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
32	Date/time setup	Date setup **.**.** = month/day/year	DATE**.**.**		
		Time setup **:** = hour/minute	TIME**:**		
33	Label setup	**---** = comment consisting of no more than 25 characters	LABEL '**---**' or LBL '**---**'		(See note 2)
34	Label clear		LABEL:CLEAR or LBL :CLR		
35	Request for service request output	ON	EI		(See note 3)
		OFF	DI		
36	Request for measured data output	*** = measurement point The leftmost end of the graph is 1 and the rightmost end is 581. R1-R581 if the portion within square brackets is omitted	DDATA△[R***-R***]		(See note 4) (See note 5) (See note 6)

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-2 Function Codes (Continued)

No.	Category	Description	Data Format	Unit	Remarks
37	Request for memory data output	Request for MEMORY (0 to 9 and A to F) data * = 0 to 9, A to F The portion within square brackets is the same as indicated under No. 36.	DMEM*△[R***-R***]		
		Request for Acc A data The portion within square brackets is the same as indicated under No. 36.	DMEM△ACCA△[R***-R***] or DREF△[R***-R***]		
		Request for Acc B data The portion within square brackets is the same as indicated under No. 36.	DMEM△ACCB△[R***-R***] or DMEAS△[R***-R***]		
38	Request for marker position data output	The marker value set up as indicated under Nos. 16 through 18 is transferred out.	MARKER or MKR		
39	Request for FUNCTION information output		STATE		

NOTE 1: As is the case with manual setup, the acceptable stop wavelength setting = start wavelength + sweep width (SPAN). If the specified wavelength value does not satisfy the above equation, the system ignores such a value.

NOTE 2: Only the label characters (A-Z, 0-9, +, -, *, /, =, >, <, ., ,, :, ;, and space) are acceptable. The other characters are treated as blanks when displayed. Further, when the command is executed while a label is set up on the screen, the system deletes a previously displayed label and displays a newly selected label instead. When 25 or fewer characters are displayed, they are centered. However, if a space is put in front of a comment, the comment will not be centered but left-justified.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

NOTE 3: When the power is turned on or the device clear signal is received, the setup is for DI.

NOTE 4: In the DDATA state, data transfer does not take place until measurement is terminated. Further, once data transfer has started, the system does not proceed to the next measurement until the transfer is completed. If ongoing transfer is to be stopped, execute IFC.

NOTE 5: The output data setup remains effective until it is updated. When the power is turned on, the setup is for DDATA.

NOTE 6: After DATA MEMORY data subtraction command execution (No. 27), the data derived from subtraction is to be transmitted.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-3 Talker Formats

No.	Category	Format		
1	DATA	LOG	DB△number of data, data, ..., data CR LF Number of data 1 to 581 Data format ±**.**	
		LIN	AccA/AccB or AccB/AccA display state	LN△number of data, data, ..., data CR LF Number of data 1 to 581 Data format *.**** Data 0 to 4
			Other display states	LN△number of data, data, ..., data CR LF Number of data 1 to 581 Data format **.*** Data 0 to 10
2	MARKER	Level marker LOG	MARKER△LOG△ data, data, data CR LF Marker Marker 1 Marker 2 Data format ±**.**	
		Level marker LIN	MARKER△LIN△ data, data, data CR LF Marker Marker 1 Marker 2 Data format 0≦**.***≦10 The data value is determined relative to a REF LEVEL of 1.	
		Wave-length marker	MARKER△WAVE△wavelength value, data point, Marker wavelength value, data point, wavelength value, data point CR LF Marker 1 Marker 2 If the values are not specified, the associated portions are treated as blanks. Data format ****.*, ***	
3	STATE	STATE△△CTR△WL****.*, SWP△WD***.*, RESOLN**.*, REF△LEV-**, AVR****, YSCL**.*, HIGH CR LF (YSCL△LIN)(LOW△)		

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-4 STATUS Contents

Bit	Bit Contents Description
D8	0
D7	SRQ signal is transmitted.
D6	0
D5	0
D4	0
D3	Monochromator has become abnormal.
D2	Printer paper has run out.
D1	Measurement is terminated.

Table 3-5 AQ-6310 Interface Construction

Category	Code	Remarks
Source (transmission) handshake	SH1	All SH functions are provided.
Acceptor (reception) handshake	AH1	All AH functions are provided.
Talker	T6	Basic talker
		Serial polling
		Talker clear by MLA
Listener	L4	Basic listener
		Listener clear by MTA
Remote local	RL1	Local lockout function is provided.
Service request	SR1	All SR functions are provided.
Parallel polling	PP0	No PP function is provided.
Device clear	DC1	All DC functions are provided.
Device trigger	DT1	All DT functions are provided.
Controller	C0	No CO function is provided.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-56 Bus Line Description and Connector Difference

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.7 ATTENTION Display Function

The AQ-6310 causes the interrupt display area to give an indication in the form of "ATTN **" (** = any number between 0 and 37) when an inoperative switch is pressed during operation, forced setup condition change is effected by switch activation, or a switch-specified operation cannot be executed because of unacceptable conditions. Various numbers (0 through 37) are attached to the ATTN indication to show the cause of ATTN condition occurrence. These ATTN numbers are classified and defined in Tables 3-6 through 3-8.

(1) An Inoperative Switch Was Pressed

Table 3-6 ATTN Indications No. 0 to 10

No.	Cause
0	An inoperative switch was pressed during sweep operation.
1	An inoperative switch was pressed during 3-D display sweep operation.
2	An inoperative switch was pressed while 3-D display sweep operation was stopped.
3	The [MARKER→CENTER] or [MARKER→REF LEVEL] switch was pressed although marker setup was not performed.
4	Soft key <SWEEP MKR1-2> was pressed although wavelength markers 1 and 2 were set outside the preselected measurement range or not set up at all.
5	An inoperative switch was pressed while sweep was being conducted between markers 1 and 2.
8	A switch other than label character switches 0 through 9 or A through F was pressed after [SAVE] switch activation.
9	The [RECALL] and [*] switches were pressed when no waveform was stored in memory * (* = 0 through 9, A through F, Acc A, or Acc B).
10	The [PEAK→REF LEVEL] or [MARKER→REF LEVEL] switch was pressed after waveform update was implemented by [A-B A/B], [B-A B/A], [NORMALIZE], or [DOMINANT] switch activation.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(2) Setup Conditions Were Changed by Switch Activation

Table 3-7 ATTN Indications No. 21 to 27

No.	Cause
21	When the center wavelength or sweep width was changed at a resolution setting of 10 nm, the range of 400 to 569 nm was included in the sweep range so that the resolution was changed to 5 nm.
22	The Y scale is changed to 5 dB/div because the measurement sensitivity is changed to **<H-SENS OFF>** while the Y scale is being set to 10 dB/div by **<H-SENS SLOW>** or **<H-SENS FAST>**.
23	The averaging times setting is changed to 1 because the measurement sensitivity is changed to **<H-SENS OFF>** while the averaging times setting is being set to 2 by **<H-SENS SLOW>** or **<H-SENS FAST>**.
24	The reference level is changed to a acceptable value because the reference level becomes an unacceptable value in response to a change in measurement sensitivity.
25	The averaging times setting is changed to 20 because the measurement sensitivity is changed to **<H-SENS SLOW>** or **<H-SENS FAST>** while the averaging times setting is being set to 50 or more by [H-SENS OFF].
26	Problems 23 and 24 above occur at the same time because the measurement sensitivity is changed from **<H- SENS SLOW>** or **<H-SENS FAST>** to **<H-SENS OFF>**.
27	The Y scale is changed because the lowest Y axis scale mark is smaller than -99 dB when the reference level is changed in the LOG scale.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(3) A Switch-specified Operation Could Not Be Executed because of Unacceptable Conditions

Table 3-8 ATTN Indications No. 31 to 37

No.	Cause
31	As the Acc A and Acc B did not agree in center wavelength, sweep width, and resolution for waveform measurement when the [A-B A/B] or [B-A B/A] switch was pressed, the data subtraction display function was not executed.
32	Half-width search was not conducted upon [△W→SAPAN] switch activation.
33	Although the [CALIBRATION (632.8 nm)] switch was pressed to start automatic wavelength calibration, the intended calibration was not effected because the input level was too low or the wavelength error was greater than ± 5 nm.
34	Although the [PEAK→CENTER] or [PEAK→REF LEVEL] switch was pressed, the peak value was not set as the center wavelength or reference level.
35	Although the [MARKER→CENTER] or [MARKER→REF LEVEL] switch was pressed, the marker value was not set as the center wavelength or reference level.
36	As the peak value was too small, normalization was not effected upon [NORMALIZE] switch activation.
37	Visual sensitivity correction was not made upon [DOMINANT] switch activation.

NOTE: All the above-listed ATTN indications go off when the sweep or other function switch is pressed.

3.8 Video Interface

The composite video signal output is delivered to the [VIDEO OUT] terminal on the rear panel. This output signal is the same as that provided for CRT display purposes, as indicated in Fig. 3-57.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-57 Output Video Signal

When a video plotter matching the above indicated terminal and output video signal is hooked up to the AQ-6310, it is easy to obtain hard-copy printout of CRT screen contents.

Video plotter recording examples are presented in Fig. 3-4 and other figures.

Further, when the AQ-6310 is connected to a monitor TV that is compatible with the above indicated output video signal, it is possible to watch the AQ-6310 CRT screen contents on the monitor TV. Note that the monitor TV to be used must have a video signal band of 25 MHz or higher frequencies.


Before producing hard-copy printout of AQ-6310 CRT screen contents, it is necessary to make video plotter adjustment as directed below.

Turn ON only switch 1 of the video plotter rear panel DIP switch; be sure that the other switches are OFF. Adjust trimmers VR1 through VR8 as indicated in Fig. 3- 58.

The adjustment must be nearly complete now. However, satisfactory image quality may still not result because there are inherent errors in individual plotters. If such a condition is encountered, make fine adjustment as indicated in Fig. 3-59.

The plotters supplied with the AQ-6310 have been optimally adjusted at factory prior to shipment; therefore, do not tamper with the trimmers on such plotters.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



Rear Panel





Fig. 3-58 Video Plotter Adjustment

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com



VR1	Gain adjustment	VR5	Print image START adjustment
VR2	Print image density adjustment	VR6	Print image END adjustment
VR3	Video timing adjustment	VR7	Print image undesired portion elimination
VR4	Print image aspect ratio adjustment	VR8	Print image WIDTH adjustment

Fig. 3-59 Fine Adjustment

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.9 Abbreviations

The abbreviations displayed on the CRT have the following meanings.

Table 3-9 Abbreviation List

Abbreviation	Meaning
ACC	Accumulator
ATTN	Attention
AVR	Averaging Times
A-B	Accumulator A-Accumulator B
CAL	Calibration
CONT	Continue
CROSS	Cross Marker
CTR WL	Center Wavelength
DEG	Degree
DOM	Dominant
ENV	Envelope
H	High Sensitivity
HORIZE	Horizontal
INTVL	Interval
LIN	Linear Scale
LMARK	Level Marker
LOG	Logarithm Scale
MEAS	Measure
MEM	Memory
NORM	Normalize
P OFF	Printer Off
P ON	Printer On
REF	Reference Level
REF LEV	
RES	Resolution
RESOLN	

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Table 3-9 Abbreviation List (Continued)

Abbreviation	Meaning
SWP	Sweep Width
SWP WD	
THRESH	Threshold
VERT	Vertical
WMARK	Wavelength Marker
YS	Y-Scale
3D	3-Dimension
m	10^{-3}
μ	10^{-6}
n	10^{-9}
p	10^{-12}

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.10 Measuring Precautions

3.10.1

Of lights diffracted by the grating only the primary light is usually used. Being designed to permit measurement with wider range, higher sensitivity, and finer resolution, this unit provides for selective use of the primary, secondary, or tertiary light according to the wavelength chosen. With this unit the spectrum therefore appears at particular points determined by wavelengths other than the actual source wavelength. For example, in cases where measurement is made with the secondary light the primary and tertiary light waveforms also appear with the secondary light waveform.

The division of light is shown in Fig. 3-60. This figure indicates at which wavelength the spectrum appear elsewhere than the actual source wavelength.




Fig. 3-60

For example, assume that He-Ne is launched at 632.8 nm. In this case while measurement is made with the secondary light as shown in Fig. 3-60, spectra also appear at 421.9 nm for the tertiary light and at 1265.6 nm for the primary light. Also when light is launched at 421.9nm and 1265.6 nm, spectra appears at 421.9, 632.8, and 1265.6 nm.

However, when light is launched into the tertiary or secondary light the primary light is substantially reduced in level depending on the effect of cut filters, grating diffraction efficiency, and sensor response. Similarly, when the light is launched into the primary light, the secondary or tertiary light is reduced in level to a considerable extent.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

It is possible to determine at which point a spectrum appears when each divided light is launched, using the data of (1), (2), and (3) shown below.

(1) 400 to 560 nm (tertiary light) : λ, 3/2λ, 2λ

(2) 560 to 950 nm (secondary light) : 2/3λ, λ, 2λ

(3) 950 to 1750 nm (primary light) : 1/3λ, 1/2λ, λ

To obtain best results it is well to remember the characteristics above of this unit.

When measurements for loss wavelength characteristics or transfer characteristics are made over a range of 700 to 1700 nm, the AQ-4303B is used with the white light source in the wavelength range of 700 to 1000 nm. The filter built in the white light source cuts the high- order light, thereby permitting accurate measurements without any problem. (See 3.5.17.)

This unit provides for measurement throughout a range of 700 to 1700 nm, using white light source wavelength over a 700 to 1000 nm range. This is because the AQ-4303B cuts the wavelength of 700 nm or less but passes the light of 1000 nm or longer wavelength. Conventional spectroscopes pass light of integral multiples of a certain wavelength so that when they are used over a wavelength range of 1000 nm or more, they must be set in the range of 1000 to 1800 nm. This unit, however, eliminates the need of such settings because it contains the filter to cut high-order light on the longer- wavelength side.
When 1700 nm or longer wavelength measurements are to be made, set the AQ-4303B for a wavelength range of 1000 to 1800 nm. When this wavelength range setting is employed, wavelengths of 1000 to 1750 nm can be covered.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Figs. 3-61 to 3-63 show how waveforms appear when light sources satisfying the above three conditions are measured.




Fig. 3-61 $\lambda = 441.6$ nm




Fig. 3-62 $\lambda = 632.8$ nm

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com




Fig. 3-63 $\lambda = 1152.3$ nm

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

3.10.2 Spectroscope Construction and Principle

The construction and principle of the spectroscope and the reason why the phenomenon mentioned above occurs are described below.

(1) The spectroscope of the apparatus is a Czerny Turner monochromator that employs a diffraction grating.

Fig. 3-64 shows the principle of operation of the Czerny Turner monochromator.




Fig. 3-64 Principle of Czerny Turner Monochromator

The light from the point light source at the input end is converted into parallel rays by mirror 1. These rays fall upon the grating which diffracts the rays of light according to their wavelength. The light diffracted by the grating is reflacted by mirror 2 and mirror 3, passed through the output slit, and received by the sensor.

In general, only the primary light out of the diffracted rays from the grating is used for measurement. However, this apparatus uses the secondary and third in addition to the primary to widen the measurement band and speed up the measurement.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

For this reason, when light to be measured with the secondary is injected, the primary and third appears at the output in addition to the secondary, so that waveforms for the primary and third appear besides the waveform at the actual wavelength.

However, waveforms appearing at wavelengths other than the actual wavelength can be easily distinguished as their levels are low because of the low grating diffraction efficiency and cut fiber action.

(2) Principle of generation of primary, secondary, third, ... , and nth light by the grating




Fig. 3-65 Grating

As shown in Fig. 3-65, when light whose wavelength is λ falls upon the grating at an angle of α with respect to the normal, the relationship between the wavelength, incident angle, and outgoing angle is expressed by the following formula.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

$m\lambda = D(\sin\alpha \pm \sin\beta)$ — Grating basic formula

where m = 0, 1, 2, 3, ... (degree)

λ = wavelength

D = grating constant

α = incident angle

β = outgoing angle

Setting the right hand side of the formula to λo, we have

$$m\lambda = \lambda o\ ,\ \lambda = \frac{\lambda o}{m}$$

When a wavelength of λo is injected, mth light appears at $\lambda = \lambda o$, $\frac{\lambda o}{2}$, $\frac{\lambda o}{3}$ for m = 1, 2, 3, ...

For example, at λo = 1500 nm, secondary and third light appear at 750 nm and 500 nm, respectively.

Since this apparatus measures wavelengths lower than 560 nm with the third light, wavelengths between 560 nm and 950 nm with the secondary, and wavelengths higher than 950 nm with the primary, waveforms appear at the wavelengths shown in Section 3.10.1, according to the above formula.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SECTION 4
CIRCUITRY AND CONSTRUCTION

4.1 Introduction

This section outlines the circuitry and construction of the AQ-6310.

4.2 Circuitry

4.2.1 Circuitry Overview

The AQ-6310 converts the monochromator output light (consisting of isolated narrow portions of the spectrum) to electrical signals, and amplifies such signals in the input amplifier circuit. The output voltage from this circuit is subjected to A/D conversion and transferred on the data bus. The CPU circuits receives, processes, and feeds the data to the image control circuit so that the data is used for CRT Y-axis display. In such processing sequences, the wavelength sensitivity characteristics of the sensor and other devices are corrected and multiplied to suit the scale. On the other hand, the wavelengths related to the X-axis are monitored by counting the pulses fed to the DC servomotor.

4.2.2 Individual Circuit Descriptions

(1) Input Amplifier Circuit (PRE-200784)

The input light is converted to electrical signals by the sensor mounted on the monochromator, and then amplified in accordance with the preselected reference level.

(2) CPU Circuit I (PRE-407698)

This circuit contains ROMs, RAMs, and 16-bit CPU to exercise control over the entire system and carry out data processing.

(3) CPU Circuit II (PRE-407699)

This circuit contains ROMs, RAMs, and 16-bit CPU, and mainly provides control over the monochromator section.

(4) CPU Circuit III (PRE-407737)

This circuit contains ROMs, RAMs, and 16-bit CPU, and mainly controls image processing operations.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

(5) A/D Conversion Circuit (PRE-906388)

This circuit is divided into a signal A/D converter and a high-sensitivity measurement circuit. The high-sensitivity measurement circuit consists of a synchronous detector and a low-pass filter.

(6) A/D Conversion Control Circuit (PRE-407702)

This circuit generates the timing for A/D conversion.

(7) Image Control Circuit (PRE-407701)

The CRT control LSI provides CRT display in accordance with the instructions from the CPU III.

(8) Unit Drive Circuit (PRE-407703)

This circuit provides control of the monochromator section in accordance with the instructions from the CPU II.

(9) KEY/GP-IB Circuit (PRE-407700)

This circuit monitors the panel switches and rotary encoder, and causes the panel LEDs to come on as needed. It also contains a GP-IB controller.

(10) Keyboard Circuit (PRE-906389)

This circuit consists of the front panel switches, LEDs, and rotary encoder.

(11) Soft Key Circuit (PRE-906395)

This circuit consists of the soft key switches and LEDs located on the lower end of the CRT.

(12) Motor Drive Circuit (PRE-407705)

This circuit drives the resolution setup motor.

(13) Printer Circuit (PRE-407704)

This circuit contains RAMs and 8-bit one-chip CPU to exercise control over the printer operations.

(14) Printer Hookup Circuit (PRE-906506)

This is the junction circuit for connecting the printer to the printer circuit (13).

(15) Motherboard Circuit (PRE-906395)

This circuit interconnects PC board terminals.

4.3 Construction

The AQ-6310 external views are presented in Attached Drawings ASD-61712-1 1/3 through 3/3.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

SECTION 5
MAINTENANCE

5.1 Introduction

This section describes the pre-operational checkout and mechanical inspection which are performed to verify the normal operations of the AQ-6310. If any abnormality is found, however, proceed as indicated in sections 2.3 and 2.4.

5.2 Pre-operational Checkout


The AQ-6310 performs internal memory checkout upon power turn-on. If any abnormality is detected, the system stops its operations and displays the message "***** CHECK ERROR" appears at the center of the CRT screen (***** = ROM, RAM, NVRAM, DROM, OROM, DPRAM1, or DPRAM2).

5.3 Mechanical Inspection


Visually check the operating controls, connectors, and other assembled parts exposed to view to ensure that they are free of damage and deformation. Also check them for looseness or rough movement to be sure that their mechanical performance is normal.

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


OUTSIDE VIEW OF AQ-6310 SPECTRUM ANALYZER (EXTERNAL VIEW)




TOP VIEW



REAR VIEW




FRONT VIEW



SIDE VIEW

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com


OUTSIDE VIEW OF AQ-6310 SPECTRUM ANALYZER (FRONT VIEW)



DIMENSIONS SHOWN IN mm

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

OUTSIDE VIEW OF AQ-6310 SPECTRUM ANALYZER (REAR VIEW)



ASD-61712-1-3/3

Artisan Technology Group - Quality Instrumentation ... Guaranteed | (888) 88-SOURCE | www.artisantg.com

Artisan Technology Group is an independent supplier of quality pre-owned equipment

Gold-standard solutions

Extend the life of your critical industrial, commercial, and military systems with our superior service and support.

We buy equipment

Planning to upgrade your current equipment? Have surplus equipment taking up shelf space? We'll give it a new home.

Learn more!

Visit us at **artisantg.com** for more info on price quotes, drivers, technical specifications, manuals, and documentation.

Artisan Scientific Corporation dba Artisan Technology Group is not an affiliate, representative, or authorized distributor for any manufacturer listed herein.

We're here to make your life easier. How can we help you today?

(217) 352-9330 | sales@artisantg.com | artisantg.com

ARTISAN® TECHNOLOGY GROUP