

Pyris 6 DSC

Differential Scanning Calorimeter

OPERATION

DSC Type	Disc-type heat flux design	
DSC Cell	Low mass furnace with rust-free solid metal	
Sensor	Precisely machined disc of Chromel alloy (90% Nickel/10% Chromium)	
Atmosphere	Static or dynamic, including nitrogen, argon, helium, carbon dioxide, air, oxygen or other inert or active gases, over full temperature range.	
Calorimetry	Accuracy	Better than $\pm 2\%$
	Precision	Better than $\pm 1\%$
Sensitivity	Digital resolution	0.6 μW
	Detectable peak	2 μW
Dynamic range	± 250 mW	
Isothermal Drift (10 min)	100°C	< 3 μW
	400°C	< 5 μW
Signal Response	1 mg Indium, 10°C/min, Nitrogen purge	
	Peak Height	7 mW
	Width at half height	0.6°C
	H/W Ratio	11.7 mW/°C
Temperature range	-120°C to +450°C	
Temperature	Accuracy	$\pm 0.2^\circ\text{C}$
	Precision	$\pm 0.1^\circ\text{C}$
Scan Rates	Heating	0.1°C to 100°C/min
	Cooling	0.1°C to 100°C/min
Cooling Time	300°C to 50°C	
	Under 7 minutes using C6 Chiller	
Built-in Gas Controller	Precise mass flow controller, and gas switch from gas A to gas B	
	Gas type	Pre-calibrated for N ₂ , O ₂ and air
	Pressure range	0.5 to 4 bar
	Flow rate range	1 to 200 ml/min
	Flow rate accuracy	± 1 ml/min
	Time constant	< 1 second
Cooling Options	Circulating Liquid (C6 Chiller)	
	Cryostat	
	Liquid N ₂ Gas	
User Control	Pyris™ Software for Windows®	
Autosampler	Up to 45 samples can be run unattended, and samples can be added anytime. Customize your sample runs through our Pyris Player software to increase productivity.	

Sample Pans

Type	Material	Pressure Rating
Standard	Aluminum Gold, Platinum, Copper, Graphite Aluminum oxide	Ambient
Sealed	Aluminum, Gold	2 atmospheres
Large Volume	Stainless Steel	24 atmospheres
High Pressure	Stainless Steel Gold plated Stainless Steel, Titanium	150 atmospheres
Robotic	Aluminum, Gold	Ambient, vented 1 & 2 atmospheres

Quality Assurance	Developed under ISO 9001	
Dimensions (HxWxD)	17 cm x 36 cm x 40 cm (6.7 x 14.2 x 15.7 in)	
Weight	13 kg (29 lbs)	
Power Requirement	100-240 Volts, 50/60 Hz	

Ordering Information

Part Number	Description
N520-0080	Pyris 6 DSC Instrument (100/115/230v)
N537-0070	Pyris 6 DSC Lab System (100/115/230v)
N537-0071	Pyris 6 DSC with AS6 Autosampler Lab System (100/115/230v)

To order or request additional information on the Pyris 6 DSC, please contact PerkinElmer or your local representative.



PerkinElmer Instruments
761 Main Avenue
Norwalk, CT 06859-0010 USA
Tel: 800-762-4000 or
(+1) 203-762-4000
Fax: (+1) 203-762-4228

PerkinElmer is a trademark of PerkinElmer, Inc.
Pyris is a trademark of PerkinElmer LLC.

